

Data Validation Package

**September 2014
Groundwater and Surface Water
Sampling at the Slick Rock, Colorado,
Processing Sites**

December 2014



U.S. DEPARTMENT OF
ENERGY

Legacy
Management

This page intentionally left blank

Contents

Sampling Event Summary	1
Slick Rock, Colorado, Processing Sites, Sample Location Map	5
Data Assessment Summary.....	7
Water Sampling Field Activities Verification Checklist	9
Laboratory Performance Assessment	11
Sampling Quality Control Assessment.....	21
Certification	24

Attachment 1—Assessment of Anomalous Data

Potential Outliers Report

Attachment 2—Data Presentation

Groundwater Quality Data
Surface Water Quality Data
Equipment Blank and Trip Blank Data
Static Water Level Data
Hydrographs
Groundwater Time-Concentration Graphs
Surface Water Time-Concentration Graphs
Dolores River Discharge Graph

Attachment 3—Sampling and Analysis Work Order

Attachment 4—Trip Report

This page intentionally left blank

Sampling Event Summary

Site: Slick Rock, Colorado, Processing Sites

Sampling Period: September 8–9, 2014

The Slick Rock, Colorado, Processing Sites are referred to as the Slick Rock West Processing Site (SRK05) and the Slick Rock East Processing Site (SRK06). This annual event involved sampling both sites for a total of 16 monitoring wells and 7 surface water locations as required by the 2006 *Draft Final Ground Water Compliance Action Plan for the Slick Rock, Colorado, Processing Sites* (GCAP). Water levels were measured at all sampled wells.

The proposed compliance strategy for the Slick Rock sites is natural flushing in conjunction with institutional controls and compliance monitoring. Contaminant concentrations at the Slick Rock sites are compared to their respective maximum concentration limit (MCL) to assess compliance with Title 40, *Code of Federal Regulations*, Part 192 (40 CFR 192), with the exception of manganese and selenium. Manganese concentrations are compared to the maximum historical background concentration of 4.2 milligrams per liter (mg/L) to assess compliance because manganese does not have an MCL. A human-health risk-based alternate concentration limit of 0.18 mg/L has been proposed to assess compliance for selenium because groundwater modeling predicts that selenium concentrations at the Slick Rock West Processing Site will not be reduced to below the MCL within 100 years.

The constituents of potential concern (COPCs) defined in the GCAP for the West Processing Site are manganese, molybdenum, nitrate, selenium, and uranium. Additional COPCs (radium-226, radium-228, benzene, toluene, ethylbenzene, and xylenes) are isolated to one well (0319). As shown in Table 1, results from this sampling event demonstrate elevated concentrations for most contaminants at West Processing Site locations.

Selenium and uranium are the COPCs at the East Processing Site. Uranium concentrations exceed the MCL at most East Processing Site groundwater locations. The selenium contamination is isolated to the onsite well 0305. Wells with analyte concentrations that exceeded applicable groundwater standards are listed in Table 1.

Table 2 lists the drinking water maximum contaminant levels and results for benzene, toluene, ethyl benzene, and xylenes (total) in well 0319. The radium-226 plus radium-228 concentration has decreased in this well since 2006, and remains below the maximum contaminant level of 5 picocuries per liter.

Table 1. Slick Rock Wells with Samples that Exceeded Standards in September 2014

Analyte	Standard (mg/L)	Site	Location	Concentration (mg/L)
Manganese ^a	4.2	West	0340	4.4
Molybdenum	0.1	West	0317	0.18
			0318A	0.76
			0339	0.92
			0340	1.5
			0508	1.10
			0510	0.89
Nitrate + Nitrite as Nitrogen	10	West	0318A	46
			0339	54
			0340	210
			0508	140
			0510	180
Selenium ^b	0.18	West	0318A	4.7
			0339	2.8
			0340	2.0
			0508	1.3
			0510	0.87
	0.01	East	0305	0.024
Uranium	0.044	West	0508	0.055
			0510	0.100
		East	0300	0.045
			0303	1.2
			0305	0.72
			0307	0.42
			0311	0.077

Standards are listed in 40 CFR 192.02 Table 1 to Subpart A; concentrations are in milligrams per liter (mg/L).

^a Manganese standard is the maximum historical background concentration observed in well SRK06 0300.

^b Selenium standard for the West Processing Site is the proposed Alternate Concentration Limit.

Table 2. BTEX ^a Maximum Contaminant Levels and Results for Well 0319 in September 2014

Analyte	Maximum Contaminant Level (mg/L)	Concentration in Well 0319 (mg/L)
Benzene	0.005	2.2
Ethyl benzene	0.7	0.13
Toluene	1	0.37
Xylenes, Total	10	3.67

Maximum Contaminant Levels are listed in the 2009 *National Primary Drinking Water Regulations*.

(EPA 816-F-09-0004, May 2009); concentrations are in milligrams per liter (mg/L).

^a BTEX = Benzene, toluene, ethyl benzene, and xylenes (total).

Surface water results from Dolores River locations downstream of and adjacent to the processing sites were compared to statistical background threshold values (BTVs) derived using historical data (from 1997 to present) at background river locations. The background locations are 0693, which is located upstream of the West Processing Site, but downstream of the East Processing Site, and 0696 which is located upstream of the East Processing Site.

Surface water location 0692 at the East Processing Site is monitored for uranium because it is the predicted location where the centroid of the uranium plume will intersect the river. The uranium concentrations at this location and at 0700, which is farther downstream, remain well below the BTV concentration for background location 0696, as shown in Table 3.

Table 3. Comparison of Slick Rock East Processing Site September 2014 Surface Water Concentrations to Historical Upgradient BTVs

Analyte	BTV for 0696 (mg/L)	0692 Concentration (mg/L)	0700 Concentration (mg/L)
Uranium	0.0023	0.00084	0.00099

West Processing Site surface water locations in the Dolores River are monitored to verify that the compliance strategy is protective of the environment. The potential for environmental exposure to site contaminants exists in the Dolores River because it receives groundwater discharge from the contaminated alluvial aquifer. As shown in Table 4, BTVs were exceeded for nitrate + nitrite as N at location 0694, and manganese, molybdenum, nitrate + nitrite as N, selenium, and uranium at locations 0349 during this event. Location 0349 is the predicted location where the centroid of the contaminant plumes will intersect the river. The concentrations observed are likely related to the elevated discharge rate of the Dolores River at the time of sampling (see Discharge Graph in Attachment 2).

Table 4. Comparison of Slick Rock West Processing Site September 2014 Surface Water Concentrations to Historical Upgradient BTVs

Analyte	BTV for 0693 (mg/L)	0347 Concentration (mg/L)	0349 Concentration (mg/L)	0694 Concentration (mg/L)
Manganese	0.024	0.020	0.170	0.0035
Molybdenum	0.008	0.0079	0.019	0.0004
Nitrate + Nitrite as N	0.47	0.22	0.52	0.66
Selenium	0.0047	0.0035	0.022	0.00018
Uranium	0.0043	0.0028	0.007	0.00039


David Traub
 David Traub
 Site Lead, The S.M. Stoller Corporation,
 a wholly owned subsidiary of Huntington Ingalls Industries,

12-9-14
 Date

This page intentionally left blank



- LEGEND**
- WELL TO BE SAMPLED
 - SURFACE LOCATION TO BE SAMPLED
 - - - SITE BOUNDARY



U.S. DEPARTMENT OF ENERGY GRAND JUNCTION, COLORADO	Work Performed by S.M. Stoller Corporation Under DOE Contract No. DE-AM01-07LM00060
Planned Sampling Map Slick Rock East and West, CO, Processing Sites September 2014	
DATE PREPARED: July 29, 2014	FILENAME: S1207200

M:\LTS\111\0001\16\000\S12072\S1207200_11x17.mxd smithw 07/29/2014 4:02:19 PM

Slick Rock, Colorado, Processing Sites, Sample Location Map

This page intentionally left blank

Data Assessment Summary

This page intentionally left blank

Water Sampling Field Activities Verification Checklist

Project	<u>Slick Rock, Colorado</u>	Date(s) of Water Sampling	<u>September 8–9, 2014</u>
Date(s) of Verification	<u>November 17, 2014</u>	Name of Verifier	<u>Alison Kuhlman</u>

	Response (Yes, No, NA)	Comments
1. Is the SAP the primary document directing field procedures? List any Program Directives or other documents, SOPs, instructions.	<u>Yes</u>	<u>Work Order letter dated August 5, 2014.</u>
2. Were the sampling locations specified in the planning documents sampled?	<u>No</u>	<u>Location 0312 was dry.</u>
3. Were calibrations conducted as specified in the above-named documents?	<u>Yes</u>	<u>Calibrations were performed on September 5, 2014.</u>
4. Was an operational check of the field equipment conducted daily? Did the operational checks meet criteria?	<u>Yes</u> <u>No</u>	 <u>In the post-trip operational check, the two lowest turbidity standards failed low. All associated turbidities are qualified with "J" flags as estimated values.</u>
5. Were the number and types (alkalinity, temperature, specific conductance, pH, turbidity, DO, ORP) of field measurements taken as specified?	<u>Yes</u>	
6. Were wells categorized correctly?	<u>Yes</u>	
7. Were the following conditions met when purging a Category I well: Was one pump/tubing volume purged prior to sampling?	<u>Yes</u>	
Did the water level stabilize prior to sampling?	<u>Yes</u>	
Did pH, specific conductance, and turbidity measurements meet criteria prior to sampling?	<u>Yes</u>	
Was the flow rate less than 500 mL/min?	<u>Yes</u>	

Water Sampling Field Activities Verification Checklist (continued)

	Response (Yes, No, NA)	Comments
8. Were the following conditions met when purging a Category II well: Was the flow rate less than 500 mL/min? Was one pump/tubing volume removed prior to sampling?	NA	All wells were Category I.
9. Were duplicates taken at a frequency of one per 20 samples?	Yes	Duplicate samples were collected from locations 0300 and 0510.
10. Were equipment blanks taken at a frequency of one per 20 samples that were collected with non-dedicated equipment?	Yes	One equipment blank was collected.
11. Were trip blanks prepared and included with each shipment of VOC samples?	Yes	One trip blank was collected.
12. Were the true identities of the QC samples documented?	Yes	
13. Were samples collected in the containers specified?	Yes	
14. Were samples filtered and preserved as specified?	Yes	
15. Were the number and types of samples collected as specified?	Yes	
16. Were chain of custody records completed and was sample custody maintained?	Yes	
17. Was all pertinent information documented on the field data sheets?	Yes	
18. Was the presence or absence of ice in the cooler documented at every sample location?	Yes	Sample chilling was confirmed when required.
19. Were water levels measured at the locations specified in the planning documents?	Yes	

Laboratory Performance Assessment

General Information

Report Number (RIN): 14096456
 Sample Event: September 8–9, 2014
 Site(s): Slick Rock, Colorado, Processing Sites
 Laboratory: ALS Laboratory Group, Fort Collins, Colorado
 Work Order No.: 1409156
 Analysis: Metals, Organics, Wet Chemistry, and Radiochemistry
 Validator: Alison Kuhlman
 Review Date: November 13, 2014

This validation was performed according to the *Environmental Procedures Catalog* (LMS/POL/S04325, continually updated), “Standard Practice for Validation of Environmental Data.” The procedure was applied at Level 3, Data Validation. All analyses were successfully completed. The samples were prepared and analyzed using accepted procedures based on methods specified by line item code, which are listed in Table 1.

Table 1. Analytes and Methods

Analyte	Line Item Code	Prep Method	Analytical Method
Manganese	LMM-01	SW-846 3005A	SW-846 6010B
Molybdenum, Selenium, Uranium	LMM-02	SW-846 3005A	SW-846 6020A
Nitrite + Nitrate as N	WCH-A-022	MCAWW 353.2	MCAWW 353.2
Radium-226	ASP-A-016	SOP 783	SOP 783, EPA 903.1m
Radium-228	GPC-A-020	SOP 749	SOP 724
Volatile Organics	VOA-A-009	SW-846 5030C	SW-846 8260

Data Qualifier Summary

Analytical results were qualified as listed in Table 2. Refer to the sections below for an explanation of the data qualifiers applied.

Table 2. Data Qualifier Summary

Sample Number	Location	Analyte	Flag	Reason
1409156-7	0347	Manganese	J	Serial dilution percent difference criteria not met
1409156-14	0300 Duplicate	Manganese	J	Duplicate relative percent difference criteria not met
1409156-18	0300	Manganese	J	Duplicate relative percent difference criteria not met
1409156-17	Equipment Blank	Manganese	U	Less than 5 times the calibration blank
1409156-8	0349	Selenium	J	Serial dilution percent difference criteria not met
1409156-18	0300	Radium-226	J	Recoveries adjusted to meet criteria
All	All	Turbidity	J	Calibration check not met

Sample Shipping/Receiving

ALS Laboratory Group in Fort Collins, Colorado, received 27 water samples on September 11, 2014, accompanied by a Chain of Custody (COC) form. Copies of the three air bills were included in the receiving documentation. The COC form was checked to confirm that all of the samples were listed with sample collection dates and times, and that signatures and dates were present indicating sample relinquishment and receipt. The COC form was complete with no errors or omissions.

Preservation and Holding Times

The sample shipment was received intact with the temperature inside the iced cooler at 1.6 °C, which complies with requirements. The other two coolers were received at ambient temperature, which complies with requirements. All samples were received in the correct container types and had been preserved correctly. All samples were analyzed within the applicable holding times.

Detection and Quantitation Limits

The method detection limit (MDL) was reported for all metal, organic, and wet chemical analytes as required. The MDL, as defined in 40 CFR 136, is the minimum concentration of an analyte that can be measured and reported with 99 percent confidence that the analyte concentration is greater than zero. The practical quantitation limit (PQL) for these analytes is the lowest concentration that can be reliably measured, and is defined as 5 times the MDL.

For radiochemical analytes (those measured by radiometric counting) the MDL and PQL are not applicable, and these results are evaluated using the minimum detectable concentration (MDC), Decision Level Concentration (DLC), and Determination Limit (DL). The MDC is a measure of radiochemical method performance and was calculated and reported as specified in *Quality Systems for Analytical Services*. The DLC is the minimum concentration of an analyte that can be measured and reported with 99 percent confidence that the analyte concentration is greater than zero, and is estimated as 3 times the one-sigma total propagated uncertainty. Results that are greater than the MDC, but less than the DLC are qualified with a “U” flag (not detected). The DL for radiochemical results is the lowest concentration that can be reliably measured, and is defined as 3 times the MDC. Results not previously “U” qualified that are less than the DL are qualified with a “J” flag as estimated values.

The reported MDLs for all metal, organic, and wet chemical analytes, and MDCs for radiochemical analytes demonstrate compliance with contractual requirements.

Laboratory Instrument Calibration

Compliance requirements for satisfactory instrument calibration are established to ensure that the instrument is capable of producing acceptable qualitative and quantitative data for all analytes. Initial calibration demonstrates that the instrument is capable of acceptable performance in the beginning of the analytical run and of producing a linear curve. Compliance requirements for continuing calibration checks are established to ensure that the instrument continues to be capable of producing acceptable qualitative and quantitative data. All laboratory instrument

calibrations were performed correctly in accordance with the cited methods. All calibration and laboratory spike standards were prepared from independent sources.

Method MCAWW 353.2, Nitrate+Nitrite as N

Calibrations for nitrate + nitrite as N were performed using seven calibration standards on September 24, 2014. The calibration curve correlation coefficient values were greater than 0.995 and the absolute values of the intercepts were less than 3 times the MDL. Initial and continuing calibration verification checks were made at the required frequency with all calibration check results within the acceptance range.

Method SW-846 6010B, Manganese

Calibration for manganese was performed on September 25, 2014, using three calibration standards. The calibration curve correlation coefficient value was greater than 0.995 and the absolute value of the intercept was only slightly greater than 3 times the MDL. Initial and continuing calibration verification checks were made at the required frequency with all calibration checks meeting the acceptance criteria. Reporting limit verification checks were made at the required frequency to verify the linearity of the calibration curve near the PQL and all results were within the acceptance range.

Method SW-846 6020, Molybdenum, Selenium, Uranium

Calibrations were performed on September 24, 2014, using four calibration standards. The calibration curve correlation coefficient values were greater than 0.995 and the absolute values of the intercepts were less than 3 times the MDL. Initial and continuing calibration verification checks were made at the required frequency with all calibration checks meeting the acceptance criteria. Reporting limit verification checks were made at the required frequency to verify the linearity of the calibration curve near the PQL and all results were within the acceptance range. Mass calibration and resolution verifications were performed at the beginning of each analytical run in accordance with the analytical procedure. Internal standard recoveries associated with requested analytes were stable and within acceptable ranges.

Method SW-846 8260, Volatiles

The initial calibrations for benzene, ethylbenzene, toluene, and xylenes were performed using six calibration standards on July 26, 2014. Calibration curves are established using linear regression, quadratic regression, or the average response factor approach. Calibrations using average response factors had relative standard deviations of less than 15 percent. Initial and continuing calibration verification checks were made at the required frequency. The verification checks met all acceptance criteria. The mass spectrometer calibration and resolution were checked at the beginning of each analytical run in accordance with the procedure.

Radiochemical Analysis

Radium-226

Emanation cell plateau voltage determinations and cell efficiency calibrations were performed in October 2013. Daily instrument checks performed on September 29, 2014, met the acceptance criteria. All sample chemical recoveries were within the acceptance range of 40 to 110 percent. Chemical recoveries for several of the samples were adjusted by the laboratory to minimize possible low biases. The results for these samples are qualified with a “J” flag as estimated values.

Radium-228

Plateau voltage determinations were performed in October 2013 and detector efficiency calibrations were performed in May 2013. Background determinations were performed on September 24, 2014. The daily instrument checks performed on September 27, 2014, met the acceptance criteria. All sample chemical recoveries were within the acceptance range of 40 to 110 percent.

Method and Calibration Blanks

Method blanks are analyzed to assess any contamination that may have occurred during sample preparation. Calibration blanks are analyzed to assess instrument contamination prior to and during sample analysis.

Metals and Wet Chemistry

All method blank and calibration blank results associated with the samples were below the PQLs for all analytes. In cases where a blank concentration exceeds the MDL, the associated sample results are qualified with a “U” flag (not detected) when the sample result is greater than the MDL but less than 5 times the blank concentration. For manganese the method blank was negative and the absolute value was greater than the PQL. All associated manganese results were greater than 5 times the MDL, not requiring qualification.

Volatile Organics

The method blank results were below the MDLs for all target compounds.

Radiochemistry

The radiochemical method blank results were below the DLC.

Inductively Coupled Plasma (ICP) Interference Check Sample (ICS) Analysis

ICP interference check samples ICSA and ICSAB were analyzed at the required frequency to verify the instrumental interelement and background correction factors. All check sample results met the acceptance criteria.

Matrix Spike Analysis

Matrix spike and matrix spike duplicate (MS/MSD) pairs were analyzed for metals and nitrate + nitrite as N as a measure of method performance in the sample matrix. The MS/MSD data are not evaluated when the concentration of the unspiked sample is greater than 4 times the spike concentration. MS/MSD pairs were not analyzed for volatile organics, due to insufficient sample. The spike recoveries met the recovery and precision criteria for all analytes evaluated.

Laboratory Replicate Analysis

Laboratory replicate analyses are used to determine laboratory precision for each sample matrix. The relative percent difference for non-radiochemical replicate results that are greater than 5 times the PQL should be less than 20 percent (or less than the laboratory-derived control limits for organics). For results that are less than 5 times the PQL, the range should be no greater than

the PQL. The replicate results met these criteria. The relative error ratio for radiochemical replicate results (calculated using the one-sigma total propagated uncertainty) was less than 3, indicating acceptable precision.

Laboratory Control Sample

Laboratory control samples were analyzed at the correct frequency to provide information on the accuracy of the analytical method and the overall laboratory performance, including sample preparation. All control sample results were acceptable.

Metals Serial Dilution

Serial dilutions were prepared and analyzed for the metals analyses to monitor chemical or physical interferences in the sample matrix. Serial dilution data are evaluated when the concentration of the undiluted sample is greater than 50 times the MDL. All evaluated serial dilution data were acceptable, with the exception of manganese and selenium. The serial dilution percent difference for manganese and selenium were greater than 10 percent and the associated sample results are qualified with “J” flags as estimated values.

Volatile Organics Internal Standard and Surrogate Recovery

Laboratory performance for individual samples is evaluated by means of surrogate spikes. All samples are spiked with surrogate compounds prior to sample preparation. Surrogate recoveries are used to monitor factors such as interference and high concentrations of analytes. Surrogate recoveries may also be influenced by the success in recoveries of the internal standards. Internal standard recoveries were stable and within acceptance ranges. All surrogate recoveries were within the acceptance ranges.

Chromatography Peak Integration

The integration of analyte peaks was reviewed for all volatile organics data. All peak integrations were satisfactory.

Completeness

Results were reported in the correct units for all analytes requested using contract-required laboratory qualifiers. The analytical report included the MDL (MDC for radiochemistry) and PQL for all analytes and all required supporting documentation.

Electronic Data Deliverable (EDD) File

The EDD file arrived on October 1, 2014. The Sample Management System EDD validation module was used to verify that the EDD file was complete and in compliance with requirements. The module compares the contents of the file to the requested analyses to ensure all and only the requested data are delivered. The contents of the EDD were manually examined to verify that the sample results accurately reflect the data contained in the sample data package.

SAMPLE MANAGEMENT SYSTEM

General Data Validation Report

RIN: 14096456 Lab Code: PAR Validator: Alison Kuhlman Validation Date: 11/13/2014
Project: Slick Rock Analysis Type: Metals General Chem Rad Organics
of Samples: 27 Matrix: WATER Requested Analysis Completed: Yes

Chain of Custody

Present: OK Signed: OK Dated: OK

Sample

Integrity: OK Preservation: OK Temperature: OK

Select Quality Parameters

- Holding Times
- Detection Limits
- Field/Trip Blanks
- Field Duplicates

All analyses were completed within the applicable holding times.

The reported detection limits are equal to or below contract requirements.

There were 2 trip/equipment blanks evaluated.

There were 2 duplicates evaluated.

SAMPLE MANAGEMENT SYSTEM
Metals Data Validation Worksheet

RIN: 14096456 Lab Code: PAR Date Due: 10/9/2014
 Matrix: Water Site Code: SRK01 Date Completed: 10/2/2014

Analyte	Method Type	Date Analyzed	CALIBRATION				Method Blank	LCS %R	MS %R	MSD %R	Dup. RPD	ICSAB %R	Serial Dil. %R	CRI %R
			Int.	R^2	CCV	CCB								
Manganese	ICP/ES	09/25/2014	0.0000	1.0000	OK	0.496	-0.78	104.0	97.0	101.0	4.0	99.0	4.0	107.0
Manganese	ICP/ES	09/25/2014	0.0000	1.0000	OK	OK	-0.72	105.0	94.0	98.0	3.0	99.0	13.0	109.0
Molybdenum	ICP/MS	09/24/2014	0.0000	1.0000	OK	OK	OK	111.0		107.0	1.0		2.0	
Molybdenum	ICP/MS	09/24/2014	0.0000	1.0000	OK	OK	OK	106.0		110.0	1.0			
Selenium	ICP/MS	09/24/2014	0.0000	1.0000	OK	OK	OK	115.0		113.0	2.0		11.0	
Selenium	ICP/MS	09/24/2014	0.0000	1.0000	OK	OK	OK	105.0		115.0	16.0			
Uranium	ICP/MS	09/24/2014	0.0000	1.0000	OK	OK	OK	103.0		104.0	9.0		1.0	
Uranium	ICP/MS	09/24/2014	0.0000	1.0000	OK	OK	OK	96.0		101.0	3.0		3.0	

SAMPLE MANAGEMENT SYSTEM
Organics Data Validation Summary

RIN: 14096456

Project: Slick Rock

Lab Code: PAR

Validation Date: 11/17/2014

LCS Recovery: All LCS recoveries were within the laboratory acceptance limits.

Method Blank(s): All method blanks results were below the method detection limit.

MS/MSD Recovery: All MS/MSD recoveries were within the laboratory acceptance limits.

Surrogate Recovery: All surrogate recoveries were within the laboratory acceptance limits.

SAMPLE MANAGEMENT SYSTEM
Radiochemistry Data Validation Worksheet

RIN: 14096456 **Lab Code:** PAR **Date Due:** 10/9/2014
Matrix: Water **Site Code:** SRK01 **Date Completed:** 10/2/2014

Sample	Analyte	Date Analyzed	Result	Flag	Tracer %R	LCS %R	MS %R	Duplicate RER
0300	Radium-226	09/29/2014			94.2			
0319	Radium-226	09/29/2014			79.5			
2498	Radium-226	09/29/2014			95.5			
Blank_Spike	Radium-226	09/29/2014			99.8	101		
Blank_Spike_Du	Radium-226	09/29/2014			97.2	99		0.11
Blank	Radium-226	09/29/2014	0.0244	U	101			
0300	Radium-228	09/27/2014			99.1			
Blank	Radium-228	09/27/2014	0.3100	U	96.1			
Blank_Spike_Du	Radium-228	09/27/2014			97.9	94.1		0.99
Blank_Spike	Radium-228	09/27/2014			95.3	111		
2498	Radium-228	09/27/2014			90.8			
0319	Radium-228	09/27/2014			92.2			

SAMPLE MANAGEMENT SYSTEM
Wet Chemistry Data Validation Worksheet

RIN: 14096456 Lab Code: PAR Date Due: 10/9/2014
 Matrix: Water Site Code: SRK01 Date Completed: 10/2/2014

Analyte	Date Analyzed	CALIBRATION				Method Blank	LCS %R	MS %R	MSD %R	DUP RPD	Serial Dil. %R
		Int.	R^2	CCV	CCB						
Nitrate+Nitrite as N	09/24/2014	0.000	1.0000	OK	OK	OK	99.00	96.0	96.0	0	

Sampling Quality Control Assessment

The following information summarizes and assesses quality control for this sampling event.

Sampling Protocol

Surface water locations were sampled using a peristaltic pump and tubing reel or container immersion. Monitoring wells were sampled using a peristaltic pump and dedicated tubing. All monitoring wells met the Category I low-flow sampling criteria. Sample results for these wells were qualified with an “F” flag in the database, indicating the wells were purged and sampled using the low-flow sampling method. The daily calibration from the morning of October 9, 2014, and the post-trip calibration check from later that day, show the turbidimeter reading out of calibration most likely due to the rain. All associated turbidity results are qualified with a “J” flag as estimated values.

Equipment Blank

Equipment blanks are prepared and analyzed to document contamination attributable to the sample collection process. An equipment blank (field ID 2676) was taken from the tubing reel used to collect the surface water samples. This blank was filtered before being containerized and preserved according to analytical requirements. Manganese, molybdenum, and uranium were detected in the equipment blank. The manganese result has been previously qualified as non-detect. All other associated sample results that are greater than 5 times the equipment blank concentration require no further action.

Trip Blank Assessment

A trip blank (field ID 2500) was prepared and analyzed for volatile organics to document contamination attributable to shipping and field handling procedures. There were no target analytes detected in the trip blank.

Field Duplicate Assessment

Field duplicate samples are collected and analyzed as an indication of overall precision of the measurement process. The precision observed includes both field and laboratory precision and has more variability than laboratory duplicates, which measure only laboratory performance. The relative percent difference for duplicate results that are greater than 5 times the PQL should be more than 20 percent. For results that are less than 5 times the PQL, the range should be no greater than the PQL. Duplicate samples were collected from locations 0300 and 0510 (field duplicate IDs 2498 and 0510). The non-radiochemical duplicate results met the criteria, with the exception of selenium at location 0300. The associated samples are both qualified with “J” flags as estimated values. The relative error ratio for radiochemical duplicate results (calculated using the one-sigma total propagated uncertainty) was less than 3, indicating acceptable precision.

SAMPLE MANAGEMENT SYSTEM
Validation Report: Equipment/Trip Blanks

Page 1 of 1

RIN: 14096456 Lab Code: PAR Project: Slick Rock Validation Date: 11/13/2014

Blank Data

Blank Type	Lab Sample ID	Lab Method	Analyte Name	Result	Qualifier	MDL	Units
Equipment Blank	1409156-17	SW6010	Manganese	2.5	B	0.11	UG/L

Sample ID	Sample Ticket	Location	Result	Dilution Factor	Lab Qualifier	Validation Qualifier
1409156-12	MKT 201	0693	5.6	1		J
1409156-13	MKT 188	0694	3.5	1	B	J
1409156-7	MKT 186	0347	20	1	E	
1409156-8	MKT 187	0349	170	1		

Blank Data

Blank Type	Lab Sample ID	Lab Method	Analyte Name	Result	Qualifier	MDL	Units
Equipment Blank	1409156-17	SW6020	Molybdenum	0.059	B	0.032	UG/L

Sample ID	Sample Ticket	Location	Result	Dilution Factor	Lab Qualifier	Validation Qualifier
1409156-12	MKT 201	0693	4.9	5		
1409156-13	MKT 188	0694	0.4	1		
1409156-7	MKT 186	0347	7.9	5		
1409156-8	MKT 187	0349	19	5		

Blank Data

Blank Type	Lab Sample ID	Lab Method	Analyte Name	Result	Qualifier	MDL	Units
Equipment Blank	1409156-17	SW6020	Uranium	0.003	B	0.0029	UG/L

Sample ID	Sample Ticket	Location	Result	Dilution Factor	Lab Qualifier	Validation Qualifier
1409156-12	MKT 201	0693	2.1	5		
1409156-13	MKT 188	0694	0.39	1		
1409156-7	MKT 186	0347	2.8	5		
1409156-8	MKT 187	0349	7	5		

SAMPLE MANAGEMENT SYSTEM
Validation Report: Field Duplicates

RIN: 14096456 Lab Code: PAR Project: Slick Rock Validation Date: 11/13/2014

Duplicate: 2498

Sample: 0300

Analyte	Sample				Duplicate				RPD	RER	Units
	Result	Flag	Error	Dilution	Result	Flag	Error	Dilution			
Manganese	2300			1	2200			1	4.44		UG/L
Molybdenum	10			5	9.3			5	7.25		UG/L
Nitrate+Nitrite as N	0.017			1	0.014			1			MG/L
Radium-226	0.0823	U	0.105	1	0.0817	U	0.0884	1		0	pCi/L
Radium-228	0.623	U	0.431	1	0.604	U	0.475	1		0.1	pCi/L
Selenium	2.1			5	1.6			5	27.03		UG/L
Uranium	45			5	46			5	2.20		UG/L

Duplicate: 2533

Sample: 0510

Analyte	Sample				Duplicate				RPD	RER	Units
	Result	Flag	Error	Dilution	Result	Flag	Error	Dilution			
Manganese	3700			1	3700			1	0		UG/L
Molybdenum	890			100	1000			10	11.64		UG/L
Nitrate+Nitrite as N	180			200	190			200	5.41		MG/L
Selenium	870			100	970			10	10.87		UG/L
Uranium	100			100	100			10	0		UG/L

Certification

All laboratory analytical quality control criteria were met except as qualified in this report. The data qualifiers listed on the SEEPro database reports are defined on the last page of each report. All data in this package are considered validated and available for use.

Laboratory Coordinator: Stephen Donovan 12-8-2014
Stephen Donovan Date

Data Validation Lead: Alison Kuhlman 12/18/2014
Alison Kuhlman Date

Attachment 1
Assessment of Anomalous Data

This page intentionally left blank

Potential Outliers Report

This page intentionally left blank

Potential Outliers Report

Potential outliers are measurements that are extremely large or small relative to the rest of the data and, therefore, are suspected of misrepresenting the population from which they were collected. Potential outliers can result from transcription errors, data-coding errors, or measurement system problems. However, outliers can also represent true extreme values of a distribution and can indicate more variability in the population than was expected.

Statistical outlier tests give probabilistic evidence that an extreme value does not "fit" with the distribution of the remainder of the data and is therefore a statistical outlier. These tests should only be used to identify data points that require further investigation. The tests alone cannot determine whether a statistical outlier should be discarded or corrected within a data set.

There are three steps involved in identifying extreme values or outliers:

1. **Identify extreme values that may be potential outliers.** Do this by generating the Outliers Report using the Sample Management System from data in the environmental database. The application compares the new data set (in standard environmental database units) with historical data and lists the new data that fall outside the historical data range. A determination is also made as to whether the data are normally distributed using the Shapiro-Wilk Test.
2. **Apply the appropriate statistical test.** Dixon's Test for extreme values is used to test for statistical outliers when the sample size is less than or equal to 25. This test considers both extreme values that are much smaller than the rest of the data (case 1) and extreme values that are much larger than the rest of the data (case 2). This test is valid only if the data without the suspected outlier are normally distributed. Rosner's Test is a parametric test that is used to detect outliers for sample sizes of 25 or more. This test also assumes that the data without the suspected outliers are normally distributed.
3. **Scientifically review statistical outliers and decide on their disposition.** The review should include an evaluation of any notable trends in the data that may indicate the outliers represent true extreme values.

Three laboratory results were identified as potential outliers. There is no evidence that these data are the result of sampling or laboratory error and the data from this event are acceptable as qualified.

Potential anomalies in the field parameters were also examined for patterns of repeated high or low bias, which suggest a systematic error due to instrument malfunction. No such patterns were found in the field data from this event. One field parameter was identified as a potential outlier. There is no evidence that this data is the result of sampling error and the data from this event are acceptable as qualified.

Data Validation Outliers Report - No Field Parameters

Comparison: All Historical Data

Laboratory: ALS Laboratory Group

RIN: 14096456

Report Date: 11/17/2014

Site Code	Location Code	Sample ID	Sample Date	Analyte	Current			Historical Maximum			Historical Minimum			Number of Data Points		Statistical Outlier	
					Result	Qualifiers		Result	Qualifiers		Result	Qualifiers		N	N Below Detect		
						Lab	Data		Lab	Data		Lab	Data				
SRK05	0319	N001	09/09/2014	Toluene	370			13700			425		F	31	0	No	
SRK05	0320	N001	09/09/2014	Uranium	0.00910			0.0300		F	0.00920		F	21	0	No	
SRK05	0339	N001	09/09/2014	Molybdenum	0.920			1.30		F	0.950		F	6	0	No	
SRK05	0339	N001	09/09/2014	Selenium	2.80			2.10		F	1.80		F	6	0	Yes	
SRK05	0340	N001	09/09/2014	Manganese	4.40			6.00		F	5.10		F	5	0	No	
SRK05	0340	N001	09/09/2014	Nitrate + Nitrite as Nitrogen	210			400		F	290		F	5	0	No	
SRK05	0340	N001	09/09/2014	Uranium	0.0410			0.0530		F	0.0430		F	5	0	No	
SRK05	0349	0001	09/09/2014	Manganese	0.170			0.140			0.00220		B	17	0	No	
SRK05	0349	0001	09/09/2014	Molybdenum	0.0190			0.0110			0.00110			17	5	No	
SRK05	0349	0001	09/09/2014	Selenium	0.0220		E	0.01000			0.00028		B	17	3	NA	
SRK05	0349	0001	09/09/2014	Uranium	0.00700			0.00380			0.00049		B	17	1	No	
SRK05	0508	N001	09/09/2014	Manganese	2.60			7.49			2.70		F	35	0	No	
SRK05	0508	N001	09/09/2014	Uranium	0.0550			0.139		F	0.0580		F	35	0	NA	
SRK05	0694	0001	09/09/2014	Molybdenum	0.0004			0.1000		U	0.00094		B	44	25	NA	
SRK05	0694	0001	09/09/2014	Selenium	0.00018			0.0500		UI	0.0002		U	43	22	NA	
SRK06	0300	N001	09/08/2014	Molybdenum	0.01000			0.00980		F	0.00300		U	F	7	3	No
SRK06	0300	N001	09/08/2014	Selenium	0.00210			0.00110		F	0.0001		U	8	5	No	
SRK06	0300	N002	09/08/2014	Selenium	0.00160			0.00110		F	0.0001		U	8	5	No	

Data Validation Outliers Report - No Field Parameters

Comparison: All Historical Data

Laboratory: ALS Laboratory Group

RIN: 14096456

Report Date: 11/17/2014

Site Code	Location Code	Sample ID	Sample Date	Analyte	Current	Qualifiers		Historical Maximum			Historical Minimum			Number of Data Points		Statistical Outlier
					Result	Lab	Data	Result	Lab	Data	Result	Lab	Data	N	N Below Detect	
SRK06	0300	N001	09/08/2014	Uranium	0.0450			0.0200		F	0.00660		F	8	0	Yes
SRK06	0300	N002	09/08/2014	Uranium	0.0460			0.0200		F	0.00660		F	8	0	Yes
SRK06	0307	N001	09/08/2014	Selenium	0.000088	B		0.00290		F	0.0001	U	F	19	11	NA
SRK06	0309	N001	09/08/2014	Uranium	0.0360			0.340		F	0.0370		F	19	0	No

Data Validation Outliers Report - Field Parameters Only

Comparison: All Historical Data

Laboratory: Field Measurements

RIN: 14096456

Report Date: 11/17/2014

Site Code	Location Code	Sample ID	Sample Date	Analyte	Current	Qualifiers		Historical Maximum	Qualifiers		Historical Minimum	Qualifiers		Number of Data Points		Statistical Outlier
					Result	Lab	Data	Result	Lab	Data	Result	Lab	Data	N	N Below Detect	
SRK05	0319	N001	09/09/2014	Oxidation Reduction Potential	-100			-112			-274		F	19	0	NA
SRK05	0319	N001	09/09/2014	Specific Conductance	3299			12640			3661		F	19	0	No
SRK05	0349	N001	09/09/2014	pH	7.39			8.52			7.75			17	0	No
SRK05	0349	N001	09/09/2014	Specific Conductance	2310			1885			325			17	0	NA
SRK05	0508	N001	09/09/2014	Specific Conductance	3166			5710			3300			31	0	No
SRK05	0684	N001	09/09/2014	Alkalinity, Total (as CaCO ₃)	156			228		F	157			36	0	NA
SRK05	0694	N001	09/09/2014	Specific Conductance	107			1673			158			35	0	NA
SRK06	0300	N001	09/08/2014	Alkalinity, Total (as CaCO ₃)	769			760		F	506		F	11	0	NA
SRK06	0305	N001	09/08/2014	Alkalinity, Total (as CaCO ₃)	527			504		F	238		F	24	0	NA
SRK06	0309	N001	09/08/2014	Alkalinity, Total (as CaCO ₃)	622			1095			682		F	24	0	No
SRK06	0309	N001	09/08/2014	Specific Conductance	1813			6675		F	1859		F	19	0	No
SRK06	0696	N001	09/09/2014	pH	7.15			8.73			7.74			24	0	Yes

STATISTICAL TESTS:

The distribution of the data is tested for normality or lognormality using the Shapiro-Wilk Test

Outliers are identified using Dixon's Test when there are 25 or fewer data points.

Outliers are identified using Rosner's Test when there are 26 or more data points.

See Data Quality Assessment: Statistical Methods for Practitioners, EPA QC/G-9S, February 2006.

NA: Data are not normally or lognormally distributed.

Attachment 2

Data Presentation

This page intentionally left blank

Groundwater Quality Data

This page intentionally left blank

Groundwater Quality Data by Location (USEE100) FOR SITE SRK05, Slick Rock West Processing Site

REPORT DATE: 11/24/2014

Location: 0317 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)		Result	Qualifiers			Detection Limit	Uncertainty
							Lab	Data	QA		
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	N001	19.46	- 39.52	278		F	#		
Molybdenum	mg/L	09/09/2014	N001	19.46	- 39.52	0.18		F	#	0.00032	
Oxidation Reduction Potential	mV	09/09/2014	N001	19.46	- 39.52	177.1		F	#		
pH	s.u.	09/09/2014	N001	19.46	- 39.52	7.19		F	#		
Selenium	mg/L	09/09/2014	N001	19.46	- 39.52	0.0065		F	#	0.00032	
Specific Conductance	umhos/cm	09/09/2014	N001	19.46	- 39.52	2741		F	#		
Temperature	C	09/09/2014	N001	19.46	- 39.52	13.76		F	#		
Turbidity	NTU	09/09/2014	N001	19.46	- 39.52	2.62		FJ	#		

Groundwater Quality Data by Location (USEE100) FOR SITE SRK05, Slick Rock West Processing Site

REPORT DATE: 11/24/2014

Location: 0318A WELL Replacement well for 0318

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)			Result	Qualifiers			Detection Limit	Uncertainty
							Lab	Data	QA			
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	N001	9.2	-	14.2	261		F	#		
Manganese	mg/L	09/09/2014	N001	9.2	-	14.2	0.9		F	#	0.00011	
Molybdenum	mg/L	09/09/2014	N001	9.2	-	14.2	0.76		F	#	0.0032	
Nitrate + Nitrite as Nitrogen	mg/L	09/09/2014	N001	9.2	-	14.2	46		F	#	0.5	
Oxidation Reduction Potential	mV	09/09/2014	N001	9.2	-	14.2	159.9		F	#		
pH	s.u.	09/09/2014	N001	9.2	-	14.2	6.89		F	#		
Selenium	mg/L	09/09/2014	N001	9.2	-	14.2	4.7		F	#	0.0032	
Specific Conductance	umhos/cm	09/09/2014	N001	9.2	-	14.2	2006		F	#		
Temperature	C	09/09/2014	N001	9.2	-	14.2	17.44		F	#		
Turbidity	NTU	09/09/2014	N001	9.2	-	14.2	8.8		FJ	#		
Uranium	mg/L	09/09/2014	N001	9.2	-	14.2	0.029		F	#	0.00029	

Groundwater Quality Data by Location (USEE100) FOR SITE SRK05, Slick Rock West Processing Site

REPORT DATE: 11/24/2014

Location: 0319 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)		Result	Qualifiers Lab	Data	QA	Detection Limit	Uncertainty
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	N001	4.55	- 14.58	1042		F	#		
Benzene	ug/L	09/09/2014	N001	4.55	- 14.58	2200		F	#	15	
Ethylbenzene	ug/L	09/09/2014	N001	4.55	- 14.58	130		F	#	15	
m,p-Xylene	ug/L	09/09/2014	N001	4.55	- 14.58	3000		F	#	15	
o-Xylene	ug/L	09/09/2014	N001	4.55	- 14.58	670		F	#	15	
Oxidation Reduction Potential	mV	09/09/2014	N001	4.55	- 14.58	-100		F	#		
pH	s.u.	09/09/2014	N001	4.55	- 14.58	6.93		F	#		
Radium-226	pCi/L	09/09/2014	N001	4.55	- 14.58	2.02		F	#	0.18	0.62
Radium-228	pCi/L	09/09/2014	N001	4.55	- 14.58	1.42		F	#	0.7	0.586
Selenium	mg/L	09/09/2014	N001	4.55	- 14.58	0.0011		F	#	0.00016	
Specific Conductance	umhos/cm	09/09/2014	N001	4.55	- 14.58	3299		F	#		
Temperature	C	09/09/2014	N001	4.55	- 14.58	16.78		F	#		
Toluene	ug/L	09/09/2014	N001	4.55	- 14.58	370		F	#	15	
Turbidity	NTU	09/09/2014	N001	4.55	- 14.58	6.4		FJ	#		

Groundwater Quality Data by Location (USEE100) FOR SITE SRK05, Slick Rock West Processing Site

REPORT DATE: 11/24/2014

Location: 0320 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)			Result	Qualifiers			Detection Limit	Uncertainty
							Lab	Data	QA			
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	N001	4.92	-	9.96	352		F	#		
Manganese	mg/L	09/09/2014	N001	4.92	-	9.96	0.49		F	#	0.00011	
Molybdenum	mg/L	09/09/2014	N001	4.92	-	9.96	0.011		F	#	0.000032	
Nitrate + Nitrite as Nitrogen	mg/L	09/09/2014	N001	4.92	-	9.96	0.011		F	#	0.01	
Oxidation Reduction Potential	mV	09/09/2014	N001	4.92	-	9.96	-45.8		F	#		
pH	s.u.	09/09/2014	N001	4.92	-	9.96	7.06		F	#		
Selenium	mg/L	09/09/2014	N001	4.92	-	9.96	0.000052	B	F	#	0.000032	
Specific Conductance	umhos/cm	09/09/2014	N001	4.92	-	9.96	852		F	#		
Temperature	C	09/09/2014	N001	4.92	-	9.96	15.74		F	#		
Turbidity	NTU	09/09/2014	N001	4.92	-	9.96	9.36		FJ	#		
Uranium	mg/L	09/09/2014	N001	4.92	-	9.96	0.0091		F	#	0.0000029	

Groundwater Quality Data by Location (USEE100) FOR SITE SRK05, Slick Rock West Processing Site

REPORT DATE: 11/24/2014

Location: 0339 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)		Result	Qualifiers Lab	Data	QA	Detection Limit	Uncertainty
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	N001	11	- 14	282		F	#		
Manganese	mg/L	09/09/2014	N001	11	- 14	1.8		F	#	0.00011	
Molybdenum	mg/L	09/09/2014	N001	11	- 14	0.92		F	#	0.0032	
Nitrate + Nitrite as Nitrogen	mg/L	09/09/2014	N001	11	- 14	54		F	#	0.5	
Oxidation Reduction Potential	mV	09/09/2014	N001	11	- 14	161.3		F	#		
pH	s.u.	09/09/2014	N001	11	- 14	6.91		F	#		
Selenium	mg/L	09/09/2014	N001	11	- 14	2.8		F	#	0.0032	
Specific Conductance	umhos/cm	09/09/2014	N001	11	- 14	2017		F	#		
Temperature	C	09/09/2014	N001	11	- 14	16.36		F	#		
Turbidity	NTU	09/09/2014	N001	11	- 14	5.32		FJ	#		
Uranium	mg/L	09/09/2014	N001	11	- 14	0.033		F	#	0.00029	

Groundwater Quality Data by Location (USEE100) FOR SITE SRK05, Slick Rock West Processing Site

REPORT DATE: 11/24/2014

Location: 0340 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)			Result	Qualifiers			Detection Limit	Uncertainty
							Lab	Data	QA			
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	N001	6.51	-	11.51	286		F	#		
Manganese	mg/L	09/09/2014	N001	6.51	-	11.51	4.4		F	#	0.00011	
Molybdenum	mg/L	09/09/2014	N001	6.51	-	11.51	1.5		F	#	0.0032	
Nitrate + Nitrite as Nitrogen	mg/L	09/09/2014	N001	6.51	-	11.51	210		F	#	2	
Oxidation Reduction Potential	mV	09/09/2014	N001	6.51	-	11.51	164.5		F	#		
pH	s.u.	09/09/2014	N001	6.51	-	11.51	6.63		F	#		
Selenium	mg/L	09/09/2014	N001	6.51	-	11.51	2		F	#	0.0032	
Specific Conductance	umhos/cm	09/09/2014	N001	6.51	-	11.51	3537		F	#		
Temperature	C	09/09/2014	N001	6.51	-	11.51	17.92		F	#		
Turbidity	NTU	09/09/2014	N001	6.51	-	11.51	6.05		FJ	#		
Uranium	mg/L	09/09/2014	N001	6.51	-	11.51	0.041		F	#	0.00029	

Groundwater Quality Data by Location (USEE100) FOR SITE SRK05, Slick Rock West Processing Site

REPORT DATE: 11/24/2014

Location: 0508 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)	Result	Qualifiers Lab	Data	QA	Detection Limit	Uncertainty
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	N001	1.01 - 11.01	255		F	#		
Manganese	mg/L	09/09/2014	N001	1.01 - 11.01	2.6		F	#	0.00011	
Molybdenum	mg/L	09/09/2014	N001	1.01 - 11.01	1.1		F	#	0.0032	
Nitrate + Nitrite as Nitrogen	mg/L	09/09/2014	N001	1.01 - 11.01	140		F	#	2	
Oxidation Reduction Potential	mV	09/09/2014	N001	1.01 - 11.01	96		F	#		
pH	s.u.	09/09/2014	N001	1.01 - 11.01	6.69		F	#		
Selenium	mg/L	09/09/2014	N001	1.01 - 11.01	1.3		F	#	0.0032	
Specific Conductance	umhos/cm	09/09/2014	N001	1.01 - 11.01	3166		F	#		
Temperature	C	09/09/2014	N001	1.01 - 11.01	18.64		F	#		
Turbidity	NTU	09/09/2014	N001	1.01 - 11.01	3.56		FJ	#		
Uranium	mg/L	09/09/2014	N001	1.01 - 11.01	0.055		F	#	0.00029	

Groundwater Quality Data by Location (USEE100) FOR SITE SRK05, Slick Rock West Processing Site

REPORT DATE: 11/24/2014

Location: 0510 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)		Result	Qualifiers			Detection Limit	Uncertainty
							Lab	Data	QA		
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	N001	4.92	- 13.92	275		F	#		
Manganese	mg/L	09/09/2014	N001	4.92	- 13.92	3.7		F	#	0.00011	
Manganese	mg/L	09/09/2014	N002	4.92	- 13.92	3.7		F	#	0.00011	
Molybdenum	mg/L	09/09/2014	N001	4.92	- 13.92	0.89		F	#	0.0032	
Molybdenum	mg/L	09/09/2014	N002	4.92	- 13.92	1		F	#	0.00032	
Nitrate + Nitrite as Nitrogen	mg/L	09/09/2014	N001	4.92	- 13.92	180		F	#	2	
Nitrate + Nitrite as Nitrogen	mg/L	09/09/2014	N002	4.92	- 13.92	190		F	#	2	
Oxidation Reduction Potential	mV	09/09/2014	N001	4.92	- 13.92	187.4		F	#		
pH	s.u.	09/09/2014	N001	4.92	- 13.92	6.57		F	#		
Selenium	mg/L	09/09/2014	N001	4.92	- 13.92	0.87		F	#	0.0032	
Selenium	mg/L	09/09/2014	N002	4.92	- 13.92	0.97		F	#	0.00032	
Specific Conductance	umhos/cm	09/09/2014	N001	4.92	- 13.92	3673		F	#		
Temperature	C	09/09/2014	N001	4.92	- 13.92	17.09		F	#		
Turbidity	NTU	09/09/2014	N001	4.92	- 13.92	2.27		FJ	#		
Uranium	mg/L	09/09/2014	N001	4.92	- 13.92	0.1		F	#	0.00029	
Uranium	mg/L	09/09/2014	N002	4.92	- 13.92	0.1		F	#	0.000029	

Groundwater Quality Data by Location (USEE100) FOR SITE SRK05, Slick Rock West Processing Site

REPORT DATE: 11/24/2014

Location: 0684 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)		Result	Qualifiers			Detection Limit	Uncertainty
							Lab	Data	QA		
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	N001	11	- 21	156		F	#		
Manganese	mg/L	09/09/2014	N001	11	- 21	0.64		F	#	0.00011	
Molybdenum	mg/L	09/09/2014	N001	11	- 21	0.0065		F	#	0.000032	
Nitrate + Nitrite as Nitrogen	mg/L	09/09/2014	N001	11	- 21	0.01	U	F	#	0.01	
Oxidation Reduction Potential	mV	09/09/2014	N001	11	- 21	57.6		F	#		
pH	s.u.	09/09/2014	N001	11	- 21	7.22		F	#		
Selenium	mg/L	09/09/2014	N001	11	- 21	0.000066	B	F	#	0.000032	
Specific Conductance	umhos/cm	09/09/2014	N001	11	- 21	761		F	#		
Temperature	C	09/09/2014	N001	11	- 21	13.48		F	#		
Turbidity	NTU	09/09/2014	N001	11	- 21	5.11		FJ	#		
Uranium	mg/L	09/09/2014	N001	11	- 21	0.0087		F	#	0.000029	

Groundwater Quality Data by Location (USEE100) FOR SITE SRK06, Slick Rock East Processing Site

REPORT DATE: 11/24/2014

Location: 0300 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)		Result	Qualifiers			Detection Limit	Uncertainty
							Lab	Data	QA		
Alkalinity, Total (as CaCO ₃)	mg/L	09/08/2014	N001	9.5	- 19.5	769		F	#		
Manganese	mg/L	09/08/2014	N001	9.5	- 19.5	2.3		JF	#	0.00011	
Manganese	mg/L	09/08/2014	N002	9.5	- 19.5	2.2		JF	#	0.00011	
Molybdenum	mg/L	09/08/2014	N001	9.5	- 19.5	0.01		F	#	0.00016	
Molybdenum	mg/L	09/08/2014	N002	9.5	- 19.5	0.0093		F	#	0.00016	
Nitrate + Nitrite as Nitrogen	mg/L	09/08/2014	N001	9.5	- 19.5	0.017		F	#	0.01	
Nitrate + Nitrite as Nitrogen	mg/L	09/08/2014	N002	9.5	- 19.5	0.014		F	#	0.01	
Oxidation Reduction Potential	mV	09/08/2014	N001	9.5	- 19.5	-72.6		F	#		
pH	s.u.	09/08/2014	N001	9.5	- 19.5	6.75		F	#		
Radium-226	pCi/L	09/08/2014	N001	9.5	- 19.5	0.17	U	JF	#	0.17	0.105
Radium-226	pCi/L	09/08/2014	N002	9.5	- 19.5	0.13	U	F	#	0.13	0.0884
Radium-228	pCi/L	09/08/2014	N001	9.5	- 19.5	0.64	U	F	#	0.64	0.431
Radium-228	pCi/L	09/08/2014	N002	9.5	- 19.5	0.72	U	F	#	0.72	0.475
Selenium	mg/L	09/08/2014	N001	9.5	- 19.5	0.0021		F	#	0.00016	
Selenium	mg/L	09/08/2014	N002	9.5	- 19.5	0.0016		F	#	0.00016	
Specific Conductance	umhos /cm	09/08/2014	N001	9.5	- 19.5	11962		F	#		
Temperature	C	09/08/2014	N001	9.5	- 19.5	14.45		F	#		
Turbidity	NTU	09/08/2014	N001	9.5	- 19.5	3.72		FJ	#		
Uranium	mg/L	09/08/2014	N001	9.5	- 19.5	0.045		F	#	0.000015	
Uranium	mg/L	09/08/2014	N002	9.5	- 19.5	0.046		F	#	0.000015	

Groundwater Quality Data by Location (USEE100) FOR SITE SRK06, Slick Rock East Processing Site

REPORT DATE: 11/24/2014

Location: 0303 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)			Result	Qualifiers			Detection Limit	Uncertainty
							Lab	Data	QA			
Alkalinity, Total (as CaCO ₃)	mg/L	09/08/2014	N001	4.3	-	14.3	508		F	#		
Oxidation Reduction Potential	mV	09/08/2014	N001	4.3	-	14.3	-74		F	#		
pH	s.u.	09/08/2014	N001	4.3	-	14.3	7.18		F	#		
Specific Conductance	umhos/cm	09/08/2014	N001	4.3	-	14.3	3516		F	#		
Temperature	C	09/08/2014	N001	4.3	-	14.3	17.4		F	#		
Turbidity	NTU	09/08/2014	N001	4.3	-	14.3	8.3		FJ	#		
Uranium	mg/L	09/08/2014	N001	4.3	-	14.3	1.2		F	#	0.00029	

Groundwater Quality Data by Location (USEE100) FOR SITE SRK06, Slick Rock East Processing Site

REPORT DATE: 11/24/2014

Location: 0305 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)			Result	Qualifiers			Detection Limit	Uncertainty
							Lab	Data	QA			
Alkalinity, Total (as CaCO ₃)	mg/L	09/08/2014	N001	8.7	-	18.7	527		F	#		
Oxidation Reduction Potential	mV	09/08/2014	N001	8.7	-	18.7	77		F	#		
pH	s.u.	09/08/2014	N001	8.7	-	18.7	7.11		F	#		
Selenium	mg/L	09/08/2014	N001	8.7	-	18.7	0.024		F	#	0.0016	
Specific Conductance	umhos/cm	09/08/2014	N001	8.7	-	18.7	2874		F	#		
Temperature	C	09/08/2014	N001	8.7	-	18.7	15.98		F	#		
Turbidity	NTU	09/08/2014	N001	8.7	-	18.7	4.5		FJ	#		
Uranium	mg/L	09/08/2014	N001	8.7	-	18.7	0.72		F	#	0.00015	

Groundwater Quality Data by Location (USEE100) FOR SITE SRK06, Slick Rock East Processing Site

REPORT DATE: 11/24/2014

Location: 0307 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)			Result	Qualifiers			Detection Limit	Uncertainty
							Lab	Data	QA			
Alkalinity, Total (as CaCO ₃)	mg/L	09/08/2014	N001	4.4	-	14.4	686		F	#		
Oxidation Reduction Potential	mV	09/08/2014	N001	4.4	-	14.4	-51.5		F	#		
pH	s.u.	09/08/2014	N001	4.4	-	14.4	7.09		F	#		
Selenium	mg/L	09/08/2014	N001	4.4	-	14.4	0.000088	B	F	#	0.000032	
Specific Conductance	umhos/cm	09/08/2014	N001	4.4	-	14.4	5587		F	#		
Temperature	C	09/08/2014	N001	4.4	-	14.4	15.11		F	#		
Turbidity	NTU	09/08/2014	N001	4.4	-	14.4	9.67		FJ	#		
Uranium	mg/L	09/08/2014	N001	4.4	-	14.4	0.42		F	#	0.000015	

Groundwater Quality Data by Location (USEE100) FOR SITE SRK06, Slick Rock East Processing Site

REPORT DATE: 11/24/2014

Location: 0309 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)	Result	Qualifiers			Detection Limit	Uncertainty
						Lab	Data	QA		
Alkalinity, Total (as CaCO ₃)	mg/L	09/08/2014	N001	10.2 - 20.2	622		F	#		
Oxidation Reduction Potential	mV	09/08/2014	N001	10.2 - 20.2	-118		F	#		
pH	s.u.	09/08/2014	N001	10.2 - 20.2	7.57		F	#		
Specific Conductance	umhos/cm	09/08/2014	N001	10.2 - 20.2	1813		F	#		
Temperature	C	09/08/2014	N001	10.2 - 20.2	14.28		F	#		
Turbidity	NTU	09/08/2014	N001	10.2 - 20.2	9.62		FJ	#		
Uranium	mg/L	09/08/2014	N001	10.2 - 20.2	0.036		F	#	0.000029	

Groundwater Quality Data by Location (USEE100) FOR SITE SRK06, Slick Rock East Processing Site

REPORT DATE: 11/24/2014

Location: 0310 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)			Result	Qualifiers			Detection Limit	Uncertainty
								Lab	Data	QA		
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	N001	14.7	-	19.7	188		F	#		
Oxidation Reduction Potential	mV	09/09/2014	N001	14.7	-	19.7	-39.5		F	#		
pH	s.u.	09/09/2014	N001	14.7	-	19.7	7.31		F	#		
Specific Conductance	umhos/cm	09/09/2014	N001	14.7	-	19.7	949		F	#		
Temperature	C	09/09/2014	N001	14.7	-	19.7	14.97		F	#		
Turbidity	NTU	09/09/2014	N001	14.7	-	19.7	9.49		FJ	#		
Uranium	mg/L	09/09/2014	N001	14.7	-	19.7	0.021		F	#	0.000029	

Groundwater Quality Data by Location (USEE100) FOR SITE SRK06, Slick Rock East Processing Site

REPORT DATE: 11/24/2014

Location: 0311 WELL

Parameter	Units	Sample Date	ID	Depth Range (Ft BLS)		Result	Qualifiers			Detection Limit	Uncertainty
							Lab	Data	QA		
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	N001	14.1	- 19.1	288		F	#		
Oxidation Reduction Potential	mV	09/09/2014	N001	14.1	- 19.1	66.7		F	#		
pH	s.u.	09/09/2014	N001	14.1	- 19.1	6.88		F	#		
Specific Conductance	umhos /cm	09/09/2014	N001	14.1	- 19.1	1469		F	#		
Temperature	C	09/09/2014	N001	14.1	- 19.1	15.85		F	#		
Turbidity	NTU	09/09/2014	N001	14.1	- 19.1	8.25		FJ	#		
Uranium	mg/L	09/09/2014	N001	14.1	- 19.1	0.077		F	#	0.000029	

SAMPLE ID CODES: 000X = Filtered sample (0.45 µm). N00X = Unfiltered sample. X = replicate number.

LAB QUALIFIERS:

- * Replicate analysis not within control limits.
- > Result above upper detection limit.
- A TIC is a suspected aldol-condensation product.
- B Inorganic: Result is between the IDL and CRDL. Organic: Analyte also found in method blank.
- C Pesticide result confirmed by GC-MS.
- D Analyte determined in diluted sample.
- E Inorganic: Estimate value because of interference, see case narrative. Organic: Analyte exceeded calibration range of the GC-MS.
- H Holding time expired, value suspect.
- I Increased detection limit due to required dilution.
- J Estimated
- N Inorganic or radiochemical: Spike sample recovery not within control limits. Organic: Tentatively identified compound (TIC).
- P > 25% difference in detected pesticide or Aroclor concentrations between 2 columns.
- U Analytical result below detection limit.

DATA QUALIFIERS:

- F Low flow sampling method used.
- L Less than 3 bore volumes purged prior to sampling.
- U Parameter analyzed for but was not detected.
- G Possible grout contamination, pH > 9.
- Q Qualitative result due to sampling technique.
- X Location is undefined.
- J Estimated value.
- R Unusable result.

QA QUALIFIER:

- # Validated according to quality assurance guidelines.

Surface Water Quality Data

This page intentionally left blank

Surface Water Quality Data by Location (USEE102) FOR SITE SRK05, Slick Rock West Processing Site

REPORT DATE: 11/17/2014

Location: 0347 SURFACE LOCATION

Parameter	Units	Sample Date	ID	Result	Qualifiers Lab Data QA	Detection Limit	Uncertainty
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	N001	162		#	
Manganese	mg/L	09/09/2014	0001	0.02	E J	#	0.00011
Molybdenum	mg/L	09/09/2014	0001	0.0079		#	0.00016
Nitrate + Nitrite as Nitrogen	mg/L	09/09/2014	0001	0.22		#	0.01
Oxidation Reduction Potential	mV	09/09/2014	N001	124		#	
pH	s.u.	09/09/2014	N001	7.96		#	
Selenium	mg/L	09/09/2014	0001	0.0035		#	0.00016
Specific Conductance	umhos/cm	09/09/2014	N001	1108		#	
Temperature	C	09/09/2014	N001	21.83		#	
Turbidity	NTU	09/09/2014	N001	1000	>	#	
Uranium	mg/L	09/09/2014	0001	0.0028		#	0.000015

Surface Water Quality Data by Location (USEE102) FOR SITE SRK05, Slick Rock West Processing Site

REPORT DATE: 11/17/2014

Location: 0349 SURFACE LOCATION

Parameter	Units	Sample Date	ID	Result	Qualifiers Lab	Data	QA	Detection Limit	Uncertainty
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	N001	90			#		
Manganese	mg/L	09/09/2014	0001	0.17			#	0.00011	
Molybdenum	mg/L	09/09/2014	0001	0.019			#	0.00016	
Nitrate + Nitrite as Nitrogen	mg/L	09/09/2014	0001	0.52			#	0.01	
Oxidation Reduction Potential	mV	09/09/2014	N001	160.2			#		
pH	s.u.	09/09/2014	N001	7.39			#		
Selenium	mg/L	09/09/2014	0001	0.022	E	J	#	0.00016	
Specific Conductance	umhos/cm	09/09/2014	N001	2310			#		
Temperature	C	09/09/2014	N001	21.6			#		
Turbidity	NTU	09/09/2014	N001	1000	>		#		
Uranium	mg/L	09/09/2014	0001	0.007			#	0.000015	

Surface Water Quality Data by Location (USEE102) FOR SITE SRK05, Slick Rock West Processing Site

REPORT DATE: 11/17/2014

Location: 0693 SURFACE LOCATION

Parameter	Units	Sample Date	ID	Result	Qualifiers Lab	Data	QA	Detection Limit	Uncertainty
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	0001	60			#		
Manganese	mg/L	09/09/2014	0001	0.0056			#	0.00011	
Molybdenum	mg/L	09/09/2014	0001	0.0049			#	0.00016	
Nitrate + Nitrite as Nitrogen	mg/L	09/09/2014	0001	0.47			#	0.01	
Oxidation Reduction Potential	mV	09/09/2014	N001	173.1			#		
pH	s.u.	09/09/2014	N001	7.96			#		
Selenium	mg/L	09/09/2014	0001	0.0015			#	0.00016	
Specific Conductance	umhos/cm	09/09/2014	N001	730			#		
Temperature	C	09/09/2014	N001	17.05			#		
Turbidity	NTU	09/09/2014	N001	1000	>		#		
Uranium	mg/L	09/09/2014	0001	0.0021			#	0.000015	

Surface Water Quality Data by Location (USEE102) FOR SITE SRK05, Slick Rock West Processing Site

REPORT DATE: 11/17/2014

Location: 0694 SURFACE LOCATION

Parameter	Units	Sample Date	ID	Result	Qualifiers Lab Data QA	Detection Limit	Uncertainty
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	N001	58			#
Manganese	mg/L	09/09/2014	0001	0.0035	B		# 0.00011
Molybdenum	mg/L	09/09/2014	0001	0.0004			# 0.000032
Nitrate + Nitrite as Nitrogen	mg/L	09/09/2014	0001	0.066			# 0.01
Oxidation Reduction Potential	mV	09/09/2014	N001	37.4			#
pH	s.u.	09/09/2014	N001	8.33			#
Selenium	mg/L	09/09/2014	0001	0.00018			# 0.000032
Specific Conductance	umhos/cm	09/09/2014	N001	107			#
Temperature	C	09/09/2014	N001	17.83			#
Turbidity	NTU	09/09/2014	N001	1000	>		#
Uranium	mg/L	09/09/2014	0001	0.00039			# 0.0000029

Surface Water Quality Data by Location (USEE102) FOR SITE SRK06, Slick Rock East Processing Site

REPORT DATE: 11/17/2014

Location: 0692 SURFACE LOCATION

Parameter	Units	Sample		Result	Qualifiers			Detection Limit	Uncertainty
		Date	ID		Lab	Data	QA		
Alkalinity, Total (as CaCO ₃)	mg/L	09/08/2014	N001	110			#		
Oxidation Reduction Potential	mV	09/08/2014	N001	22.2			#		
pH	s.u.	09/08/2014	N001	8.35			#		
Specific Conductance	umhos/cm	09/08/2014	N001	335			#		
Temperature	C	09/08/2014	N001	19.94			#		
Turbidity	NTU	09/08/2014	N001	94.8			#		
Uranium	mg/L	09/08/2014	0001	0.00084			#	0.000029	

Surface Water Quality Data by Location (USEE102) FOR SITE SRK06, Slick Rock East Processing Site

REPORT DATE: 11/17/2014

Location: 0696 SURFACE LOCATION WQD, KNOWNS

Parameter	Units	Sample Date	ID	Result	Qualifiers			Detection Limit	Uncertainty
					Lab	Data	QA		
Alkalinity, Total (as CaCO ₃)	mg/L	09/09/2014	0001	76			#		
Oxidation Reduction Potential	mV	09/09/2014	N001	174			#		
pH	s.u.	09/09/2014	N001	7.15			#		
Specific Conductance	umhos/cm	09/09/2014	N001	842			#		
Temperature	C	09/09/2014	N001	17.77			#		
Turbidity	NTU	09/09/2014	N001	1000	>		#		
Uranium	mg/L	09/09/2014	0001	0.0024			#	0.000029	

Surface Water Quality Data by Location (USEE102) FOR SITE SRK06, Slick Rock East Processing Site

REPORT DATE: 11/17/2014

Location: 0700 SURFACE LOCATION

Parameter	Units	Sample Date	ID	Result	Qualifiers			Detection Limit	Uncertainty
					Lab	Data	QA		
Alkalinity, Total (as CaCO ₃)	mg/L	09/08/2014	N001	121			#		
Oxidation Reduction Potential	mV	09/08/2014	N001	-19.9			#		
pH	s.u.	09/08/2014	N001	8.1			#		
Specific Conductance	umhos/cm	09/08/2014	N001	339			#		
Temperature	C	09/08/2014	N001	20.6			#		
Turbidity	NTU	09/08/2014	N001	190			#		
Uranium	mg/L	09/08/2014	0001	0.00099			#	0.000029	

SAMPLE ID CODES: 000X = Filtered sample (0.45 µm). N00X = Unfiltered sample. X = replicate number.

LAB QUALIFIERS:

- * Replicate analysis not within control limits.
- > Result above upper detection limit.
- A TIC is a suspected aldol-condensation product.
- B Inorganic: Result is between the IDL and CRDL. Organic: Analyte also found in method blank.
- C Pesticide result confirmed by GC-MS.
- D Analyte determined in diluted sample.
- E Inorganic: Estimate value because of interference, see case narrative. Organic: Analyte exceeded calibration range of the GC-MS.
- H Holding time expired, value suspect.
- I Increased detection limit due to required dilution.
- J Estimated
- N Inorganic or radiochemical: Spike sample recovery not within control limits. Organic: Tentatively identified compound (TIC).
- P > 25% difference in detected pesticide or Aroclor concentrations between 2 columns.
- U Analytical result below detection limit.

DATA QUALIFIERS:

- F Low flow sampling method used.
- L Less than 3 bore volumes purged prior to sampling.
- U Parameter analyzed for but was not detected.
- G Possible grout contamination, pH > 9.
- Q Qualitative result due to sampling technique.
- X Location is undefined.
- J Estimated value.
- R Unusable result.

QA QUALIFIER:

- # Validated according to quality assurance guidelines.

This page intentionally left blank

Equipment Blank and Trip Blank Data

This page intentionally left blank

BLANKS REPORT

LAB: PARAGON/ALS LABORATORY GROUP (Fort Collins, CO)

RIN: 14096456

Report Date: 11/17/2014

Parameter	Site Code	Location ID	Sample Date	ID	Units	Result	Qualifiers Lab	Data	Detection Limit	Uncertainty	Sample Type
Benzene	SRK05	0999	09/09/2014	N001	ug/L	0.3	U		0.3		TB
Ethylbenzene	SRK05	0999	09/09/2014	N001	ug/L	0.3	U		0.3		TB
m,p-Xylene	SRK05	0999	09/09/2014	N001	ug/L	0.3	U		0.3		TB
Manganese	SRK05	0999	09/09/2014	N002	mg/L	0.0025	B	U	0.00011		E
Molybdenum	SRK05	0999	09/09/2014	N002	mg/L	0.000059	B		0.000032		E
Nitrate + Nitrite as Nitrogen	SRK05	0999	09/09/2014	N002	mg/L	0.01	U		0.01		E
o-Xylene	SRK05	0999	09/09/2014	N001	ug/L	0.3	U		0.3		TB
Selenium	SRK05	0999	09/09/2014	N002	mg/L	0.000032	U		0.000032		E
Toluene	SRK05	0999	09/09/2014	N001	ug/L	0.3	U		0.3		TB
Uranium	SRK05	0999	09/09/2014	N002	mg/L	0.000003	B		0.000029		E

SAMPLE ID CODES: 000X = Filtered sample (0.45 µm). N00X = Unfiltered sample. X = replicate number.

LAB QUALIFIERS:

- * Replicate analysis not within control limits.
- > Result above upper detection limit.
- A TIC is a suspected aldol-condensation product.
- B Inorganic: Result is between the IDL and CRDL. Organic: Analyte also found in method blank.
- C Pesticide result confirmed by GC-MS.
- D Analyte determined in diluted sample.
- E Inorganic: Estimate value because of interference, see case narrative. Organic: Analyte exceeded calibration range of the GC-MS.
- H Holding time expired, value suspect.
- I Increased detection limit due to required dilution.
- J Estimated
- N Inorganic or radiochemical: Spike sample recovery not within control limits. Organic: Tentatively identified compound (TIC).
- P > 25% difference in detected pesticide or Aroclor concentrations between 2 columns.
- U Analytical result below detection limit.

DATA QUALIFIERS:

- | | | | | | |
|---|--|---|---|---|------------------|
| F | Low flow sampling method used. | G | Possible grout contamination, pH > 9. | J | Estimated value. |
| L | Less than 3 bore volumes purged prior to sampling. | Q | Qualitative result due to sampling technique. | R | Unusable result. |
| U | Parameter analyzed for but was not detected. | X | Location is undefined. | | |

SAMPLE TYPES:

- | | |
|----|------------------|
| E | Equipment Blank. |
| TB | Trip Blank |

Static Water Level Data

This page intentionally left blank

STATIC WATER LEVELS (USEE700) FOR SITE SRK06, Slick Rock East Processing Site
REPORT DATE: 11/17/2014

Location Code	Flow Code	Top of Casing Elevation (Ft)	Measurement Date	Time	Depth From Top of Casing (Ft)	Water Elevation (Ft)	Water Level Flag
0300	U	5467.35	09/08/2014	17:00:38	14.31	5453.04	
0303	O	5446.91	09/08/2014	14:10:57	10.21	5436.7	
0305	O	5448.75	09/08/2014	14:45:33	13.04	5435.71	
0307	O	5447.1	09/08/2014	15:15:35	11.83	5435.27	
0309	O	5450.18	09/08/2014	15:50:20	15.58	5434.6	
0310	D	5450.56	09/09/2014	16:35:47	17.78	5432.78	
0311	D	5450.7	09/09/2014	17:10:39	18.68	5432.02	
0312	D	5451.06	09/09/2014	17:32:00			D

STATIC WATER LEVELS (USEE700) FOR SITE SRK05, Slick Rock West Processing Site
REPORT DATE: 11/17/2014

Location Code	Flow Code	Top of Casing Elevation (Ft)	Measurement Date	Time	Depth From Top of Casing (Ft)	Water Elevation (Ft)	Water Level Flag
0317		5435.18	09/09/2014	15:00:55	12.1	5423.08	
0318A		5435.77	09/09/2014	14:30:10	12.67	5423.1	
0319	O	5430.66	09/09/2014	11:35:14	9.09	5421.57	
0320	O	5427.4	09/09/2014	12:30:17	5.85	5421.55	
0339		5434.47	09/09/2014	14:15:30	11.49	5422.98	
0340		5433.09	09/09/2014	13:50:54	10.05	5423.04	
0508	O	5430.2	09/09/2014	13:00:17	7.1	5423.1	
0510	O	5427.87	09/09/2014	15:40:03	5.73	5422.14	
0684	D	5432.68	09/09/2014	10:55:34	16.55	5416.13	


FLOW CODES: B BACKGROUND C CROSS GRADIENT D DOWNGRADIENT F OFFSITE
 N UNKNOWN O ONSITE U UPGRADIENT

WATER LEVEL FLAGS: D Dry F Flowing B Below top of pump


Hydrographs

This page intentionally left blank

Slick Rock West Processing Site Hydrograph



Slick Rock East Processing Site Hydrograph




Groundwater Time-Concentration Graphs


This page intentionally left blank

Slick Rock West Processing Site Manganese Concentration


Maximum Background Concentration (MBC) = 4.2 mg/L




Slick Rock West Processing Site
Molybdenum Concentration
 Maximum Concentration Limit (MCL) = 0.10 mg/L




Slick Rock West Processing Site
Nitrate + Nitrite as Nitrogen Concentration
 Maximum Concentration Limit (MCL) = 10 mg/L




Slick Rock West Processing Site
Radium-226 and Radium-228 Concentrations in Well 0319
 Maximum Concentration Limit (MCL) = 5 pCi/L for Ra-226+228




**Slick Rock West Processing Site
Selenium Concentration**
Alternate Concentration Limit (ACL) = 0.18 mg/L




**Slick Rock East Processing Site
Selenium Concentration**
Maximum Concentration Limit (MCL) = 0.01 mg/L



Slick Rock West Processing Site
Uranium Concentration
 Maximimum Concentration Limit (MCL) = 0.044 mg/L




**Slick Rock East Processing Site
Uranium Concentration**
Maximum Concentration Limit (MCL) = 0.044 mg/L




Surface Water Time-Concentration Graphs

This page intentionally left blank


Slick Rock West Processing Site
Manganese Concentration
Background Threshold Value = 0.024 mg/L



Slick Rock West Processing Site
Molybdenum Concentration
Background Treshold Value = 0.008 mg/L




Slick Rock West Processing Site
Nitrate + Nitrite as Nitrogen Concentration
Background Threshold Value = 0.47 mg/L




Slick Rock West Processing Site Selenium Concentration


Background Threshold Value = 0.0047 mg/L



Slick Rock West Processing Site
Uranium Concentration
Background Threshold Value = 0.0043 mg/L



**Slick Rock East Processing Site
Uranium Concentration**
Background Threshold Value = 0.0023 mg/L




Dolores River Discharge Graph

This page intentionally left blank

Discharge, cubic feet per second

Most recent instantaneous value: 25 12-08-2014 13:00 MST

USGS 09168730 DOLORES RIVER NEAR SLICK ROCK, CO



△ Median daily statistic (8 years) — Period of approved data
— Discharge

This page intentionally left blank

Attachment 3
Sampling and Analysis Work Order

This page intentionally left blank



August 5, 2014

Task Assignment 501
Control Number 14-0788

U.S. Department of Energy
Office of Legacy Management
ATTN: Jason Nguyen
Site Manager
2597 Legacy Way
Grand Junction, CO 81503

SUBJECT: Contract No. DE-LM0000415, The S.M. Stoller Corporation, a wholly owned subsidiary of Huntington Ingalls Industries (Stoller)
Task Assignment 501 LTS&M 1
September 2014 Environmental Sampling at the Slick Rock, Colorado, Processing Sites

REFERENCE: Task Assignment 501-02-120-402, Slick Rock, Colorado, Processing Sites

Dear Mr. Nguyen:

The purpose of this letter is to inform you of the upcoming sampling event at Slick Rock, Colorado. Enclosed are the map and tables specifying sample locations and analytes for monitoring at the Slick Rock processing sites. Water quality data will be collected at this site as part of the routine environmental sampling currently scheduled to begin the week of September 8, 2014.

The following lists show the monitoring wells (along with associated zone of completion) scheduled for sampling during this event.

MONITORING WELLS

West Site

317 Je 319 Al 320 Al 339 Al 340 Al 508 Al 510 Al 684 Al
318A Al

East Site

300 Al 303 Al 305 Al 307 Al 309 Al 310 Al 311 Al 312 Al

*NOTE: Al = Alluvium; Je = Jurassic Entrada Sandstone

SURFACE WATER

West Site

347 349 693 694

A SUBSIDIARY OF HUNTINGTON INGALLS INDUSTRIES

2597 Legacy Way • Grand Junction, CO 81503-1789 • Telephone (970) 248-6000 • Fax (970) 248-6040

Jason Nguyen
Control Number 14-0788
Page 2

East Site

692 696 700

All samples will be collected as directed in the *Sampling and Analysis Plan for U.S. Department of Energy Office of Legacy Management Sites*. Access agreements are being reviewed and are expected to be complete by the beginning of fieldwork.

Please contact me at (970) 248-6557 if you have any questions.

Sincerely,



David Traub
Site Lead

DT/lcg/lb

Enclosures (3)

cc: (electronic)

Christina Pennal, DOE
Steve Donovan, Stoller
Bev Gallagher, Stoller
Lauren Goodknight, Stoller
David Traub, Stoller
EDD Delivery
rc-grand.junction
File: SRE 410.02(A)
 SRW 410.02(A)

Sampling Frequencies for Locations at Slick Rock, Colorado

Location ID	Quarterly	Semiannually	Annually	Biennially	Not Sampled	Notes
Monitoring Wells						
WEST						
317			X			
318A			X			
319			X			
320			X			
339			X			
340			X			
508			X			
510			X			
684			X			
EAST						
300			X			
303			X			
305			X			
307			X			
309			X			
310			X			
311			X			
312			X			
Surface Locations						
WEST						
347			X			
349			X			
693			X			
694			X			
EAST						
692			X			
696			X			
700			X			

Sampling conducted in September

Constituent Sampling Breakdown

Site	Slick Rock		Required Detection Limit (mg/L)	Analytical Method	Line Item Code
Analyte	Groundwater	Surface Water			
Approx. No. Samples/yr	14	7			
<i>Field Measurements</i>					
Alkalinity	X	X			
Dissolved Oxygen					
Redox Potential	X	X			
pH	X	X			
Specific Conductance	X	X			
Turbidity	X	X			
Temperature	X	X			
<i>Laboratory Measurements</i>					
Aluminum					
Ammonia as N (NH ₃ -N)					
Calcium					
Iron					
Lead					
Magnesium					
Manganese	0300, 0318A, 0320, 0339, 0340, 0508, 0510, 0684	0347, 0349, 0693, 0694	0.005	SW-846 6010	LMM-01
Molybdenum	0300, 0317, 0318A, 0320, 0339, 0340, 0508, 0510, 0684	0347, 0349, 0693, 0694	0.003	SW-846 6020	LMM-02
Nickel					
Nickel-63					
Nitrate + Nitrite as N (NO ₃ +NO ₂)-N	0300, 0318A, 0320, 0339, 0340, 0508, 0510, 0684	0347, 0349, 0693, 0694	0.05	EPA 353.1	WCH-A-022
Potassium					
Radium-226	0300, '0319		1 pCi/L	Gas Proportional Counter	GPC-A-018
Radium-228	0300, '0319		1 pCi/L	Gas Proportional Counter	GPC-A-020
Selenium	0300, 0305, 0307, 0317, 0318A, 0319, 0320, 0339, 0340, 0508, 0510, 0684	0347, 0349, 0693, 0694	0.0001	SW-846 6020	LMM-02
Total Dissolved Solids					
Total Organic Carbon					
Uranium	0300, 0303, 0305, 0307, 0309, 0310, 0311, 0312,	X	0.0001	SW-846 6020	LMM-02

Constituent Sampling Breakdown

Site	Slick Rock		Required Detection Limit (mg/L)	Analytical Method	Line Item Code
Analyte	Groundwater	Surface Water			
	0318A, 0320, 0339, 0340, 0508, 0510, 0684				
Vanadium					
VOCs (BETX)	0319 only		0.005	SW-846 8260	VOA-A- 009
Zinc					
Total No. of Analytes	8	5			

Notes: All private well samples are to be unfiltered. The total number of analytes does not include field parameters.

This page intentionally left blank

Attachment 4 Trip Report

This page intentionally left blank



Memorandum

DATE: September 25, 2014

TO: David Traub

FROM: Dan Sellers

SUBJECT: Trip Report

Site: Slick Rock, Colorado, West and East Processing Sites

Dates of Sampling Event: September 8 and 9, 2013

Team Members: Alison Kuhlman and Dan Sellers

Number of Locations Sampled: Samples were collected from 23 of the 24 locations identified on the sampling notification letter as follows:

SRK05 (West Site)–9 monitoring wells and 4 surface locations.

SRK06 (East Site)–7 monitoring wells and 3 surface locations.

Locations Not Sampled/Reason: Well 0312 (East Site) was dry.

Location Specific Information:

Location IDs	Comments
West Site: 0319	VOCs were collected by reverse flow as follows: <i>After purging and collecting non-VOC samples, tubing was pulled from the well with volume of water retained in it; 40 mL vials were filled by reversing the flow on the pump.</i>
River Surface Locations	All surface locations were filtered. West Site; Surface locations along the Dolores River were collected during and after heavy rain storm. Surface location 0349 was sampled upstream from Map ID location due to accessibility issues (heavy vegetation) and heavy rain.

Quality Control Sample Cross Reference: The following are the false identifications assigned to the quality control samples.

False ID	True ID	Sample Type	Ticket Number	Associated Matrix
2498	0300	Duplicate (Radium, Metals and Nitrate)	MKT 191	Water
2500	-----	Trip Blank (VOCs))	MKT 194	Water
2533	0510	Duplicate (Metals and Nitrates)	MKT 200	Water
2676	Associated with all surface water locations	Equipment Blank (Metals and Nitrate)	MKT 195	Water

Requisition Identification Number (RIN) Assigned: 14096456. Field data sheets can be found in Crow\sms\14096456 in the Field Data folder.

Sample Shipment: Samples were shipped from Grand Junction to ALS Laboratory Group on September 10, 2014.

Water Level Measurements: Water levels were measured in all sampled wells.

Well Inspection Summary: No issues were identified.

Field Variance: None.

Equipment: Turbidity meter was out of calibration on the morning of 09/09/2014 but was fixed by cleaning light and drying meter out – it then calibrated. In the evening, at 1720, a post calibration of meter passed on the high standard but failed on the low side of the low and medium standards (>-10%). All other equipment functioned properly. Wells were sampled with a peristaltic pump and dedicated tubing. Surface water samples were collected using a peristaltic pump and tubing with an attached weight or by container immersion and then filtered. An equipment blank was collected after decontamination of the tubing reel.

Stakeholder/Regulatory: Nothing to note.

Institutional Controls:

Fences, Gates, and Locks: All gates were locked and in good condition.

Signs: OK.

Trespassing/Site Disturbances: None noted.

Site Issues:

Disposal Cell/Drainage Structure Integrity: N/A

Vegetation/Noxious Weed Concerns: Very difficult to access several surface locations along the Dolores River due to heavy vegetation (willows).

Maintenance Requirements: The roads to the sample location on the East Site have eroded from previous rainstorms and need to be reconstructed.

Safety Issues: N/A

Access Issues: Because of recent heavy rains at the West Site, access was difficult to and from the following sample locations: Wells locations 0309, 0310, 0311, 0312, and surface location 0700. An ATV was used to transport sampling equipment to wells 0303, 0305, 0307, and 0309, and surface locations 0692 and 0700.

Corrective Action Required/Taken: See Maintenance Requirements above.

cc: (electronic)
Jason Nguyen, DOE
Steve Donovan, Stoller
Dave Traub, Stoller
EDD Delivery