Pages have been deleted from this laboratory report file to reduce file size. The deleted pages contain raw data and instrument calibrations. If the full laboratory report is needed, contact Jack.Craig.@lm.doe.gov.

ANALYTICAL REPORT

Job Number: 280-1272-1

SDG Number: 10022892

Job Description: PINELLAS MONITORING

For:
S.M. Stoller Corporation
2597 B 3/4 Road
Grand Junction, CO 81503

Attention: Mr. Steve Donivan

Approved for release.
Kae E Yoder
Project Manager II
4/1/2010 12:52 PM

Kae E Yoder
Project Manager II
kae.yoder@testamericainc.com
04/01/2010

The test results in this report relate only to the samples in this report and meet all requirements of NELAC, with any exceptions noted. Pursuant to NELAP, this report shall not be reproduced except in full, without the written approval of the laboratory. All questions regarding this report should be directed to the TestAmerica Denver Project Manager.

The Lab Certification ID# is E87667.

Reporting limits are adjusted for sample size used, dilutions and moisture content if applicable.

Table of Contents

Cover Title Page 1

Data Summaries 5

Report Narrative 5

Data Qualifiers 6

Executive Summary 7

Method Summary 8

Method / Analyst Summary 9

Sample Summary 10

Sample Datasheets 11

Surrogate Summary 19

QC Data Summary 21

QC Association Summary 33

Lab Chronicle 35

Manual Integration Summary 38

Organic Sample Data 44

GC/MS VOA 44

Method 8260B 44

Method 8260B QC Summary 45

Method 8260B Sample Data 55

Standards Data 75

Method 8260B ICAL Data 75

Method 8260B CCAL Data 141

Raw QC Data 163

Method 8260B Tune Data 163

Method 8260B Blank Data 175

Method 8260B LCS/LCSD Data 182

Table of Contents

Method 8260B MS/MSD Data 190

Method 8260B Run Logs 198

Method 8260B SIM 212

Method 8260B SIM QC Summary 213

Method 8260B SIM Sample Data 222

Standards Data 228

Method 8260B SIM ICAL Data 228

Method 8260B SIM CCAL Data 244

Raw QC Data 250

Method 8260B SIM Tune Data 250

Method 8260B SIM Blank Data 262

Method 8260B SIM LCS/LCSD Data 265

Method 8260B SIM MS/MSD Data 268

Method 8260B SIM Run Logs 274

Inorganic Sample Data 277

Metals Data 277

Met Cover Page 278

Met Sample Data 279

Met QC Data 281

Met ICV/CCV 281

Met CRQL 286

Met Blanks 287

Met ICSA/ICSAB 291

Met MS/MSD/PDS 295

Met LCS/LCSD 297

Met Serial Dilution 299

Table of Contents

Met MDL 300
Met IECF 301
Met Linear Ranges 309
Met Preparation Log 310
Met Analysis Run Log 311
Met Raw Data 318
Met Prep Data 1422
Shipping and Receiving Documents 1423
Client Chain of Custody 1424
Sample Receipt Checklist 1425

CASE NARRATIVE

Client: S.M. Stoller Corporation

Project: PINELLAS MONITORING - 10022892

Report Number: 280-1272-1

With exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. All laboratory quality control samples analyzed in conjunction with the samples in this project were within established control limits, with any exceptions noted. Calculations are performed before rounding to avoid round-off errors in calculated results.

This report includes reporting limits (RLs) less than TestAmerica Denver's practical quantitation limits. These reporting limits are being used specifically at the client's request to meet the needs of this project. Please note that data are not normally reported to these levels without qualification, since they are inherently less reliable and potentially less defensible than required by the current NELAC standards.

Results between the method detection limit (MDL) and reporting limit (RL) are flagged with a "J" qualifier to indicate an estimated value. These results are statistically less reliable than results greater than or equal to the RL and should be considered a qualitative value.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

RECEIPT

The samples were received on 3/11/2010; the samples arrived in good condition, properly preserved and on ice. The temperature of the cooler at receipt was 0.7 C.

The chain-of-custody was not relinquished by TestAmerica's Tampa laboratory, the forwarding laboratory. The client was notified on 3/12/2010.

GC/MS VOLATILES - SW846 8260B

Acetone, a common laboratory contaminant, was detected in the method blank associated with batch 280-7794 at a level that was above the method detection limit but below the reporting limit. The value should be considered an estimate, and has been flagged "J". If the associated samples reported a result above the MDL and/or RL, the result has been "B" flagged.

Due to insufficient sample volume submitted by the client, Pinellas sample specific MS and MSD analysis could not be performed. Standard batch MS and MSD data have been provided.

No other anomalies were encountered.

GC/MS VOLATILES - SW846 8260B-SIM

No anomalies were encountered.

TOTAL METALS - SW846 6010B

The MDL date is incorrect on the CLP forms. The correct MDL date is 1/29/2010.

No other anomalies were encountered.

DATA REPORTING QUALIFIERS

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Lab Section	Qualifier	Description
GC/MS VOA		
	B	Compound was found in the blank and sample.
	U	Indicates the analyte was analyzed for but not detected.
	J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
Metals		
	U	Indicates the analyte was analyzed for but not detected.
	J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1272-1	**PIN99-2892**					
Acetone		4.1	J B	10	ug/L	8260B
Methylene Chloride		1.1		1.0	ug/L	8260B
Tetrachloroethene		0.29	J	1.0	ug/L	8260B
Toluene		0.41	J	1.0	ug/L	8260B
280-1272-2	**PIN99-2893**					
Acetone		3.3	J B	10	ug/L	8260B
Methylene Chloride		1.1		1.0	ug/L	8260B
Tetrachloroethene		0.30	J	1.0	ug/L	8260B
Toluene		0.39	J	1.0	ug/L	8260B
Aluminum		31	J	100	ug/L	6010B
Iron		24	J	100	ug/L	6010B

METHOD SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Description	Lab Location	Method	Preparation Method
Matrix: Water			
Volatile Organic Compounds (GC/MS)	TAL DEN	SW846 8260B	
Purge and Trap	TAL DEN		SW846 5030B
Volatile Organic Compounds (GC/MS-SIM)	TAL DEN	SW846 8260B SIM	
Purge and Trap	TAL DEN		SW846 5030B
Metals (ICP)	TAL DEN	SW846 6010B	
Preparation, Total Metals	TAL DEN		SW846 3010A

Lab References:

TAL DEN = TestAmerica Denver

Method References:

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

METHOD / ANALYST SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Method	Analyst	Analyst ID
SW846 8260B	Waterland, Hayley E	HEW
SW846 8260B SIM	Wolfe, Ashley E	AEW
SW846 6010B	Trudell, Lynn-Anne	LT

SAMPLE SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
280-1272-1	PIN99-2892	Water	03/09/2010 1425	03/11/2010 1000
280-1272-1MS	PIN99-2892	Water	03/09/2010 1425	03/11/2010 1000
280-1272-1MSD	PIN99-2892	Water	03/09/2010 1425	03/11/2010 1000
280-1272-2	PIN99-2893	Water	03/09/2010 1410	03/11/2010 1000

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2892**

Lab Sample ID: 280-1272-1
Client Matrix: Water

Date Sampled: 03/09/2010 1425
Date Received: 03/11/2010 1000

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7794	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7342.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/16/2010 1840		Final Weight/Volume:	20 mL
Date Prepared:	03/16/2010 1840			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	4.1	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	1.1		0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2892**

Lab Sample ID: 280-1272-1
Client Matrix: Water

Date Sampled: 03/09/2010 1425
Date Received: 03/11/2010 1000

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7794	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7342.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/16/2010 1840		Final Weight/Volume:	20 mL
Date Prepared:	03/16/2010 1840			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.29	J	0.20	1.0
Toluene	0.41	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	100		70 - 127
Toluene-d8 (Surr)	103		80 - 125
4-Bromofluorobenzene (Surr)	99		78 - 118
Dibromofluoromethane (Surr)	110		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2893**

Lab Sample ID: 280-1272-2
Client Matrix: Water

Date Sampled: 03/09/2010 1410
Date Received: 03/11/2010 1000

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7794	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7343.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/16/2010 1859		Final Weight/Volume:	20 mL
Date Prepared:	03/16/2010 1859			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	3.3	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	1.1		0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2893**

Lab Sample ID: 280-1272-2
Client Matrix: Water

Date Sampled: 03/09/2010 1410
Date Received: 03/11/2010 1000

8260B Volatile Organic Compounds (GC/MS)

Method: 8260B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 03/16/2010 1859
Date Prepared: 03/16/2010 1859

Analysis Batch: 280-7794

Instrument ID: MSV_R1
Lab File ID: R7343.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.30	J	0.20	1.0
Toluene	0.39	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	100		70 - 127
Toluene-d8 (Surr)	103		80 - 125
4-Bromofluorobenzene (Surr)	100		78 - 118
Dibromofluoromethane (Surr)	110		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2892**

Lab Sample ID: 280-1272-1
Client Matrix: Water

Date Sampled: 03/09/2010 1425
Date Received: 03/11/2010 1000

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-7423	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33916.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/15/2010 1241		Final Weight/Volume:	20 mL
Date Prepared:	03/15/2010 1241			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2893**

Lab Sample ID: 280-1272-2
Client Matrix: Water

Date Sampled: 03/09/2010 1410
Date Received: 03/11/2010 1000

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-7423	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33915.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/15/2010 1216		Final Weight/Volume:	20 mL
Date Prepared:	03/15/2010 1216			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	106		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2892**

Lab Sample ID: 280-1272-1
Client Matrix: Water

Date Sampled: 03/09/2010 1425
Date Received: 03/11/2010 1000

6010B Metals (ICP)

Method:	6010B	Analysis Batch: 280-7274		Instrument ID:	MT_025
Preparation:	3010A	Prep Batch: 280-7097		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/12/2010 2338			Final Weight/Volume:	50 mL
Date Prepared:	03/12/2010 0730				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Aluminum	18	U	18	100

Method:	6010B	Analysis Batch: 280-7325		Instrument ID:	MT_025
Preparation:	3010A	Prep Batch: 280-7097		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/13/2010 1335			Final Weight/Volume:	50 mL
Date Prepared:	03/12/2010 0730				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Iron	22	U	22	100

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2893**

Lab Sample ID: 280-1272-2
Client Matrix: Water

Date Sampled: 03/09/2010 1410
Date Received: 03/11/2010 1000

6010B Metals (ICP)

Method:	6010B	Analysis Batch: 280-7274		Instrument ID:	MT_025
Preparation:	3010A	Prep Batch: 280-7097		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/12/2010 2347			Final Weight/Volume:	50 mL
Date Prepared:	03/12/2010 0730				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Aluminum	31	J	18	100

Method:	6010B	Analysis Batch: 280-7325		Instrument ID:	MT_025
Preparation:	3010A	Prep Batch: 280-7097		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/13/2010 1344			Final Weight/Volume:	50 mL
Date Prepared:	03/12/2010 0730				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Iron	24	J	22	100

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B Volatile Organic Compounds (GC/MS)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DBFM %Rec	DCA %Rec	TOL %Rec	BFB %Rec
280-1272-1	PIN99-2892	110	100	103	99
280-1272-2	PIN99-2893	110	100	103	100
MB 280-7794/6		109	98	105	100
LCS 280-7794/4		106	92	105	100
LCSD 280-7794/5		108	96	105	102
280-1175-C-1 MS		109	96	105	99
280-1175-C-1 MSD		108	95	105	99

Surrogate	Acceptance Limits
DBFM = Dibromofluoromethane (Surr)	77-119
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127
TOL = Toluene-d8 (Surr)	80-125
BFB = 4-Bromofluorobenzene (Surr)	78-118

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DCA %Rec
280-1272-1	PIN99-2892	101
280-1272-2	PIN99-2893	106
MB 280-7423/4		97
LCS 280-7423/3		101
280-1272-1 MS	PIN99-2892 MS	103
280-1272-1 MSD	PIN99-2892 MSD	98

Surrogate	Acceptance Limits
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Method Blank - Batch: 280-7794

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-7794/6	Analysis Batch:	280-7794	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7320.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/16/2010 1019			Final Weight/Volume:	20 mL
Date Prepared:	03/16/2010 1019				

Analyte	Result	Qual	MDL	RL
Acetone	1.98	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Method Blank - Batch: 280-7794

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-7794/6	Analysis Batch:	280-7794	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7320.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/16/2010 1019			Final Weight/Volume:	20 mL
Date Prepared:	03/16/2010 1019				

Analyte	Result	Qual	MDL	RL
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	98	70 - 127
Toluene-d8 (Surr)	105	80 - 125
4-Bromofluorobenzene (Surr)	100	78 - 118
Dibromofluoromethane (Surr)	109	77 - 119

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Lab Control Sample/
Lab Control Sample Duplicate Recovery Report - Batch: 280-7794

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-7794/4	Analysis Batch:	280-7794	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7318.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/16/2010 0939			Final Weight/Volume:	20 mL
Date Prepared:	03/16/2010 0939				

LCSD Lab Sample ID:	LCSD 280-7794/5	Analysis Batch:	280-7794	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7319.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/16/2010 0959			Final Weight/Volume:	20 mL
Date Prepared:	03/16/2010 0959				

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	90	90	77 - 120	0	20		
Bromodichloromethane	85	88	78 - 120	4	20		
Carbon tetrachloride	92	90	80 - 120	3	21		
Chlorobenzene	88	88	78 - 120	1	20		
Chloroform	90	90	78 - 120	0	20		
1,3-Dichlorobenzene	88	87	75 - 120	1	20		
1,1-Dichloroethane	91	91	77 - 120	0	21		
trans-1,2-Dichloroethene	93	93	80 - 120	0	24		
1,1-Dichloroethene	98	95	68 - 133	3	20		
1,2-Dichloropropane	89	90	76 - 120	2	20		
Ethylbenzene	90	87	78 - 120	4	26		
Methylene Chloride	90	94	71 - 120	4	20		
Tetrachloroethene	89	86	77 - 120	4	20		
Toluene	87	87	73 - 120	0	20		
1,1,1-Trichloroethane	90	89	78 - 120	2	20		
Trichloroethene	91	90	78 - 122	1	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	92	96	70 - 127
Toluene-d8 (Surr)	105	105	80 - 125
4-Bromofluorobenzene (Surr)	100	102	78 - 118
Dibromofluoromethane (Surr)	106	108	77 - 119

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Laboratory Control/
Laboratory Duplicate Data Report - Batch: 280-7794

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-7794/4	Units:	ug/L	LCSD Lab Sample ID:	LCSD 280-7794/5
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/16/2010 0939			Date Analyzed:	03/16/2010 0959
Date Prepared:	03/16/2010 0939			Date Prepared:	03/16/2010 0959

Analyte	LCS Spike Amount	LCSD Spike Amount	LCS Result/Qual	LCSD Result/Qual
Benzene	5.00	5.00	4.51	4.49
Bromodichloromethane	5.00	5.00	4.24	4.42
Carbon tetrachloride	5.00	5.00	4.62	4.50
Chlorobenzene	5.00	5.00	4.38	4.41
Chloroform	5.00	5.00	4.51	4.52
1,3-Dichlorobenzene	5.00	5.00	4.39	4.35
1,1-Dichloroethane	5.00	5.00	4.53	4.55
trans-1,2-Dichloroethene	5.00	5.00	4.65	4.63
1,1-Dichloroethene	5.00	5.00	4.92	4.77
1,2-Dichloropropane	5.00	5.00	4.44	4.52
Ethylbenzene	5.00	5.00	4.52	4.35
Methylene Chloride	5.00	5.00	4.50	4.68
Tetrachloroethene	5.00	5.00	4.47	4.31
Toluene	5.00	5.00	4.37	4.37
1,1,1-Trichloroethane	5.00	5.00	4.50	4.43
Trichloroethene	5.00	5.00	4.54	4.48

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-7794

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1175-C-1 MS	Analysis Batch:	280-7794	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7322.D
Dilution:	1.0			Initial Weight/Volume:	0.04 mL
Date Analyzed:	03/16/2010 1204			Final Weight/Volume:	20 mL
Date Prepared:	03/16/2010 1204				

MSD Lab Sample ID:	280-1175-C-1 MSD	Analysis Batch:	280-7794	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7323.D
Dilution:	1.0			Initial Weight/Volume:	0.04 mL
Date Analyzed:	03/16/2010 1224			Final Weight/Volume:	20 mL
Date Prepared:	03/16/2010 1224				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	91	92	77 - 120	1	20		
Bromodichloromethane	89	90	78 - 120	2	20		
Carbon tetrachloride	92	87	80 - 120	5	21		
Chlorobenzene	91	91	78 - 120	0	20		
Chloroform	94	93	78 - 120	2	20		
1,3-Dichlorobenzene	89	90	75 - 120	1	20		
1,1-Dichloroethane	92	91	77 - 120	1	21		
trans-1,2-Dichloroethene	92	91	80 - 120	0	24		
1,1-Dichloroethene	98	92	68 - 133	6	20		
1,2-Dichloropropane	90	90	76 - 120	0	20		
Ethylbenzene	90	90	78 - 120	0	26		
Methylene Chloride	95	94	71 - 120	1	20		
Tetrachloroethene	89	84	77 - 120	6	20		
Toluene	88	89	73 - 120	1	20		
1,1,1-Trichloroethane	91	87	78 - 120	5	20		
Trichloroethene	91	89	78 - 122	1	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	96	95	70 - 127
Toluene-d8 (Surr)	105	105	80 - 125
4-Bromofluorobenzene (Surr)	99	99	78 - 118
Dibromofluoromethane (Surr)	109	108	77 - 119

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-7794

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1175-C-1 MS	Units: ug/L	MSD Lab Sample ID:	280-1175-C-1 MSD
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/16/2010 1204		Date Analyzed:	03/16/2010 1224
Date Prepared:	03/16/2010 1204		Date Prepared:	03/16/2010 1224

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	570		2500	2500	2850	2870
Bromodichloromethane	85	U	2500	2500	2220	2260
Carbon tetrachloride	95	U	2500	2500	2290	2180
Chlorobenzene	85	U	2500	2500	2280	2270
Chloroform	80	U	2500	2500	2360	2310
1,3-Dichlorobenzene	65	U	2500	2500	2220	2250
1,1-Dichloroethane	130	J	2500	2500	2440	2410
trans-1,2-Dichloroethene	75	U	2500	2500	2290	2280
1,1-Dichloroethene	120	U	2500	2500	2460	2310
1,2-Dichloropropane	90	U	2500	2500	2250	2260
Ethylbenzene	100	J	2500	2500	2350	2350
Methylene Chloride	160	U	2500	2500	2380	2340
Tetrachloroethene	100	U	2500	2500	2220	2100
Toluene	460	J	2500	2500	2660	2680
1,1,1-Trichloroethane	80	U	2500	2500	2280	2180
Trichloroethene	80	U	2500	2500	2270	2240

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Method Blank - Batch: 280-7423

Method: 8260B SIM
Preparation: 5030B

Lab Sample ID:	MB 280-7423/4	Analysis Batch:	280-7423	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_33914.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/15/2010 1150			Final Weight/Volume:	20 mL
Date Prepared:	03/15/2010 1150				

Analyte	Result	Qual	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	97	70 - 127

Lab Control Sample - Batch: 280-7423

Method: 8260B SIM
Preparation: 5030B

Lab Sample ID:	LCS 280-7423/3	Analysis Batch:	280-7423	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_33909.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/15/2010 0934			Final Weight/Volume:	20 mL
Date Prepared:	03/15/2010 0934				

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
1,4-Dioxane	5.00	5.68	114	25 - 141	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101	70 - 127

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-7423

Method: 8260B SIM
Preparation: 5030B

MS Lab Sample ID:	280-1272-1	Analysis Batch:	280-7423	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_33917.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/15/2010 1307			Final Weight/Volume:	20 mL
Date Prepared:	03/15/2010 1307				

MSD Lab Sample ID:	280-1272-1	Analysis Batch:	280-7423	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_33918.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/15/2010 1332			Final Weight/Volume:	20 mL
Date Prepared:	03/15/2010 1332				

Analyte	% Rec. MS	% Rec. MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
1,4-Dioxane	101	92	25 - 141	10	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103	98	70 - 127

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-7423

Method: 8260B SIM
Preparation: 5030B

MS Lab Sample ID:	280-1272-1	Units:	ug/L	MSD Lab Sample ID:	280-1272-1
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/15/2010 1307			Date Analyzed:	03/15/2010 1332
Date Prepared:	03/15/2010 1307			Date Prepared:	03/15/2010 1332

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
1,4-Dioxane	0.75	U	5.00	5.00	5.07	4.59

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Method Blank - Batch: 280-7097

Method: 6010B
Preparation: 3010A

Lab Sample ID:	MB 280-7097/1-A	Analysis Batch:	280-7274	Instrument ID:	MT_025
Client Matrix:	Water	Prep Batch:	280-7097	Lab File ID:	N/A
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	50 mL
Date Analyzed:	03/12/2010 2333			Final Weight/Volume:	50 mL
Date Prepared:	03/12/2010 0730				

Analyte	Result	Qual	MDL	RL
Aluminum	18	U	18	100

Method Blank - Batch: 280-7097

Method: 6010B
Preparation: 3010A

Lab Sample ID:	MB 280-7097/1-A	Analysis Batch:	280-7325	Instrument ID:	MT_025
Client Matrix:	Water	Prep Batch:	280-7097	Lab File ID:	N/A
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	50 mL
Date Analyzed:	03/13/2010 1330			Final Weight/Volume:	50 mL
Date Prepared:	03/12/2010 0730				

Analyte	Result	Qual	MDL	RL
Iron	22	U	22	100

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Lab Control Sample - Batch: 280-7097

Method: 6010B
Preparation: 3010A

Lab Sample ID:	LCS 280-7097/2-A	Analysis Batch:	280-7274	Instrument ID:	MT_025
Client Matrix:	Water	Prep Batch:	280-7097	Lab File ID:	N/A
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	50 mL
Date Analyzed:	03/12/2010 2336			Final Weight/Volume:	50 mL
Date Prepared:	03/12/2010 0730				

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Aluminum	2000	1920	96	87 - 111	

Lab Control Sample - Batch: 280-7097

Method: 6010B
Preparation: 3010A

Lab Sample ID:	LCS 280-7097/2-A	Analysis Batch:	280-7325	Instrument ID:	MT_025
Client Matrix:	Water	Prep Batch:	280-7097	Lab File ID:	N/A
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	50 mL
Date Analyzed:	03/13/2010 1333			Final Weight/Volume:	50 mL
Date Prepared:	03/12/2010 0730				

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Iron	1000	1010	101	89 - 115	

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-7097

Method: 6010B
Preparation: 3010A

MS Lab Sample ID:	280-1272-1	Analysis Batch:	280-7274	Instrument ID:	MT_025
Client Matrix:	Water	Prep Batch:	280-7097	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/12/2010 2343			Final Weight/Volume:	50 mL
Date Prepared:	03/12/2010 0730				

MSD Lab Sample ID:	280-1272-1	Analysis Batch:	280-7274	Instrument ID:	MT_025
Client Matrix:	Water	Prep Batch:	280-7097	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/12/2010 2345			Final Weight/Volume:	50 mL
Date Prepared:	03/12/2010 0730				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Aluminum	97	98	83 - 119	1	25		

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-7097

Method: 6010B
Preparation: 3010A

MS Lab Sample ID:	280-1272-1	Analysis Batch:	280-7325	Instrument ID:	MT_025
Client Matrix:	Water	Prep Batch:	280-7097	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/13/2010 1340			Final Weight/Volume:	50 mL
Date Prepared:	03/12/2010 0730				

MSD Lab Sample ID:	280-1272-1	Analysis Batch:	280-7325	Instrument ID:	MT_025
Client Matrix:	Water	Prep Batch:	280-7097	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/13/2010 1342			Final Weight/Volume:	50 mL
Date Prepared:	03/12/2010 0730				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Iron	100	102	52 - 155	2	25		

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-7097

Method: 6010B
Preparation: 3010A

MS Lab Sample ID:	280-1272-1	Units: ug/L	MSD Lab Sample ID:	280-1272-1
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/12/2010 2343		Date Analyzed:	03/12/2010 2345
Date Prepared:	03/12/2010 0730		Date Prepared:	03/12/2010 0730

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Aluminum	18	U	2000	2000	1930	1950

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-7097

Method: 6010B
Preparation: 3010A

MS Lab Sample ID:	280-1272-1	Units: ug/L	MSD Lab Sample ID:	280-1272-1
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/13/2010 1340		Date Analyzed:	03/13/2010 1342
Date Prepared:	03/12/2010 0730		Date Prepared:	03/12/2010 0730

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Iron	22	U	1000	1000	1000	1020

Calculations are performed before rounding to avoid round-off errors in calculated results.

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:280-7423					
LCS 280-7423/3	Lab Control Sample	T	Water	8260B SIM	
MB 280-7423/4	Method Blank	T	Water	8260B SIM	
280-1272-1	PIN99-2892	T	Water	8260B SIM	
280-1272-1MS	Matrix Spike	T	Water	8260B SIM	
280-1272-1MSD	Matrix Spike Duplicate	T	Water	8260B SIM	
280-1272-2	PIN99-2893	T	Water	8260B SIM	
Analysis Batch:280-7794					
LCS 280-7794/4	Lab Control Sample	T	Water	8260B	
LCSD 280-7794/5	Lab Control Sample Duplicate	T	Water	8260B	
MB 280-7794/6	Method Blank	T	Water	8260B	
280-1175-C-1 MS	Matrix Spike	T	Water	8260B	
280-1175-C-1 MSD	Matrix Spike Duplicate	T	Water	8260B	
280-1272-1	PIN99-2892	T	Water	8260B	
280-1272-2	PIN99-2893	T	Water	8260B	

Report Basis
T = Total

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
Metals					
Prep Batch: 280-7097					
LCS 280-7097/2-A	Lab Control Sample	T	Water	3010A	
MB 280-7097/1-A	Method Blank	T	Water	3010A	
280-1272-1	PIN99-2892	T	Water	3010A	
280-1272-1MS	Matrix Spike	T	Water	3010A	
280-1272-1MSD	Matrix Spike Duplicate	T	Water	3010A	
280-1272-2	PIN99-2893	T	Water	3010A	
Analysis Batch:280-7274					
LCS 280-7097/2-A	Lab Control Sample	T	Water	6010B	280-7097
MB 280-7097/1-A	Method Blank	T	Water	6010B	280-7097
280-1272-1	PIN99-2892	T	Water	6010B	280-7097
280-1272-1MS	Matrix Spike	T	Water	6010B	280-7097
280-1272-1MSD	Matrix Spike Duplicate	T	Water	6010B	280-7097
280-1272-2	PIN99-2893	T	Water	6010B	280-7097
Analysis Batch:280-7325					
LCS 280-7097/2-A	Lab Control Sample	T	Water	6010B	280-7097
MB 280-7097/1-A	Method Blank	T	Water	6010B	280-7097
280-1272-1	PIN99-2892	T	Water	6010B	280-7097
280-1272-1MS	Matrix Spike	T	Water	6010B	280-7097
280-1272-1MSD	Matrix Spike Duplicate	T	Water	6010B	280-7097
280-1272-2	PIN99-2893	T	Water	6010B	280-7097

Report Basis
T = Total

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
SDG: 10022892

Laboratory Chronicle

Lab ID: **280-1272-1** **Client ID:** **PIN99-2892**

Sample Date/Time: 03/09/2010 14:25 Received Date/Time: 03/11/2010 10:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1272-E-1		280-7794		03/16/2010 18:40	1	TAL DEN	HEW
A:8260B	280-1272-E-1		280-7794		03/16/2010 18:40	1	TAL DEN	HEW
P:5030B	280-1272-C-1		280-7423		03/15/2010 12:41	1	TAL DEN	AEW
A:8260B SIM	280-1272-C-1		280-7423		03/15/2010 12:41	1	TAL DEN	AEW
P:3010A	280-1272-A-1-A		280-7274	280-7097	03/12/2010 07:30	1	TAL DEN	JW
A:6010B	280-1272-A-1-A		280-7274	280-7097	03/12/2010 23:38	1	TAL DEN	LT
P:3010A	280-1272-A-1-A		280-7325	280-7097	03/12/2010 07:30	1	TAL DEN	JW
A:6010B	280-1272-A-1-A		280-7325	280-7097	03/13/2010 13:35	1	TAL DEN	LT

Lab ID: **280-1272-1 MS** **Client ID:** **PIN99-2892**

Sample Date/Time: 03/09/2010 14:25 Received Date/Time: 03/11/2010 10:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1272-D-1 MS		280-7423		03/15/2010 13:07	1	TAL DEN	AEW
A:8260B SIM	280-1272-D-1 MS		280-7423		03/15/2010 13:07	1	TAL DEN	AEW
P:3010A	280-1272-A-1-B MS		280-7274	280-7097	03/12/2010 07:30	1	TAL DEN	JW
A:6010B	280-1272-A-1-B MS		280-7274	280-7097	03/12/2010 23:43	1	TAL DEN	LT
P:3010A	280-1272-A-1-B MS		280-7325	280-7097	03/12/2010 07:30	1	TAL DEN	JW
A:6010B	280-1272-A-1-B MS		280-7325	280-7097	03/13/2010 13:40	1	TAL DEN	LT

Lab ID: **280-1272-1 MSD** **Client ID:** **PIN99-2892**

Sample Date/Time: 03/09/2010 14:25 Received Date/Time: 03/11/2010 10:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1272-D-1 MSD		280-7423		03/15/2010 13:32	1	TAL DEN	AEW
A:8260B SIM	280-1272-D-1 MSD		280-7423		03/15/2010 13:32	1	TAL DEN	AEW
P:3010A	280-1272-A-1-C MSD		280-7274	280-7097	03/12/2010 07:30	1	TAL DEN	JW
A:6010B	280-1272-A-1-C MSD		280-7274	280-7097	03/12/2010 23:45	1	TAL DEN	LT
P:3010A	280-1272-A-1-C MSD		280-7325	280-7097	03/12/2010 07:30	1	TAL DEN	JW
A:6010B	280-1272-A-1-C MSD		280-7325	280-7097	03/13/2010 13:42	1	TAL DEN	LT

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
SDG: 10022892

Laboratory Chronicle

Lab ID: **280-1272-2** **Client ID:** **PIN99-2893**

Sample Date/Time: 03/09/2010 14:10 Received Date/Time: 03/11/2010 10:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1272-D-2		280-7794		03/16/2010 18:59	1	TAL DEN	HEW
A:8260B	280-1272-D-2		280-7794		03/16/2010 18:59	1	TAL DEN	HEW
P:5030B	280-1272-E-2		280-7423		03/15/2010 12:16	1	TAL DEN	AEW
A:8260B SIM	280-1272-E-2		280-7423		03/15/2010 12:16	1	TAL DEN	AEW
P:3010A	280-1272-A-2-A		280-7274	280-7097	03/12/2010 07:30	1	TAL DEN	JW
A:6010B	280-1272-A-2-A		280-7274	280-7097	03/12/2010 23:47	1	TAL DEN	LT
P:3010A	280-1272-A-2-A		280-7325	280-7097	03/12/2010 07:30	1	TAL DEN	JW
A:6010B	280-1272-A-2-A		280-7325	280-7097	03/13/2010 13:44	1	TAL DEN	LT

Lab ID: **MB** **Client ID:** **N/A**

Sample Date/Time: N/A Received Date/Time: N/A

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	MB 280-7794/6		280-7794		03/16/2010 10:19	1	TAL DEN	HEW
A:8260B	MB 280-7794/6		280-7794		03/16/2010 10:19	1	TAL DEN	HEW
P:5030B	MB 280-7423/4		280-7423		03/15/2010 11:50	1	TAL DEN	AEW
A:8260B SIM	MB 280-7423/4		280-7423		03/15/2010 11:50	1	TAL DEN	AEW
P:3010A	MB 280-7097/1-A		280-7274	280-7097	03/12/2010 07:30	1	TAL DEN	JW
A:6010B	MB 280-7097/1-A		280-7274	280-7097	03/12/2010 23:33	1	TAL DEN	LT
P:3010A	MB 280-7097/1-A		280-7325	280-7097	03/12/2010 07:30	1	TAL DEN	JW
A:6010B	MB 280-7097/1-A		280-7325	280-7097	03/13/2010 13:30	1	TAL DEN	LT

Lab ID: **LCS** **Client ID:** **N/A**

Sample Date/Time: N/A Received Date/Time: N/A

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	LCS 280-7794/4		280-7794		03/16/2010 09:39	1	TAL DEN	HEW
A:8260B	LCS 280-7794/4		280-7794		03/16/2010 09:39	1	TAL DEN	HEW
P:5030B	LCS 280-7423/3		280-7423		03/15/2010 09:34	1	TAL DEN	AEW
A:8260B SIM	LCS 280-7423/3		280-7423		03/15/2010 09:34	1	TAL DEN	AEW
P:3010A	LCS 280-7097/2-A		280-7274	280-7097	03/12/2010 07:30	1	TAL DEN	JW
A:6010B	LCS 280-7097/2-A		280-7274	280-7097	03/12/2010 23:36	1	TAL DEN	LT
P:3010A	LCS 280-7097/2-A		280-7325	280-7097	03/12/2010 07:30	1	TAL DEN	JW
A:6010B	LCS 280-7097/2-A		280-7325	280-7097	03/13/2010 13:33	1	TAL DEN	LT

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
SDG: 10022892

Laboratory Chronicle

Lab ID: **LCSD** **Client ID:** **N/A**

Sample Date/Time: N/A Received Date/Time: N/A

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	LCSD 280-7794/5		280-7794		03/16/2010 09:59	1	TAL DEN	HEW
A:8260B	LCSD 280-7794/5		280-7794		03/16/2010 09:59	1	TAL DEN	HEW

Lab ID: **MS** **Client ID:** **N/A**

Sample Date/Time: 03/09/2010 10:15 Received Date/Time: 03/09/2010 14:31

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1175-C-1 MS		280-7794		03/16/2010 12:04	1	TAL DEN	HEW
A:8260B	280-1175-C-1 MS		280-7794		03/16/2010 12:04	1	TAL DEN	HEW

Lab ID: **MSD** **Client ID:** **N/A**

Sample Date/Time: 03/09/2010 10:15 Received Date/Time: 03/09/2010 14:31

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1175-C-1 MSD		280-7794		03/16/2010 12:24	1	TAL DEN	HEW
A:8260B	280-1175-C-1 MSD		280-7794		03/16/2010 12:24	1	TAL DEN	HEW

Lab References:
TAL DEN = TestAmerica Denver

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1272-1

SDG No.: 10022892

Instrument ID: MSV_R1 Analysis Batch Number: 5135

Lab Sample ID: IC 280-5135/4 Client Sample ID:

Date Analyzed: 02/19/10 13:54 Lab File ID: R6582.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Nitropropane	8.79	Analyte not Identified by the Data System	garciaj	02/19/10 17:15
Tetrahydrothiophene	9.89	Analyte not Identified by the Data System	garciaj	02/19/10 17:15
cis-1,4-Dichloro-2-butene	11.05	Analyte not Identified by the Data System	garciaj	02/19/10 17:16

MANUAL INTEGRATION SUMMARY REPORT

Sample Name	Data File	Lot No.	Compound	Code
BFB	030810.b/g2_33845.D	BFB	No Manual Integrations	
IC1.5	030810.b/g2_33848.D	IC1.5	No Manual Integrations	
IC2.0	030810.b/g2_33849.D	IC2.0	No Manual Integrations	
IC2.5	030810.b/g2_33850.D	IC2.5	No Manual Integrations	
IC3.5	030810.b/g2_33851.D	IC3.5	No Manual Integrations	
IC5	030810.b/g2_33852.D	IC5	No Manual Integrations	
IC7.5	030810.b/g2_33853.D	IC7.5	No Manual Integrations	
IC17.5	030810.b/g2_33855.D	IC17.5	No Manual Integrations	
BFB	030910.b/g2_33860.D	BFB	No Manual Integrations	
ICV5	030910.b/g2_33862.D	10479	No Manual Integrations	

JM 3/10/10

Legend

BID - Baseline ID
BAS - Baseline Event
SP - Split Peak
TAIL - Peak Tailing or Fronting
NOID - Analyte not Identified by the Data System
MSID - Analyte Misidentified by the Data System
MSIN - Mis-Integrated

MANUAL INTEGRATION SUMMARY REPORT

Sample Name	Data File	Lot No.	Compound	Code
BFB	021910i.b/R6579.D		No Manual Integrations	
IC 0.3	021910i.b/R6581.D	IC 0.3	No Manual Integrations	
IC 1.0	021910i.b/R6582.D	IC 1.0	2-nitropropane	NOID
IC 1.0	021910i.b/R6582.D	IC 1.0	cis-1,4-dichloro-2-butene	NOID
IC 1.0	021910i.b/R6582.D	IC 1.0	Tetrahydrothiophene	NOID
IC 2.0	021910i.b/R6583.D	IC 2.0	No Manual Integrations	
IC 5.0	021910i.b/R6584.D	IC 5.0	No Manual Integrations	
IC 10.0	021910i.b/R6585.D	IC 10.0	No Manual Integrations	
IC 30.0	021910i.b/R6586.D	IC 30.0	No Manual Integrations	
IC 60.0	021910i.b/R6587.D	IC 60.0	No Manual Integrations	
ICV SA	021910i.b/R6588.D	ICV SA	No Manual Integrations	
ICV SA	021910i.b/R6589.D	ICV SA	No Manual Integrations	

Thu 02/22

Legend

BID - Baseline ID
BAS - Baseline Event
SP - Split Peak
TAIL - Peak Tailing or Fronting
NOID - Analyte not Identified by the Data System
MSID - Analyte Misidentified by the Data System
MSIN - Mis-Integrated

MANUAL INTEGRATION SUMMARY REPORT

Sample Name	Data File	Lot No.	Compound	Code
BFB	022510i.b/R6800.D		No Manual Integrations	
IC	022510i.b/R6801.D	M 0.3	No Manual Integrations	
IC	022510i.b/R6802.D	M 1.0	No Manual Integrations	
IC	022510i.b/R6803.D	M 2.0	No Manual Integrations	
IC	022510i.b/R6804.D	M 5.0	No Manual Integrations	
IC	022510i.b/R6805.D	M 10.0	No Manual Integrations	
IC	022510i.b/R6806.D	M 30.0	No Manual Integrations	
IC	022510i.b/R6807.D	M 60.0	No Manual Integrations	
ICV	022510i.b/R6809.D	MA	No Manual Integrations	
ICV	022510i.b/R6810.D	MB	No Manual Integrations	
ICV	022510i.b/R6811.D	MC	No Manual Integrations	
ICV	022510i.b/R6812.D	MD	No Manual Integrations	

Legend

BID - Baseline ID
BAS - Baseline Event
SP - Split Peak
TAIL - Peak Tailing or Fronting
NOID - Analyte not Identified by the Data System
MSID - Analyte Misidentified by the Data System
MSIN - Mis-Integrated

2/25/10

MANUAL INTEGRATION SUMMARY REPORT

Sample Name	Data File	Lot No.	Compound	Code
BFB	031510.b/g2_33905.D	BFB	No Manual Integrations	
CCV3.5	031510.b/g2_33906x.D		No Manual Integrations	
CCV3.5	031510.b/g2_33907.D		No Manual Integrations	
CCV3.5	031510.b/g2_33908x.D	CONFIRM	No Manual Integrations	
LCS5	031510.b/g2_33909.D		No Manual Integrations	
SCREEN	031510.b/g2_33910s.D	1272-1	No Manual Integrations	
SCREEN	031510.b/g2_33911s.D	1272-1	No Manual Integrations	
SCREEN	031510.b/g2_33912s.D	1272-2	No Manual Integrations	
VBLK	031510.b/g2_33914.D	VBLK	No Manual Integrations	
280-1272-E-2	031510.b/g2_33915.D	pH<2	No Manual Integrations	
280-1272-C-1	031510.b/g2_33916.D	pH<2	No Manual Integrations	
280-1272-d-1 MS	031510.b/g2_33917.D	pH<2	No Manual Integrations	
280-1272-d-1 MS	031510.b/g2_33918.D	pH<2	No Manual Integrations	

4/1/10

Legend

BID - Baseline ID
BAS - Baseline Event
SP - Split Peak
TAIL - Peak Tailing or Fronting
NOID - Analyte not Identified by the Data System
MSID - Analyte Misidentified by the Data System
MSIN - Mis-Integrated

MANUAL INTEGRATION SUMMARY REPORT

Sample Name	Data File	Lot No.	Compound	Code
BFB	031610.b/R7315.D		No Manual Integrations	
CCV	031610.b/R7316.D	M	No Manual Integrations	
CCV	031610.b/R7317.D	S	No Manual Integrations	
LCS	031610.b/R7318.D		No Manual Integrations	
LCSD	031610.b/R7319.D		No Manual Integrations	
MB	031610.b/R7320.D		No Manual Integrations	
280-1175-C-1	031610.b/R7321.D	pH<2	No Manual Integrations	
280-1175-c-1MS	031610.b/R7322.D	pH<2	No Manual Integrations	
280-1175-c-1MSD	031610.b/R7323.D	pH<2	No Manual Integrations	
280-1175-B-5	031610.b/R7324.D	pH<2	No Manual Integrations	
280-1175-C-6	031610.b/R7325.D	pH<2	No Manual Integrations	
280-1301-A-1	031610.b/R7326.D	pH7	No Manual Integrations	
280-1177-B-14	031610.b/R7333.D	pH<2	No Manual Integrations	
280-1177-A-15	031610.b/R7334.D	pH<2	No Manual Integrations	
280-1177-A-15	031610.b/R7335.D	pH<2	No Manual Integrations	
280-1177-A-16	031610.b/R7336.D	pH<2	No Manual Integrations	
280-1178-A-7	031610.b/R7337.D	pH<2	No Manual Integrations	
280-1178-A-8	031610.b/R7338.D	pH<2	No Manual Integrations	
280-1184-B-1	031610.b/R7339.D	pH<2	No Manual Integrations	
280-1184-A-2	031610.b/R7340.D	pH<2	No Manual Integrations	
280-1338-B-1	031610.b/R7341.D	pH<2	No Manual Integrations	
280-1272-E-1	031610.b/R7342.D	pH<2	No Manual Integrations	
280-1272-D-2	031610.b/R7343.D	pH<2	No Manual Integrations	
280-1198-A-2	031610.b/R7344.D	pH<2	No Manual Integrations	
280-1198-L-1	031610.b/R7345.D	pH<2	No Manual Integrations	

UK 3/22/10

Legend

BID - Baseline ID
BAS - Baseline Event
SP - Split Peak
TAIL - Peak Tailing or Fronting
NOID - Analyte not Identified by the Data System
MSID - Analyte Misidentified by the Data System
MSIN - Mis-Integrated

Shipping and Receiving Documents

0.7°C
3-11-10

Stoller
Legacy Management Team

Chain of Custody / Sample Submittal Form

RIN: 10022892 COC: 10022892.1.9

PO #: 3864 Cost Number: 1-502-1-06-509-4-02 Sampler(s): Julian Caballero

Project: Pinellas Monitoring Laboratory: TestAmerica Denver

Address: 4955 Yarrow Street Arvada, Colorado 80002 Phone: 303 736 0100 Fax: 303 431 7171

Matrix: WA - Water Turnaround (Days): 28

Ship #	Ticket Number	Date / Time Sampled	Site	Location	Container	# Cont.	Preservation	Matrix	Comp.	Grab	Filtered	QC	Analysis
0	IDW 248		PIN99	PIN99-2880	Glass 40 mL	3	4 C, HCl	WA			N		VOA
0	IDW 249		PIN99	PIN99-2881	Glass 40 mL	3	4 C, HCl	WA			N		VOA
0	IDW 250		PIN99	PIN99-2882	Glass 40 mL	3	4 C, HCl	WA			N		VOA
0	IDW 251		PIN99	PIN99-2883	Glass 40 mL	3	4 C, HCl	WA			N		VOA
0	IDW 252		PIN99	PIN99-2884	Glass 40 mL	3	4 C, HCl	WA			N		VOA
0	IDW 253		PIN99	PIN99-2885	Glass 40 mL	3	4 C, HCl	WA			N		VOA
0	IDW 254		PIN99	PIN99-2886	Glass 40 mL	3	4 C, HCl	WA			N		VOA
0	IDW 255		PIN99	PIN99-2887	Glass 40 mL	3	4 C, HCl	WA			N		VOA
0	IDW 256		PIN99	PIN99-2888	Glass 40 mL	3	4 C, HCl	WA			N		VOA
0	IDW 257		PIN99	PIN99-2889	Glass 40 mL	3	4 C, HCl	WA			N		VOA
0	IDW 258		PIN99	PIN99-2890	Glass 40 mL	3	4 C, HCl	WA			N		VOA
0	IDW 259		PIN99	PIN99-2891	Glass 40 mL	3	4 C, HCl	WA			N		VOA
0	IDW 260	3-9-10/1425	PIN99	PIN99-2892	Glass 40 mL	4	4 C, HCl	WA		✓	N		VOA + 1,4-Dioxane
0	IDW 261	3-9-10/1410	PIN99	PIN99-2893	Glass 40 mL	4	4 C, HCl	WA		✓	N		VOA + 1,4-Dioxane

IDW260, 3-9-10/1425, PIN99 PIN99-2892 Plastic 125ml 1 4C, HNO_3 WA G N Fe, Al

IDW261, 3-9-10/1410, PIN99 PIN99-2893 Plastic 125ml 1 4C, HNO_3 WA G N Fe, Al

Relinquished by: Julian P. Caballero 3-9-10 1645

Received by: Charles E Voby 3-9-10 1645, RECEIVED 3/11/10 10:00

Login Sample Receipt Check List

Client: S.M. Stoller Corporation

Job Number: 280-1272-1
SDG Number: 10022892

Login Number: 1272
Creator: Harrington, Nicholas
List Number: 1

List Source: TestAmerica Denver

Question	T / F/ NA	Comment
Radioactivity either was not measured or, if measured, is at or below background	True	
The cooler's custody seal, if present, is intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the sample IDs on the containers and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
VOA sample vials do not have headspace or bubble is <6mm (1/4") in diameter.	True	
If necessary, staff have been informed of any short hold time or quick TAT needs	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Is the Field Sampler's name present on COC?	True	
Sample Preservation Verified	True	

27

52

FedEx Express — *US Airbill*

FedEx Tracking Number: 8712 8177 1977

1 From This portion can be removed for Recipient's records.

Date: 3/10/10 FedEx Tracking Number: 871281771977

Sender's Name: CUSTODY Phone: 813 885 7427

Company: TESTAMERICA TAMPA

Address: 4712 BENJAMIN RD STE 100 Dept./Floor/Suite/Room

City: TAMPA State: FL ZIP: 33634-4403

2 Your Internal Billing Reference

3 To

Recipient's Name: SANDRA KEGINE Phone: 303 736 0100

Company: TEST AMERICA

HOLD Weekday — Print FedEx location address below. NOT available for FedEx First Overnight.

HOLD Saturday. Print FedEx location address below. Available ONLY for FedEx Priority Overnight and FedEx 2Day to select locations.

Address: 4955 E. 49TH AVE Dept./Floor/Suite/Room

We cannot deliver to P.O. boxes or P.O. ZIP codes.

Address: Print FedEx location address here if HOLD option is selected.

City: ARVADA State: CO ZIP: 80002

0414646911

8712 8177 1977

RECIPIENT: PEEL HERE

fedex.com 1.800.GoFedEx 1.800.463.3339

SPH32

Form ID No. 0215 — Recipient's Copy

4a Express Package Service — Packages up to 150 lbs.

- [x] FedEx Priority Overnight — Next business morning.* Friday shipments will be delivered on Monday
- [ ] FedEx Standard Overnight — Next business afternoon.* Saturday Delivery NOT available.
- [ ] FedEx First Overnight — Earliest next business morning delivery to select locations.*

*To most locations.

FedEx TRK# 0215 8712 8177 1977

THU - 11 MAR A1
PRIORITY OVERNIGHT

80002
CO-US
DEN

XH WHHA

Emp# 631384 10MAR10 TPFA

One box must be checked.

- [x] No
- [ ] Yes — As per attached Shipper's Declaration.
- [ ] Yes — Shipper's Declaration not required.
- [ ] Dry Ice — Dry ice, 9, UN 1845 ____ x ____ kg
- [ ] Cargo Aircraft Only

Dangerous goods (including dry ice) cannot be shipped in FedEx packaging or placed in a FedEx Express Drop Box.

7 Payment Bill to: Enter FedEx Acct. No. or Credit Card No. below.

- [ ] Sender — Acct. No. in Section 1 will be billed.
- [x] Recipient
- [ ] Third Party
- [ ] Credit Card
- [ ] Cash/Check

Obtain Recip. Acct. No. [ ]

Total Packages: 1 Total Weight: 31 lbs. Credit Card Auth.

[1] Our liability is limited to $100 unless you declare a higher value. See the current FedEx Service Guide for details.

553

Rev. Date 2/08•Part #158279•©1994-2008 FedEx•PRINTED IN U.S.A.•SRS

ANALYTICAL REPORT

Job Number: 280-1377-1

SDG Number: 10022892

Job Description: PINELLAS MONITORING

For:
S.M. Stoller Corporation
2597 B 3/4 Road
Grand Junction, CO 81503

Attention: Mr. Steve Donivan

Approved for release.
Kae E Yoder
Project Manager II
4/9/2010 4:07 PM

Kae E Yoder
Project Manager II
kae.yoder@testamericainc.com
04/09/2010

The test results in this report relate only to the samples in this report and meet all requirements of NELAC, with any exceptions noted. Pursuant to NELAP, this report shall not be reproduced except in full, without the written approval of the laboratory. All questions regarding this report should be directed to the TestAmerica Denver Project Manager.

The Lab Certification ID# is E87667.

Reporting limits are adjusted for sample size used, dilutions and moisture content if applicable.

Table of Contents

Cover Title Page 1

Data Summaries 4

Report Narrative 4

Data Qualifiers 5

Executive Summary 6

Method Summary 11

Method / Analyst Summary 12

Sample Summary 13

Sample Datasheets 14

Surrogate Summary 99

QC Data Summary 102

QC Association Summary 145

Lab Chronicle 149

Manual Integration Summary 160

Organic Sample Data 174

GC/MS VOA 174

Method 8260B 174

Method 8260B QC Summary 175

Method 8260B Sample Data 234

Standards Data 564

Method 8260B ICAL Data 564

Method 8260B CCAL Data 891

Raw QC Data 1025

Method 8260B Tune Data 1025

Method 8260B Blank Data 1089

Method 8260B LCS/LCSD Data 1133

Table of Contents

Method 8260B MS/MSD Data 1193

Method 8260B Run Logs 1259

Inorganic Sample Data 1417

Metals Data 1417

Met Cover Page 1418

Met Sample Data 1419

Met QC Data 1426

Met ICV/CCV 1426

Met CRQL 1430

Met Blanks 1431

Met ICSA/ICSAB 1433

Met MS/MSD/PDS 1435

Met LCS/LCSD 1437

Met Serial Dilution 1438

Met MDL 1439

Met IECF 1440

Met Linear Ranges 1446

Met Preparation Log 1447

Met Analysis Run Log 1448

Met Raw Data 1452

Met Prep Data 1873

Shipping and Receiving Documents 1875

Client Chain of Custody 1876

Sample Receipt Checklist 1879

CASE NARRATIVE

Client: S.M. Stoller Corporation

Project: PINELLAS MONITORING - 10022892

Report Number: 280-1377-1

With exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. All laboratory quality control samples analyzed in conjunction with the samples in this project were within established control limits, with any exceptions noted. Calculations are performed before rounding to avoid round-off errors in calculated results.

This report includes reporting limits (RLs) less than TestAmerica Denver's practical quantitation limits. These reporting limits are being used specifically at the client's request to meet the needs of this project. Please note that data are not normally reported to these levels without qualification, since they are inherently less reliable and potentially less defensible than required by the current NELAC standards.

Results between the method detection limit (MDL) and reporting limit (RL) are flagged with a "J" qualifier to indicate an estimated value. These results are statistically less reliable than results greater than or equal to the RL and should be considered a qualitative value.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

RECEIPT

The samples were received on 3/16/2010; the samples arrived in good condition, properly preserved and on ice. The temperature of the cooler at receipt was 4.1 C.

One of the three 40 mL vials submitted for sample PIN20-M003 (IDW 219) requesting VOA analysis was received at the laboratory broken. Sufficient volume remained to complete the requested analysis. The client was notified on 3/16/2010.

GC/MS VOLATILES - SW846 8260B

Due to analytes present above the linear calibration curve, a reduced aliquot size had to be used for the preparation/analysis of samples PIN12-S30B (IDW 181), PIN12-S33C (IDW 182), PIN12-S35B (IDW 183), PIN15-0430 (IDW 207), PIN15-2873 (IDW 241) and PIN20-0502 (IDW 216). To provide the lowest possible detection limits, multiple runs are reported.

The LCS associated with batch 280-7614 failed the recovery criteria high for Hexachlorobutadiene at 125% (upper limit is 123%) and sec-Butylbenzene at 123% (upper limit is 120%), and the associated sample results have been flagged “*”. These are not standard method control analytes, and no detectable concentrations of these constituents are present in the associated samples; therefore, data are reported as is.

The LCS or LCSD associated with batch 280-7993 failed the recovery criteria high for Hexachlorobutadiene at 125% (upper limit is 123%) and sec-Butylbenzene at 121% (upper limit is 120%), and the associated sample results have been flagged “*”. These are not standard method control analytes, and no detectable concentrations of these constituents are present in the associated sample; therefore, data are reported as is.

Trichloroethene and Methylene chloride were detected in the method blanks associated with batches 280-7614 and 280-7816, respectively, at levels that were above the method detection limit but below the reporting limit. The values should be considered estimates, and have been flagged “J”. If the associated samples reported a result above the MDL and/or RL, the result has been “B” flagged.

The Continuing Calibration Verification (CCV) standard associated with samples in batch 280-7816 exhibited the %Difference (%D) value >35%, biased low, for Methyl tert-butyl ether (-36.5%). All CCC and SPCC compounds are in control; therefore, method criteria have been met.

No other anomalies were encountered.

TOTAL METALS - SW846 6010B

The Linear Range date listed on the CLP forms for ICP instrument 26 is incorrect. The correct Linear range date is 1/25/2010.

No other anomalies were encountered.

DATA REPORTING QUALIFIERS

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Lab Section	Qualifier	Description
GC/MS VOA		
	B	Compound was found in the blank and sample.
	U	Indicates the analyte was analyzed for but not detected.
	*	LCS or LCSD exceeds the control limits
	F	MS or MSD exceeds the control limits
	4	MS, MSD: The analyte present in the original sample is 4 times greater than the matrix spike concentration; therefore, control limits are not applicable.
	J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
Metals		
	U	Indicates the analyte was analyzed for but not detected.
	J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1377-1	**PIN12-0550-1**					
Acetone		3.8	J	10	ug/L	8260B
280-1377-3	**PIN12-0550-3**					
Acetone		3.4	J	10	ug/L	8260B
Toluene		0.19	J	1.0	ug/L	8260B
280-1377-4	**PIN12-0551-1**					
Acetone		3.2	J	10	ug/L	8260B
280-1377-6	**PIN12-0551-3**					
Acetone		3.8	J	10	ug/L	8260B
280-1377-7	**PIN12-0553A**					
1,2-Dichlorobenzene		0.25	J	1.0	ug/L	8260B
280-1377-9	**PIN12-0553C**					
1,1-Dichloroethane		0.74	J	1.0	ug/L	8260B
cis-1,2-Dichloroethene		0.72	J	1.0	ug/L	8260B
280-1377-10	**PIN12-S30B**					
1,3-Dichlorobenzene		0.18	J	1.0	ug/L	8260B
Dichlorodifluoromethane		1.1		1.0	ug/L	8260B
1,1-Dichloroethane		5.0		1.0	ug/L	8260B
cis-1,2-Dichloroethene		150		5.0	ug/L	8260B
trans-1,2-Dichloroethene		15		1.0	ug/L	8260B
1,1-Dichloroethene		3.3		1.0	ug/L	8260B
2-Hexanone		2.9	J	5.0	ug/L	8260B
Trichloroethene		39	B	1.0	ug/L	8260B
Vinyl chloride		81		5.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Lab Sample ID	Client Sample ID			Reporting		
Analyte		Result / Qualifier		Limit	Units	Method
280-1377-11	**PIN12-S33C**					
1,1-Dichloroethane		0.23	J	1.0	ug/L	8260B
cis-1,2-Dichloroethene		1000		20	ug/L	8260B
trans-1,2-Dichloroethene		20		1.0	ug/L	8260B
1,1-Dichloroethene		54		1.0	ug/L	8260B
Toluene		0.24	J	1.0	ug/L	8260B
Trichloroethene		76		20	ug/L	8260B
Vinyl chloride		590		20	ug/L	8260B
280-1377-12	**PIN12-S35B**					
cis-1,2-Dichloroethene		37000		2000	ug/L	8260B
trans-1,2-Dichloroethene		5900		200	ug/L	8260B
1,1-Dichloroethene		790		200	ug/L	8260B
Trichloroethene		10000		200	ug/L	8260B
Vinyl chloride		8700		200	ug/L	8260B
280-1377-15	**PIN12-S71C**					
1,1-Dichloroethane		0.48	J	1.0	ug/L	8260B
cis-1,2-Dichloroethene		6.3		1.0	ug/L	8260B
trans-1,2-Dichloroethene		3.9		1.0	ug/L	8260B
Vinyl chloride		12		1.0	ug/L	8260B
280-1377-16	**PIN12-S71D**					
1,1-Dichloroethane		1.0		1.0	ug/L	8260B
cis-1,2-Dichloroethene		14		1.0	ug/L	8260B
trans-1,2-Dichloroethene		8.5		1.0	ug/L	8260B
1,1-Dichloroethene		0.30	J	1.0	ug/L	8260B
Vinyl chloride		25		1.0	ug/L	8260B
280-1377-17	**PIN15-0520**					
Aluminum		250		100	ug/L	6010B
Iron		520		100	ug/L	6010B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1377-18	**PIN15-0530**					
Benzene		0.80	J	1.0	ug/L	8260B
Chloroform		0.40	J	1.0	ug/L	8260B
cis-1,2-Dichloroethene		220		20	ug/L	8260B
trans-1,2-Dichloroethene		18		1.0	ug/L	8260B
1,1-Dichloroethene		0.63	J	1.0	ug/L	8260B
Styrene		0.38	J	1.0	ug/L	8260B
Vinyl chloride		790		20	ug/L	8260B
Aluminum		1300		100	ug/L	6010B
Iron		3600		100	ug/L	6010B
280-1377-19	**PIN15-0534**					
Aluminum		480		100	ug/L	6010B
Iron		320		100	ug/L	6010B
280-1377-20	**PIN15-0535**					
Benzene		0.31	J	1.0	ug/L	8260B
cis-1,2-Dichloroethene		0.24	J	1.0	ug/L	8260B
Styrene		0.44	J	1.0	ug/L	8260B
Vinyl chloride		16		1.0	ug/L	8260B
Aluminum		720		100	ug/L	6010B
Iron		330		100	ug/L	6010B
280-1377-21	**PIN15-0568**					
Styrene		0.41	J	1.0	ug/L	8260B
Aluminum		290		100	ug/L	6010B
Iron		750		100	ug/L	6010B
280-1377-22	**PIN15-0569**					
Styrene		0.43	J	1.0	ug/L	8260B
Vinyl chloride		1.0		1.0	ug/L	8260B
Aluminum		2200		100	ug/L	6010B
Iron		3100		100	ug/L	6010B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1377-23	**PIN15-2873**					
Benzene		0.74	J	1.0	ug/L	8260B
cis-1,2-Dichloroethene		210		20	ug/L	8260B
trans-1,2-Dichloroethene		18		1.0	ug/L	8260B
1,1-Dichloroethene		0.60	J	1.0	ug/L	8260B
Styrene		0.40	J	1.0	ug/L	8260B
Vinyl chloride		740		20	ug/L	8260B
Aluminum		1600		100	ug/L	6010B
Iron		3500		100	ug/L	6010B
280-1377-24	**PIN15-E001**					
4-Isopropyltoluene		0.47	J	1.0	ug/L	8260B
Styrene		0.45	J	1.0	ug/L	8260B
280-1377-25	**PIN20-0502**					
cis-1,2-Dichloroethene		66		6.7	ug/L	8260B
trans-1,2-Dichloroethene		0.78	J	1.0	ug/L	8260B
1,1-Dichloroethene		0.38	J	1.0	ug/L	8260B
1,1-Dichloropropene		0.44	J	1.0	ug/L	8260B
Styrene		0.44	J	1.0	ug/L	8260B
Vinyl chloride		130		6.7	ug/L	8260B
280-1377-26	**PIN20-0503**					
cis-1,2-Dichloroethene		0.20	J	1.0	ug/L	8260B
Styrene		0.41	J	1.0	ug/L	8260B
Vinyl chloride		0.52	J	1.0	ug/L	8260B
280-1377-27	**PIN20-2867**					
cis-1,2-Dichloroethene		2.3		1.0	ug/L	8260B
trans-1,2-Dichloroethene		0.31	J	1.0	ug/L	8260B
Trichloroethene		0.23	J	1.0	ug/L	8260B
Vinyl chloride		3.6		1.0	ug/L	8260B
280-1377-28	**PIN20-M003**					
trans-1,3-Dichloropropene		1.5		1.0	ug/L	8260B
4-Isopropyltoluene		1.1		1.0	ug/L	8260B
Styrene		0.39	J	1.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1377-29	**PIN20-M005**					
Styrene		0.39	J	1.0	ug/L	8260B
280-1377-30	**PIN20-M015**					
cis-1,2-Dichloroethene		2.5		1.0	ug/L	8260B
Vinyl chloride		5.5		1.0	ug/L	8260B
280-1377-31	**PIN20-M035**					
cis-1,2-Dichloroethene		2.1		1.0	ug/L	8260B
trans-1,2-Dichloroethene		0.30	J	1.0	ug/L	8260B
Trichloroethene		0.22	J	1.0	ug/L	8260B
Vinyl chloride		3.1		1.0	ug/L	8260B

METHOD SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Description	Lab Location	Method	Preparation Method
Matrix: Water			
Volatile Organic Compounds (GC/MS)	TAL DEN	SW846 8260B	
Purge and Trap	TAL DEN		SW846 5030B
Metals (ICP)	TAL DEN	SW846 6010B	
Preparation, Total Metals	TAL DEN		SW846 3010A

Lab References:

TAL DEN = TestAmerica Denver

Method References:

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

METHOD / ANALYST SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method	Analyst	Analyst ID
SW846 8260B	Garcia, James D	JDG
SW846 8260B	Ilczyszyn, Dennis P	DPI
SW846 8260B	Jackson, Todd D	TDJ
SW846 8260B	Waterland, Hayley E	HEW
SW846 6010B	Wells, David	DW

SAMPLE SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
280-1377-1	PIN12-0550-1	Water	03/11/2010 0928	03/16/2010 0930
280-1377-2	PIN12-0550-2	Water	03/11/2010 1033	03/16/2010 0930
280-1377-3	PIN12-0550-3	Water	03/11/2010 1158	03/16/2010 0930
280-1377-4	PIN12-0551-1	Water	03/11/2010 1346	03/16/2010 0930
280-1377-5	PIN12-0551-2	Water	03/11/2010 1424	03/16/2010 0930
280-1377-6	PIN12-0551-3	Water	03/11/2010 1503	03/16/2010 0930
280-1377-7	PIN12-0553A	Water	03/12/2010 1245	03/16/2010 0930
280-1377-8	PIN12-0553B	Water	03/12/2010 1145	03/16/2010 0930
280-1377-9	PIN12-0553C	Water	03/12/2010 1025	03/16/2010 0930
280-1377-10	PIN12-S30B	Water	03/12/2010 1220	03/16/2010 0930
280-1377-11	PIN12-S33C	Water	03/12/2010 1050	03/16/2010 0930
280-1377-12	PIN12-S35B	Water	03/12/2010 1302	03/16/2010 0930
280-1377-13	PIN12-S36B	Water	03/12/2010 1136	03/16/2010 0930
280-1377-14	PIN12-S71B	Water	03/12/2010 1230	03/16/2010 0930
280-1377-15	PIN12-S71C	Water	03/12/2010 0950	03/16/2010 0930
280-1377-16	PIN12-S71D	Water	03/12/2010 1105	03/16/2010 0930
280-1377-17	PIN15-0520	Water	03/11/2010 1345	03/16/2010 0930
280-1377-17MS	PIN15-0520	Water	03/11/2010 1345	03/16/2010 0930
280-1377-17MSD	PIN15-0520	Water	03/11/2010 1345	03/16/2010 0930
280-1377-18	PIN15-0530	Water	03/11/2010 1015	03/16/2010 0930
280-1377-19	PIN15-0534	Water	03/11/2010 1430	03/16/2010 0930
280-1377-20	PIN15-0535	Water	03/11/2010 1100	03/16/2010 0930
280-1377-21	PIN15-0568	Water	03/11/2010 1610	03/16/2010 0930
280-1377-22	PIN15-0569	Water	03/11/2010 1515	03/16/2010 0930
280-1377-23	PIN15-2873	Water	03/11/2010 1200	03/16/2010 0930
280-1377-24	PIN15-E001	Water	03/11/2010 1115	03/16/2010 0930
280-1377-25	PIN20-0502	Water	03/11/2010 0920	03/16/2010 0930
280-1377-26	PIN20-0503	Water	03/11/2010 1020	03/16/2010 0930
280-1377-27	PIN20-2867	Water	03/11/2010 1625	03/16/2010 0930
280-1377-28	PIN20-M003	Water	03/11/2010 1525	03/16/2010 0930
280-1377-29	PIN20-M005	Water	03/11/2010 1610	03/16/2010 0930
280-1377-30	PIN20-M015	Water	03/11/2010 1100	03/16/2010 0930
280-1377-31	PIN20-M035	Water	03/11/2010 1425	03/16/2010 0930
280-1377-32	PIN20-M38D	Water	03/11/2010 1335	03/16/2010 0930
280-1377-33	PIN99-2874	Water	03/11/2010 0806	03/16/2010 0930
280-1377-34	PIN99-2881	Water	03/11/2010 0800	03/16/2010 0930
280-1377-35	PIN99-2888	Water	03/11/2010 0800	03/16/2010 0930
280-1377-36	PIN12-0554A	Water	03/12/2010 1505	03/16/2010 0930

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0550-1**

Lab Sample ID: 280-1377-1
Client Matrix: Water

Date Sampled: 03/11/2010 0928
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7816	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7394.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1640		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1640			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	3.8	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0550-1**

Lab Sample ID: 280-1377-1
Client Matrix: Water

Date Sampled: 03/11/2010 0928
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7816	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7394.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1640		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1640			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	111		70 - 127
Toluene-d8 (Surr)	100		80 - 125
4-Bromofluorobenzene (Surr)	97		78 - 118
Dibromofluoromethane (Surr)	114		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0550-2**

Lab Sample ID: 280-1377-2
Client Matrix: Water

Date Sampled: 03/11/2010 1033
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method: 8260B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 03/17/2010 1700
Date Prepared: 03/17/2010 1700

Analysis Batch: 280-7816

Instrument ID: MSV_R1
Lab File ID: R7395.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0550-2**

Lab Sample ID: 280-1377-2
Client Matrix: Water

Date Sampled: 03/11/2010 1033
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7816	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7395.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1700		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1700			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	115		70 - 127
Toluene-d8 (Surr)	100		80 - 125
4-Bromofluorobenzene (Surr)	94		78 - 118
Dibromofluoromethane (Surr)	116		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0550-3**

Lab Sample ID: 280-1377-3
Client Matrix: Water

Date Sampled: 03/11/2010 1158
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7816	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7396.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1720		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1720			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	3.4	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0550-3**

Lab Sample ID: 280-1377-3
Client Matrix: Water

Date Sampled: 03/11/2010 1158
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7816	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7396.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1720		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1720			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.19	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	115		70 - 127
Toluene-d8 (Surr)	100		80 - 125
4-Bromofluorobenzene (Surr)	94		78 - 118
Dibromofluoromethane (Surr)	116		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0551-1**

Lab Sample ID: 280-1377-4
Client Matrix: Water

Date Sampled: 03/11/2010 1346
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7816	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7397.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1739		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1739			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	3.2	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0551-1**

Lab Sample ID: 280-1377-4
Client Matrix: Water

Date Sampled: 03/11/2010 1346
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7816	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7397.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1739		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1739			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	116		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	95		78 - 118
Dibromofluoromethane (Surr)	118		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0551-2**

Lab Sample ID: 280-1377-5
Client Matrix: Water

Date Sampled: 03/11/2010 1424
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method: 8260B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 03/17/2010 1800
Date Prepared: 03/17/2010 1800

Analysis Batch: 280-7816

Instrument ID: MSV_R1
Lab File ID: R7398.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0551-2**

Lab Sample ID: 280-1377-5
Client Matrix: Water

Date Sampled: 03/11/2010 1424
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7816	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7398.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1800		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1800			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	115		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	94		78 - 118
Dibromofluoromethane (Surr)	119		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0551-3**

Lab Sample ID: 280-1377-6
Client Matrix: Water

Date Sampled: 03/11/2010 1503
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7816	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7399.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1820		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1820			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	3.8	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0551-3**

Lab Sample ID: 280-1377-6
Client Matrix: Water

Date Sampled: 03/11/2010 1503
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7816	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7399.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1820		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1820			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	113		70 - 127
Toluene-d8 (Surr)	100		80 - 125
4-Bromofluorobenzene (Surr)	94		78 - 118
Dibromofluoromethane (Surr)	119		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0553A**

Lab Sample ID: 280-1377-7
Client Matrix: Water

Date Sampled: 03/12/2010 1245
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7614	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G2975.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1615		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1615			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U *	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.25	J	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U *	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0553A**

Lab Sample ID: 280-1377-7
Client Matrix: Water

Date Sampled: 03/12/2010 1245
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method: 8260B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 03/17/2010 1615
Date Prepared: 03/17/2010 1615

Analysis Batch: 280-7614

Instrument ID: MSV_G
Lab File ID: G2975.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	87		70 - 127
Toluene-d8 (Surr)	85		80 - 125
4-Bromofluorobenzene (Surr)	85		78 - 118
Dibromofluoromethane (Surr)	85		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0553B**

Lab Sample ID: 280-1377-8
Client Matrix: Water

Date Sampled: 03/12/2010 1145
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7614	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G2976.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1635		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1635			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U *	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U *	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0553B**

Lab Sample ID: 280-1377-8
Client Matrix: Water

Date Sampled: 03/12/2010 1145
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7614	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G2976.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1635		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1635			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	91		70 - 127
Toluene-d8 (Surr)	93		80 - 125
4-Bromofluorobenzene (Surr)	87		78 - 118
Dibromofluoromethane (Surr)	92		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0553C**

Lab Sample ID: 280-1377-9
Client Matrix: Water

Date Sampled: 03/12/2010 1025
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7614	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G2977.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1656		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1656			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U *	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.74	J	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.72	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U *	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0553C**

Lab Sample ID: 280-1377-9
Client Matrix: Water

Date Sampled: 03/12/2010 1025
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7614	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G2977.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1656		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1656			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	90		70 - 127
Toluene-d8 (Surr)	89		80 - 125
4-Bromofluorobenzene (Surr)	88		78 - 118
Dibromofluoromethane (Surr)	90		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S30B**

Lab Sample ID: 280-1377-10
Client Matrix: Water

Date Sampled: 03/12/2010 1220
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7614	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G2978.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1716		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1716			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U *	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.18	J	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	1.1		0.31	1.0
1,1-Dichloroethane	5.0		0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
trans-1,2-Dichloroethene	15		0.15	1.0
1,1-Dichloroethene	3.3		0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U *	0.12	1.0
2-Hexanone	2.9	J	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S30B**

Lab Sample ID: 280-1377-10
Client Matrix: Water

Date Sampled: 03/12/2010 1220
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7614	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G2978.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1716		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1716			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	39	B	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	90		70 - 127
Toluene-d8 (Surr)	90		80 - 125
4-Bromofluorobenzene (Surr)	89		78 - 118
Dibromofluoromethane (Surr)	91		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S30B**

Lab Sample ID: 280-1377-10
Client Matrix: Water

Date Sampled: 03/12/2010 1220
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8149		Instrument ID:	MSV_G
Preparation:	5030B			Lab File ID:	G3019.D
Dilution:	1.0			Initial Weight/Volume:	4 mL
Date Analyzed:	03/18/2010 0913	Run Type:	DL	Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0913				

Analyte	Result (ug/L)	Qualifier	MDL	RL
cis-1,2-Dichloroethene	150		0.75	5.0
Vinyl chloride	81		2.0	5.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	91		70 - 127
Toluene-d8 (Surr)	95		80 - 125
4-Bromofluorobenzene (Surr)	94		78 - 118
Dibromofluoromethane (Surr)	95		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S33C**

Lab Sample ID: 280-1377-11
Client Matrix: Water

Date Sampled: 03/12/2010 1050
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7614	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G2979.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1737		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1737			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U *	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.23	J	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
trans-1,2-Dichloroethene	20		0.15	1.0
1,1-Dichloroethene	54		0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U *	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S33C**

Lab Sample ID: 280-1377-11
Client Matrix: Water

Date Sampled: 03/12/2010 1050
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7614	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G2979.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1737		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1737			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.24	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	94		70 - 127
Toluene-d8 (Surr)	94		80 - 125
4-Bromofluorobenzene (Surr)	91		78 - 118
Dibromofluoromethane (Surr)	94		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S33C**

Lab Sample ID: 280-1377-11
Client Matrix: Water

Date Sampled: 03/12/2010 1050
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8149		Instrument ID:	MSV_G
Preparation:	5030B			Lab File ID:	G3020.D
Dilution:	1.0			Initial Weight/Volume:	1 mL
Date Analyzed:	03/18/2010 0933	Run Type:	DL	Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0933				

Analyte	Result (ug/L)	Qualifier	MDL	RL
cis-1,2-Dichloroethene	1000		3.0	20
Trichloroethene	76		3.2	20
Vinyl chloride	590		8.0	20

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	90		70 - 127
Toluene-d8 (Surr)	92		80 - 125
4-Bromofluorobenzene (Surr)	91		78 - 118
Dibromofluoromethane (Surr)	93		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S35B**

Lab Sample ID: 280-1377-12
Client Matrix: Water

Date Sampled: 03/12/2010 1302
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7833	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3010.D
Dilution:	1.0		Initial Weight/Volume:	0.1 mL
Date Analyzed:	03/18/2010 0421		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0421			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	380	U	380	2000
Benzene	32	U	32	200
Bromobenzene	34	U	34	200
Bromochloromethane	20	U	20	200
Bromodichloromethane	34	U	34	200
Bromoform	38	U	38	200
Bromomethane	42	U	42	200
2-Butanone (MEK)	400	U	400	1000
n-Butylbenzene	28	U	28	200
sec-Butylbenzene	34	U	34	200
tert-Butylbenzene	32	U	32	200
Carbon disulfide	90	U	90	200
Carbon tetrachloride	38	U	38	200
Chlorobenzene	34	U	34	200
Dibromochloromethane	34	U	34	200
Chloroethane	82	U	82	200
Chloroform	32	U	32	200
Chloromethane	60	U	60	200
2-Chlorotoluene	34	U	34	200
4-Chlorotoluene	42	U	42	200
1,2-Dibromo-3-Chloropropane	94	U	94	200
Dibromomethane	34	U	34	200
1,2-Dichlorobenzene	30	U	30	200
1,3-Dichlorobenzene	26	U	26	200
1,4-Dichlorobenzene	32	U	32	200
Dichlorodifluoromethane	62	U	62	200
1,1-Dichloroethane	44	U	44	200
1,2-Dichloroethane	26	U	26	200
trans-1,2-Dichloroethene	5900		30	200
1,1-Dichloroethene	790		46	200
1,2-Dichloropropane	36	U	36	200
1,3-Dichloropropane	44	U	44	200
2,2-Dichloropropane	36	U	36	200
cis-1,3-Dichloropropene	32	U	32	200
trans-1,3-Dichloropropene	38	U	38	200
1,1-Dichloropropene	38	U	38	200
Ethylbenzene	32	U	32	200
Hexachlorobutadiene	24	U	24	200
2-Hexanone	340	U	340	1000
Isopropylbenzene	38	U	38	200
4-Isopropyltoluene	40	U	40	200
Methylene Chloride	64	U	64	200
4-Methyl-2-pentanone	200	U	200	1000
Naphthalene	44	U	44	200
n-Propylbenzene	32	U	32	200
Styrene	34	U	34	200

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S35B**

Lab Sample ID: 280-1377-12
Client Matrix: Water

Date Sampled: 03/12/2010 1302
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7833	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3010.D
Dilution:	1.0		Initial Weight/Volume:	0.1 mL
Date Analyzed:	03/18/2010 0421		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0421			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,1,1,2-Tetrachloroethane	42	U	42	200
1,1,2,2-Tetrachloroethane	42	U	42	200
Tetrachloroethene	40	U	40	200
Toluene	34	U	34	200
1,2,3-Trichlorobenzene	42	U	42	200
1,2,4-Trichlorobenzene	42	U	42	200
1,1,1-Trichloroethane	32	U	32	200
1,1,2-Trichloroethane	54	U	54	200
Trichloroethene	10000		32	200
Trichlorofluoromethane	58	U	58	200
1,2,3-Trichloropropane	66	U	66	200
1,2,4-Trimethylbenzene	30	U	30	200
1,3,5-Trimethylbenzene	32	U	32	200
Vinyl chloride	8700		80	200
Xylenes, Total	38	U	38	200
1,2-Dibromoethane	36	U	36	200

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	89		70 - 127
Toluene-d8 (Surr)	86		80 - 125
4-Bromofluorobenzene (Surr)	89		78 - 118
Dibromofluoromethane (Surr)	90		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S35B**

Lab Sample ID: 280-1377-12
Client Matrix: Water

Date Sampled: 03/12/2010 1302
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7833	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3011.D
Dilution:	1.0		Initial Weight/Volume:	0.01 mL
Date Analyzed:	03/18/2010 0441		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0441			

Analyte	Result (ug/L)	Qualifier	MDL	RL
cis-1,2-Dichloroethene	37000		300	2000

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	89		70 - 127
Toluene-d8 (Surr)	90		80 - 125
4-Bromofluorobenzene (Surr)	89		78 - 118
Dibromofluoromethane (Surr)	91		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S36B**

Lab Sample ID: 280-1377-13
Client Matrix: Water

Date Sampled: 03/12/2010 1136
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4789.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1853		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1853			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S36B**

Lab Sample ID: 280-1377-13
Client Matrix: Water

Date Sampled: 03/12/2010 1136
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4789.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1853		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1853			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	109		70 - 127
Toluene-d8 (Surr)	98		80 - 125
4-Bromofluorobenzene (Surr)	101		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S71B**

Lab Sample ID: 280-1377-14
Client Matrix: Water

Date Sampled: 03/12/2010 1230
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4793.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2015		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2015			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S71B**

Lab Sample ID: 280-1377-14
Client Matrix: Water

Date Sampled: 03/12/2010 1230
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4793.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2015		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2015			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	107		70 - 127
Toluene-d8 (Surr)	98		80 - 125
4-Bromofluorobenzene (Surr)	100		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S71C**

Lab Sample ID: 280-1377-15
Client Matrix: Water

Date Sampled: 03/12/2010 0950
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4794.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2035		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2035			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.48	J	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	6.3		0.15	1.0
trans-1,2-Dichloroethene	3.9		0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S71C**

Lab Sample ID: 280-1377-15
Client Matrix: Water

Date Sampled: 03/12/2010 0950
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4794.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2035		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2035			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	12		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	111		70 - 127
Toluene-d8 (Surr)	97		80 - 125
4-Bromofluorobenzene (Surr)	105		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S71D**

Lab Sample ID: 280-1377-16
Client Matrix: Water

Date Sampled: 03/12/2010 1105
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4795.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2056		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2056			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	1.0		0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	14		0.15	1.0
trans-1,2-Dichloroethene	8.5		0.15	1.0
1,1-Dichloroethene	0.30	J	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S71D**

Lab Sample ID: 280-1377-16
Client Matrix: Water

Date Sampled: 03/12/2010 1105
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4795.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2056		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2056			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	25		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	108		70 - 127
Toluene-d8 (Surr)	98		80 - 125
4-Bromofluorobenzene (Surr)	100		78 - 118
Dibromofluoromethane (Surr)	102		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0520**

Lab Sample ID: 280-1377-17
Client Matrix: Water

Date Sampled: 03/11/2010 1345
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4796.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2117		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2117			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0520**

Lab Sample ID: 280-1377-17
Client Matrix: Water

Date Sampled: 03/11/2010 1345
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4796.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2117		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2117			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101		70 - 127
Toluene-d8 (Surr)	104		80 - 125
4-Bromofluorobenzene (Surr)	102		78 - 118
Dibromofluoromethane (Surr)	95		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0530**

Lab Sample ID: 280-1377-18
Client Matrix: Water

Date Sampled: 03/11/2010 1015
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4797.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2137		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2137			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.80	J	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.40	J	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
trans-1,2-Dichloroethene	18		0.15	1.0
1,1-Dichloroethene	0.63	J	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.38	J	0.17	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0530**

Lab Sample ID: 280-1377-18
Client Matrix: Water

Date Sampled: 03/11/2010 1015
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4797.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2137		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2137			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	110		70 - 127
Toluene-d8 (Surr)	99		80 - 125
4-Bromofluorobenzene (Surr)	105		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0530**

Lab Sample ID: 280-1377-18
Client Matrix: Water

Date Sampled: 03/11/2010 1015
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8159		Instrument ID:	MSV_G
Preparation:	5030B			Lab File ID:	G3090.D
Dilution:	1.0			Initial Weight/Volume:	1 mL
Date Analyzed:	03/19/2010 1910	Run Type:	DL	Final Weight/Volume:	20 mL
Date Prepared:	03/19/2010 1910				

Analyte	Result (ug/L)	Qualifier	MDL	RL
cis-1,2-Dichloroethene	220		3.0	20
Vinyl chloride	790		8.0	20

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	86		70 - 127
Toluene-d8 (Surr)	87		80 - 125
4-Bromofluorobenzene (Surr)	87		78 - 118
Dibromofluoromethane (Surr)	90		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0534**

Lab Sample ID: 280-1377-19
Client Matrix: Water

Date Sampled: 03/11/2010 1430
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8149	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3021.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1003		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1003			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0534**

Lab Sample ID: 280-1377-19
Client Matrix: Water

Date Sampled: 03/11/2010 1430
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8149	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3021.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1003		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1003			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	87		70 - 127
Toluene-d8 (Surr)	89		80 - 125
4-Bromofluorobenzene (Surr)	87		78 - 118
Dibromofluoromethane (Surr)	88		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0535**

Lab Sample ID: 280-1377-20
Client Matrix: Water

Date Sampled: 03/11/2010 1100
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4798.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2157		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2157			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.31	J	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.24	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0535**

Lab Sample ID: 280-1377-20
Client Matrix: Water

Date Sampled: 03/11/2010 1100
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4798.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2157		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2157			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.44	J	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	16		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	100		70 - 127
Toluene-d8 (Surr)	98		80 - 125
4-Bromofluorobenzene (Surr)	103		78 - 118
Dibromofluoromethane (Surr)	106		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0568**

Lab Sample ID: 280-1377-21
Client Matrix: Water

Date Sampled: 03/11/2010 1610
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4799.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2218		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2218			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0568**

Lab Sample ID: 280-1377-21
Client Matrix: Water

Date Sampled: 03/11/2010 1610
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4799.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2218		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2218			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.41	J	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	105		70 - 127
Toluene-d8 (Surr)	100		80 - 125
4-Bromofluorobenzene (Surr)	100		78 - 118
Dibromofluoromethane (Surr)	102		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0569**

Lab Sample ID: 280-1377-22
Client Matrix: Water

Date Sampled: 03/11/2010 1515
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4800.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2238		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2238			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0569**

Lab Sample ID: 280-1377-22
Client Matrix: Water

Date Sampled: 03/11/2010 1515
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4800.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2238		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2238			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.43	J	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	1.0		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	107		70 - 127
Toluene-d8 (Surr)	100		80 - 125
4-Bromofluorobenzene (Surr)	106		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-2873**

Lab Sample ID: 280-1377-23
Client Matrix: Water

Date Sampled: 03/11/2010 1200
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4802.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2320		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2320			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.74	J	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
trans-1,2-Dichloroethene	18		0.15	1.0
1,1-Dichloroethene	0.60	J	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.40	J	0.17	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-2873**

Lab Sample ID: 280-1377-23
Client Matrix: Water

Date Sampled: 03/11/2010 1200
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4802.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2320		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2320			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	106		70 - 127
Toluene-d8 (Surr)	102		80 - 125
4-Bromofluorobenzene (Surr)	108		78 - 118
Dibromofluoromethane (Surr)	101		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-2873**

Lab Sample ID: 280-1377-23
Client Matrix: Water

Date Sampled: 03/11/2010 1200
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8159		Instrument ID:	MSV_G
Preparation:	5030B			Lab File ID:	G3091.D
Dilution:	1.0			Initial Weight/Volume:	1 mL
Date Analyzed:	03/19/2010 1931	Run Type:	DL	Final Weight/Volume:	20 mL
Date Prepared:	03/19/2010 1931				

Analyte	Result (ug/L)	Qualifier	MDL	RL
cis-1,2-Dichloroethene	210		3.0	20
Vinyl chloride	740		8.0	20

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	85		70 - 127
Toluene-d8 (Surr)	91		80 - 125
4-Bromofluorobenzene (Surr)	91		78 - 118
Dibromofluoromethane (Surr)	93		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-E001**

Lab Sample ID: 280-1377-24
Client Matrix: Water

Date Sampled: 03/11/2010 1115
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4803.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2340		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2340			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.47	J	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-E001**

Lab Sample ID: 280-1377-24
Client Matrix: Water

Date Sampled: 03/11/2010 1115
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4803.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 2340		Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2340			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.45	J	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101		70 - 127
Toluene-d8 (Surr)	104		80 - 125
4-Bromofluorobenzene (Surr)	105		78 - 118
Dibromofluoromethane (Surr)	99		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-0502**

Lab Sample ID: 280-1377-25
Client Matrix: Water

Date Sampled: 03/11/2010 0920
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4805.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0022		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0022			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
trans-1,2-Dichloroethene	0.78	J	0.15	1.0
1,1-Dichloroethene	0.38	J	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.44	J	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.44	J	0.17	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-0502**

Lab Sample ID: 280-1377-25
Client Matrix: Water

Date Sampled: 03/11/2010 0920
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4805.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0022		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0022			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	109		70 - 127
Toluene-d8 (Surr)	104		80 - 125
4-Bromofluorobenzene (Surr)	106		78 - 118
Dibromofluoromethane (Surr)	105		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-0502**

Lab Sample ID: 280-1377-25
Client Matrix: Water

Date Sampled: 03/11/2010 0920
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7993	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3055.D
Dilution:	1.0		Initial Weight/Volume:	3 mL
Date Analyzed:	03/18/2010 2356		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 2356			

Analyte	Result (ug/L)	Qualifier	MDL	RL
cis-1,2-Dichloroethene	66		1.0	6.7
Vinyl chloride	130		2.7	6.7

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	94		70 - 127
Toluene-d8 (Surr)	95		80 - 125
4-Bromofluorobenzene (Surr)	91		78 - 118
Dibromofluoromethane (Surr)	96		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-0503**

Lab Sample ID: 280-1377-26
Client Matrix: Water

Date Sampled: 03/11/2010 1020
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4806.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0043		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0043			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.20	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-0503**

Lab Sample ID: 280-1377-26
Client Matrix: Water

Date Sampled: 03/11/2010 1020
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4806.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0043		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0043			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.41	J	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.52	J	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	111		70 - 127
Toluene-d8 (Surr)	99		80 - 125
4-Bromofluorobenzene (Surr)	104		78 - 118
Dibromofluoromethane (Surr)	102		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-2867**

Lab Sample ID: 280-1377-27
Client Matrix: Water

Date Sampled: 03/11/2010 1625
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4807.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0103		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0103			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	2.3		0.15	1.0
trans-1,2-Dichloroethene	0.31	J	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-2867**

Lab Sample ID: 280-1377-27
Client Matrix: Water

Date Sampled: 03/11/2010 1625
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4807.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0103		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0103			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.23	J	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	3.6		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	112		70 - 127
Toluene-d8 (Surr)	97		80 - 125
4-Bromofluorobenzene (Surr)	109		78 - 118
Dibromofluoromethane (Surr)	106		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M003**

Lab Sample ID: 280-1377-28
Client Matrix: Water

Date Sampled: 03/11/2010 1525
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4809.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0144		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0144			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	1.5		0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	1.1		0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M003**

Lab Sample ID: 280-1377-28
Client Matrix: Water

Date Sampled: 03/11/2010 1525
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4809.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0144		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0144			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.39	J	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	109		70 - 127
Toluene-d8 (Surr)	100		80 - 125
4-Bromofluorobenzene (Surr)	108		78 - 118
Dibromofluoromethane (Surr)	100		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M005**

Lab Sample ID: 280-1377-29
Client Matrix: Water

Date Sampled: 03/11/2010 1610
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4810.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0204		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0204			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M005**

Lab Sample ID: 280-1377-29
Client Matrix: Water

Date Sampled: 03/11/2010 1610
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4810.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0204		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0204			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.39	J	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	115		70 - 127
Toluene-d8 (Surr)	100		80 - 125
4-Bromofluorobenzene (Surr)	108		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M015**

Lab Sample ID: 280-1377-30
Client Matrix: Water

Date Sampled: 03/11/2010 1100
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7993	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3056.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/19/2010 0016		Final Weight/Volume:	20 mL
Date Prepared:	03/19/2010 0016			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U *	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	2.5		0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U *	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M015**

Lab Sample ID: 280-1377-30
Client Matrix: Water

Date Sampled: 03/11/2010 1100
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7993	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3056.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/19/2010 0016		Final Weight/Volume:	20 mL
Date Prepared:	03/19/2010 0016			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	5.5		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	88		70 - 127
Toluene-d8 (Surr)	93		80 - 125
4-Bromofluorobenzene (Surr)	90		78 - 118
Dibromofluoromethane (Surr)	93		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M035**

Lab Sample ID: 280-1377-31
Client Matrix: Water

Date Sampled: 03/11/2010 1425
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4812.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0245		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0245			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	2.1		0.15	1.0
trans-1,2-Dichloroethene	0.30	J	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M035**

Lab Sample ID: 280-1377-31
Client Matrix: Water

Date Sampled: 03/11/2010 1425
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4812.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0245		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0245			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.22	J	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	3.1		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	113		70 - 127
Toluene-d8 (Surr)	99		80 - 125
4-Bromofluorobenzene (Surr)	98		78 - 118
Dibromofluoromethane (Surr)	105		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M38D**

Lab Sample ID: 280-1377-32
Client Matrix: Water

Date Sampled: 03/11/2010 1335
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4813.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0305		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0305			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M38D**

Lab Sample ID: 280-1377-32
Client Matrix: Water

Date Sampled: 03/11/2010 1335
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-7831	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR4813.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0305		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0305			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	116		70 - 127
Toluene-d8 (Surr)	97		80 - 125
4-Bromofluorobenzene (Surr)	101		78 - 118
Dibromofluoromethane (Surr)	106		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2874**

Lab Sample ID: 280-1377-33
Client Matrix: Water

Date Sampled: 03/11/2010 0806
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8149	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3022.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1024		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1024			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2874**

Lab Sample ID: 280-1377-33
Client Matrix: Water

Date Sampled: 03/11/2010 0806
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8149	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3022.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1024		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1024			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	76		70 - 127
Toluene-d8 (Surr)	93		80 - 125
4-Bromofluorobenzene (Surr)	89		78 - 118
Dibromofluoromethane (Surr)	84		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2881**

Lab Sample ID: 280-1377-34
Client Matrix: Water

Date Sampled: 03/11/2010 0800
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8149	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3023.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1045		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1045			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2881**

Lab Sample ID: 280-1377-34
Client Matrix: Water

Date Sampled: 03/11/2010 0800
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8149	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3023.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1045		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1045			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	83		70 - 127
Toluene-d8 (Surr)	94		80 - 125
4-Bromofluorobenzene (Surr)	89		78 - 118
Dibromofluoromethane (Surr)	88		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2888**

Lab Sample ID: 280-1377-35
Client Matrix: Water

Date Sampled: 03/11/2010 0800
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8149	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3024.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1105		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1105			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2888**

Lab Sample ID: 280-1377-35
Client Matrix: Water

Date Sampled: 03/11/2010 0800
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8149	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3024.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1105		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1105			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	82		70 - 127
Toluene-d8 (Surr)	90		80 - 125
4-Bromofluorobenzene (Surr)	87		78 - 118
Dibromofluoromethane (Surr)	86		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0554A**

Lab Sample ID: 280-1377-36
Client Matrix: Water

Date Sampled: 03/12/2010 1505
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8149	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3025.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1126		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1126			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0554A**

Lab Sample ID: 280-1377-36
Client Matrix: Water

Date Sampled: 03/12/2010 1505
Date Received: 03/16/2010 0930

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8149	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3025.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1126		Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1126			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	87		70 - 127
Toluene-d8 (Surr)	89		80 - 125
4-Bromofluorobenzene (Surr)	89		78 - 118
Dibromofluoromethane (Surr)	88		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0520**

Lab Sample ID: 280-1377-17
Client Matrix: Water

Date Sampled: 03/11/2010 1345
Date Received: 03/16/2010 0930

6010B Metals (ICP)

Method:	6010B	Analysis Batch: 280-7860		Instrument ID:	MT_026
Preparation:	3010A	Prep Batch: 280-7561		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/18/2010 2238			Final Weight/Volume:	50 mL
Date Prepared:	03/17/2010 1530				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Aluminum	250		18	100
Iron	520		22	100

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0530**

Lab Sample ID: 280-1377-18
Client Matrix: Water

Date Sampled: 03/11/2010 1015
Date Received: 03/16/2010 0930

6010B Metals (ICP)

Method:	6010B	Analysis Batch: 280-7860		Instrument ID:	MT_026
Preparation:	3010A	Prep Batch: 280-7561		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/18/2010 2300			Final Weight/Volume:	50 mL
Date Prepared:	03/17/2010 1530				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Aluminum	1300		18	100
Iron	3600		22	100

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0534**

Lab Sample ID: 280-1377-19
Client Matrix: Water

Date Sampled: 03/11/2010 1430
Date Received: 03/16/2010 0930

6010B Metals (ICP)

Method:	6010B	Analysis Batch: 280-7860		Instrument ID:	MT_026
Preparation:	3010A	Prep Batch: 280-7561		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/18/2010 2303			Final Weight/Volume:	50 mL
Date Prepared:	03/17/2010 1530				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Aluminum	480		18	100
Iron	320		22	100

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0535**

Lab Sample ID: 280-1377-20
Client Matrix: Water

Date Sampled: 03/11/2010 1100
Date Received: 03/16/2010 0930

6010B Metals (ICP)

Method:	6010B	Analysis Batch: 280-7860		Instrument ID:	MT_026
Preparation:	3010A	Prep Batch: 280-7561		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/18/2010 2306			Final Weight/Volume:	50 mL
Date Prepared:	03/17/2010 1530				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Aluminum	720		18	100
Iron	330		22	100

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0568**

Lab Sample ID: 280-1377-21
Client Matrix: Water

Date Sampled: 03/11/2010 1610
Date Received: 03/16/2010 0930

6010B Metals (ICP)

Method:	6010B	Analysis Batch: 280-7860		Instrument ID:	MT_026
Preparation:	3010A	Prep Batch: 280-7561		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/18/2010 2309			Final Weight/Volume:	50 mL
Date Prepared:	03/17/2010 1530				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Aluminum	290		18	100
Iron	750		22	100

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0569**

Lab Sample ID: 280-1377-22
Client Matrix: Water

Date Sampled: 03/11/2010 1515
Date Received: 03/16/2010 0930

6010B Metals (ICP)

Method:	6010B	Analysis Batch: 280-7860		Instrument ID:	MT_026
Preparation:	3010A	Prep Batch: 280-7561		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/18/2010 2312			Final Weight/Volume:	50 mL
Date Prepared:	03/17/2010 1530				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Aluminum	2200		18	100
Iron	3100		22	100

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Client Sample ID: **PIN15-2873**

Lab Sample ID: 280-1377-23
Client Matrix: Water

Date Sampled: 03/11/2010 1200
Date Received: 03/16/2010 0930

6010B Metals (ICP)

Method:	6010B	Analysis Batch: 280-7860		Instrument ID:	MT_026
Preparation:	3010A	Prep Batch: 280-7561		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/18/2010 2315			Final Weight/Volume:	50 mL
Date Prepared:	03/17/2010 1530				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Aluminum	1600		18	100
Iron	3500		22	100

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B Volatile Organic Compounds (GC/MS)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DBFM %Rec	DCA %Rec	TOL %Rec	BFB %Rec
280-1377-1	PIN12-0550-1	114	111	100	97
280-1377-2	PIN12-0550-2	116	115	100	94
280-1377-3	PIN12-0550-3	116	115	100	94
280-1377-4	PIN12-0551-1	118	116	101	95
280-1377-5	PIN12-0551-2	119	115	101	94
280-1377-6	PIN12-0551-3	119	113	100	94
280-1377-7	PIN12-0553A	85	87	85	85
280-1377-8	PIN12-0553B	92	91	93	87
280-1377-9	PIN12-0553C	90	90	89	88
280-1377-10	PIN12-S30B	91	90	90	89
280-1377-10 DL	PIN12-S30B DL	95	91	95	94
280-1377-11	PIN12-S33C	94	94	94	91
280-1377-11 DL	PIN12-S33C DL	93	90	92	91
280-1377-12	PIN12-S35B	90	89	86	89
280-1377-12	PIN12-S35B	91	89	90	89
280-1377-13	PIN12-S36B	103	109	98	101
280-1377-14	PIN12-S71B	103	107	98	100
280-1377-15	PIN12-S71C	104	111	97	105
280-1377-16	PIN12-S71D	102	108	98	100
280-1377-17	PIN15-0520	95	101	104	102
280-1377-18 DL	PIN15-0530 DL	90	86	87	87
280-1377-18	PIN15-0530	103	110	99	105
280-1377-19	PIN15-0534	88	87	89	87
280-1377-20	PIN15-0535	106	100	98	103
280-1377-21	PIN15-0568	102	105	100	100
280-1377-22	PIN15-0569	103	107	100	106
280-1377-23 DL	PIN15-2873 DL	93	85	91	91
280-1377-23	PIN15-2873	101	106	102	108
280-1377-24	PIN15-E001	99	101	104	105

Surrogate	Acceptance Limits
DBFM = Dibromofluoromethane (Surr)	77-119
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127
TOL = Toluene-d8 (Surr)	80-125
BFB = 4-Bromofluorobenzene (Surr)	78-118

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B Volatile Organic Compounds (GC/MS)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DBFM %Rec	DCA %Rec	TOL %Rec	BFB %Rec
280-1377-25	PIN20-0502	96	94	95	91
280-1377-25	PIN20-0502	105	109	104	106
280-1377-26	PIN20-0503	102	111	99	104
280-1377-27	PIN20-2867	106	112	97	109
280-1377-28	PIN20-M003	100	109	100	108
280-1377-29	PIN20-M005	104	115	100	108
280-1377-30	PIN20-M015	93	88	93	90
280-1377-31	PIN20-M035	105	113	99	98
280-1377-32	PIN20-M38D	106	116	97	101
280-1377-33	PIN99-2874	84	76	93	89
280-1377-34	PIN99-2881	88	83	94	89
280-1377-35	PIN99-2888	86	82	90	87
280-1377-36	PIN12-0554A	88	87	89	89
MB 280-7614/6		88	86	86	88
MB 280-7816/6		110	98	102	94
MB 280-7831/6		103	105	102	102
MB 280-7833/5		86	84	88	85
MB 280-7993/6		85	80	92	90
MB 280-8149/6		91	87	92	91
MB 280-8159/6		91	88	90	89
LCS 280-7614/4		91	90	89	91
LCS 280-7816/4		111	96	106	97
LCS 280-7831/4		101	106	109	106
LCS 280-7833/4		89	86	89	90
LCS 280-7993/4		84	78	88	90
LCS 280-8149/4		86	82	89	88
LCS 280-8159/4		95	96	92	95
LCSD 280-7614/29		88	85	89	89
LCSD 280-7816/5		111	98	104	97
LCSD 280-7831/5		100	104	99	101

Surrogate	Acceptance Limits
DBFM = Dibromofluoromethane (Surr)	77-119
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127
TOL = Toluene-d8 (Surr)	80-125
BFB = 4-Bromofluorobenzene (Surr)	78-118

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B Volatile Organic Compounds (GC/MS)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DBFM %Rec	DCA %Rec	TOL %Rec	BFB %Rec
LCSD 280-7993/5		90	85	94	90
LCSD 280-8149/5		93	90	95	93
LCSD 280-8159/5		92	93	89	90
280-1377-10 MS	PIN12-S30B MS	86	83	87	88
280-1377-13 MS	PIN12-S36B MS	101	106	102	105
280-1404-B-1 MS		89	87	90	90
280-1177-C-13 MS		112	98	103	98
280-1281-G-2 MS		88	87	90	91
280-1443-C-6 MS		87	81	94	92
280-1469-F-1 MS		89	85	87	85
280-1377-10 MSD	PIN12-S30B MSD	89	86	91	89
280-1377-13 MSD	PIN12-S36B MSD	103	108	98	102
280-1404-B-1 MSD		89	89	86	87
280-1177-C-13 MSD		112	98	103	100
280-1281-G-2 MSD		89	85	91	94
280-1443-C-6 MSD		86	81	93	93
280-1469-F-1 MSD		89	87	88	95

Surrogate	Acceptance Limits
DBFM = Dibromofluoromethane (Surr)	77-119
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127
TOL = Toluene-d8 (Surr)	80-125
BFB = 4-Bromofluorobenzene (Surr)	78-118

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-7614

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-7614/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/17/2010 1003
Date Prepared: 03/17/2010 1003

Analysis Batch: 280-7614
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_G
Lab File ID: G2957.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-7614

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-7614/6	Analysis Batch:	280-7614	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G2957.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1003			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1003				

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.199	J	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	86	70 - 127
Toluene-d8 (Surr)	86	80 - 125
4-Bromofluorobenzene (Surr)	88	78 - 118
Dibromofluoromethane (Surr)	88	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Lab Control Sample/
Lab Control Sample Duplicate Recovery Report - Batch: 280-7614

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-7614/4	Analysis Batch:	280-7614	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G2955.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 0921			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 0921				

LCSD Lab Sample ID:	LCSD 280-7614/29	Analysis Batch:	280-7614	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G2954.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 0901			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 0901				

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	91	95	77 - 120	5	20		
Bromodichloromethane	89	84	78 - 120	6	20		
Carbon tetrachloride	96	108	80 - 120	12	21		
Chlorobenzene	99	100	78 - 120	1	20		
Chloroform	92	92	78 - 120	1	20		
1,3-Dichlorobenzene	103	103	75 - 120	0	20		
1,1-Dichloroethane	93	95	77 - 120	3	21		
trans-1,2-Dichloroethene	92	101	80 - 120	9	24		
1,1-Dichloroethene	102	116	68 - 133	12	20		
1,2-Dichloropropane	93	87	76 - 120	6	20		
Ethylbenzene	100	107	78 - 120	7	26		
Methylene Chloride	94	88	71 - 120	6	20		
Tetrachloroethene	96	108	77 - 120	12	20		
Toluene	92	94	73 - 120	3	20		
1,1,1-Trichloroethane	93	103	78 - 120	10	20		
Trichloroethene	95	102	78 - 122	7	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	90	85	70 - 127
Toluene-d8 (Surr)	89	89	80 - 125
4-Bromofluorobenzene (Surr)	91	89	78 - 118
Dibromofluoromethane (Surr)	91	88	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Laboratory Control/
Laboratory Duplicate Data Report - Batch: 280-7614

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-7614/4	Units:	ug/L	LCSD Lab Sample ID:	LCSD 280-7614/29
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/17/2010 0921			Date Analyzed:	03/17/2010 0901
Date Prepared:	03/17/2010 0921			Date Prepared:	03/17/2010 0901

Analyte	LCS Spike Amount	LCSD Spike Amount	LCS Result/Qual	LCSD Result/Qual
Benzene	5.00	5.00	4.54	4.75
Bromodichloromethane	5.00	5.00	4.44	4.18
Carbon tetrachloride	5.00	5.00	4.80	5.40
Chlorobenzene	5.00	5.00	4.97	5.02
Chloroform	5.00	5.00	4.62	4.59
1,3-Dichlorobenzene	5.00	5.00	5.13	5.13
1,1-Dichloroethane	5.00	5.00	4.64	4.77
trans-1,2-Dichloroethene	5.00	5.00	4.61	5.07
1,1-Dichloroethene	5.00	5.00	5.10	5.78
1,2-Dichloropropane	5.00	5.00	4.63	4.36
Ethylbenzene	5.00	5.00	5.01	5.35
Methylene Chloride	5.00	5.00	4.68	4.40
Tetrachloroethene	5.00	5.00	4.81	5.39
Toluene	5.00	5.00	4.59	4.72
1,1,1-Trichloroethane	5.00	5.00	4.67	5.17
Trichloroethene	5.00	5.00	4.75	5.11

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-7614

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1404-B-1 MS	Analysis Batch:	280-7614	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G2960.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1105			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1105				

MSD Lab Sample ID:	280-1404-B-1 MSD	Analysis Batch:	280-7614	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G2961.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1125			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1125				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	94	96	77 - 120	1	20		
Bromodichloromethane	87	90	78 - 120	4	20		
Carbon tetrachloride	105	108	80 - 120	2	21		
Chlorobenzene	101	102	78 - 120	1	20		
Chloroform	93	95	78 - 120	1	20		
1,3-Dichlorobenzene	103	102	75 - 120	1	20		
1,1-Dichloroethane	98	100	77 - 120	2	21		
trans-1,2-Dichloroethene	96	100	80 - 120	4	24		
1,1-Dichloroethene	105	107	68 - 133	1	20		
1,2-Dichloropropane	91	91	76 - 120	1	20		
Ethylbenzene	106	105	78 - 120	1	26		
Methylene Chloride	81	83	71 - 120	2	20		
Tetrachloroethene	105	104	77 - 120	1	20		
Toluene	95	96	73 - 120	0	20		
1,1,1-Trichloroethane	100	103	78 - 120	2	20		
Trichloroethene	99	95	78 - 122	3	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	87	89	70 - 127
Toluene-d8 (Surr)	90	86	80 - 125
4-Bromofluorobenzene (Surr)	90	87	78 - 118
Dibromofluoromethane (Surr)	89	89	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-7614

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1404-B-1 MS	Units: ug/L	MSD Lab Sample ID:	280-1404-B-1 MSD
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/17/2010 1105		Date Analyzed:	03/17/2010 1125
Date Prepared:	03/17/2010 1105		Date Prepared:	03/17/2010 1125

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	0.16	U	5.00	5.00	4.72	4.79
Bromodichloromethane	0.17	U	5.00	5.00	4.34	4.50
Carbon tetrachloride	0.19	U	5.00	5.00	5.25	5.38
Chlorobenzene	0.17	U	5.00	5.00	5.06	5.11
Chloroform	0.94	J	5.00	5.00	5.60	5.69
1,3-Dichlorobenzene	0.13	U	5.00	5.00	5.15	5.12
1,1-Dichloroethane	0.22	U	5.00	5.00	4.91	5.02
trans-1,2-Dichloroethene	0.15	U	5.00	5.00	4.82	5.00
1,1-Dichloroethene	6.3		5.00	5.00	11.6	11.7
1,2-Dichloropropane	0.18	U	5.00	5.00	4.54	4.57
Ethylbenzene	0.16	U	5.00	5.00	5.31	5.26
Methylene Chloride	0.32	U	5.00	5.00	4.05	4.15
Tetrachloroethene	0.26	J	5.00	5.00	5.52	5.48
Toluene	0.17	U	5.00	5.00	4.76	4.78
1,1,1-Trichloroethane	1.3		5.00	5.00	6.25	6.39
Trichloroethene	0.44	J	5.00	5.00	5.38	5.20

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-7816

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-7816/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/17/2010 1039
Date Prepared: 03/17/2010 1039

Analysis Batch: 280-7816
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_R1
Lab File ID: R7378.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.486	J	0.32	1.0

Quality Control Results

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-7816

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-7816/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/17/2010 1039
Date Prepared: 03/17/2010 1039

Analysis Batch: 280-7816
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_R1
Lab File ID: R7378.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	98	70 - 127
Toluene-d8 (Surr)	102	80 - 125
4-Bromofluorobenzene (Surr)	94	78 - 118
Dibromofluoromethane (Surr)	110	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Lab Control Sample/
Lab Control Sample Duplicate Recovery Report - Batch: 280-7816

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-7816/4	Analysis Batch:	280-7816	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7376.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 0959			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 0959				

LCSD Lab Sample ID:	LCSD 280-7816/5	Analysis Batch:	280-7816	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7377.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1019			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1019				

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	87	89	77 - 120	3	20		
Bromodichloromethane	85	88	78 - 120	3	20		
Carbon tetrachloride	91	94	80 - 120	4	21		
Chlorobenzene	87	89	78 - 120	3	20		
Chloroform	90	92	78 - 120	3	20		
1,3-Dichlorobenzene	85	88	75 - 120	4	20		
1,1-Dichloroethane	89	91	77 - 120	2	21		
trans-1,2-Dichloroethene	90	92	80 - 120	2	24		
1,1-Dichloroethene	90	94	68 - 133	5	20		
1,2-Dichloropropane	84	87	76 - 120	4	20		
Ethylbenzene	85	87	78 - 120	2	26		
Methylene Chloride	91	99	71 - 120	9	20		
Tetrachloroethene	86	89	77 - 120	4	20		
Toluene	84	88	73 - 120	5	20		
1,1,1-Trichloroethane	87	91	78 - 120	4	20		
Trichloroethene	86	88	78 - 122	3	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	96	98	70 - 127
Toluene-d8 (Surr)	106	104	80 - 125
4-Bromofluorobenzene (Surr)	97	97	78 - 118
Dibromofluoromethane (Surr)	111	111	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Laboratory Control/
Laboratory Duplicate Data Report - Batch: 280-7816

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-7816/4	Units:	ug/L	LCSD Lab Sample ID:	LCSD 280-7816/5
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/17/2010 0959			Date Analyzed:	03/17/2010 1019
Date Prepared:	03/17/2010 0959			Date Prepared:	03/17/2010 1019

Analyte	LCS Spike Amount	LCSD Spike Amount	LCS Result/Qual	LCSD Result/Qual
Benzene	5.00	5.00	4.35	4.47
Bromodichloromethane	5.00	5.00	4.25	4.40
Carbon tetrachloride	5.00	5.00	4.53	4.69
Chlorobenzene	5.00	5.00	4.34	4.45
Chloroform	5.00	5.00	4.48	4.62
1,3-Dichlorobenzene	5.00	5.00	4.24	4.40
1,1-Dichloroethane	5.00	5.00	4.45	4.53
trans-1,2-Dichloroethene	5.00	5.00	4.51	4.60
1,1-Dichloroethene	5.00	5.00	4.50	4.72
1,2-Dichloropropane	5.00	5.00	4.20	4.36
Ethylbenzene	5.00	5.00	4.25	4.35
Methylene Chloride	5.00	5.00	4.54	4.96
Tetrachloroethene	5.00	5.00	4.31	4.47
Toluene	5.00	5.00	4.19	4.39
1,1,1-Trichloroethane	5.00	5.00	4.36	4.54
Trichloroethene	5.00	5.00	4.30	4.42

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-7816

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1177-C-13 MS	Analysis Batch:	280-7816	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7385.D
Dilution:	1.0			Initial Weight/Volume:	0.5 mL
Date Analyzed:	03/17/2010 1342			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1342				

MSD Lab Sample ID:	280-1177-C-13 MSD	Analysis Batch:	280-7816	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7386.D
Dilution:	1.0			Initial Weight/Volume:	0.5 mL
Date Analyzed:	03/17/2010 1401			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1401				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	91	88	77 - 120	4	20		
Bromodichloromethane	89	87	78 - 120	3	20		
Carbon tetrachloride	94	90	80 - 120	4	21		
Chlorobenzene	90	86	78 - 120	5	20		
Chloroform	94	90	78 - 120	4	20		
1,3-Dichlorobenzene	88	86	75 - 120	3	20		
1,1-Dichloroethane	94	100	77 - 120	3	21		
trans-1,2-Dichloroethene	90	89	80 - 120	2	24		
1,1-Dichloroethene	84	87	68 - 133	2	20		
1,2-Dichloropropane	90	86	76 - 120	4	20		
Ethylbenzene	88	83	78 - 120	5	26		
Methylene Chloride	94	92	71 - 120	3	20		
Tetrachloroethene	87	82	77 - 120	5	20		
Toluene	89	85	73 - 120	4	20		
1,1,1-Trichloroethane	85	93	78 - 120	3	20		
Trichloroethene	89	86	78 - 122	3	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	98	98	70 - 127
Toluene-d8 (Surr)	103	103	80 - 125
4-Bromofluorobenzene (Surr)	98	100	78 - 118
Dibromofluoromethane (Surr)	112	112	77 - 119

Quality Control Results

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-7816

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1177-C-13 MS	Units: ug/L	MSD Lab Sample ID:	280-1177-C-13 MSD
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/17/2010 1342		Date Analyzed:	03/17/2010 1401
Date Prepared:	03/17/2010 1342		Date Prepared:	03/17/2010 1401

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	0.64	U	200	200	182	175
Bromodichloromethane	0.68	U	200	200	179	174
Carbon tetrachloride	0.76	U	200	200	187	181
Chlorobenzene	0.68	U	200	200	181	172
Chloroform	4.4		200	200	192	185
1,3-Dichlorobenzene	0.52	U	200	200	177	171
1,1-Dichloroethane	280		200	200	472	484
trans-1,2-Dichloroethene	2.5	J	200	200	183	180
1,1-Dichloroethene	120		200	200	291	298
1,2-Dichloropropane	4.0	J	200	200	183	176
Ethylbenzene	0.64	U	200	200	175	167
Methylene Chloride	1.3	J	200	200	190	185
Tetrachloroethene	23		200	200	198	187
Toluene	0.68	U	200	200	178	171
1,1,1-Trichloroethane	250		200	200	425	439
Trichloroethene	15		200	200	193	187

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-7831

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-7831/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/17/2010 1833
Date Prepared: 03/17/2010 1833

Analysis Batch: 280-7831
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_R2
Lab File ID: RR4788.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Quality Control Results

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-7831

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-7831/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/17/2010 1833
Date Prepared: 03/17/2010 1833

Analysis Batch: 280-7831
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_R2
Lab File ID: RR4788.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	105	70 - 127
Toluene-d8 (Surr)	102	80 - 125
4-Bromofluorobenzene (Surr)	102	78 - 118
Dibromofluoromethane (Surr)	103	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Lab Control Sample/
Lab Control Sample Duplicate Recovery Report - Batch: 280-7831

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-7831/4	Analysis Batch:	280-7831	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR4785.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1732			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1732				

LCSD Lab Sample ID:	LCSD 280-7831/5	Analysis Batch:	280-7831	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR4786.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1753			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1753				

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	84	92	77 - 120	9	20		
Bromodichloromethane	87	97	78 - 120	11	20		
Carbon tetrachloride	106	114	80 - 120	7	21		
Chlorobenzene	83	93	78 - 120	11	20		
Chloroform	93	105	78 - 120	12	20		
1,3-Dichlorobenzene	84	95	75 - 120	13	20		
1,1-Dichloroethane	93	103	77 - 120	10	21		
trans-1,2-Dichloroethene	92	102	80 - 120	11	24		
1,1-Dichloroethene	103	113	68 - 133	10	20		
1,2-Dichloropropane	85	97	76 - 120	14	20		
Ethylbenzene	90	98	78 - 120	9	26		
Methylene Chloride	81	92	71 - 120	12	20		
Tetrachloroethene	92	96	77 - 120	3	20		
Toluene	91	102	73 - 120	11	20		
1,1,1-Trichloroethane	108	115	78 - 120	7	20		
Trichloroethene	93	104	78 - 122	11	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	106	104	70 - 127
Toluene-d8 (Surr)	109	99	80 - 125
4-Bromofluorobenzene (Surr)	106	101	78 - 118
Dibromofluoromethane (Surr)	101	100	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Laboratory Control/
Laboratory Duplicate Data Report - Batch: 280-7831

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-7831/4	Units:	ug/L	LCSD Lab Sample ID:	LCSD 280-7831/5
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/17/2010 1732			Date Analyzed:	03/17/2010 1753
Date Prepared:	03/17/2010 1732			Date Prepared:	03/17/2010 1753

Analyte	LCS Spike Amount	LCSD Spike Amount	LCS Result/Qual	LCSD Result/Qual
Benzene	5.00	5.00	4.21	4.61
Bromodichloromethane	5.00	5.00	4.34	4.86
Carbon tetrachloride	5.00	5.00	5.28	5.68
Chlorobenzene	5.00	5.00	4.17	4.67
Chloroform	5.00	5.00	4.66	5.23
1,3-Dichlorobenzene	5.00	5.00	4.20	4.77
1,1-Dichloroethane	5.00	5.00	4.67	5.17
trans-1,2-Dichloroethene	5.00	5.00	4.59	5.11
1,1-Dichloroethene	5.00	5.00	5.13	5.65
1,2-Dichloropropane	5.00	5.00	4.24	4.86
Ethylbenzene	5.00	5.00	4.50	4.91
Methylene Chloride	5.00	5.00	4.06	4.58
Tetrachloroethene	5.00	5.00	4.62	4.78
Toluene	5.00	5.00	4.57	5.09
1,1,1-Trichloroethane	5.00	5.00	5.39	5.77
Trichloroethene	5.00	5.00	4.65	5.19

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-7831

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1377-13	Analysis Batch:	280-7831	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR4790.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1914			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1914				

MSD Lab Sample ID:	280-1377-13	Analysis Batch:	280-7831	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR4791.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/17/2010 1934			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 1934				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	82	92	77 - 120	12	20		
Bromodichloromethane	85	94	78 - 120	10	20		
Carbon tetrachloride	101	112	80 - 120	10	21		
Chlorobenzene	84	91	78 - 120	8	20		
Chloroform	95	103	78 - 120	8	20		
1,3-Dichlorobenzene	83	90	75 - 120	8	20		
1,1-Dichloroethane	98	105	77 - 120	7	21		
trans-1,2-Dichloroethene	91	102	80 - 120	11	24		
1,1-Dichloroethene	106	116	68 - 133	9	20		
1,2-Dichloropropane	84	92	76 - 120	10	20		
Ethylbenzene	86	94	78 - 120	8	26		
Methylene Chloride	82	95	71 - 120	15	20		
Tetrachloroethene	87	93	77 - 120	7	20		
Toluene	90	100	73 - 120	11	20		
1,1,1-Trichloroethane	106	118	78 - 120	10	20		
Trichloroethene	93	101	78 - 122	7	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	106	108	70 - 127
Toluene-d8 (Surr)	102	98	80 - 125
4-Bromofluorobenzene (Surr)	105	102	78 - 118
Dibromofluoromethane (Surr)	101	103	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-7831

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1377-13	Units: ug/L	MSD Lab Sample ID:	280-1377-13
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/17/2010 1914		Date Analyzed:	03/17/2010 1934
Date Prepared:	03/17/2010 1914		Date Prepared:	03/17/2010 1934

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	0.16	U	5.00	5.00	4.08	4.58
Bromodichloromethane	0.17	U	5.00	5.00	4.24	4.70
Carbon tetrachloride	0.19	U	5.00	5.00	5.03	5.58
Chlorobenzene	0.17	U	5.00	5.00	4.22	4.55
Chloroform	0.16	U	5.00	5.00	4.74	5.16
1,3-Dichlorobenzene	0.13	U	5.00	5.00	4.17	4.50
1,1-Dichloroethane	0.22	U	5.00	5.00	4.90	5.27
trans-1,2-Dichloroethene	0.15	U	5.00	5.00	4.56	5.12
1,1-Dichloroethene	0.23	U	5.00	5.00	5.31	5.80
1,2-Dichloropropane	0.18	U	5.00	5.00	4.18	4.61
Ethylbenzene	0.16	U	5.00	5.00	4.30	4.68
Methylene Chloride	0.32	U	5.00	5.00	4.09	4.77
Tetrachloroethene	0.20	U	5.00	5.00	4.34	4.64
Toluene	0.17	U	5.00	5.00	4.49	5.00
1,1,1-Trichloroethane	0.16	U	5.00	5.00	5.32	5.88
Trichloroethene	0.16	U	5.00	5.00	4.67	5.03

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-7833

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-7833/5
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/17/2010 2049
Date Prepared: 03/17/2010 2049

Analysis Batch: 280-7833
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_G
Lab File ID: G2988.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-7833

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-7833/5
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/17/2010 2049
Date Prepared: 03/17/2010 2049

Analysis Batch: 280-7833
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_G
Lab File ID: G2988.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	84	70 - 127
Toluene-d8 (Surr)	88	80 - 125
4-Bromofluorobenzene (Surr)	85	78 - 118
Dibromofluoromethane (Surr)	86	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Lab Control Sample - Batch: 280-7833

Method: 8260B
Preparation: 5030B

Lab Sample ID: LCS 280-7833/4
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/17/2010 2009
Date Prepared: 03/17/2010 2009

Analysis Batch: 280-7833
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_G
Lab File ID: G2986.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Benzene	5.00	4.79	96	77 - 120	
Bromodichloromethane	5.00	4.43	89	78 - 120	
Carbon tetrachloride	5.00	5.26	105	80 - 120	
Chlorobenzene	5.00	5.12	102	78 - 120	
Chloroform	5.00	4.76	95	78 - 120	
1,3-Dichlorobenzene	5.00	5.28	106	75 - 120	
1,1-Dichloroethane	5.00	4.82	96	77 - 120	
trans-1,2-Dichloroethene	5.00	5.08	102	80 - 120	
1,1-Dichloroethene	5.00	5.85	117	68 - 133	
1,2-Dichloropropane	5.00	4.70	94	76 - 120	
Ethylbenzene	5.00	5.34	107	78 - 120	
Methylene Chloride	5.00	4.05	81	71 - 120	
Tetrachloroethene	5.00	5.58	112	77 - 120	
Toluene	5.00	4.83	97	73 - 120	
1,1,1-Trichloroethane	5.00	5.18	104	78 - 120	
Trichloroethene	5.00	4.88	98	78 - 122	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	86	70 - 127
Toluene-d8 (Surr)	89	80 - 125
4-Bromofluorobenzene (Surr)	90	78 - 118
Dibromofluoromethane (Surr)	89	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-7833

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1281-G-2 MS	Analysis Batch:	280-7833	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G2991.D
Dilution:	1.0			Initial Weight/Volume:	0.5 mL
Date Analyzed:	03/17/2010 2150			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2150				

MSD Lab Sample ID:	280-1281-G-2 MSD	Analysis Batch:	280-7833	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G2992.D
Dilution:	1.0			Initial Weight/Volume:	0.5 mL
Date Analyzed:	03/17/2010 2211			Final Weight/Volume:	20 mL
Date Prepared:	03/17/2010 2211				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	98	97	77 - 120	1	20		
Bromodichloromethane	87	87	78 - 120	1	20		
Carbon tetrachloride	108	106	80 - 120	2	21		
Chlorobenzene	102	102	78 - 120	0	20		
Chloroform	95	96	78 - 120	1	20		
1,3-Dichlorobenzene	106	107	75 - 120	1	20		
1,1-Dichloroethane	99	97	77 - 120	2	21		
trans-1,2-Dichloroethene	104	102	80 - 120	2	24		
1,1-Dichloroethene	122	115	68 - 133	6	20		
1,2-Dichloropropane	93	93	76 - 120	0	20		
Ethylbenzene	108	108	78 - 120	0	26		
Methylene Chloride	82	78	71 - 120	5	20		
Tetrachloroethene	113	110	77 - 120	2	20		
Toluene	97	97	73 - 120	1	20		
1,1,1-Trichloroethane	106	103	78 - 120	3	20		
Trichloroethene	70	66	78 - 122	1	20	4	4

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	87	85	70 - 127
Toluene-d8 (Surr)	90	91	80 - 125
4-Bromofluorobenzene (Surr)	91	94	78 - 118
Dibromofluoromethane (Surr)	88	89	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-7833

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1281-G-2 MS	Units: ug/L	MSD Lab Sample ID:	280-1281-G-2 MSD
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/17/2010 2150		Date Analyzed:	03/17/2010 2211
Date Prepared:	03/17/2010 2150		Date Prepared:	03/17/2010 2211

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual		MSD Result/Qual	
Benzene	6.4	U	200	200	197		194	
Bromodichloromethane	6.8	U	200	200	173		174	
Carbon tetrachloride	7.6	U	200	200	216		212	
Chlorobenzene	6.8	U	200	200	204		204	
Chloroform	6.4	U	200	200	191		192	
1,3-Dichlorobenzene	5.2	U	200	200	212		215	
1,1-Dichloroethane	8.8	U	200	200	198		195	
trans-1,2-Dichloroethene	6.0	U	200	200	208		204	
1,1-Dichloroethene	9.2	U	200	200	244		231	
1,2-Dichloropropane	7.2	U	200	200	187		187	
Ethylbenzene	6.4	U	200	200	215		215	
Methylene Chloride	13	U	200	200	163		155	
Tetrachloroethene	8.0	U	200	200	225		221	
Toluene	6.8	U	200	200	195		193	
1,1,1-Trichloroethane	6.4	U	200	200	213		207	
Trichloroethene	1000		200	200	1170	4	1160	4

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-7993

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-7993/6	Analysis Batch:	280-7993	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3036.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1727			Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1727				

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-7993

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-7993/6	Analysis Batch:	280-7993	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3036.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1727			Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1727				

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	80	70 - 127
Toluene-d8 (Surr)	92	80 - 125
4-Bromofluorobenzene (Surr)	90	78 - 118
Dibromofluoromethane (Surr)	85	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Lab Control Sample/
Lab Control Sample Duplicate Recovery Report - Batch: 280-7993

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-7993/4	Analysis Batch:	280-7993	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3033.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1625			Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1625				

LCSD Lab Sample ID:	LCSD 280-7993/5	Analysis Batch:	280-7993	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3034.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1646			Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1646				

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	93	95	77 - 120	2	20		
Bromodichloromethane	82	84	78 - 120	2	20		
Carbon tetrachloride	99	99	80 - 120	0	21		
Chlorobenzene	100	100	78 - 120	0	20		
Chloroform	90	92	78 - 120	3	20		
1,3-Dichlorobenzene	103	103	75 - 120	1	20		
1,1-Dichloroethane	93	95	77 - 120	2	21		
trans-1,2-Dichloroethene	96	98	80 - 120	2	24		
1,1-Dichloroethene	111	115	68 - 133	3	20		
1,2-Dichloropropane	90	90	76 - 120	0	20		
Ethylbenzene	107	105	78 - 120	1	26		
Methylene Chloride	75	83	71 - 120	10	20		
Tetrachloroethene	107	107	77 - 120	0	20		
Toluene	93	94	73 - 120	2	20		
1,1,1-Trichloroethane	98	98	78 - 120	1	20		
Trichloroethene	93	96	78 - 122	3	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	78	85	70 - 127
Toluene-d8 (Surr)	88	94	80 - 125
4-Bromofluorobenzene (Surr)	90	90	78 - 118
Dibromofluoromethane (Surr)	84	90	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Laboratory Control/
Laboratory Duplicate Data Report - Batch: 280-7993

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-7993/4	Units:	ug/L	LCSD Lab Sample ID:	LCSD 280-7993/5
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/18/2010 1625			Date Analyzed:	03/18/2010 1646
Date Prepared:	03/18/2010 1625			Date Prepared:	03/18/2010 1646

Analyte	LCS Spike Amount	LCSD Spike Amount	LCS Result/Qual	LCSD Result/Qual
Benzene	5.00	5.00	4.64	4.75
Bromodichloromethane	5.00	5.00	4.11	4.20
Carbon tetrachloride	5.00	5.00	4.96	4.97
Chlorobenzene	5.00	5.00	4.99	5.00
Chloroform	5.00	5.00	4.48	4.62
1,3-Dichlorobenzene	5.00	5.00	5.17	5.13
1,1-Dichloroethane	5.00	5.00	4.65	4.73
trans-1,2-Dichloroethene	5.00	5.00	4.81	4.90
1,1-Dichloroethene	5.00	5.00	5.56	5.74
1,2-Dichloropropane	5.00	5.00	4.48	4.50
Ethylbenzene	5.00	5.00	5.33	5.27
Methylene Chloride	5.00	5.00	3.75	4.15
Tetrachloroethene	5.00	5.00	5.37	5.37
Toluene	5.00	5.00	4.64	4.72
1,1,1-Trichloroethane	5.00	5.00	4.88	4.91
Trichloroethene	5.00	5.00	4.66	4.79

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-7993

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1443-C-6 MS	Analysis Batch:	280-7993	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3038.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1807			Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1807				

MSD Lab Sample ID:	280-1443-C-6 MSD	Analysis Batch:	280-7993	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3039.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 1828			Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1828				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	95	95	77 - 120	0	20		
Bromodichloromethane	82	82	78 - 120	0	20		
Carbon tetrachloride	101	99	80 - 120	2	21		
Chlorobenzene	102	100	78 - 120	2	20		
Chloroform	91	91	78 - 120	1	20		
1,3-Dichlorobenzene	101	104	75 - 120	2	20		
1,1-Dichloroethane	94	94	77 - 120	0	21		
trans-1,2-Dichloroethene	99	98	80 - 120	2	24		
1,1-Dichloroethene	119	101	68 - 133	16	20		
1,2-Dichloropropane	90	90	76 - 120	0	20		
Ethylbenzene	108	106	78 - 120	2	26		
Methylene Chloride	77	75	71 - 120	3	20		
Tetrachloroethene	111	108	77 - 120	3	20		
Toluene	96	93	73 - 120	3	20		
1,1,1-Trichloroethane	100	96	78 - 120	4	20		
Trichloroethene	98	94	78 - 122	3	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	81	81	70 - 127
Toluene-d8 (Surr)	94	93	80 - 125
4-Bromofluorobenzene (Surr)	92	93	78 - 118
Dibromofluoromethane (Surr)	87	86	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-7993

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1443-C-6 MS	Units: ug/L	MSD Lab Sample ID:	280-1443-C-6 MSD
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/18/2010 1807		Date Analyzed:	03/18/2010 1828
Date Prepared:	03/18/2010 1807		Date Prepared:	03/18/2010 1828

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	0.16	U	5.00	5.00	4.76	4.74
Bromodichloromethane	0.17	U	5.00	5.00	4.09	4.09
Carbon tetrachloride	0.19	U	5.00	5.00	5.06	4.94
Chlorobenzene	0.17	U	5.00	5.00	5.10	4.98
Chloroform	0.16	U	5.00	5.00	4.57	4.55
1,3-Dichlorobenzene	0.13	U	5.00	5.00	5.05	5.18
1,1-Dichloroethane	0.22	U	5.00	5.00	4.70	4.70
trans-1,2-Dichloroethene	0.15	U	5.00	5.00	4.97	4.89
1,1-Dichloroethene	0.23	U	5.00	5.00	5.94	5.06
1,2-Dichloropropane	0.18	U	5.00	5.00	4.52	4.51
Ethylbenzene	0.16	U	5.00	5.00	5.39	5.30
Methylene Chloride	0.32	U	5.00	5.00	3.85	3.74
Tetrachloroethene	0.20	U	5.00	5.00	5.55	5.40
Toluene	0.17	U	5.00	5.00	4.79	4.66
1,1,1-Trichloroethane	0.16	U	5.00	5.00	4.99	4.81
Trichloroethene	0.16	U	5.00	5.00	4.88	4.72

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-8149

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-8149/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/18/2010 0852
Date Prepared: 03/18/2010 0852

Analysis Batch: 280-8149
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_G
Lab File ID: G3018.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-8149

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-8149/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/18/2010 0852
Date Prepared: 03/18/2010 0852

Analysis Batch: 280-8149
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_G
Lab File ID: G3018.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	87	70 - 127
Toluene-d8 (Surr)	92	80 - 125
4-Bromofluorobenzene (Surr)	91	78 - 118
Dibromofluoromethane (Surr)	91	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Lab Control Sample/
Lab Control Sample Duplicate Recovery Report - Batch: 280-8149

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-8149/4	Analysis Batch:	280-8149	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3015.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0750			Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0750				

LCSD Lab Sample ID:	LCSD 280-8149/5	Analysis Batch:	280-8149	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3016.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/18/2010 0811			Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 0811				

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	90	105	77 - 120	15	20		
Bromodichloromethane	81	94	78 - 120	15	20		
Carbon tetrachloride	99	115	80 - 120	15	21		
Chlorobenzene	95	109	78 - 120	14	20		
Chloroform	86	102	78 - 120	16	20		
1,3-Dichlorobenzene	98	112	75 - 120	13	20		
1,1-Dichloroethane	90	106	77 - 120	16	21		
trans-1,2-Dichloroethene	96	108	80 - 120	12	24		
1,1-Dichloroethene	113	126	68 - 133	11	20		
1,2-Dichloropropane	87	99	76 - 120	13	20		
Ethylbenzene	101	114	78 - 120	12	26		
Methylene Chloride	79	94	71 - 120	17	20		
Tetrachloroethene	103	119	77 - 120	14	20		
Toluene	89	105	73 - 120	17	20		
1,1,1-Trichloroethane	96	111	78 - 120	15	20		
Trichloroethene	92	106	78 - 122	14	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	82	90	70 - 127
Toluene-d8 (Surr)	89	95	80 - 125
4-Bromofluorobenzene (Surr)	88	93	78 - 118
Dibromofluoromethane (Surr)	86	93	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Laboratory Control/
Laboratory Duplicate Data Report - Batch: 280-8149

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-8149/4	Units:	ug/L	LCSD Lab Sample ID:	LCSD 280-8149/5
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/18/2010 0750			Date Analyzed:	03/18/2010 0811
Date Prepared:	03/18/2010 0750			Date Prepared:	03/18/2010 0811

Analyte	LCS Spike Amount	LCSD Spike Amount	LCS Result/Qual	LCSD Result/Qual
Benzene	10.0	10.0	9.02	10.5
Bromodichloromethane	10.0	10.0	8.12	9.40
Carbon tetrachloride	10.0	10.0	9.89	11.5
Chlorobenzene	10.0	10.0	9.53	10.9
Chloroform	10.0	10.0	8.65	10.2
1,3-Dichlorobenzene	10.0	10.0	9.82	11.2
1,1-Dichloroethane	10.0	10.0	9.04	10.6
trans-1,2-Dichloroethene	10.0	10.0	9.55	10.8
1,1-Dichloroethene	10.0	10.0	11.3	12.6
1,2-Dichloropropane	10.0	10.0	8.69	9.93
Ethylbenzene	10.0	10.0	10.1	11.4
Methylene Chloride	10.0	10.0	7.89	9.35
Tetrachloroethene	10.0	10.0	10.3	11.9
Toluene	10.0	10.0	8.89	10.5
1,1,1-Trichloroethane	10.0	10.0	9.58	11.1
Trichloroethene	10.0	10.0	9.18	10.6

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-8149

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1377-10	Analysis Batch:	280-8149	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3027.D
Dilution:	1.0			Initial Weight/Volume:	4 mL
Date Analyzed:	03/18/2010 1310			Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1310				

MSD Lab Sample ID:	280-1377-10	Analysis Batch:	280-8149	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3028.D
Dilution:	1.0			Initial Weight/Volume:	4 mL
Date Analyzed:	03/18/2010 1331			Final Weight/Volume:	20 mL
Date Prepared:	03/18/2010 1331				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	96	97	77 - 120	1	20		
Bromodichloromethane	86	87	78 - 120	1	20		
Carbon tetrachloride	103	104	80 - 120	1	21		
Chlorobenzene	101	103	78 - 120	3	20		
Chloroform	93	95	78 - 120	2	20		
1,3-Dichlorobenzene	105	105	75 - 120	0	20		
1,1-Dichloroethane	96	96	77 - 120	0	21		
trans-1,2-Dichloroethene	102	102	80 - 120	1	24		
1,1-Dichloroethene	116	119	68 - 133	2	20		
1,2-Dichloropropane	92	94	76 - 120	3	20		
Ethylbenzene	105	106	78 - 120	0	26		
Methylene Chloride	90	91	71 - 120	1	20		
Tetrachloroethene	109	110	77 - 120	2	20		
Toluene	96	95	73 - 120	0	20		
1,1,1-Trichloroethane	102	104	78 - 120	1	20		
Trichloroethene	97	97	78 - 122	0	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	83	86	70 - 127
Toluene-d8 (Surr)	87	91	80 - 125
4-Bromofluorobenzene (Surr)	88	89	78 - 118
Dibromofluoromethane (Surr)	86	89	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-8149

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1377-10	Units: ug/L	MSD Lab Sample ID:	280-1377-10
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/18/2010 1310		Date Analyzed:	03/18/2010 1331
Date Prepared:	03/18/2010 1310		Date Prepared:	03/18/2010 1331

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	0.80	U	25.0	25.0	24.0	24.2
Bromodichloromethane	0.85	U	25.0	25.0	21.5	21.8
Carbon tetrachloride	0.95	U	25.0	25.0	25.7	26.0
Chlorobenzene	0.85	U	25.0	25.0	25.2	25.8
Chloroform	0.80	U	25.0	25.0	23.2	23.7
1,3-Dichlorobenzene	0.65	U	25.0	25.0	26.3	26.2
1,1-Dichloroethane	5.4		25.0	25.0	29.3	29.5
trans-1,2-Dichloroethene	15		25.0	25.0	40.8	41.0
1,1-Dichloroethene	3.0	J	25.0	25.0	32.1	32.8
1,2-Dichloropropane	0.90	U	25.0	25.0	22.9	23.5
Ethylbenzene	0.80	U	25.0	25.0	26.4	26.5
Methylene Chloride	1.6	U	25.0	25.0	22.6	22.8
Tetrachloroethene	1.0	U	25.0	25.0	27.1	27.6
Toluene	0.85	U	25.0	25.0	23.9	23.9
1,1,1-Trichloroethane	0.80	U	25.0	25.0	25.6	25.9
Trichloroethene	41		25.0	25.0	65.1	65.1

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-8159

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-8159/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/19/2010 1850
Date Prepared: 03/19/2010 1850

Analysis Batch: 280-8159
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_G
Lab File ID: G3089.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-8159

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-8159/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/19/2010 1850
Date Prepared: 03/19/2010 1850

Analysis Batch: 280-8159
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_G
Lab File ID: G3089.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	88	70 - 127
Toluene-d8 (Surr)	90	80 - 125
4-Bromofluorobenzene (Surr)	89	78 - 118
Dibromofluoromethane (Surr)	91	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Lab Control Sample/
Lab Control Sample Duplicate Recovery Report - Batch: 280-8159

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-8159/4	Analysis Batch:	280-8159	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3086.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/19/2010 1748			Final Weight/Volume:	20 mL
Date Prepared:	03/19/2010 1748				

LCSD Lab Sample ID:	LCSD 280-8159/5	Analysis Batch:	280-8159	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3087.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/19/2010 1809			Final Weight/Volume:	20 mL
Date Prepared:	03/19/2010 1809				

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	93	91	77 - 120	2	20		
Bromodichloromethane	88	85	78 - 120	3	20		
Carbon tetrachloride	94	94	80 - 120	0	21		
Chlorobenzene	99	95	78 - 120	4	20		
Chloroform	94	89	78 - 120	6	20		
1,3-Dichlorobenzene	100	96	75 - 120	4	20		
1,1-Dichloroethane	93	92	77 - 120	2	21		
trans-1,2-Dichloroethene	92	91	80 - 120	1	24		
1,1-Dichloroethene	107	105	68 - 133	2	20		
1,2-Dichloropropane	93	89	76 - 120	4	20		
Ethylbenzene	101	98	78 - 120	3	26		
Methylene Chloride	90	79	71 - 120	13	20		
Tetrachloroethene	100	97	77 - 120	3	20		
Toluene	94	92	73 - 120	2	20		
1,1,1-Trichloroethane	98	96	78 - 120	1	20		
Trichloroethene	91	92	78 - 122	1	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	96	93	70 - 127
Toluene-d8 (Surr)	92	89	80 - 125
4-Bromofluorobenzene (Surr)	95	90	78 - 118
Dibromofluoromethane (Surr)	95	92	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Laboratory Control/
Laboratory Duplicate Data Report - Batch: 280-8159

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-8159/4	Units:	ug/L	LCSD Lab Sample ID:	LCSD 280-8159/5
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/19/2010 1748			Date Analyzed:	03/19/2010 1809
Date Prepared:	03/19/2010 1748			Date Prepared:	03/19/2010 1809

Analyte	LCS Spike Amount	LCSD Spike Amount	LCS Result/Qual	LCSD Result/Qual
Benzene	5.00	5.00	4.67	4.56
Bromodichloromethane	5.00	5.00	4.40	4.26
Carbon tetrachloride	5.00	5.00	4.71	4.72
Chlorobenzene	5.00	5.00	4.94	4.76
Chloroform	5.00	5.00	4.72	4.44
1,3-Dichlorobenzene	5.00	5.00	5.00	4.82
1,1-Dichloroethane	5.00	5.00	4.66	4.58
trans-1,2-Dichloroethene	5.00	5.00	4.61	4.56
1,1-Dichloroethene	5.00	5.00	5.36	5.23
1,2-Dichloropropane	5.00	5.00	4.65	4.45
Ethylbenzene	5.00	5.00	5.07	4.92
Methylene Chloride	5.00	5.00	4.49	3.93
Tetrachloroethene	5.00	5.00	5.00	4.84
Toluene	5.00	5.00	4.69	4.59
1,1,1-Trichloroethane	5.00	5.00	4.88	4.81
Trichloroethene	5.00	5.00	4.56	4.60

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-8159

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1469-F-1 MS	Analysis Batch:	280-8159	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3095.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/19/2010 2053			Final Weight/Volume:	20 mL
Date Prepared:	03/19/2010 2053				

MSD Lab Sample ID:	280-1469-F-1 MSD	Analysis Batch:	280-8159	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3096.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/19/2010 2114			Final Weight/Volume:	20 mL
Date Prepared:	03/19/2010 2114				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	96	94	77 - 120	3	20		
Bromodichloromethane	86	83	78 - 120	3	20		
Carbon tetrachloride	104	99	80 - 120	5	21		
Chlorobenzene	99	98	78 - 120	1	20		
Chloroform	95	92	78 - 120	4	20		
1,3-Dichlorobenzene	101	99	75 - 120	2	20		
1,1-Dichloroethane	98	95	77 - 120	4	21		
trans-1,2-Dichloroethene	98	94	80 - 120	5	24		
1,1-Dichloroethene	120	114	68 - 133	5	20		
1,2-Dichloropropane	92	88	76 - 120	5	20		
Ethylbenzene	105	102	78 - 120	3	26		
Methylene Chloride	78	80	71 - 120	3	20		
Tetrachloroethene	107	104	77 - 120	3	20		
Toluene	98	95	73 - 120	4	20		
1,1,1-Trichloroethane	105	101	78 - 120	4	20		
Trichloroethene	98	94	78 - 122	4	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	85	87	70 - 127
Toluene-d8 (Surr)	87	88	80 - 125
4-Bromofluorobenzene (Surr)	85	95	78 - 118
Dibromofluoromethane (Surr)	89	89	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-8159

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1469-F-1 MS	Units: ug/L	MSD Lab Sample ID:	280-1469-F-1 MSD
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/19/2010 2053		Date Analyzed:	03/19/2010 2114
Date Prepared:	03/19/2010 2053		Date Prepared:	03/19/2010 2114

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	0.16	U	5.00	5.00	4.82	4.69
Bromodichloromethane	0.17	U	5.00	5.00	4.29	4.17
Carbon tetrachloride	0.19	U	5.00	5.00	5.18	4.95
Chlorobenzene	0.17	U	5.00	5.00	4.97	4.91
Chloroform	0.16	U	5.00	5.00	4.77	4.60
1,3-Dichlorobenzene	0.13	U	5.00	5.00	5.04	4.94
1,1-Dichloroethane	0.22	U	5.00	5.00	4.91	4.73
trans-1,2-Dichloroethene	0.15	U	5.00	5.00	4.92	4.69
1,1-Dichloroethene	0.71	J	5.00	5.00	6.73	6.41
1,2-Dichloropropane	0.18	U	5.00	5.00	4.62	4.41
Ethylbenzene	0.16	U	5.00	5.00	5.27	5.10
Methylene Chloride	0.32	U	5.00	5.00	3.88	4.01
Tetrachloroethene	0.20	U	5.00	5.00	5.35	5.21
Toluene	0.17	U	5.00	5.00	4.92	4.74
1,1,1-Trichloroethane	0.16	U	5.00	5.00	5.26	5.04
Trichloroethene	0.16	U	5.00	5.00	4.91	4.71

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Method Blank - Batch: 280-7561

Method: 6010B
Preparation: 3010A

Lab Sample ID: MB 280-7561/1-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/18/2010 2233
Date Prepared: 03/17/2010 1530

Analysis Batch: 280-7860
Prep Batch: 280-7561
Units: ug/L

Instrument ID: MT_026
Lab File ID: N/A
Initial Weight/Volume: 50 mL
Final Weight/Volume: 50 mL

Analyte	Result	Qual	MDL	RL
Aluminum	18	U	18	100
Iron	22	U	22	100

Lab Control Sample - Batch: 280-7561

Method: 6010B
Preparation: 3010A

Lab Sample ID: LCS 280-7561/2-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/18/2010 2235
Date Prepared: 03/17/2010 1530

Analysis Batch: 280-7860
Prep Batch: 280-7561
Units: ug/L

Instrument ID: MT_026
Lab File ID: N/A
Initial Weight/Volume: 50 mL
Final Weight/Volume: 50 mL

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Aluminum	2000	2020	101	87 - 111	
Iron	1000	993	99	89 - 115	

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-7561

Method: 6010B
Preparation: 3010A

MS Lab Sample ID:	280-1377-17	Analysis Batch:	280-7860	Instrument ID:	MT_026
Client Matrix:	Water	Prep Batch:	280-7561	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/18/2010 2243			Final Weight/Volume:	50 mL
Date Prepared:	03/17/2010 1530				

MSD Lab Sample ID:	280-1377-17	Analysis Batch:	280-7860	Instrument ID:	MT_026
Client Matrix:	Water	Prep Batch:	280-7561	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/18/2010 2246			Final Weight/Volume:	50 mL
Date Prepared:	03/17/2010 1530				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Aluminum	107	110	83 - 119	2	25		
Iron	92	94	52 - 155	2	25		

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-7561

Method: 6010B
Preparation: 3010A

MS Lab Sample ID:	280-1377-17	Units: ug/L		MSD Lab Sample ID:	280-1377-17
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/18/2010 2243			Date Analyzed:	03/18/2010 2246
Date Prepared:	03/17/2010 1530			Date Prepared:	03/17/2010 1530

Analyte	Sample Result/Qual	MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Aluminum	250	2000	2000	2400	2450
Iron	520	1000	1000	1440	1460

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:280-7614					
LCS 280-7614/4	Lab Control Sample	T	Water	8260B	
LCSD 280-7614/29	Lab Control Sample Duplicate	T	Water	8260B	
MB 280-7614/6	Method Blank	T	Water	8260B	
280-1377-7	PIN12-0553A	T	Water	8260B	
280-1377-8	PIN12-0553B	T	Water	8260B	
280-1377-9	PIN12-0553C	T	Water	8260B	
280-1377-10	PIN12-S30B	T	Water	8260B	
280-1377-11	PIN12-S33C	T	Water	8260B	
280-1404-B-1 MS	Matrix Spike	T	Water	8260B	
280-1404-B-1 MSD	Matrix Spike Duplicate	T	Water	8260B	
Analysis Batch:280-7816					
LCS 280-7816/4	Lab Control Sample	T	Water	8260B	
LCSD 280-7816/5	Lab Control Sample Duplicate	T	Water	8260B	
MB 280-7816/6	Method Blank	T	Water	8260B	
280-1177-C-13 MS	Matrix Spike	T	Water	8260B	
280-1177-C-13 MSD	Matrix Spike Duplicate	T	Water	8260B	
280-1377-1	PIN12-0550-1	T	Water	8260B	
280-1377-2	PIN12-0550-2	T	Water	8260B	
280-1377-3	PIN12-0550-3	T	Water	8260B	
280-1377-4	PIN12-0551-1	T	Water	8260B	
280-1377-5	PIN12-0551-2	T	Water	8260B	
280-1377-6	PIN12-0551-3	T	Water	8260B	

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:280-7831					
LCS 280-7831/4	Lab Control Sample	T	Water	8260B	
LCSD 280-7831/5	Lab Control Sample Duplicate	T	Water	8260B	
MB 280-7831/6	Method Blank	T	Water	8260B	
280-1377-13	PIN12-S36B	T	Water	8260B	
280-1377-13MS	Matrix Spike	T	Water	8260B	
280-1377-13MSD	Matrix Spike Duplicate	T	Water	8260B	
280-1377-14	PIN12-S71B	T	Water	8260B	
280-1377-15	PIN12-S71C	T	Water	8260B	
280-1377-16	PIN12-S71D	T	Water	8260B	
280-1377-17	PIN15-0520	T	Water	8260B	
280-1377-18	PIN15-0530	T	Water	8260B	
280-1377-20	PIN15-0535	T	Water	8260B	
280-1377-21	PIN15-0568	T	Water	8260B	
280-1377-22	PIN15-0569	T	Water	8260B	
280-1377-23	PIN15-2873	T	Water	8260B	
280-1377-24	PIN15-E001	T	Water	8260B	
280-1377-25	PIN20-0502	T	Water	8260B	
280-1377-26	PIN20-0503	T	Water	8260B	
280-1377-27	PIN20-2867	T	Water	8260B	
280-1377-28	PIN20-M003	T	Water	8260B	
280-1377-29	PIN20-M005	T	Water	8260B	
280-1377-31	PIN20-M035	T	Water	8260B	
280-1377-32	PIN20-M38D	T	Water	8260B	
Analysis Batch:280-7833					
LCS 280-7833/4	Lab Control Sample	T	Water	8260B	
MB 280-7833/5	Method Blank	T	Water	8260B	
280-1281-G-2 MS	Matrix Spike	T	Water	8260B	
280-1281-G-2 MSD	Matrix Spike Duplicate	T	Water	8260B	
280-1377-12	PIN12-S35B	T	Water	8260B	
Analysis Batch:280-7993					
LCS 280-7993/4	Lab Control Sample	T	Water	8260B	
LCSD 280-7993/5	Lab Control Sample Duplicate	T	Water	8260B	
MB 280-7993/6	Method Blank	T	Water	8260B	
280-1377-25	PIN20-0502	T	Water	8260B	
280-1377-30	PIN20-M015	T	Water	8260B	
280-1443-C-6 MS	Matrix Spike	T	Water	8260B	
280-1443-C-6 MSD	Matrix Spike Duplicate	T	Water	8260B	

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:280-8149					
LCS 280-8149/4	Lab Control Sample	T	Water	8260B	
LCSD 280-8149/5	Lab Control Sample Duplicate	T	Water	8260B	
MB 280-8149/6	Method Blank	T	Water	8260B	
280-1377-10DL	PIN12-S30B	T	Water	8260B	
280-1377-10MS	Matrix Spike	T	Water	8260B	
280-1377-10MSD	Matrix Spike Duplicate	T	Water	8260B	
280-1377-11DL	PIN12-S33C	T	Water	8260B	
280-1377-19	PIN15-0534	T	Water	8260B	
280-1377-33	PIN99-2874	T	Water	8260B	
280-1377-34	PIN99-2881	T	Water	8260B	
280-1377-35	PIN99-2888	T	Water	8260B	
280-1377-36	PIN12-0554A	T	Water	8260B	
Analysis Batch:280-8159					
LCS 280-8159/4	Lab Control Sample	T	Water	8260B	
LCSD 280-8159/5	Lab Control Sample Duplicate	T	Water	8260B	
MB 280-8159/6	Method Blank	T	Water	8260B	
280-1377-18DL	PIN15-0530	T	Water	8260B	
280-1377-23DL	PIN15-2873	T	Water	8260B	
280-1469-F-1 MS	Matrix Spike	T	Water	8260B	
280-1469-F-1 MSD	Matrix Spike Duplicate	T	Water	8260B	

Report Basis
T = Total

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
Metals					
Prep Batch: 280-7561					
LCS 280-7561/2-A	Lab Control Sample	T	Water	3010A	
MB 280-7561/1-A	Method Blank	T	Water	3010A	
280-1377-17	PIN15-0520	T	Water	3010A	
280-1377-17MS	Matrix Spike	T	Water	3010A	
280-1377-17MSD	Matrix Spike Duplicate	T	Water	3010A	
280-1377-18	PIN15-0530	T	Water	3010A	
280-1377-19	PIN15-0534	T	Water	3010A	
280-1377-20	PIN15-0535	T	Water	3010A	
280-1377-21	PIN15-0568	T	Water	3010A	
280-1377-22	PIN15-0569	T	Water	3010A	
280-1377-23	PIN15-2873	T	Water	3010A	
Analysis Batch:280-7860					
LCS 280-7561/2-A	Lab Control Sample	T	Water	6010B	280-7561
MB 280-7561/1-A	Method Blank	T	Water	6010B	280-7561
280-1377-17	PIN15-0520	T	Water	6010B	280-7561
280-1377-17MS	Matrix Spike	T	Water	6010B	280-7561
280-1377-17MSD	Matrix Spike Duplicate	T	Water	6010B	280-7561
280-1377-18	PIN15-0530	T	Water	6010B	280-7561
280-1377-19	PIN15-0534	T	Water	6010B	280-7561
280-1377-20	PIN15-0535	T	Water	6010B	280-7561
280-1377-21	PIN15-0568	T	Water	6010B	280-7561
280-1377-22	PIN15-0569	T	Water	6010B	280-7561
280-1377-23	PIN15-2873	T	Water	6010B	280-7561

Report Basis
T = Total

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1377-1 **Client ID:** PIN12-0550-1

Sample Date/Time: 03/11/2010 09:28 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-1		280-7816		03/17/2010 16:40	1	TAL DEN	HEW
A:8260B	280-1377-B-1		280-7816		03/17/2010 16:40	1	TAL DEN	HEW

Lab ID: 280-1377-2 **Client ID:** PIN12-0550-2

Sample Date/Time: 03/11/2010 10:33 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-2		280-7816		03/17/2010 17:00	1	TAL DEN	HEW
A:8260B	280-1377-B-2		280-7816		03/17/2010 17:00	1	TAL DEN	HEW

Lab ID: 280-1377-3 **Client ID:** PIN12-0550-3

Sample Date/Time: 03/11/2010 11:58 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-3		280-7816		03/17/2010 17:20	1	TAL DEN	HEW
A:8260B	280-1377-B-3		280-7816		03/17/2010 17:20	1	TAL DEN	HEW

Lab ID: 280-1377-4 **Client ID:** PIN12-0551-1

Sample Date/Time: 03/11/2010 13:46 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-4		280-7816		03/17/2010 17:39	1	TAL DEN	HEW
A:8260B	280-1377-B-4		280-7816		03/17/2010 17:39	1	TAL DEN	HEW

Lab ID: 280-1377-5 **Client ID:** PIN12-0551-2

Sample Date/Time: 03/11/2010 14:24 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-A-5		280-7816		03/17/2010 18:00	1	TAL DEN	HEW
A:8260B	280-1377-A-5		280-7816		03/17/2010 18:00	1	TAL DEN	HEW

Lab ID: 280-1377-6 **Client ID:** PIN12-0551-3

Sample Date/Time: 03/11/2010 15:03 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-6		280-7816		03/17/2010 18:20	1	TAL DEN	HEW
A:8260B	280-1377-B-6		280-7816		03/17/2010 18:20	1	TAL DEN	HEW

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1377-7 **Client ID:** PIN12-0553A

Sample Date/Time: 03/12/2010 12:45 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-7		280-7614		03/17/2010 16:15	1	TAL DEN	TDJ
A:8260B	280-1377-B-7		280-7614		03/17/2010 16:15	1	TAL DEN	TDJ

Lab ID: 280-1377-8 **Client ID:** PIN12-0553B

Sample Date/Time: 03/12/2010 11:45 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-C-8		280-7614		03/17/2010 16:35	1	TAL DEN	TDJ
A:8260B	280-1377-C-8		280-7614		03/17/2010 16:35	1	TAL DEN	TDJ

Lab ID: 280-1377-9 **Client ID:** PIN12-0553C

Sample Date/Time: 03/12/2010 10:25 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-9		280-7614		03/17/2010 16:56	1	TAL DEN	TDJ
A:8260B	280-1377-B-9		280-7614		03/17/2010 16:56	1	TAL DEN	TDJ

Lab ID: 280-1377-10 **Client ID:** PIN12-S30B

Sample Date/Time: 03/12/2010 12:20 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-10		280-7614		03/17/2010 17:16	1	TAL DEN	TDJ
A:8260B	280-1377-B-10		280-7614		03/17/2010 17:16	1	TAL DEN	TDJ
P:5030B	280-1377-C-10	DL	280-8149		03/18/2010 09:13	1	TAL DEN	TDJ
A:8260B	280-1377-C-10	DL	280-8149		03/18/2010 09:13	1	TAL DEN	TDJ

Lab ID: 280-1377-10 MS **Client ID:** PIN12-S30B

Sample Date/Time: 03/12/2010 12:20 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-C-10 MS		280-8149		03/18/2010 13:10	1	TAL DEN	TDJ
A:8260B	280-1377-C-10 MS		280-8149		03/18/2010 13:10	1	TAL DEN	TDJ

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1377-10 MSD **Client ID:** PIN12-S30B

Sample Date/Time: 03/12/2010 12:20 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-C-10 MSD		280-8149		03/18/2010 13:31	1	TAL DEN	TDJ
A:8260B	280-1377-C-10 MSD		280-8149		03/18/2010 13:31	1	TAL DEN	TDJ

Lab ID: 280-1377-11 **Client ID:** PIN12-S33C

Sample Date/Time: 03/12/2010 10:50 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-11		280-7614		03/17/2010 17:37	1	TAL DEN	TDJ
A:8260B	280-1377-B-11		280-7614		03/17/2010 17:37	1	TAL DEN	TDJ
P:5030B	280-1377-C-11	DL	280-8149		03/18/2010 09:33	1	TAL DEN	TDJ
A:8260B	280-1377-C-11	DL	280-8149		03/18/2010 09:33	1	TAL DEN	TDJ

Lab ID: 280-1377-12 **Client ID:** PIN12-S35B

Sample Date/Time: 03/12/2010 13:02 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-12		280-7833		03/18/2010 04:21	1	TAL DEN	DPI
A:8260B	280-1377-B-12		280-7833		03/18/2010 04:21	1	TAL DEN	DPI
P:5030B	280-1377-B-12		280-7833		03/18/2010 04:41	1	TAL DEN	DPI
A:8260B	280-1377-B-12		280-7833		03/18/2010 04:41	1	TAL DEN	DPI

Lab ID: 280-1377-13 **Client ID:** PIN12-S36B

Sample Date/Time: 03/12/2010 11:36 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-13		280-7831		03/17/2010 18:53	1	TAL DEN	JDG
A:8260B	280-1377-B-13		280-7831		03/17/2010 18:53	1	TAL DEN	JDG

Lab ID: 280-1377-13 MS **Client ID:** PIN12-S36B

Sample Date/Time: 03/12/2010 11:36 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-13 MS		280-7831		03/17/2010 19:14	1	TAL DEN	JDG
A:8260B	280-1377-B-13 MS		280-7831		03/17/2010 19:14	1	TAL DEN	JDG

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1377-13 MSD **Client ID:** PIN12-S36B

Sample Date/Time: 03/12/2010 11:36 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-13 MSD		280-7831		03/17/2010 19:34	1	TAL DEN	JDG
A:8260B	280-1377-B-13 MSD		280-7831		03/17/2010 19:34	1	TAL DEN	JDG

Lab ID: 280-1377-14 **Client ID:** PIN12-S71B

Sample Date/Time: 03/12/2010 12:30 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-C-14		280-7831		03/17/2010 20:15	1	TAL DEN	JDG
A:8260B	280-1377-C-14		280-7831		03/17/2010 20:15	1	TAL DEN	JDG

Lab ID: 280-1377-15 **Client ID:** PIN12-S71C

Sample Date/Time: 03/12/2010 09:50 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-15		280-7831		03/17/2010 20:35	1	TAL DEN	JDG
A:8260B	280-1377-B-15		280-7831		03/17/2010 20:35	1	TAL DEN	JDG

Lab ID: 280-1377-16 **Client ID:** PIN12-S71D

Sample Date/Time: 03/12/2010 11:05 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-16		280-7831		03/17/2010 20:56	1	TAL DEN	JDG
A:8260B	280-1377-B-16		280-7831		03/17/2010 20:56	1	TAL DEN	JDG

Lab ID: 280-1377-17 **Client ID:** PIN15-0520

Sample Date/Time: 03/11/2010 13:45 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-C-17		280-7831		03/17/2010 21:17	1	TAL DEN	JDG
A:8260B	280-1377-C-17		280-7831		03/17/2010 21:17	1	TAL DEN	JDG
P:3010A	280-1377-A-17-A		280-7860	280-7561	03/17/2010 15:30	1	TAL DEN	CGG
A:6010B	280-1377-A-17-A		280-7860	280-7561	03/18/2010 22:38	1	TAL DEN	DW

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1377-17 **Client ID:** PIN15-0520

Sample Date/Time: 03/11/2010 13:45 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:3010A	280-1377-A-17-B MS		280-7860	280-7561	03/17/2010 15:30	1	TAL DEN	CGG
A:6010B	280-1377-A-17-B MS		280-7860	280-7561	03/18/2010 22:43	1	TAL DEN	DW

Lab ID: 280-1377-17 **Client ID:** PIN15-0520

Sample Date/Time: 03/11/2010 13:45 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:3010A	280-1377-A-17-C MSD		280-7860	280-7561	03/17/2010 15:30	1	TAL DEN	CGG
A:6010B	280-1377-A-17-C MSD		280-7860	280-7561	03/18/2010 22:46	1	TAL DEN	DW

Lab ID: 280-1377-18 **Client ID:** PIN15-0530

Sample Date/Time: 03/11/2010 10:15 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-C-18		280-7831		03/17/2010 21:37	1	TAL DEN	JDG
A:8260B	280-1377-C-18		280-7831		03/17/2010 21:37	1	TAL DEN	JDG
P:5030B	280-1377-D-18	DL	280-8159		03/19/2010 19:10	1	TAL DEN	DPI
A:8260B	280-1377-D-18	DL	280-8159		03/19/2010 19:10	1	TAL DEN	DPI
P:3010A	280-1377-A-18-A		280-7860	280-7561	03/17/2010 15:30	1	TAL DEN	CGG
A:6010B	280-1377-A-18-A		280-7860	280-7561	03/18/2010 23:00	1	TAL DEN	DW

Lab ID: 280-1377-19 **Client ID:** PIN15-0534

Sample Date/Time: 03/11/2010 14:30 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-C-19		280-8149		03/18/2010 10:03	1	TAL DEN	TDJ
A:8260B	280-1377-C-19		280-8149		03/18/2010 10:03	1	TAL DEN	TDJ
P:3010A	280-1377-A-19-A		280-7860	280-7561	03/17/2010 15:30	1	TAL DEN	CGG
A:6010B	280-1377-A-19-A		280-7860	280-7561	03/18/2010 23:03	1	TAL DEN	DW

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1377-20 **Client ID:** PIN15-0535

Sample Date/Time: 03/11/2010 11:00 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-D-20		280-7831		03/17/2010 21:57	1	TAL DEN	JDG
A:8260B	280-1377-D-20		280-7831		03/17/2010 21:57	1	TAL DEN	JDG
P:3010A	280-1377-A-20-A		280-7860	280-7561	03/17/2010 15:30	1	TAL DEN	CGG
A:6010B	280-1377-A-20-A		280-7860	280-7561	03/18/2010 23:06	1	TAL DEN	DW

Lab ID: 280-1377-21 **Client ID:** PIN15-0568

Sample Date/Time: 03/11/2010 16:10 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-C-21		280-7831		03/17/2010 22:18	1	TAL DEN	JDG
A:8260B	280-1377-C-21		280-7831		03/17/2010 22:18	1	TAL DEN	JDG
P:3010A	280-1377-A-21-A		280-7860	280-7561	03/17/2010 15:30	1	TAL DEN	CGG
A:6010B	280-1377-A-21-A		280-7860	280-7561	03/18/2010 23:09	1	TAL DEN	DW

Lab ID: 280-1377-22 **Client ID:** PIN15-0569

Sample Date/Time: 03/11/2010 15:15 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-22		280-7831		03/17/2010 22:38	1	TAL DEN	JDG
A:8260B	280-1377-B-22		280-7831		03/17/2010 22:38	1	TAL DEN	JDG
P:3010A	280-1377-A-22-A		280-7860	280-7561	03/17/2010 15:30	1	TAL DEN	CGG
A:6010B	280-1377-A-22-A		280-7860	280-7561	03/18/2010 23:12	1	TAL DEN	DW

Lab ID: 280-1377-23 **Client ID:** PIN15-2873

Sample Date/Time: 03/11/2010 12:00 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-C-23		280-7831		03/17/2010 23:20	1	TAL DEN	JDG
A:8260B	280-1377-C-23		280-7831		03/17/2010 23:20	1	TAL DEN	JDG
P:5030B	280-1377-D-23	DL	280-8159		03/19/2010 19:31	1	TAL DEN	DPI
A:8260B	280-1377-D-23	DL	280-8159		03/19/2010 19:31	1	TAL DEN	DPI
P:3010A	280-1377-A-23-A		280-7860	280-7561	03/17/2010 15:30	1	TAL DEN	CGG
A:6010B	280-1377-A-23-A		280-7860	280-7561	03/18/2010 23:15	1	TAL DEN	DW

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1377-24 **Client ID:** PIN15-E001

Sample Date/Time: 03/11/2010 11:15 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-C-24		280-7831		03/17/2010 23:40	1	TAL DEN	JDG
A:8260B	280-1377-C-24		280-7831		03/17/2010 23:40	1	TAL DEN	JDG

Lab ID: 280-1377-25 **Client ID:** PIN20-0502

Sample Date/Time: 03/11/2010 09:20 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-A-25		280-7831		03/18/2010 00:22	1	TAL DEN	JDG
A:8260B	280-1377-A-25		280-7831		03/18/2010 00:22	1	TAL DEN	JDG
P:5030B	280-1377-B-25		280-7993		03/18/2010 23:56	1	TAL DEN	DPI
A:8260B	280-1377-B-25		280-7993		03/18/2010 23:56	1	TAL DEN	DPI

Lab ID: 280-1377-26 **Client ID:** PIN20-0503

Sample Date/Time: 03/11/2010 10:20 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-A-26		280-7831		03/18/2010 00:43	1	TAL DEN	JDG
A:8260B	280-1377-A-26		280-7831		03/18/2010 00:43	1	TAL DEN	JDG

Lab ID: 280-1377-27 **Client ID:** PIN20-2867

Sample Date/Time: 03/11/2010 16:25 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-C-27		280-7831		03/18/2010 01:03	1	TAL DEN	JDG
A:8260B	280-1377-C-27		280-7831		03/18/2010 01:03	1	TAL DEN	JDG

Lab ID: 280-1377-28 **Client ID:** PIN20-M003

Sample Date/Time: 03/11/2010 15:25 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-A-28		280-7831		03/18/2010 01:44	1	TAL DEN	JDG
A:8260B	280-1377-A-28		280-7831		03/18/2010 01:44	1	TAL DEN	JDG

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1377-29 **Client ID:** PIN20-M005

Sample Date/Time: 03/11/2010 16:10 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-B-29		280-7831		03/18/2010 02:04	1	TAL DEN	JDG
A:8260B	280-1377-B-29		280-7831		03/18/2010 02:04	1	TAL DEN	JDG

Lab ID: 280-1377-30 **Client ID:** PIN20-M015

Sample Date/Time: 03/11/2010 11:00 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-C-30		280-7993		03/19/2010 00:16	1	TAL DEN	DPI
A:8260B	280-1377-C-30		280-7993		03/19/2010 00:16	1	TAL DEN	DPI

Lab ID: 280-1377-31 **Client ID:** PIN20-M035

Sample Date/Time: 03/11/2010 14:25 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-C-31		280-7831		03/18/2010 02:45	1	TAL DEN	JDG
A:8260B	280-1377-C-31		280-7831		03/18/2010 02:45	1	TAL DEN	JDG

Lab ID: 280-1377-32 **Client ID:** PIN20-M38D

Sample Date/Time: 03/11/2010 13:35 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-C-32		280-7831		03/18/2010 03:05	1	TAL DEN	JDG
A:8260B	280-1377-C-32		280-7831		03/18/2010 03:05	1	TAL DEN	JDG

Lab ID: 280-1377-33 **Client ID:** PIN99-2874

Sample Date/Time: 03/11/2010 08:06 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-A-33		280-8149		03/18/2010 10:24	1	TAL DEN	TDJ
A:8260B	280-1377-A-33		280-8149		03/18/2010 10:24	1	TAL DEN	TDJ

Lab ID: 280-1377-34 **Client ID:** PIN99-2881

Sample Date/Time: 03/11/2010 08:00 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-A-34		280-8149		03/18/2010 10:45	1	TAL DEN	TDJ
A:8260B	280-1377-A-34		280-8149		03/18/2010 10:45	1	TAL DEN	TDJ

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
SDG: 10022892

Laboratory Chronicle

Lab ID: **280-1377-35** **Client ID:** **PIN99-2888**

Sample Date/Time: 03/11/2010 08:00 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-A-35		280-8149		03/18/2010 11:05	1	TAL DEN	TDJ
A:8260B	280-1377-A-35		280-8149		03/18/2010 11:05	1	TAL DEN	TDJ

Lab ID: **280-1377-36** **Client ID:** **PIN12-0554A**

Sample Date/Time: 03/12/2010 15:05 Received Date/Time: 03/16/2010 09:30

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1377-A-36		280-8149		03/18/2010 11:26	1	TAL DEN	TDJ
A:8260B	280-1377-A-36		280-8149		03/18/2010 11:26	1	TAL DEN	TDJ

Lab ID: **MB** **Client ID:** **N/A**

Sample Date/Time: N/A Received Date/Time: N/A

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	MB 280-7614/6		280-7614		03/17/2010 10:03	1	TAL DEN	TDJ
A:8260B	MB 280-7614/6		280-7614		03/17/2010 10:03	1	TAL DEN	TDJ
P:5030B	MB 280-7816/6		280-7816		03/17/2010 10:39	1	TAL DEN	HEW
A:8260B	MB 280-7816/6		280-7816		03/17/2010 10:39	1	TAL DEN	HEW
P:5030B	MB 280-7831/6		280-7831		03/17/2010 18:33	1	TAL DEN	JDG
A:8260B	MB 280-7831/6		280-7831		03/17/2010 18:33	1	TAL DEN	JDG
P:5030B	MB 280-7833/5		280-7833		03/17/2010 20:49	1	TAL DEN	DPI
A:8260B	MB 280-7833/5		280-7833		03/17/2010 20:49	1	TAL DEN	DPI
P:5030B	MB 280-8149/6		280-8149		03/18/2010 08:52	1	TAL DEN	TDJ
A:8260B	MB 280-8149/6		280-8149		03/18/2010 08:52	1	TAL DEN	TDJ
P:5030B	MB 280-7993/6		280-7993		03/18/2010 17:27	1	TAL DEN	DPI
A:8260B	MB 280-7993/6		280-7993		03/18/2010 17:27	1	TAL DEN	DPI
P:5030B	MB 280-8159/6		280-8159		03/19/2010 18:50	1	TAL DEN	DPI
A:8260B	MB 280-8159/6		280-8159		03/19/2010 18:50	1	TAL DEN	DPI
P:3010A	MB 280-7561/1-A		280-7860	280-7561	03/17/2010 15:30	1	TAL DEN	CGG
A:6010B	MB 280-7561/1-A		280-7860	280-7561	03/18/2010 22:33	1	TAL DEN	DW

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
SDG: 10022892

Laboratory Chronicle

Lab ID: **LCS** **Client ID:** **N/A**

Sample Date/Time: N/A Received Date/Time: N/A

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	LCS 280-7614/4		280-7614		03/17/2010 09:21	1	TAL DEN	TDJ
A:8260B	LCS 280-7614/4		280-7614		03/17/2010 09:21	1	TAL DEN	TDJ
P:5030B	LCS 280-7816/4		280-7816		03/17/2010 09:59	1	TAL DEN	HEW
A:8260B	LCS 280-7816/4		280-7816		03/17/2010 09:59	1	TAL DEN	HEW
P:5030B	LCS 280-7831/4		280-7831		03/17/2010 17:32	1	TAL DEN	JDG
A:8260B	LCS 280-7831/4		280-7831		03/17/2010 17:32	1	TAL DEN	JDG
P:5030B	LCS 280-7833/4		280-7833		03/17/2010 20:09	1	TAL DEN	DPI
A:8260B	LCS 280-7833/4		280-7833		03/17/2010 20:09	1	TAL DEN	DPI
P:5030B	LCS 280-8149/4		280-8149		03/18/2010 07:50	1	TAL DEN	TDJ
A:8260B	LCS 280-8149/4		280-8149		03/18/2010 07:50	1	TAL DEN	TDJ
P:5030B	LCS 280-7993/4		280-7993		03/18/2010 16:25	1	TAL DEN	DPI
A:8260B	LCS 280-7993/4		280-7993		03/18/2010 16:25	1	TAL DEN	DPI
P:5030B	LCS 280-8159/4		280-8159		03/19/2010 17:48	1	TAL DEN	DPI
A:8260B	LCS 280-8159/4		280-8159		03/19/2010 17:48	1	TAL DEN	DPI
P:3010A	LCS 280-7561/2-A		280-7860	280-7561	03/17/2010 15:30	1	TAL DEN	CGG
A:6010B	LCS 280-7561/2-A		280-7860	280-7561	03/18/2010 22:35	1	TAL DEN	DW

Lab ID: **LCSD** **Client ID:** **N/A**

Sample Date/Time: N/A Received Date/Time: N/A

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	LCSD 280-7614/29		280-7614		03/17/2010 09:01	1	TAL DEN	TDJ
A:8260B	LCSD 280-7614/29		280-7614		03/17/2010 09:01	1	TAL DEN	TDJ
P:5030B	LCSD 280-7816/5		280-7816		03/17/2010 10:19	1	TAL DEN	HEW
A:8260B	LCSD 280-7816/5		280-7816		03/17/2010 10:19	1	TAL DEN	HEW
P:5030B	LCSD 280-7831/5		280-7831		03/17/2010 17:53	1	TAL DEN	JDG
A:8260B	LCSD 280-7831/5		280-7831		03/17/2010 17:53	1	TAL DEN	JDG
P:5030B	LCSD 280-8149/5		280-8149		03/18/2010 08:11	1	TAL DEN	TDJ
A:8260B	LCSD 280-8149/5		280-8149		03/18/2010 08:11	1	TAL DEN	TDJ
P:5030B	LCSD 280-7993/5		280-7993		03/18/2010 16:46	1	TAL DEN	DPI
A:8260B	LCSD 280-7993/5		280-7993		03/18/2010 16:46	1	TAL DEN	DPI
P:5030B	LCSD 280-8159/5		280-8159		03/19/2010 18:09	1	TAL DEN	DPI
A:8260B	LCSD 280-8159/5		280-8159		03/19/2010 18:09	1	TAL DEN	DPI

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
SDG: 10022892

Laboratory Chronicle

Lab ID: **MS** **Client ID:** **N/A**

Sample Date/Time: 03/15/2010 16:30 Received Date/Time: 03/16/2010 12:09

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1404-B-1 MS		280-7614		03/17/2010 11:05	1	TAL DEN	TDJ
A:8260B	280-1404-B-1 MS		280-7614		03/17/2010 11:05	1	TAL DEN	TDJ
P:5030B	280-1177-C-13 MS		280-7816		03/17/2010 13:42	1	TAL DEN	HEW
A:8260B	280-1177-C-13 MS		280-7816		03/17/2010 13:42	1	TAL DEN	HEW
P:5030B	280-1281-G-2 MS		280-7833		03/17/2010 21:50	1	TAL DEN	DPI
A:8260B	280-1281-G-2 MS		280-7833		03/17/2010 21:50	1	TAL DEN	DPI
P:5030B	280-1443-C-6 MS		280-7993		03/18/2010 18:07	1	TAL DEN	DPI
A:8260B	280-1443-C-6 MS		280-7993		03/18/2010 18:07	1	TAL DEN	DPI
P:5030B	280-1469-F-1 MS		280-8159		03/19/2010 20:53	1	TAL DEN	DPI
A:8260B	280-1469-F-1 MS		280-8159		03/19/2010 20:53	1	TAL DEN	DPI

Lab ID: **MSD** **Client ID:** **N/A**

Sample Date/Time: 03/15/2010 16:30 Received Date/Time: 03/16/2010 12:09

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1404-B-1 MSD		280-7614		03/17/2010 11:25	1	TAL DEN	TDJ
A:8260B	280-1404-B-1 MSD		280-7614		03/17/2010 11:25	1	TAL DEN	TDJ
P:5030B	280-1177-C-13 MSD		280-7816		03/17/2010 14:01	1	TAL DEN	HEW
A:8260B	280-1177-C-13 MSD		280-7816		03/17/2010 14:01	1	TAL DEN	HEW
P:5030B	280-1281-G-2 MSD		280-7833		03/17/2010 22:11	1	TAL DEN	DPI
A:8260B	280-1281-G-2 MSD		280-7833		03/17/2010 22:11	1	TAL DEN	DPI
P:5030B	280-1443-C-6 MSD		280-7993		03/18/2010 18:28	1	TAL DEN	DPI
A:8260B	280-1443-C-6 MSD		280-7993		03/18/2010 18:28	1	TAL DEN	DPI
P:5030B	280-1469-F-1 MSD		280-8159		03/19/2010 21:14	1	TAL DEN	DPI
A:8260B	280-1469-F-1 MSD		280-8159		03/19/2010 21:14	1	TAL DEN	DPI

Lab References:
TAL DEN = TestAmerica Denver

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1377-1

SDG No.: 10022892

Instrument ID: MSV_G Analysis Batch Number: 7372

Lab Sample ID: IC 280-7372/5 Client Sample ID:

Date Analyzed: 03/15/10 11:16 Lab File ID: G2840.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
1,4-Dioxane	8.71	Analyte not Identified by the Data System	reinhardt j	03/15/10 13:39

Lab Sample ID: IC 280-7372/7 Client Sample ID:

Date Analyzed: 03/15/10 11:57 Lab File ID: G2842.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Acrylonitrile	6.25	Split Peak	reinhardt j	03/15/10 13:27

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1377-1

SDG No.: 10022892

Instrument ID: MSV_G Analysis Batch Number: 8149

Lab Sample ID: CCV 280-8149/2 Client Sample ID:

Date Analyzed: 03/18/10 07:04 Lab File ID: G3013.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
n-Butanol	8.18	Not specified	jacksonto	03/22/10 14:25

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1377-1

SDG No.: 10022892

Instrument ID: MSV_R1 Analysis Batch Number: 5135

Lab Sample ID: IC 280-5135/4 Client Sample ID:

Date Analyzed: 02/19/10 13:54 Lab File ID: R6582.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Nitropropane	8.79	Analyte not Identified by the Data System	garciaj	02/19/10 17:15
Tetrahydrothiophene	9.89	Analyte not Identified by the Data System	garciaj	02/19/10 17:15
cis-1,4-Dichloro-2-butene	11.05	Analyte not Identified by the Data System	garciaj	02/19/10 17:16

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1377-1

SDG No.: 10022892

Instrument ID: MSV_R1 Analysis Batch Number: 8905

Lab Sample ID: IC 280-8905/2 Client Sample ID:

Date Analyzed: 03/26/10 07:45 Lab File ID: R7591.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
trans-1,3-Dichloropropene	9.34	Analyte not Identified by the Data System	tinkhams	03/26/10 11:59

Lab Sample ID: IC 280-8905/4 Client Sample ID:

Date Analyzed: 03/26/10 08:27 Lab File ID: R7593.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Ethanol	5.33	Analyte not Identified by the Data System	tinkhams	03/26/10 11:58

Lab Sample ID: IC 280-8905/10 Client Sample ID:

Date Analyzed: 03/26/10 10:39 Lab File ID: R7599.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Chloroethyl vinyl ether	8.79	Analyte not Identified by the Data System	waterlah	03/30/10 08:37

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1377-1

SDG No.: 10022892

Instrument ID: MSV_R2 Analysis Batch Number: 4173

Lab Sample ID: IC 280-4173/10 Client Sample ID:

Date Analyzed: 02/04/10 13:00 Lab File ID: RR3509.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Ethanol	4.04	Analyte not Identified by the Data System	waterlah	02/04/10 15:29

Lab Sample ID: IC 280-4173/11 Client Sample ID:

Date Analyzed: 02/04/10 13:21 Lab File ID: RR3510.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Ethanol	4.02	Split Peak	waterlah	02/04/10 15:29

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1377-1

SDG No.: 10022892

Instrument ID: MSV_R2 Analysis Batch Number: 6717

Lab Sample ID: IC 280-6717/5 Client Sample ID:

Date Analyzed: 03/08/10 10:28 Lab File ID: RR4519.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Isobutyl alcohol	6.17	Not specified	waterlah	03/08/10 12:44

Lab Sample ID: IC 280-6717/8 Client Sample ID:

Date Analyzed: 03/08/10 11:31 Lab File ID: RR4522.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Isobutyl alcohol	6.16	Not specified	waterlah	03/08/10 12:44

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1377-1

SDG No.: 10022892

Instrument ID: MSV_R2 Analysis Batch Number: 7831

Lab Sample ID: 280-1377-18 Client Sample ID: PIN15-0530

Date Analyzed: 03/17/10 21:37 Lab File ID: RR4797.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Trichloroethene	6.88	Not specified	rhoadesw	03/22/10 10:15

8260B

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1377-1

SDG No.: 10022892

Instrument ID: MSV_G Analysis Batch Number: 7372

Lab Sample ID: IC 280-7372/5 Client Sample ID:

Date Analyzed: 03/15/10 11:16 Lab File ID: G2840.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
1,4-Dioxane	8.71	Analyte not Identified by the Data System	reinhardt j	03/15/10 13:39

Lab Sample ID: IC 280-7372/7 Client Sample ID:

Date Analyzed: 03/15/10 11:57 Lab File ID: G2842.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Acrylonitrile	6.25	Split Peak	reinhardt j	03/15/10 13:27

4/6

8260B

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1377-1

SDG No.: 10022892

Instrument ID: MSV_G Analysis Batch Number: 8149

Lab Sample ID: CCV 280-8149/2 Client Sample ID:

Date Analyzed: 03/18/10 07:04 Lab File ID: G3013.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
n-Butanol	8.18	Not specified	jacksonto	03/22/10 14:25

4/6

8260B

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1377-1

SDG No.: 10022892

Instrument ID: MSV_R1 Analysis Batch Number: 5135

Lab Sample ID: IC 280-5135/4 Client Sample ID:

Date Analyzed: 02/19/10 13:54 Lab File ID: R6582.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Nitropropane	8.79	Analyte not Identified by the Data System	garciaj	02/19/10 17:15
Tetrahydrothiophene	9.89	Analyte not Identified by the Data System	garciaj	02/19/10 17:15
cis-1,4-Dichloro-2-butene	11.05	Analyte not Identified by the Data System	garciaj	02/19/10 17:16

4/6

8260B

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1377-1

SDG No.: 10022892

Instrument ID: MSV_R1 Analysis Batch Number: 8905

Lab Sample ID: IC 280-8905/2 Client Sample ID:

Date Analyzed: 03/26/10 07:45 Lab File ID: R7591.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
trans-1,3-Dichloropropene	9.34	Analyte not Identified by the Data System	tinkhams	03/26/10 11:59

Lab Sample ID: IC 280-8905/4 Client Sample ID:

Date Analyzed: 03/26/10 08:27 Lab File ID: R7593.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Ethanol	5.33	Analyte not Identified by the Data System	tinkhams	03/26/10 11:58

Lab Sample ID: IC 280-8905/10 Client Sample ID:

Date Analyzed: 03/26/10 10:39 Lab File ID: R7599.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Chloroethyl vinyl ether	8.79	Analyte not Identified by the Data System	waterlah	03/30/10 08:37

4/6

8260B

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1377-1

SDG No.: 10022892

Instrument ID: MSV_R2 Analysis Batch Number: 4173

Lab Sample ID: IC 280-4173/10 Client Sample ID:

Date Analyzed: 02/04/10 13:00 Lab File ID: RR3509.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Ethanol	4.04	Analyte not Identified by the Data System	waterlah	02/04/10 15:29

Lab Sample ID: IC 280-4173/11 Client Sample ID:

Date Analyzed: 02/04/10 13:21 Lab File ID: RR3510.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Ethanol	4.02	Split Peak	waterlah	02/04/10 15:29

4/6

8260B

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1377-1

SDG No.: 10022892

Instrument ID: MSV_R2 Analysis Batch Number: 6717

Lab Sample ID: IC 280-6717/5 Client Sample ID:

Date Analyzed: 03/08/10 10:28 Lab File ID: RR4519.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Isobutyl alcohol	6.17	Not specified	waterlah	03/08/10 12:44

Lab Sample ID: IC 280-6717/8 Client Sample ID:

Date Analyzed: 03/08/10 11:31 Lab File ID: RR4522.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Isobutyl alcohol	6.16	Not specified	waterlah	03/08/10 12:44

8260B

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1377-1

SDG No.: 10022892

Instrument ID: MSV_R2 Analysis Batch Number: 7831

Lab Sample ID: 280-1377-18 Client Sample ID: PIN15-0530

Date Analyzed: 03/17/10 21:37 Lab File ID: RR4797.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Trichloroethene	6.88	Not specified	rhoadesw	03/22/10 10:15

4/9/6

8260B

4.1 IRI
Sm 3/16/10

Stoller
Legacy Management Team

Chain of Custody / Sample Submittal Form

RIN: 10022892
COC: 10022892.2.1
Sampler(s): baer, moe, atkinson, lombardi, walters, caballero

Project: Pinellas Monitoring
Purchase Order: 3864
Cost Number: 1-502-1-06-509-4-02
Turnaround (Days): 28
Matrix: WA - Water

Laboratory: TestAmerica Denver
Address: 4955 Yarrow Street
Arvada, Colorado 80002
Phone: 303 736 0100

Ship #	Ticket	Sample Date	Time	Site	Location	Container	# Cont.	Preservation	Matrix	Comp.	Grab	Filtered	QC	Analysis
2	IDW 130	03/11/2010	9:28	PIN12	PIN12-0550-1	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 131	03/11/2010	10:33	PIN12	PIN12-0550-2	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 132	03/11/2010	11:58	PIN12	PIN12-0550-3	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 133	03/11/2010	13:46	PIN12	PIN12-0551-1	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 134	03/11/2010	14:24	PIN12	PIN12-0551-2	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 135	03/11/2010	15:03	PIN12	PIN12-0551-3	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 139	03/12/2010	12:45	PIN12	PIN12-0553A	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 140	03/12/2010	11:45	PIN12	PIN12-0553B	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 141	03/12/2010	10:25	PIN12	PIN12-0553C	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 181	03/12/2010	12:20	PIN12	PIN12-S30B	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 182	03/12/2010	10:50	PIN12	PIN12-S33C	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 183	03/12/2010	13:02	PIN12	PIN12-S35B	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 184	03/12/2010	11:36	PIN12	PIN12-S36B	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 197	03/12/2010	12:30	PIN12	PIN12-S71B	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 198	03/12/2010	9:50	PIN12	PIN12-S71C	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 199	03/12/2010	11:05	PIN12	PIN12-S71D	Glass 40 mL	3	4 C, HCl	WA			N		VOA

Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time
[signature]	3-12-10	1700	Amanda Hanson	3/15/10	1700			
Received by (signature)	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**
Amanda Hanson	3/12/10	1700	[signature]	3/16/10	0930			

Stoller
Legacy Management Team

Chain of Custody / Sample Submittal Form

RIN: 10022892
COC: 10022892.2.2
Sampler(s): baer, moe, atkinson, lombardi, walters, caballero

Project: Pinellas Monitoring
Purchase Order: 3864
Cost Number: 1-502-1-06-509-4-02
Turnaround (Days): 28
Matrix: WA - Water

Laboratory: TestAmerica Denver
Address: 4955 Yarrow Street
Arvada, Colorado 80002
Phone: 303 736 0100

Ship #	Ticket	Sample Date	Time	Site	Location	Container	# Cont.	Preservation	Matrix	Comp.	Grab	Filtered	QC	Analysis
2	IDW 206	03/11/2010	13:45	PIN15	PIN15-0520	HDPE 250 mL	1	HNO3	WA			N		Al,Fe
2	IDW 206	03/11/2010	13:45	PIN15	PIN15-0520	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 207	03/11/2010	10:15	PIN15	PIN15-0530	HDPE 250 mL	1	HNO3	WA			N		Al,Fe
2	IDW 207	03/11/2010	10:15	PIN15	PIN15-0530	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 208	03/11/2010	14:30	PIN15	PIN15-0534	HDPE 250 mL	1	HNO3	WA			N		Al,Fe
2	IDW 208	03/11/2010	14:30	PIN15	PIN15-0534	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 209	03/11/2010	11:00	PIN15	PIN15-0535	HDPE 250 mL	1	HNO3	WA			N		Al,Fe
2	IDW 209	03/11/2010	11:00	PIN15	PIN15-0535	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 211	03/11/2010	16:10	PIN15	PIN15-0568	HDPE 250 mL	1	HNO3	WA			N		Al,Fe
2	IDW 211	03/11/2010	16:10	PIN15	PIN15-0568	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 212	03/11/2010	15:15	PIN15	PIN15-0569	HDPE 250 mL	1	HNO3	WA			N		Al,Fe
2	IDW 212	03/11/2010	15:15	PIN15	PIN15-0569	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 241	03/11/2010	12:00	PIN15	PIN15-2873	HDPE 250 mL	1	HNO3	WA			N		Al,Fe
2	IDW 241	03/11/2010	12:00	PIN15	PIN15-2873	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 215	03/11/2010	11:15	PIN15	PIN15-E001	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 216	03/11/2010	9:20	PIN20	PIN20-0502	Glass 40 mL	3	4 C, HCl	WA			N		VOA

Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time
[signature]	3-12-10	1700	[signature]	3/15/10	1700			
Received by (signature)	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**
[signature]	3/10/10	1700	[signature]	3/16/10	0930			

Stoller
Legacy Management Team

Chain of Custody / Sample Submittal Form

RIN: 10022892
COC: 10022892.2.3
Sampler(s): baer, moe, atkinson, lombardi, walters, caballero

Project: Pinellas Monitoring
Purchase Order: 3864
Cost Number: 1-502-1-06-509-4-02
Turnaround (Days): 28
Matrix: WA - Water

Laboratory: TestAmerica Denver
Address: 4955 Yarrow Street
Arvada, Colorado 80002
Phone: 303 736 0100

Ship #	Ticket	Sample Date	Time	Site	Location	Container	# Cont.	Preservation	Matrix	Comp.	Grab	Filtered	QC	Analysis
2	IDW 217	03/11/2010	10:20	PIN20	PIN20-0503	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 235	03/11/2010	16:25	PIN20	PIN20-2867	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 219	03/11/2010	15:25	PIN20	PIN20-M003	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 220	03/11/2010	16:10	PIN20	PIN20-M005	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 221	03/11/2010	11:00	PIN20	PIN20-M015	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 222	03/11/2010	14:25	PIN20	PIN20-M035	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 228	03/11/2010	13:35	PIN20	PIN20-M38D	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 242	03/11/2010	8:06	PIN99	PIN99-2874	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 249	03/11/2010	8:00	PIN99	PIN99-2881	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 256	03/11/2010	8:00	PIN99	PIN99-2888	Glass 40 mL	3	4 C, HCl	WA			N		VOA
2	IDW 142	3/12/2010	1505	PIN12	PIN12-0554A	"	3	4°C HCL	WA			N		VOA

Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time
[signature]	3-12-10	1700	[signature]	3-12-10		[signature]	3/15/10	1700
Received by (signature)	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**
[signature]	3/12/10	1700	[signature]	3/16/10	0930			

Login Sample Receipt Check List

Client: S.M. Stoller Corporation

Job Number: 280-1377-1
SDG Number: 10022892

Login Number: 1377
Creator: Green, Angel L
List Number: 1

List Source: TestAmerica Denver

Question	T / F/ NA	Comment
Radioactivity either was not measured or, if measured, is at or below background	True	
The cooler's custody seal, if present, is intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the sample IDs on the containers and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	False	1 0F 3 VIALS FOR PIN20-M003 WAS RECEIVED BROKEN
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
VOA sample vials do not have headspace or bubble is <6mm (1/4") in diameter.	True	
If necessary, staff have been informed of any short hold time or quick TAT needs	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Is the Field Sampler's name present on COC?	True	
Sample Preservation Verified	True	

10

52

FedEx. US Airbill — FedEx Tracking Number 8712 8177 1808

ess

1 This portion can be removed for Recipient's records.

Date 3/15/10 — FedEx Tracking Number 871281771808

Sender's Name Gus Spoon — Phone 813 885 7427

Company TESTAMERICA TAMPA

Address 4712 BENJAMIN RD STE 100 — Dept./Floor/Suite/Room

City TAMPA — State FL — ZIP 33634-4404

RECIPIENT: PEEL HERE

fedex.com 1.800.GoFedEx 1.800.463.33

2 Your Internal Billing Reference

3 To

Recipient's Name Sample Review — Phone 303 736 0100

Company TEST AMERICA

HOLD Weekday: Print FedEx location address below. NOT available for FedEx First Overnight.

HOLD Saturday: Print FedEx location address below. Available ONLY for FedEx Priority Overnight and FedEx 2Day to select locations.

Address — Dept./Floor/Suite/Room

We cannot deliver to P.O. boxes or P.O. ZIP codes.

Address 4455 YARROW ST

Print FedEx location address here if HOLD option is selected.

City ARVADA — State CO — ZIP 80002

0414546911

Form ID No. 0215 — SPH32 — Recipient's Copy

4a Express Freight

4b Express Freight

5

6

TUE - 16 MAR A1
PRIORITY OVERNIGHT

FedEx TRK# 0215 8712 8177 1808

80002
CO-US
DEN

XH WHHA

Emp# 631384 15MAR10 TPFA

4 No Signature Required: Package may be left without obtaining a signature for delivery.

Direct Signature: Someone at recipient's address may sign for delivery. *Fee applies.*

If no one is available at recipient's address, someone at a neighboring address may sign for delivery. For residential deliveries only. *Fee applies.*

Does this shipment contain dangerous goods?

One box must be checked.

[X] No

[ ] Yes — As per attached Shipper's Declaration.

[ ] Yes — Shipper's Declaration not required.

[ ] Dry Ice — Dry Ice, 9, UN 1845 ____ x ____ kg

[ ] Cargo Aircraft Only

Dangerous goods (including dry ice) cannot be shipped in FedEx packaging or placed in a FedEx Express Drop Box.

7 Payment *Bill to:*

Enter FedEx Acct. No. or Credit Card No. below.

[ ] Sender — Acct. No. in Section 1 will be billed.

[X] Recipient

[ ] Third Party

[ ] Credit Card

[ ] Cash/Check

Obtain Recip. Acct. No. [ ]

Credit Card Auth.

Total Packages 1 — Total Weight 53 LBS

Our liability is limited to $100 unless you declare a higher value. See the current FedEx Service Guide for details.

553

Rev. Date 2/08•Part #158279•©1994–2008 FedEx•PRINTED IN U.S.A.•SRS

ANALYTICAL REPORT

Job Number: 280-1489-1

SDG Number: 10022892

Job Description: PINELLAS MONITORING

For:
S.M. Stoller Corporation
2597 B 3/4 Road
Grand Junction, CO 81503

Attention: Mr. Steve Donivan

Approved for release.
Kae E Yoder
Project Manager II
4/12/2010 3:19 PM

Kae E Yoder
Project Manager II
kae.yoder@testamericainc.com
04/12/2010

The test results in this report relate only to the samples in this report and meet all requirements of NELAC, with any exceptions noted. Pursuant to NELAP, this report shall not be reproduced except in full, without the written approval of the laboratory. All questions regarding this report should be directed to the TestAmerica Denver Project Manager.

The Lab Certification ID# is E87667.

Reporting limits are adjusted for sample size used, dilutions and moisture content if applicable.

TestAmerica Laboratories, Inc.
TestAmerica Denver 4955 Yarrow Street, Arvada, CO 80002
Tel (303) 736-0100 Fax (303) 431-7171 www.testamericainc.com

Table of Contents

Cover Title Page 1

Data Summaries 5

Report Narrative 5

Data Qualifiers 6

Executive Summary 7

Method Summary 13

Method / Analyst Summary 14

Sample Summary 15

Sample Datasheets 16

Surrogate Summary 107

QC Data Summary 111

QC Association Summary 144

Lab Chronicle 147

Manual Integration Summary 158

Organic Sample Data 160

GC/MS VOA 160

Method 8260B 160

Method 8260B QC Summary 161

Method 8260B Sample Data 203

Standards Data 556

Method 8260B ICAL Data 556

Method 8260B CCAL Data 786

Raw QC Data 909

Method 8260B Tune Data 909

Method 8260B Blank Data 953

Method 8260B LCS/LCSD Data 984

Table of Contents

Method 8260B MS/MSD Data 1020

Method 8260B Run Logs 1062

Method 8260B SIM 1226

Method 8260B SIM QC Summary 1227

Method 8260B SIM Sample Data 1236

Standards Data 1266

Method 8260B SIM ICAL Data 1266

Method 8260B SIM CCAL Data 1282

Raw QC Data 1288

Method 8260B SIM Tune Data 1288

Method 8260B SIM Blank Data 1300

Method 8260B SIM LCS/LCSD Data 1303

Method 8260B SIM MS/MSD Data 1306

Method 8260B SIM Run Logs 1312

Inorganic Sample Data 1315

Metals Data 1315

Met Cover Page 1316

Met Sample Data 1317

Met QC Data 1320

Met ICV/CCV 1320

Met CRQL 1323

Met Blanks 1324

Met ICSA/ICSAB 1326

Met MS/MSD/PDS 1328

Met LCS/LCSD 1330

Met Serial Dilution 1331

Table of Contents

Met MDL 1332
Met IECF 1333
Met Linear Ranges 1341
Met Preparation Log 1342
Met Analysis Run Log 1343
Met Raw Data 1345
Met Prep Data 2080
Shipping and Receiving Documents 2081
Client Chain of Custody 2082
Sample Receipt Checklist 2085

CASE NARRATIVE

Client: S.M. Stoller Corporation

Project: PINELLAS MONITORING - 10022892

Report Number: 280-1489-1

With exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. All laboratory quality control samples analyzed in conjunction with the samples in this project were within established control limits, with any exceptions noted. Calculations are performed before rounding to avoid round-off errors in calculated results.

This report includes reporting limits (RLs) less than TestAmerica Denver's practical quantitation limits. These reporting limits are being used specifically at the client's request to meet the needs of this project. Please note that data are not normally reported to these levels without qualification, since they are inherently less reliable and potentially less defensible than required by the current NELAC standards.

Results between the method detection limit (MDL) and reporting limit (RL) are flagged with a "J" qualifier to indicate an estimated value. These results are statistically less reliable than results greater than or equal to the RL and should be considered a qualitative value.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

RECEIPT

The samples were received on 3/18/2010; the samples arrived in good condition, properly preserved and on ice. The temperature of the cooler at receipt was 2.4 C.

Sample PIN15-0593 (IDW 213), requesting Total Metals 6010B analysis, was received at the laboratory improperly preserved. The laboratory adjusted the pH of the sample and proceeded with the requested analysis. The client was notified on 3/19/2010.

GC/MS VOLATILES - SW846 8260B

Due to analytes present above the linear calibration curve, a reduced aliquot size had to be used for the preparation/analysis of samples PIN12-0524 (IDW 104), PIN15-0537 (IDW 210), PIN15-0594 (IDW 214) and PIN20-M001 (IDW 218). To provide the lowest possible detection limits, multiple runs are reported.

Methylene chloride, a common laboratory contaminant, was detected in the method blank associated with batch 280-8905 at a level that was above the method detection limit but below the reporting limit. The value should be considered an estimate, and has been flagged “J”. If the associated samples reported a result above the MDL and/or RL, the result has been “B” flagged.

The MS and MSD associated with batch 280-9091 exhibited spike compound recoveries and RPD data outside the control limits. The LCS and LCSD were within control limits.

No other anomalies were encountered.

GC/MS VOLATILES - SW846 8260B-SIM

It can be noted that the MS and MSD analysis presented in this report were performed on a Pinellas sample; however, the Pinellas sample is not specific to this Job.

No other anomalies were encountered.

TOTAL METALS - SW846 6010B

No anomalies were encountered.

DATA REPORTING QUALIFIERS

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Lab Section	Qualifier	Description
GC/MS VOA		
	B	Compound was found in the blank and sample.
	U	Indicates the analyte was analyzed for but not detected.
	F	MS or MSD exceeds the control limits
	4	MS, MSD: The analyte present in the original sample is 4 times greater than the matrix spike concentration; therefore, control limits are not applicable.
	J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
	F	RPD of the MS and MSD exceeds the control limits
Metals		
	U	Indicates the analyte was analyzed for but not detected.
	J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1489-1	**PIN99-2875**					
Acetone		5.4	J	10	ug/L	8260B
Methylene Chloride		0.75	J	1.0	ug/L	8260B
280-1489-2	**PIN12-0515**					
Acetone		5.0	J	10	ug/L	8260B
280-1489-3	**PIN12-0516**					
Acetone		6.1	J	10	ug/L	8260B
280-1489-4	**PIN12-0517**					
Acetone		5.8	J	10	ug/L	8260B
280-1489-5	**PIN12-0518**					
Acetone		6.6	J	10	ug/L	8260B
Vinyl chloride		0.63	J	1.0	ug/L	8260B
280-1489-6	**PIN12-0524**					
Acetone		6.8	J	10	ug/L	8260B
Benzene		1.9		1.0	ug/L	8260B
1,1-Dichloroethane		0.42	J	1.0	ug/L	8260B
cis-1,2-Dichloroethene		150		5.0	ug/L	8260B
trans-1,2-Dichloroethene		3.4		1.0	ug/L	8260B
1,1-Dichloroethene		4.5		1.0	ug/L	8260B
Vinyl chloride		250		5.0	ug/L	8260B
280-1489-7	**PIN15-0537**					
Acetone		6.0	J	10	ug/L	8260B
Benzene		3.3		1.0	ug/L	8260B
cis-1,2-Dichloroethene		12		1.0	ug/L	8260B
Vinyl chloride		130		4.0	ug/L	8260B
Aluminum		76	J	100	ug/L	6010B
Iron		2000		100	ug/L	6010B
280-1489-8	**PIN12-0564-1**					
Acetone		6.4	J	10	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1489-9	**PIN12-0564-2**					
Acetone		6.5	J	10	ug/L	8260B
280-1489-10	**PIN12-0564-3**					
Acetone		9.1	J	10	ug/L	8260B
Naphthalene		0.37	J	1.0	ug/L	8260B
1,2,3-Trichlorobenzene		0.28	J	1.0	ug/L	8260B
280-1489-11	**PIN12-0565-1**					
Acetone		4.3	J	10	ug/L	8260B
cis-1,2-Dichloroethene		0.20	J	1.0	ug/L	8260B
280-1489-12	**PIN12-0565-2**					
2-Butanone (MEK)		4.3	J	5.0	ug/L	8260B
cis-1,2-Dichloroethene		0.54	J	1.0	ug/L	8260B
trans-1,2-Dichloroethene		0.20	J	1.0	ug/L	8260B
280-1489-13	**PIN12-0565-3**					
Acetone		3.4	J	10	ug/L	8260B
cis-1,2-Dichloroethene		0.22	J	1.0	ug/L	8260B
280-1489-14	**PIN12-0566-1**					
Acetone		6.2	J	10	ug/L	8260B
Toluene		0.59	J	1.0	ug/L	8260B
280-1489-15	**PIN12-0566-2**					
Acetone		7.2	J	10	ug/L	8260B
Toluene		0.62	J	1.0	ug/L	8260B
280-1489-16	**PIN12-0566-3**					
Acetone		11		10	ug/L	8260B
Toluene		0.75	J	1.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1489-17	**PIN12-0569-1**					
Acetone		5.5	J	10	ug/L	8260B
cis-1,2-Dichloroethene		0.24	J	1.0	ug/L	8260B
Toluene		0.24	J	1.0	ug/L	8260B
280-1489-18	**PIN15-0594**					
Benzene		34		10	ug/L	8260B
Carbon disulfide		7.9	J	10	ug/L	8260B
cis-1,2-Dichloroethene		4.5	J	10	ug/L	8260B
Ethylbenzene		10		10	ug/L	8260B
Toluene		660		100	ug/L	8260B
1,2,4-Trimethylbenzene		3.8	J	10	ug/L	8260B
Vinyl chloride		1200		100	ug/L	8260B
Xylenes, Total		23		10	ug/L	8260B
Aluminum		77	J	100	ug/L	6010B
Iron		550		100	ug/L	6010B
280-1489-19	**PIN20-2868**					
Acetone		4.2	J	10	ug/L	8260B
cis-1,2-Dichloroethene		0.57	J	1.0	ug/L	8260B
trans-1,2-Dichloroethene		2.3		1.0	ug/L	8260B
Trichloroethene		0.56	J	1.0	ug/L	8260B
Vinyl chloride		30		1.0	ug/L	8260B
280-1489-20	**PIN12-2869**					
Acetone		5.6	J	10	ug/L	8260B
Toluene		0.19	J	1.0	ug/L	8260B
280-1489-21	**PIN12-2870**					
cis-1,2-Dichloroethene		1.2		1.0	ug/L	8260B
trans-1,2-Dichloroethene		0.29	J	1.0	ug/L	8260B
Trichloroethene		0.23	J	1.0	ug/L	8260B
280-1489-22	**PIN12-2871**					
Carbon disulfide		12		1.0	ug/L	8260B
cis-1,2-Dichloroethene		0.31	J	1.0	ug/L	8260B
Styrene		0.65	J	1.0	ug/L	8260B
Toluene		2.0		1.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1489-23	**PIN99-2886**					
Methylene Chloride		1.1		1.0	ug/L	8260B
280-1489-24	**PIN99-2887**					
Methylene Chloride		0.77	J	1.0	ug/L	8260B
280-1489-25	**PIN99-2889**					
Methylene Chloride		0.92	J	1.0	ug/L	8260B
280-1489-26	**PIN20-M001**					
Acetone		23	J	40	ug/L	8260B
Benzene		1.4	J	4.0	ug/L	8260B
2-Butanone (MEK)		40		20	ug/L	8260B
Carbon disulfide		6.5		4.0	ug/L	8260B
cis-1,2-Dichloroethene		190		4.0	ug/L	8260B
trans-1,2-Dichloroethene		22		4.0	ug/L	8260B
2-Hexanone		21		20	ug/L	8260B
Methylene Chloride		2.4	J B	4.0	ug/L	8260B
Vinyl chloride		930		40	ug/L	8260B
280-1489-27	**PIN20-M065**					
Toluene		0.21	J	1.0	ug/L	8260B
280-1489-28	**PIN20-M066**					
Acetone		6.2	J	10	ug/L	8260B
280-1489-29	**PIN20-M067**					
Acetone		4.5	J	10	ug/L	8260B
280-1489-30	**PIN20-M068**					
Acetone		5.4	J	10	ug/L	8260B
cis-1,2-Dichloroethene		0.50	J	1.0	ug/L	8260B
trans-1,2-Dichloroethene		2.3		1.0	ug/L	8260B
Methylene Chloride		0.37	J B	1.0	ug/L	8260B
Trichloroethene		0.59	J	1.0	ug/L	8260B
Vinyl chloride		33		1.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1489-31	**PIN20-M069**					
cis-1,2-Dichloroethene		24		1.0	ug/L	8260B
trans-1,2-Dichloroethene		3.3		1.0	ug/L	8260B
1,1-Dichloroethene		0.35	J	1.0	ug/L	8260B
Trichloroethene		2.0		1.0	ug/L	8260B
Vinyl chloride		14		1.0	ug/L	8260B
280-1489-32	**PIN12-S69B**					
Carbon disulfide		5.9		1.0	ug/L	8260B
cis-1,2-Dichloroethene		0.16	J	1.0	ug/L	8260B
Styrene		0.45	J	1.0	ug/L	8260B
Toluene		1.1		1.0	ug/L	8260B
280-1489-33	**PIN12-S69C**					
Carbon disulfide		3.4		1.0	ug/L	8260B
cis-1,2-Dichloroethene		0.37	J	1.0	ug/L	8260B
Styrene		0.26	J	1.0	ug/L	8260B
Toluene		1.0		1.0	ug/L	8260B
280-1489-34	**PIN12-S69D**					
Acetone		5.5	J	10	ug/L	8260B
cis-1,2-Dichloroethene		0.70	J	1.0	ug/L	8260B
280-1489-35	**PIN12-S70D**					
Carbon disulfide		2.4		1.0	ug/L	8260B
1,1-Dichloroethane		6.6		1.0	ug/L	8260B
cis-1,2-Dichloroethene		21		1.0	ug/L	8260B
trans-1,2-Dichloroethene		7.4		1.0	ug/L	8260B
1,1-Dichloroethene		0.64	J	1.0	ug/L	8260B
Toluene		1.1		1.0	ug/L	8260B
Vinyl chloride		8.0		1.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1489-36	**PIN15-0593**					
Benzene		3.4		1.0	ug/L	8260B
Carbon disulfide		0.88	J	1.0	ug/L	8260B
cis-1,2-Dichloroethene		5.8		1.0	ug/L	8260B
trans-1,2-Dichloroethene		0.30	J	1.0	ug/L	8260B
Toluene		11		1.0	ug/L	8260B
Vinyl chloride		23		1.0	ug/L	8260B
Aluminum		1700		100	ug/L	6010B
Iron		15000		100	ug/L	6010B
280-1489-37	**PIN12-0525**					
cis-1,2-Dichloroethene		1.3		1.0	ug/L	8260B

METHOD SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Description	Lab Location	Method	Preparation Method
Matrix: Water			
Volatile Organic Compounds (GC/MS)	TAL DEN	SW846 8260B	
Purge and Trap	TAL DEN		SW846 5030B
Volatile Organic Compounds (GC/MS-SIM)	TAL DEN	SW846 8260B SIM	
Purge and Trap	TAL DEN		SW846 5030B
Metals (ICP)	TAL DEN	SW846 6010B	
Preparation, Total Metals	TAL DEN		SW846 3010A

Lab References:

TAL DEN = TestAmerica Denver

Method References:

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

METHOD / ANALYST SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Method	Analyst	Analyst ID
SW846 8260B	Ilczyszyn, Dennis P	DPI
SW846 8260B	Jackson, Todd D	TDJ
SW846 8260B	Tinkham, Sarah A	SAT
SW846 8260B SIM	Wolfe, Ashley E	AEW
SW846 6010B	Harre, John K	JKH

SAMPLE SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
280-1489-1	PIN99-2875	Water	03/13/2010 0800	03/18/2010 0900
280-1489-2	PIN12-0515	Water	03/13/2010 0810	03/18/2010 0900
280-1489-3	PIN12-0516	Water	03/13/2010 0850	03/18/2010 0900
280-1489-4	PIN12-0517	Water	03/13/2010 1025	03/18/2010 0900
280-1489-5	PIN12-0518	Water	03/13/2010 0935	03/18/2010 0900
280-1489-6	PIN12-0524	Water	03/13/2010 1105	03/18/2010 0900
280-1489-7	PIN15-0537	Water	03/15/2010 0910	03/18/2010 0900
280-1489-7MS	PIN15-0537	Water	03/15/2010 0910	03/18/2010 0900
280-1489-7MSD	PIN15-0537	Water	03/15/2010 0910	03/18/2010 0900
280-1489-8	PIN12-0564-1	Water	03/13/2010 1416	03/18/2010 0900
280-1489-9	PIN12-0564-2	Water	03/13/2010 1449	03/18/2010 0900
280-1489-10	PIN12-0564-3	Water	03/13/2010 1554	03/18/2010 0900
280-1489-11	PIN12-0565-1	Water	03/13/2010 0923	03/18/2010 0900
280-1489-12	PIN12-0565-2	Water	03/13/2010 1038	03/18/2010 0900
280-1489-13	PIN12-0565-3	Water	03/13/2010 1113	03/18/2010 0900
280-1489-14	PIN12-0566-1	Water	03/15/2010 0921	03/18/2010 0900
280-1489-15	PIN12-0566-2	Water	03/15/2010 0949	03/18/2010 0900
280-1489-16	PIN12-0566-3	Water	03/15/2010 1045	03/18/2010 0900
280-1489-17	PIN12-0569-1	Water	03/15/2010 1128	03/18/2010 0900
280-1489-18	PIN15-0594	Water	03/15/2010 1105	03/18/2010 0900
280-1489-19	PIN20-2868	Water	03/13/2010 1701	03/18/2010 0900
280-1489-20	PIN12-2869	Water	03/13/2010 1200	03/18/2010 0900
280-1489-21	PIN12-2870	Water	03/13/2010 1200	03/18/2010 0900
280-1489-22	PIN12-2871	Water	03/15/2010 1200	03/18/2010 0900
280-1489-23	PIN99-2886	Water	03/13/2010 0800	03/18/2010 0900
280-1489-24	PIN99-2887	Water	03/13/2010 0800	03/18/2010 0900
280-1489-25	PIN99-2889	Water	03/13/2010 0800	03/18/2010 0900
280-1489-26	PIN20-M001	Water	03/13/2010 1435	03/18/2010 0900
280-1489-27	PIN20-M065	Water	03/13/2010 1055	03/18/2010 0900
280-1489-28	PIN20-M066	Water	03/13/2010 0945	03/18/2010 0900
280-1489-29	PIN20-M067	Water	03/13/2010 1145	03/18/2010 0900
280-1489-30	PIN20-M068	Water	03/13/2010 1500	03/18/2010 0900
280-1489-31	PIN20-M069	Water	03/13/2010 1545	03/18/2010 0900
280-1489-32	PIN12-S69B	Water	03/15/2010 1030	03/18/2010 0900
280-1489-33	PIN12-S69C	Water	03/15/2010 0935	03/18/2010 0900
280-1489-34	PIN12-S69D	Water	03/15/2010 0855	03/18/2010 0900
280-1489-35	PIN12-S70D	Water	03/15/2010 1130	03/18/2010 0900
280-1489-36	PIN15-0593	Water	03/15/2010 1250	03/18/2010 0900
280-1489-37	PIN12-0525	Water	03/13/2010 1135	03/18/2010 0900

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2875**

Lab Sample ID: 280-1489-1
Client Matrix: Water

Date Sampled: 03/13/2010 0800
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method: 8260B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 03/24/2010 1529
Date Prepared: 03/24/2010 1529

Analysis Batch: 280-8529

Instrument ID: MSV_G
Lab File ID: G3174.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	5.4	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.75	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2875**

Lab Sample ID: 280-1489-1
Client Matrix: Water

Date Sampled: 03/13/2010 0800
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3174.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1529		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1529			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	86		70 - 127
Toluene-d8 (Surr)	92		80 - 125
4-Bromofluorobenzene (Surr)	91		78 - 118
Dibromofluoromethane (Surr)	91		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0515**

Lab Sample ID: 280-1489-2
Client Matrix: Water

Date Sampled: 03/13/2010 0810
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3175.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1550		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1550			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	5.0	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0515**

Lab Sample ID: 280-1489-2
Client Matrix: Water

Date Sampled: 03/13/2010 0810
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3175.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1550		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1550			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	91		70 - 127
Toluene-d8 (Surr)	90		80 - 125
4-Bromofluorobenzene (Surr)	92		78 - 118
Dibromofluoromethane (Surr)	94		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0516**

Lab Sample ID: 280-1489-3
Client Matrix: Water

Date Sampled: 03/13/2010 0850
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3176.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1610		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1610			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	6.1	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0516**

Lab Sample ID: 280-1489-3
Client Matrix: Water

Date Sampled: 03/13/2010 0850
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3176.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1610		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1610			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	92		70 - 127
Toluene-d8 (Surr)	89		80 - 125
4-Bromofluorobenzene (Surr)	90		78 - 118
Dibromofluoromethane (Surr)	95		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0517**

Lab Sample ID: 280-1489-4
Client Matrix: Water

Date Sampled: 03/13/2010 1025
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3177.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1631		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1631			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	5.8	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0517**

Lab Sample ID: 280-1489-4
Client Matrix: Water

Date Sampled: 03/13/2010 1025
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3177.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1631		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1631			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	89		70 - 127
Toluene-d8 (Surr)	91		80 - 125
4-Bromofluorobenzene (Surr)	90		78 - 118
Dibromofluoromethane (Surr)	94		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0518**

Lab Sample ID: 280-1489-5
Client Matrix: Water

Date Sampled: 03/13/2010 0935
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3178.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1651		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1651			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	6.6	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0518**

Lab Sample ID: 280-1489-5
Client Matrix: Water

Date Sampled: 03/13/2010 0935
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3178.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1651		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1651			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.63	J	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	95		70 - 127
Toluene-d8 (Surr)	91		80 - 125
4-Bromofluorobenzene (Surr)	94		78 - 118
Dibromofluoromethane (Surr)	95		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0524**

Lab Sample ID: 280-1489-6
Client Matrix: Water

Date Sampled: 03/13/2010 1105
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3179.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1712		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1712			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	6.8	J	1.9	10
Benzene	1.9		0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.42	J	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
trans-1,2-Dichloroethene	3.4		0.15	1.0
1,1-Dichloroethene	4.5		0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0524**

Lab Sample ID: 280-1489-6
Client Matrix: Water

Date Sampled: 03/13/2010 1105
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3179.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1712		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1712			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	94		70 - 127
Toluene-d8 (Surr)	92		80 - 125
4-Bromofluorobenzene (Surr)	91		78 - 118
Dibromofluoromethane (Surr)	95		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0524**

Lab Sample ID: 280-1489-6
Client Matrix: Water

Date Sampled: 03/13/2010 1105
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529		Instrument ID:	MSV_G
Preparation:	5030B			Lab File ID:	G3180.D
Dilution:	1.0			Initial Weight/Volume:	4 mL
Date Analyzed:	03/24/2010 1732	Run Type:	DL	Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1732				

Analyte	Result (ug/L)	Qualifier	MDL	RL
cis-1,2-Dichloroethene	150		0.75	5.0
Vinyl chloride	250		2.0	5.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	93		70 - 127
Toluene-d8 (Surr)	90		80 - 125
4-Bromofluorobenzene (Surr)	90		78 - 118
Dibromofluoromethane (Surr)	95		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0537**

Lab Sample ID: 280-1489-7
Client Matrix: Water

Date Sampled: 03/15/2010 0910
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3181.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1753		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1753			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	6.0	J	1.9	10
Benzene	3.3		0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	12		0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0537**

Lab Sample ID: 280-1489-7
Client Matrix: Water

Date Sampled: 03/15/2010 0910
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3181.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1753		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1753			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	95		70 - 127
Toluene-d8 (Surr)	89		80 - 125
4-Bromofluorobenzene (Surr)	91		78 - 118
Dibromofluoromethane (Surr)	95		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0537**

Lab Sample ID: 280-1489-7
Client Matrix: Water

Date Sampled: 03/15/2010 0910
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666		Instrument ID:	MSV_Q
Preparation:	5030B			Lab File ID:	Q4527.D
Dilution:	1.0			Initial Weight/Volume:	5 mL
Date Analyzed:	03/25/2010 1525	Run Type:	DL	Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1525				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Vinyl chloride	130		1.6	4.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	85		70 - 127
Toluene-d8 (Surr)	95		80 - 125
4-Bromofluorobenzene (Surr)	84		78 - 118
Dibromofluoromethane (Surr)	98		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0564-1**

Lab Sample ID: 280-1489-8
Client Matrix: Water

Date Sampled: 03/13/2010 1416
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3182.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1814		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1814			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	6.4	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0564-1**

Lab Sample ID: 280-1489-8
Client Matrix: Water

Date Sampled: 03/13/2010 1416
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3182.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1814		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1814			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	94		70 - 127
Toluene-d8 (Surr)	89		80 - 125
4-Bromofluorobenzene (Surr)	91		78 - 118
Dibromofluoromethane (Surr)	95		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0564-2**

Lab Sample ID: 280-1489-9
Client Matrix: Water

Date Sampled: 03/13/2010 1449
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3183.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1834		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1834			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	6.5	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0564-2**

Lab Sample ID: 280-1489-9
Client Matrix: Water

Date Sampled: 03/13/2010 1449
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8529	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3183.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1834		Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1834			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	95		70 - 127
Toluene-d8 (Surr)	89		80 - 125
4-Bromofluorobenzene (Surr)	82		78 - 118
Dibromofluoromethane (Surr)	94		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0564-3**

Lab Sample ID: 280-1489-10
Client Matrix: Water

Date Sampled: 03/13/2010 1554
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4530.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1629		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1629			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	9.1	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.37	J	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0564-3**

Lab Sample ID: 280-1489-10
Client Matrix: Water

Date Sampled: 03/13/2010 1554
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4530.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1629		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1629			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.28	J	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	82		70 - 127
Toluene-d8 (Surr)	95		80 - 125
4-Bromofluorobenzene (Surr)	82		78 - 118
Dibromofluoromethane (Surr)	93		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0565-1**

Lab Sample ID: 280-1489-11
Client Matrix: Water

Date Sampled: 03/13/2010 0923
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4531.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1651		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1651			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	4.3	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.20	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0565-1**

Lab Sample ID: 280-1489-11
Client Matrix: Water

Date Sampled: 03/13/2010 0923
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4531.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1651		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1651			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	84		70 - 127
Toluene-d8 (Surr)	94		80 - 125
4-Bromofluorobenzene (Surr)	81		78 - 118
Dibromofluoromethane (Surr)	94		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0565-2**

Lab Sample ID: 280-1489-12
Client Matrix: Water

Date Sampled: 03/13/2010 1038
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4532.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1713		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1713			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	4.3	J	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.54	J	0.15	1.0
trans-1,2-Dichloroethene	0.20	J	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0565-2**

Lab Sample ID: 280-1489-12
Client Matrix: Water

Date Sampled: 03/13/2010 1038
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4532.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1713		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1713			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	79		70 - 127
Toluene-d8 (Surr)	93		80 - 125
4-Bromofluorobenzene (Surr)	81		78 - 118
Dibromofluoromethane (Surr)	94		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0565-3**

Lab Sample ID: 280-1489-13
Client Matrix: Water

Date Sampled: 03/13/2010 1113
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4533.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1734		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1734			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	3.4	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.22	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0565-3**

Lab Sample ID: 280-1489-13
Client Matrix: Water

Date Sampled: 03/13/2010 1113
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4533.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1734		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1734			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	82		70 - 127
Toluene-d8 (Surr)	95		80 - 125
4-Bromofluorobenzene (Surr)	79		78 - 118
Dibromofluoromethane (Surr)	92		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0566-1**

Lab Sample ID: 280-1489-14
Client Matrix: Water

Date Sampled: 03/15/2010 0921
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3262.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2336		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2336			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	6.2	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0566-1**

Lab Sample ID: 280-1489-14
Client Matrix: Water

Date Sampled: 03/15/2010 0921
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3262.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2336		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2336			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.59	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	88		70 - 127
Toluene-d8 (Surr)	92		80 - 125
4-Bromofluorobenzene (Surr)	88		78 - 118
Dibromofluoromethane (Surr)	94		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0566-2**

Lab Sample ID: 280-1489-15
Client Matrix: Water

Date Sampled: 03/15/2010 0949
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3263.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2357		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2357			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	7.2	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0566-2**

Lab Sample ID: 280-1489-15
Client Matrix: Water

Date Sampled: 03/15/2010 0949
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3263.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2357		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2357			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.62	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	87		70 - 127
Toluene-d8 (Surr)	95		80 - 125
4-Bromofluorobenzene (Surr)	89		78 - 118
Dibromofluoromethane (Surr)	94		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0566-3**

Lab Sample ID: 280-1489-16
Client Matrix: Water

Date Sampled: 03/15/2010 1045
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3264.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0017		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0017			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	11		1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0566-3**

Lab Sample ID: 280-1489-16
Client Matrix: Water

Date Sampled: 03/15/2010 1045
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3264.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0017		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0017			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.75	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	89		70 - 127
Toluene-d8 (Surr)	90		80 - 125
4-Bromofluorobenzene (Surr)	87		78 - 118
Dibromofluoromethane (Surr)	93		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0569-1**

Lab Sample ID: 280-1489-17
Client Matrix: Water

Date Sampled: 03/15/2010 1128
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3265.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0038		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0038			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	5.5	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.24	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0569-1**

Lab Sample ID: 280-1489-17
Client Matrix: Water

Date Sampled: 03/15/2010 1128
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3265.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0038		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0038			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.24	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	96		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	94		78 - 118
Dibromofluoromethane (Surr)	102		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0594**

Lab Sample ID: 280-1489-18
Client Matrix: Water

Date Sampled: 03/15/2010 1105
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9091	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34058.D
Dilution:	1.0		Initial Weight/Volume:	2 mL
Date Analyzed:	03/29/2010 1846		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1846			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	19	U	19	100
Benzene	34		1.6	10
Bromobenzene	1.7	U	1.7	10
Bromochloromethane	1.0	U	1.0	10
Bromodichloromethane	1.7	U	1.7	10
Bromoform	1.9	U	1.9	10
Bromomethane	2.1	U	2.1	10
2-Butanone (MEK)	20	U	20	50
n-Butylbenzene	1.4	U	1.4	10
sec-Butylbenzene	1.7	U	1.7	10
tert-Butylbenzene	1.6	U	1.6	10
Carbon disulfide	7.9	J	4.5	10
Carbon tetrachloride	1.9	U	1.9	10
Chlorobenzene	1.7	U	1.7	10
Dibromochloromethane	1.7	U	1.7	10
Chloroethane	4.1	U	4.1	10
Chloroform	1.6	U	1.6	10
Chloromethane	3.0	U	3.0	10
2-Chlorotoluene	1.7	U	1.7	10
4-Chlorotoluene	2.1	U	2.1	10
1,2-Dibromo-3-Chloropropane	4.7	U	4.7	10
Dibromomethane	1.7	U	1.7	10
1,2-Dichlorobenzene	1.5	U	1.5	10
1,3-Dichlorobenzene	1.3	U	1.3	10
1,4-Dichlorobenzene	1.6	U	1.6	10
Dichlorodifluoromethane	3.1	U	3.1	10
1,1-Dichloroethane	2.2	U	2.2	10
1,2-Dichloroethane	1.3	U	1.3	10
cis-1,2-Dichloroethene	4.5	J	1.5	10
trans-1,2-Dichloroethene	1.5	U	1.5	10
1,1-Dichloroethene	2.3	U	2.3	10
1,2-Dichloropropane	1.8	U	1.8	10
1,3-Dichloropropane	2.2	U	2.2	10
2,2-Dichloropropane	1.8	U	1.8	10
cis-1,3-Dichloropropene	1.6	U	1.6	10
trans-1,3-Dichloropropene	1.9	U	1.9	10
1,1-Dichloropropene	1.9	U	1.9	10
Ethylbenzene	10		1.6	10
Hexachlorobutadiene	1.2	U	1.2	10
2-Hexanone	17	U	17	50
Isopropylbenzene	1.9	U	1.9	10
4-Isopropyltoluene	2.0	U	2.0	10
Methylene Chloride	3.2	U	3.2	10
4-Methyl-2-pentanone	9.8	U	9.8	50
Naphthalene	2.2	U	2.2	10
n-Propylbenzene	1.6	U	1.6	10

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0594**

Lab Sample ID: 280-1489-18
Client Matrix: Water

Date Sampled: 03/15/2010 1105
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9091	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34058.D
Dilution:	1.0		Initial Weight/Volume:	2 mL
Date Analyzed:	03/29/2010 1846		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1846			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	1.7	U	1.7	10
1,1,1,2-Tetrachloroethane	2.1	U	2.1	10
1,1,2,2-Tetrachloroethane	2.1	U	2.1	10
Tetrachloroethene	2.0	U	2.0	10
1,2,3-Trichlorobenzene	2.1	U	2.1	10
1,2,4-Trichlorobenzene	2.1	U	2.1	10
1,1,1-Trichloroethane	1.6	U	1.6	10
1,1,2-Trichloroethane	2.7	U	2.7	10
Trichloroethene	1.6	U	1.6	10
Trichlorofluoromethane	2.9	U	2.9	10
1,2,3-Trichloropropane	3.3	U	3.3	10
1,2,4-Trimethylbenzene	3.8	J	1.5	10
1,3,5-Trimethylbenzene	1.6	U	1.6	10
Xylenes, Total	23		1.9	10
1,2-Dibromoethane	1.8	U	1.8	10

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	94		70 - 127
Toluene-d8 (Surr)	94		80 - 125
4-Bromofluorobenzene (Surr)	92		78 - 118
Dibromofluoromethane (Surr)	89		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0594**

Lab Sample ID: 280-1489-18
Client Matrix: Water

Date Sampled: 03/15/2010 1105
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9091	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34059.D
Dilution:	1.0		Initial Weight/Volume:	0.2 mL
Date Analyzed:	03/29/2010 1907		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1907			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Toluene	660		17	100
Vinyl chloride	1200		40	100

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	94		70 - 127
Toluene-d8 (Surr)	94		80 - 125
4-Bromofluorobenzene (Surr)	90		78 - 118
Dibromofluoromethane (Surr)	87		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-2868**

Lab Sample ID: 280-1489-19
Client Matrix: Water

Date Sampled: 03/13/2010 1701
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4536.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1839		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1839			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	4.2	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.57	J	0.15	1.0
trans-1,2-Dichloroethene	2.3		0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-2868**

Lab Sample ID: 280-1489-19
Client Matrix: Water

Date Sampled: 03/13/2010 1701
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4536.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1839		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1839			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.56	J	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	30		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	87		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	82		78 - 118
Dibromofluoromethane (Surr)	97		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-2869**

Lab Sample ID: 280-1489-20
Client Matrix: Water

Date Sampled: 03/13/2010 1200
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4537.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1900		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1900			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	5.6	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-2869**

Lab Sample ID: 280-1489-20
Client Matrix: Water

Date Sampled: 03/13/2010 1200
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4537.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1900		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1900			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.19	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	89		70 - 127
Toluene-d8 (Surr)	92		80 - 125
4-Bromofluorobenzene (Surr)	82		78 - 118
Dibromofluoromethane (Surr)	97		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-2870**

Lab Sample ID: 280-1489-21
Client Matrix: Water

Date Sampled: 03/13/2010 1200
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4538.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1922		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1922			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	1.2		0.15	1.0
trans-1,2-Dichloroethene	0.29	J	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-2870**

Lab Sample ID: 280-1489-21
Client Matrix: Water

Date Sampled: 03/13/2010 1200
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4538.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1922		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1922			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.23	J	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	86		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	84		78 - 118
Dibromofluoromethane (Surr)	97		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-2871**

Lab Sample ID: 280-1489-22
Client Matrix: Water

Date Sampled: 03/15/2010 1200
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3267.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0119		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0119			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	12		0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.31	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-2871**

Lab Sample ID: 280-1489-22
Client Matrix: Water

Date Sampled: 03/15/2010 1200
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3267.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0119		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0119			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.65	J	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	2.0		0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	79		70 - 127
Toluene-d8 (Surr)	94		80 - 125
4-Bromofluorobenzene (Surr)	87		78 - 118
Dibromofluoromethane (Surr)	89		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2886**

Lab Sample ID: 280-1489-23
Client Matrix: Water

Date Sampled: 03/13/2010 0800
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4539.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1943		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1943			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	1.1		0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2886**

Lab Sample ID: 280-1489-23
Client Matrix: Water

Date Sampled: 03/13/2010 0800
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4539.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1943		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1943			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	90		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	85		78 - 118
Dibromofluoromethane (Surr)	99		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2887**

Lab Sample ID: 280-1489-24
Client Matrix: Water

Date Sampled: 03/13/2010 0800
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4540.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2005		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2005			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.77	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2887**

Lab Sample ID: 280-1489-24
Client Matrix: Water

Date Sampled: 03/13/2010 0800
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4540.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2005		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2005			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	84		70 - 127
Toluene-d8 (Surr)	95		80 - 125
4-Bromofluorobenzene (Surr)	84		78 - 118
Dibromofluoromethane (Surr)	99		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2889**

Lab Sample ID: 280-1489-25
Client Matrix: Water

Date Sampled: 03/13/2010 0800
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4541.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2027		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2027			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.92	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2889**

Lab Sample ID: 280-1489-25
Client Matrix: Water

Date Sampled: 03/13/2010 0800
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4541.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2027		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2027			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	84		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	84		78 - 118
Dibromofluoromethane (Surr)	92		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M001**

Lab Sample ID: 280-1489-26
Client Matrix: Water

Date Sampled: 03/13/2010 1435
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8905	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7612.D
Dilution:	1.0		Initial Weight/Volume:	5 mL
Date Analyzed:	03/26/2010 1516		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1516			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	23	J	7.6	40
Benzene	1.4	J	0.64	4.0
Bromobenzene	0.68	U	0.68	4.0
Bromochloromethane	0.40	U	0.40	4.0
Bromodichloromethane	0.68	U	0.68	4.0
Bromoform	0.76	U	0.76	4.0
Bromomethane	0.84	U	0.84	4.0
2-Butanone (MEK)	40		8.0	20
n-Butylbenzene	0.56	U	0.56	4.0
sec-Butylbenzene	0.68	U	0.68	4.0
tert-Butylbenzene	0.64	U	0.64	4.0
Carbon disulfide	6.5		1.8	4.0
Carbon tetrachloride	0.76	U	0.76	4.0
Chlorobenzene	0.68	U	0.68	4.0
Dibromochloromethane	0.68	U	0.68	4.0
Chloroethane	1.6	U	1.6	4.0
Chloroform	0.64	U	0.64	4.0
Chloromethane	1.2	U	1.2	4.0
2-Chlorotoluene	0.68	U	0.68	4.0
4-Chlorotoluene	0.84	U	0.84	4.0
1,2-Dibromo-3-Chloropropane	1.9	U	1.9	4.0
Dibromomethane	0.68	U	0.68	4.0
1,2-Dichlorobenzene	0.60	U	0.60	4.0
1,3-Dichlorobenzene	0.52	U	0.52	4.0
1,4-Dichlorobenzene	0.64	U	0.64	4.0
Dichlorodifluoromethane	1.2	U	1.2	4.0
1,1-Dichloroethane	0.88	U	0.88	4.0
1,2-Dichloroethane	0.52	U	0.52	4.0
cis-1,2-Dichloroethene	190		0.60	4.0
trans-1,2-Dichloroethene	22		0.60	4.0
1,1-Dichloroethene	0.92	U	0.92	4.0
1,2-Dichloropropane	0.72	U	0.72	4.0
1,3-Dichloropropane	0.88	U	0.88	4.0
2,2-Dichloropropane	0.72	U	0.72	4.0
cis-1,3-Dichloropropene	0.64	U	0.64	4.0
trans-1,3-Dichloropropene	0.76	U	0.76	4.0
1,1-Dichloropropene	0.76	U	0.76	4.0
Ethylbenzene	0.64	U	0.64	4.0
Hexachlorobutadiene	0.48	U	0.48	4.0
2-Hexanone	21		6.8	20
Isopropylbenzene	0.76	U	0.76	4.0
4-Isopropyltoluene	0.80	U	0.80	4.0
Methylene Chloride	2.4	J B	1.3	4.0
4-Methyl-2-pentanone	3.9	U	3.9	20
Naphthalene	0.88	U	0.88	4.0
n-Propylbenzene	0.64	U	0.64	4.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M001**

Lab Sample ID: 280-1489-26
Client Matrix: Water

Date Sampled: 03/13/2010 1435
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8905	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7612.D
Dilution:	1.0		Initial Weight/Volume:	5 mL
Date Analyzed:	03/26/2010 1516		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1516			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.68	U	0.68	4.0
1,1,1,2-Tetrachloroethane	0.84	U	0.84	4.0
1,1,2,2-Tetrachloroethane	0.84	U	0.84	4.0
Tetrachloroethene	0.80	U	0.80	4.0
Toluene	0.68	U	0.68	4.0
1,2,3-Trichlorobenzene	0.84	U	0.84	4.0
1,2,4-Trichlorobenzene	0.84	U	0.84	4.0
1,1,1-Trichloroethane	0.64	U	0.64	4.0
1,1,2-Trichloroethane	1.1	U	1.1	4.0
Trichloroethene	0.64	U	0.64	4.0
Trichlorofluoromethane	1.2	U	1.2	4.0
1,2,3-Trichloropropane	1.3	U	1.3	4.0
1,2,4-Trimethylbenzene	0.60	U	0.60	4.0
1,3,5-Trimethylbenzene	0.64	U	0.64	4.0
Xylenes, Total	0.76	U	0.76	4.0
1,2-Dibromoethane	0.72	U	0.72	4.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	100		70 - 127
Toluene-d8 (Surr)	99		80 - 125
4-Bromofluorobenzene (Surr)	98		78 - 118
Dibromofluoromethane (Surr)	95		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M001**

Lab Sample ID: 280-1489-26
Client Matrix: Water

Date Sampled: 03/13/2010 1435
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8905	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7613.D
Dilution:	1.0		Initial Weight/Volume:	0.5 mL
Date Analyzed:	03/26/2010 1537		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1537			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Vinyl chloride	930		16	40

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	100		70 - 127
Toluene-d8 (Surr)	102		80 - 125
4-Bromofluorobenzene (Surr)	94		78 - 118
Dibromofluoromethane (Surr)	97		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M065**

Lab Sample ID: 280-1489-27
Client Matrix: Water

Date Sampled: 03/13/2010 1055
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4542.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2048		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2048			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M065**

Lab Sample ID: 280-1489-27
Client Matrix: Water

Date Sampled: 03/13/2010 1055
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4542.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2048		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2048			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.21	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	89		70 - 127
Toluene-d8 (Surr)	92		80 - 125
4-Bromofluorobenzene (Surr)	82		78 - 118
Dibromofluoromethane (Surr)	96		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M066**

Lab Sample ID: 280-1489-28
Client Matrix: Water

Date Sampled: 03/13/2010 0945
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4543.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2110		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2110			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	6.2	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M066**

Lab Sample ID: 280-1489-28
Client Matrix: Water

Date Sampled: 03/13/2010 0945
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8666	Instrument ID:	MSV_Q
Preparation:	5030B		Lab File ID:	Q4543.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2110		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2110			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	93		70 - 127
Toluene-d8 (Surr)	90		80 - 125
4-Bromofluorobenzene (Surr)	82		78 - 118
Dibromofluoromethane (Surr)	99		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M067**

Lab Sample ID: 280-1489-29
Client Matrix: Water

Date Sampled: 03/13/2010 1145
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8905	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7611.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1456		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1456			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	4.5	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M067**

Lab Sample ID: 280-1489-29
Client Matrix: Water

Date Sampled: 03/13/2010 1145
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8905	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7611.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1456		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1456			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	90		70 - 127
Toluene-d8 (Surr)	104		80 - 125
4-Bromofluorobenzene (Surr)	98		78 - 118
Dibromofluoromethane (Surr)	92		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M068**

Lab Sample ID: 280-1489-30
Client Matrix: Water

Date Sampled: 03/13/2010 1500
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8905	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7616.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1639		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1639			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	5.4	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.50	J	0.15	1.0
trans-1,2-Dichloroethene	2.3		0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.37	J B	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M068**

Lab Sample ID: 280-1489-30
Client Matrix: Water

Date Sampled: 03/13/2010 1500
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method: 8260B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 03/26/2010 1639
Date Prepared: 03/26/2010 1639

Analysis Batch: 280-8905

Instrument ID: MSV_R1
Lab File ID: R7616.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.59	J	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	33		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	105		70 - 127
Toluene-d8 (Surr)	102		80 - 125
4-Bromofluorobenzene (Surr)	97		78 - 118
Dibromofluoromethane (Surr)	100		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M069**

Lab Sample ID: 280-1489-31
Client Matrix: Water

Date Sampled: 03/13/2010 1545
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8905	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7617.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1659		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1659			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	24		0.15	1.0
trans-1,2-Dichloroethene	3.3		0.15	1.0
1,1-Dichloroethene	0.35	J	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN20-M069**

Lab Sample ID: 280-1489-31
Client Matrix: Water

Date Sampled: 03/13/2010 1545
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8905	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7617.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1659		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1659			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	2.0		0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	14		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	106		70 - 127
Toluene-d8 (Surr)	100		80 - 125
4-Bromofluorobenzene (Surr)	95		78 - 118
Dibromofluoromethane (Surr)	98		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S69B**

Lab Sample ID: 280-1489-32
Client Matrix: Water

Date Sampled: 03/15/2010 1030
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3268.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0139		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0139			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	5.9		0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.16	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S69B**

Lab Sample ID: 280-1489-32
Client Matrix: Water

Date Sampled: 03/15/2010 1030
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3268.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0139		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0139			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.45	J	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	1.1		0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	73		70 - 127
Toluene-d8 (Surr)	97		80 - 125
4-Bromofluorobenzene (Surr)	87		78 - 118
Dibromofluoromethane (Surr)	84		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S69C**

Lab Sample ID: 280-1489-33
Client Matrix: Water

Date Sampled: 03/15/2010 0935
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3269.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0200		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0200			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	3.4		0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.37	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S69C**

Lab Sample ID: 280-1489-33
Client Matrix: Water

Date Sampled: 03/15/2010 0935
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3269.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0200		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0200			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.26	J	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	1.0		0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	85		70 - 127
Toluene-d8 (Surr)	90		80 - 125
4-Bromofluorobenzene (Surr)	87		78 - 118
Dibromofluoromethane (Surr)	92		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S69D**

Lab Sample ID: 280-1489-34
Client Matrix: Water

Date Sampled: 03/15/2010 0855
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3258.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2214		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2214			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	5.5	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.70	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S69D**

Lab Sample ID: 280-1489-34
Client Matrix: Water

Date Sampled: 03/15/2010 0855
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3258.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2214		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2214			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	88		70 - 127
Toluene-d8 (Surr)	92		80 - 125
4-Bromofluorobenzene (Surr)	89		78 - 118
Dibromofluoromethane (Surr)	94		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S70D**

Lab Sample ID: 280-1489-35
Client Matrix: Water

Date Sampled: 03/15/2010 1130
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3270.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0220		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0220			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	2.4		0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	6.6		0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	21		0.15	1.0
trans-1,2-Dichloroethene	7.4		0.15	1.0
1,1-Dichloroethene	0.64	J	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S70D**

Lab Sample ID: 280-1489-35
Client Matrix: Water

Date Sampled: 03/15/2010 1130
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3270.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0220		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0220			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	1.1		0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	8.0		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	85		70 - 127
Toluene-d8 (Surr)	92		80 - 125
4-Bromofluorobenzene (Surr)	85		78 - 118
Dibromofluoromethane (Surr)	91		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0593**

Lab Sample ID: 280-1489-36
Client Matrix: Water

Date Sampled: 03/15/2010 1250
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3280.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0545		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0545			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	3.4		0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.88	J	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	5.8		0.15	1.0
trans-1,2-Dichloroethene	0.30	J	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0593**

Lab Sample ID: 280-1489-36
Client Matrix: Water

Date Sampled: 03/15/2010 1250
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8776	Instrument ID:	MSV_G
Preparation:	5030B		Lab File ID:	G3280.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0545		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0545			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	11		0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	23		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	96		70 - 127
Toluene-d8 (Surr)	92		80 - 125
4-Bromofluorobenzene (Surr)	92		78 - 118
Dibromofluoromethane (Surr)	94		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0525**

Lab Sample ID: 280-1489-37
Client Matrix: Water

Date Sampled: 03/13/2010 1135
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8905	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7618.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1720		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1720			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	1.3		0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0525**

Lab Sample ID: 280-1489-37
Client Matrix: Water

Date Sampled: 03/13/2010 1135
Date Received: 03/18/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8905	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7618.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1720		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1720			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	108		70 - 127
Toluene-d8 (Surr)	100		80 - 125
4-Bromofluorobenzene (Surr)	97		78 - 118
Dibromofluoromethane (Surr)	100		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0564-1**

Lab Sample ID: 280-1489-8
Client Matrix: Water

Date Sampled: 03/13/2010 1416
Date Received: 03/18/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34001.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1447		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1447			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	82		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0564-2**

Lab Sample ID: 280-1489-9
Client Matrix: Water

Date Sampled: 03/13/2010 1449
Date Received: 03/18/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34002.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1511		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1511			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	83		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0564-3**

Lab Sample ID: 280-1489-10
Client Matrix: Water

Date Sampled: 03/13/2010 1554
Date Received: 03/18/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34003.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1536		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1536			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	89		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0565-1**

Lab Sample ID: 280-1489-11
Client Matrix: Water

Date Sampled: 03/13/2010 0923
Date Received: 03/18/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34004.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1601		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1601			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	88		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0565-2**

Lab Sample ID: 280-1489-12
Client Matrix: Water

Date Sampled: 03/13/2010 1038
Date Received: 03/18/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34005.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1627		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1627			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	84		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0565-3**

Lab Sample ID: 280-1489-13
Client Matrix: Water

Date Sampled: 03/13/2010 1113
Date Received: 03/18/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34006.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1652		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1652			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	80		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0566-1**

Lab Sample ID: 280-1489-14
Client Matrix: Water

Date Sampled: 03/15/2010 0921
Date Received: 03/18/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34007.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1718		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1718			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	83		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0566-2**

Lab Sample ID: 280-1489-15
Client Matrix: Water

Date Sampled: 03/15/2010 0949
Date Received: 03/18/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34008.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1743		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1743			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	85		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0566-3**

Lab Sample ID: 280-1489-16
Client Matrix: Water

Date Sampled: 03/15/2010 1045
Date Received: 03/18/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34009.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1809		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1809			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	92		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0569-1**

Lab Sample ID: 280-1489-17
Client Matrix: Water

Date Sampled: 03/15/2010 1128
Date Received: 03/18/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34010.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1834		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1834			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	85		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0537**

Lab Sample ID: 280-1489-7
Client Matrix: Water

Date Sampled: 03/15/2010 0910
Date Received: 03/18/2010 0900

6010B Metals (ICP)

Method:	6010B	Analysis Batch: 280-8196		Instrument ID:	MT_025
Preparation:	3010A	Prep Batch: 280-7923		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/22/2010 1421			Final Weight/Volume:	50 mL
Date Prepared:	03/19/2010 1530				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Aluminum	76	J	18	100
Iron	2000		22	100

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0594**

Lab Sample ID: 280-1489-18
Client Matrix: Water

Date Sampled: 03/15/2010 1105
Date Received: 03/18/2010 0900

6010B Metals (ICP)

Method:	6010B	Analysis Batch: 280-8196		Instrument ID:	MT_025
Preparation:	3010A	Prep Batch: 280-7923		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/22/2010 1430			Final Weight/Volume:	50 mL
Date Prepared:	03/19/2010 1530				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Aluminum	77	J	18	100
Iron	550		22	100

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Client Sample ID: **PIN15-0593**

Lab Sample ID: 280-1489-36
Client Matrix: Water

Date Sampled: 03/15/2010 1250
Date Received: 03/18/2010 0900

6010B Metals (ICP)

Method:	6010B	Analysis Batch: 280-8196		Instrument ID:	MT_025
Preparation:	3010A	Prep Batch: 280-7923		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/22/2010 1433			Final Weight/Volume:	50 mL
Date Prepared:	03/19/2010 1530				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Aluminum	1700		18	100
Iron	15000		22	100

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B Volatile Organic Compounds (GC/MS)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DBFM %Rec	DCA %Rec	TOL %Rec	BFB %Rec
280-1489-1	PIN99-2875	91	86	92	91
280-1489-2	PIN12-0515	94	91	90	92
280-1489-3	PIN12-0516	95	92	89	90
280-1489-4	PIN12-0517	94	89	91	90
280-1489-5	PIN12-0518	95	95	91	94
280-1489-6	PIN12-0524	95	94	92	91
280-1489-6 DL	PIN12-0524 DL	95	93	90	90
280-1489-7	PIN15-0537	95	95	89	91
280-1489-7 DL	PIN15-0537 DL	98	85	95	84
280-1489-8	PIN12-0564-1	95	94	89	91
280-1489-9	PIN12-0564-2	94	95	89	82
280-1489-10	PIN12-0564-3	93	82	95	82
280-1489-11	PIN12-0565-1	94	84	94	81
280-1489-12	PIN12-0565-2	94	79	93	81
280-1489-13	PIN12-0565-3	92	82	95	79
280-1489-14	PIN12-0566-1	94	88	92	88
280-1489-15	PIN12-0566-2	94	87	95	89
280-1489-16	PIN12-0566-3	93	89	90	87
280-1489-17	PIN12-0569-1	102	96	101	94
280-1489-19	PIN20-2868	97	87	96	82
280-1489-20	PIN12-2869	97	89	92	82
280-1489-21	PIN12-2870	97	86	96	84
280-1489-22	PIN12-2871	89	79	94	87
280-1489-23	PIN99-2886	99	90	96	85
280-1489-24	PIN99-2887	99	84	95	84
280-1489-25	PIN99-2889	92	84	96	84
280-1489-26	PIN20-M001	95	100	99	98
280-1489-26	PIN20-M001	97	100	102	94
280-1489-27	PIN20-M065	96	89	92	82

Surrogate	Acceptance Limits
DBFM = Dibromofluoromethane (Surr)	77-119
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127
TOL = Toluene-d8 (Surr)	80-125
BFB = 4-Bromofluorobenzene (Surr)	78-118

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B Volatile Organic Compounds (GC/MS)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DBFM %Rec	DCA %Rec	TOL %Rec	BFB %Rec
280-1489-28	PIN20-M066	99	93	90	82
280-1489-29	PIN20-M067	92	90	104	98
280-1489-30	PIN20-M068	100	105	102	97
280-1489-31	PIN20-M069	98	106	100	95
280-1489-32	PIN12-S69B	84	73	97	87
280-1489-33	PIN12-S69C	92	85	90	87
280-1489-34	PIN12-S69D	94	88	92	89
280-1489-35	PIN12-S70D	91	85	92	85
280-1489-36	PIN15-0593	94	96	92	92
280-1489-37	PIN12-0525	100	108	100	97
MB 280-8529/6		89	84	88	88
MB 280-8666/6		93	83	92	80
MB 280-8776/7		96	89	93	90
MB 280-8905/21		97	97	105	97
LCS 280-8529/4		87	81	85	88
LCS 280-8666/4		96	79	92	82
LCS 280-8776/5		94	87	94	88
LCS 280-8905/20		94	95	101	97
LCSD 280-8529/5		91	87	88	88
LCSD 280-8666/5		97	83	91	80
LCSD 280-8776/6		99	89	95	94
280-1489-7 MS DL	PIN15-0537 MS DL	97	87	92	81
280-1489-26 MS DL	PIN20-M001 MS DL	97	101	100	96
280-1489-34 MS	PIN12-S69D MS	91	88	89	85
280-1397-F-1 MS		95	93	88	90
280-1489-7 MSD DL	PIN15-0537 MSD DL	102	84	95	83
280-1489-26 MSD DL	PIN20-M001 MSD DL	97	99	101	97
280-1489-34 MSD	PIN12-S69D MSD	91	88	91	83
280-1397-F-1 MSD		94	91	88	90

Surrogate	Acceptance Limits
DBFM = Dibromofluoromethane (Surr)	77-119
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127
TOL = Toluene-d8 (Surr)	80-125
BFB = 4-Bromofluorobenzene (Surr)	78-118

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B Volatile Organic Compounds (GC/MS)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DCA %Rec	DBFM %Rec	TOL %Rec	BFB %Rec
280-1489-18	PIN15-0594	94	89	94	92
280-1489-18	PIN15-0594	94	87	94	90
MB 280-9091/5		104	101	108	103
LCS 280-9091/3		90	89	93	94
LCSD 280-9091/4		105	102	112	107
280-1595-F-3 MS		107	101	116	108
280-1595-F-3 MSD		92	87	96	93

Surrogate	Acceptance Limits
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127
DBFM = Dibromofluoromethane (Surr)	77-119
TOL = Toluene-d8 (Surr)	80-125
BFB = 4-Bromofluorobenzene (Surr)	78-118

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DCA %Rec
280-1489-8	PIN12-0564-1	82
280-1489-9	PIN12-0564-2	83
280-1489-10	PIN12-0564-3	89
280-1489-11	PIN12-0565-1	88
280-1489-12	PIN12-0565-2	84
280-1489-13	PIN12-0565-3	80
280-1489-14	PIN12-0566-1	83
280-1489-15	PIN12-0566-2	85
280-1489-16	PIN12-0566-3	92
280-1489-17	PIN12-0569-1	85
MB 280-8660/4		113
LCS 280-8660/3		109
280-1555-D-30 MS		88
280-1555-D-30 MSD		93

Surrogate	Acceptance Limits
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Method Blank - Batch: 280-8529

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-8529/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/24/2010 1032
Date Prepared: 03/24/2010 1032

Analysis Batch: 280-8529
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_G
Lab File ID: G3161.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Method Blank - Batch: 280-8529

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-8529/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/24/2010 1032
Date Prepared: 03/24/2010 1032

Analysis Batch: 280-8529
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_G
Lab File ID: G3161.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	84	70 - 127
Toluene-d8 (Surr)	88	80 - 125
4-Bromofluorobenzene (Surr)	88	78 - 118
Dibromofluoromethane (Surr)	89	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Lab Control Sample/
Lab Control Sample Duplicate Recovery Report - Batch: 280-8529

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-8529/4	Analysis Batch:	280-8529	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3158.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 0930			Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 0930				

LCSD Lab Sample ID:	LCSD 280-8529/5	Analysis Batch:	280-8529	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3159.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 0951			Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 0951				

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	103	108	77 - 120	4	20		
Bromodichloromethane	99	106	78 - 120	6	20		
Carbon tetrachloride	105	107	80 - 120	1	21		
Chlorobenzene	98	102	78 - 120	4	20		
Chloroform	104	107	78 - 120	3	20		
1,3-Dichlorobenzene	98	100	75 - 120	3	20		
1,1-Dichloroethane	103	105	77 - 120	3	21		
trans-1,2-Dichloroethene	104	105	80 - 120	1	24		
1,1-Dichloroethene	111	115	68 - 133	3	20		
1,2-Dichloropropane	101	106	76 - 120	5	20		
Ethylbenzene	100	102	78 - 120	2	26		
Methylene Chloride	94	108	71 - 120	13	20		
Tetrachloroethene	96	98	77 - 120	2	20		
Toluene	100	106	73 - 120	6	20		
1,1,1-Trichloroethane	104	106	78 - 120	2	20		
Trichloroethene	102	105	78 - 122	4	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	81	87	70 - 127
Toluene-d8 (Surr)	85	88	80 - 125
4-Bromofluorobenzene (Surr)	88	88	78 - 118
Dibromofluoromethane (Surr)	87	91	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Laboratory Control/
Laboratory Duplicate Data Report - Batch: 280-8529

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-8529/4	Units:	ug/L	LCSD Lab Sample ID:	LCSD 280-8529/5
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/24/2010 0930			Date Analyzed:	03/24/2010 0951
Date Prepared:	03/24/2010 0930			Date Prepared:	03/24/2010 0951

Analyte	LCS Spike Amount	LCSD Spike Amount	LCS Result/Qual	LCSD Result/Qual
Benzene	5.00	5.00	5.15	5.39
Bromodichloromethane	5.00	5.00	4.97	5.29
Carbon tetrachloride	5.00	5.00	5.26	5.33
Chlorobenzene	5.00	5.00	4.91	5.12
Chloroform	5.00	5.00	5.18	5.34
1,3-Dichlorobenzene	5.00	5.00	4.89	5.02
1,1-Dichloroethane	5.00	5.00	5.14	5.27
trans-1,2-Dichloroethene	5.00	5.00	5.21	5.27
1,1-Dichloroethene	5.00	5.00	5.57	5.74
1,2-Dichloropropane	5.00	5.00	5.03	5.29
Ethylbenzene	5.00	5.00	5.02	5.11
Methylene Chloride	5.00	5.00	4.71	5.38
Tetrachloroethene	5.00	5.00	4.79	4.89
Toluene	5.00	5.00	5.01	5.30
1,1,1-Trichloroethane	5.00	5.00	5.19	5.29
Trichloroethene	5.00	5.00	5.08	5.27

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-8529

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1397-F-1 MS	Analysis Batch:	280-8529	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3171.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1427			Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1427				

MSD Lab Sample ID:	280-1397-F-1 MSD	Analysis Batch:	280-8529	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3172.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/24/2010 1448			Final Weight/Volume:	20 mL
Date Prepared:	03/24/2010 1448				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	100	98	77 - 120	2	20		
Bromodichloromethane	105	108	78 - 120	2	20		
Carbon tetrachloride	83	81	80 - 120	3	21		
Chlorobenzene	97	98	78 - 120	2	20		
Chloroform	105	104	78 - 120	0	20		
1,3-Dichlorobenzene	94	96	75 - 120	3	20		
1,1-Dichloroethane	97	98	77 - 120	1	21		
trans-1,2-Dichloroethene	91	90	80 - 120	1	24		
1,1-Dichloroethene	87	88	68 - 133	1	20		
1,2-Dichloropropane	104	108	76 - 120	3	20		
Ethylbenzene	91	92	78 - 120	0	26		
Methylene Chloride	99	100	71 - 120	1	20		
Tetrachloroethene	80	80	77 - 120	0	20		
Toluene	97	96	73 - 120	0	20		
1,1,1-Trichloroethane	89	86	78 - 120	3	20		
Trichloroethene	92	91	78 - 122	1	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	93	91	70 - 127
Toluene-d8 (Surr)	88	88	80 - 125
4-Bromofluorobenzene (Surr)	90	90	78 - 118
Dibromofluoromethane (Surr)	95	94	77 - 119

Quality Control Results

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-8529

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1397-F-1 MS	Units: ug/L	MSD Lab Sample ID:	280-1397-F-1 MSD
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/24/2010 1427		Date Analyzed:	03/24/2010 1448
Date Prepared:	03/24/2010 1427		Date Prepared:	03/24/2010 1448

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	0.16	U	5.00	5.00	5.01	4.91
Bromodichloromethane	0.17	U	5.00	5.00	5.26	5.38
Carbon tetrachloride	0.19	U	5.00	5.00	4.14	4.03
Chlorobenzene	0.17	U	5.00	5.00	4.84	4.92
Chloroform	0.16	U	5.00	5.00	5.23	5.21
1,3-Dichlorobenzene	0.13	U	5.00	5.00	4.70	4.82
1,1-Dichloroethane	0.22	U	5.00	5.00	4.86	4.91
trans-1,2-Dichloroethene	0.15	U	5.00	5.00	4.54	4.50
1,1-Dichloroethene	0.23	U	5.00	5.00	4.37	4.39
1,2-Dichloropropane	0.18	U	5.00	5.00	5.20	5.38
Ethylbenzene	0.16	U	5.00	5.00	4.56	4.58
Methylene Chloride	0.32	U	5.00	5.00	4.95	5.01
Tetrachloroethene	0.20	U	5.00	5.00	3.99	3.99
Toluene	0.17	U	5.00	5.00	4.83	4.80
1,1,1-Trichloroethane	0.16	U	5.00	5.00	4.45	4.31
Trichloroethene	0.16	U	5.00	5.00	4.60	4.56

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Method Blank - Batch: 280-8666

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-8666/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/25/2010 1503
Date Prepared: 03/25/2010 1503

Analysis Batch: 280-8666
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_Q
Lab File ID: Q4526.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Method Blank - Batch: 280-8666

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-8666/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/25/2010 1503
Date Prepared: 03/25/2010 1503

Analysis Batch: 280-8666
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_Q
Lab File ID: Q4526.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	83	70 - 127
Toluene-d8 (Surr)	92	80 - 125
4-Bromofluorobenzene (Surr)	80	78 - 118
Dibromofluoromethane (Surr)	93	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Lab Control Sample/
Lab Control Sample Duplicate Recovery Report - Batch: 280-8666

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-8666/4	Analysis Batch:	280-8666	Instrument ID:	MSV_Q
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	Q4523.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1337			Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1337				

LCSD Lab Sample ID:	LCSD 280-8666/5	Analysis Batch:	280-8666	Instrument ID:	MSV_Q
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	Q4524.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1420			Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1420				

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	90	91	77 - 120	1	20		
Bromodichloromethane	83	84	78 - 120	1	20		
Carbon tetrachloride	81	91	80 - 120	11	21		
Chlorobenzene	85	85	78 - 120	0	20		
Chloroform	87	87	78 - 120	0	20		
1,3-Dichlorobenzene	79	83	75 - 120	6	20		
1,1-Dichloroethane	88	87	77 - 120	1	21		
trans-1,2-Dichloroethene	87	90	80 - 120	3	24		
1,1-Dichloroethene	80	94	68 - 133	17	20		
1,2-Dichloropropane	90	90	76 - 120	0	20		
Ethylbenzene	81	89	78 - 120	9	26		
Methylene Chloride	91	90	71 - 120	2	20		
Tetrachloroethene	79	87	77 - 120	9	20		
Toluene	83	87	73 - 120	4	20		
1,1,1-Trichloroethane	83	90	78 - 120	8	20		
Trichloroethene	86	90	78 - 122	5	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	79	83	70 - 127
Toluene-d8 (Surr)	92	91	80 - 125
4-Bromofluorobenzene (Surr)	82	80	78 - 118
Dibromofluoromethane (Surr)	96	97	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Laboratory Control/
Laboratory Duplicate Data Report - Batch: 280-8666

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-8666/4	Units:	ug/L	LCSD Lab Sample ID:	LCSD 280-8666/5
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/25/2010 1337			Date Analyzed:	03/25/2010 1420
Date Prepared:	03/25/2010 1337			Date Prepared:	03/25/2010 1420

Analyte	LCS Spike Amount	LCSD Spike Amount	LCS Result/Qual	LCSD Result/Qual
Benzene	5.00	5.00	4.48	4.53
Bromodichloromethane	5.00	5.00	4.13	4.19
Carbon tetrachloride	5.00	5.00	4.05	4.53
Chlorobenzene	5.00	5.00	4.23	4.23
Chloroform	5.00	5.00	4.35	4.36
1,3-Dichlorobenzene	5.00	5.00	3.93	4.17
1,1-Dichloroethane	5.00	5.00	4.39	4.37
trans-1,2-Dichloroethene	5.00	5.00	4.37	4.48
1,1-Dichloroethene	5.00	5.00	3.98	4.72
1,2-Dichloropropane	5.00	5.00	4.48	4.49
Ethylbenzene	5.00	5.00	4.06	4.43
Methylene Chloride	5.00	5.00	4.56	4.49
Tetrachloroethene	5.00	5.00	3.95	4.33
Toluene	5.00	5.00	4.15	4.33
1,1,1-Trichloroethane	5.00	5.00	4.15	4.50
Trichloroethene	5.00	5.00	4.29	4.49

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-8666

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1489-7DL	Analysis Batch:	280-8666	Instrument ID:	MSV_Q
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	Q4528.D
Dilution:	1.0	Run Type:	DL	Initial Weight/Volume:	5 mL
Date Analyzed:	03/25/2010 1546			Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1546				

MSD Lab Sample ID:	280-1489-7DL	Analysis Batch:	280-8666	Instrument ID:	MSV_Q
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	Q4529.D
Dilution:	1.0	Run Type:	DL	Initial Weight/Volume:	5 mL
Date Analyzed:	03/25/2010 1608			Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1608				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	89	89	77 - 120	0	20		
Bromodichloromethane	87	88	78 - 120	2	20		
Carbon tetrachloride	86	81	80 - 120	6	21		
Chlorobenzene	90	90	78 - 120	1	20		
Chloroform	92	103	78 - 120	11	20		
1,3-Dichlorobenzene	90	91	75 - 120	0	20		
1,1-Dichloroethane	89	90	77 - 120	2	21		
trans-1,2-Dichloroethene	91	90	80 - 120	2	24		
1,1-Dichloroethene	94	92	68 - 133	2	20		
1,2-Dichloropropane	90	95	76 - 120	5	20		
Ethylbenzene	89	93	78 - 120	4	26		
Methylene Chloride	94	95	71 - 120	1	20		
Tetrachloroethene	87	88	77 - 120	0	20		
Toluene	92	89	73 - 120	4	20		
1,1,1-Trichloroethane	86	87	78 - 120	2	20		
Trichloroethene	91	88	78 - 122	3	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	87	84	70 - 127
Toluene-d8 (Surr)	92	95	80 - 125
4-Bromofluorobenzene (Surr)	81	83	78 - 118
Dibromofluoromethane (Surr)	97	102	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-8666

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1489-7DL	Units: ug/L	MSD Lab Sample ID:	280-1489-7DL
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/25/2010 1546		Date Analyzed:	03/25/2010 1608
Date Prepared:	03/25/2010 1546		Date Prepared:	03/25/2010 1608
Run Type:	DL		Run Type:	DL

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	3.1	J	20.0	20.0	21.0	20.9
Bromodichloromethane	0.68	U	20.0	20.0	17.3	17.6
Carbon tetrachloride	0.76	U	20.0	20.0	17.3	16.3
Chlorobenzene	0.68	U	20.0	20.0	17.9	18.1
Chloroform	0.64	U	20.0	20.0	18.4	20.6
1,3-Dichlorobenzene	0.52	U	20.0	20.0	18.1	18.1
1,1-Dichloroethane	0.88	U	20.0	20.0	17.7	18.1
trans-1,2-Dichloroethene	0.60	U	20.0	20.0	18.2	17.9
1,1-Dichloroethene	0.92	U	20.0	20.0	18.7	18.4
1,2-Dichloropropane	0.72	U	20.0	20.0	18.1	19.0
Ethylbenzene	0.64	U	20.0	20.0	17.9	18.6
Methylene Chloride	1.3	U	20.0	20.0	18.8	19.0
Tetrachloroethene	0.80	U	20.0	20.0	17.5	17.5
Toluene	0.68	U	20.0	20.0	18.4	17.7
1,1,1-Trichloroethane	0.64	U	20.0	20.0	17.2	17.5
Trichloroethene	0.64	U	20.0	20.0	18.2	17.7

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Method Blank - Batch: 280-8776

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-8776/7
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/25/2010 2154
Date Prepared: 03/25/2010 2154

Analysis Batch: 280-8776
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_G
Lab File ID: G3257.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Method Blank - Batch: 280-8776

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-8776/7	Analysis Batch:	280-8776	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3257.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2154			Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2154				

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	89	70 - 127
Toluene-d8 (Surr)	93	80 - 125
4-Bromofluorobenzene (Surr)	90	78 - 118
Dibromofluoromethane (Surr)	96	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Lab Control Sample/
Lab Control Sample Duplicate Recovery Report - Batch: 280-8776

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-8776/5	Analysis Batch:	280-8776	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3254.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2052			Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2052				

LCSD Lab Sample ID:	LCSD 280-8776/6	Analysis Batch:	280-8776	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3255.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2113			Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2113				

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	97	102	77 - 120	5	20		
Bromodichloromethane	94	97	78 - 120	4	20		
Carbon tetrachloride	97	103	80 - 120	6	21		
Chlorobenzene	94	96	78 - 120	2	20		
Chloroform	96	102	78 - 120	6	20		
1,3-Dichlorobenzene	93	94	75 - 120	1	20		
1,1-Dichloroethane	94	100	77 - 120	6	21		
trans-1,2-Dichloroethene	95	102	80 - 120	7	24		
1,1-Dichloroethene	104	110	68 - 133	5	20		
1,2-Dichloropropane	94	97	76 - 120	3	20		
Ethylbenzene	95	98	78 - 120	3	26		
Methylene Chloride	89	93	71 - 120	4	20		
Tetrachloroethene	96	98	77 - 120	3	20		
Toluene	94	100	73 - 120	6	20		
1,1,1-Trichloroethane	96	101	78 - 120	6	20		
Trichloroethene	96	103	78 - 122	7	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	87	89	70 - 127
Toluene-d8 (Surr)	94	95	80 - 125
4-Bromofluorobenzene (Surr)	88	94	78 - 118
Dibromofluoromethane (Surr)	94	99	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Laboratory Control/
Laboratory Duplicate Data Report - Batch: 280-8776

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-8776/5	Units:	ug/L	LCSD Lab Sample ID:	LCSD 280-8776/6
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/25/2010 2052			Date Analyzed:	03/25/2010 2113
Date Prepared:	03/25/2010 2052			Date Prepared:	03/25/2010 2113

Analyte	LCS Spike Amount	LCSD Spike Amount	LCS Result/Qual	LCSD Result/Qual
Benzene	5.00	5.00	4.84	5.08
Bromodichloromethane	5.00	5.00	4.68	4.86
Carbon tetrachloride	5.00	5.00	4.85	5.16
Chlorobenzene	5.00	5.00	4.71	4.80
Chloroform	5.00	5.00	4.80	5.12
1,3-Dichlorobenzene	5.00	5.00	4.67	4.72
1,1-Dichloroethane	5.00	5.00	4.72	5.01
trans-1,2-Dichloroethene	5.00	5.00	4.75	5.10
1,1-Dichloroethene	5.00	5.00	5.22	5.48
1,2-Dichloropropane	5.00	5.00	4.72	4.87
Ethylbenzene	5.00	5.00	4.76	4.92
Methylene Chloride	5.00	5.00	4.47	4.64
Tetrachloroethene	5.00	5.00	4.78	4.92
Toluene	5.00	5.00	4.68	4.98
1,1,1-Trichloroethane	5.00	5.00	4.79	5.07
Trichloroethene	5.00	5.00	4.79	5.15

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-8776

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1489-34	Analysis Batch:	280-8776	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3259.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2235			Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2235				

MSD Lab Sample ID:	280-1489-34	Analysis Batch:	280-8776	Instrument ID:	MSV_G
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	G3260.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 2255			Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 2255				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	100	101	77 - 120	1	20		
Bromodichloromethane	93	95	78 - 120	2	20		
Carbon tetrachloride	105	106	80 - 120	1	21		
Chlorobenzene	95	98	78 - 120	3	20		
Chloroform	99	101	78 - 120	2	20		
1,3-Dichlorobenzene	92	94	75 - 120	2	20		
1,1-Dichloroethane	98	100	77 - 120	2	21		
trans-1,2-Dichloroethene	102	102	80 - 120	0	24		
1,1-Dichloroethene	115	115	68 - 133	0	20		
1,2-Dichloropropane	95	98	76 - 120	3	20		
Ethylbenzene	97	99	78 - 120	3	26		
Methylene Chloride	91	92	71 - 120	1	20		
Tetrachloroethene	100	102	77 - 120	2	20		
Toluene	100	103	73 - 120	3	20		
1,1,1-Trichloroethane	103	104	78 - 120	1	20		
Trichloroethene	100	102	78 - 122	1	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	88	88	70 - 127
Toluene-d8 (Surr)	89	91	80 - 125
4-Bromofluorobenzene (Surr)	85	83	78 - 118
Dibromofluoromethane (Surr)	91	91	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-8776

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1489-34	Units: ug/L	MSD Lab Sample ID:	280-1489-34
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/25/2010 2235		Date Analyzed:	03/25/2010 2255
Date Prepared:	03/25/2010 2235		Date Prepared:	03/25/2010 2255

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	0.16	U	5.00	5.00	4.99	5.03
Bromodichloromethane	0.17	U	5.00	5.00	4.63	4.74
Carbon tetrachloride	0.19	U	5.00	5.00	5.27	5.30
Chlorobenzene	0.17	U	5.00	5.00	4.74	4.89
Chloroform	0.16	U	5.00	5.00	4.95	5.05
1,3-Dichlorobenzene	0.13	U	5.00	5.00	4.58	4.68
1,1-Dichloroethane	0.22	U	5.00	5.00	4.89	5.00
trans-1,2-Dichloroethene	0.15	U	5.00	5.00	5.10	5.12
1,1-Dichloroethene	0.23	U	5.00	5.00	5.77	5.75
1,2-Dichloropropane	0.18	U	5.00	5.00	4.77	4.91
Ethylbenzene	0.16	U	5.00	5.00	4.83	4.97
Methylene Chloride	0.32	U	5.00	5.00	4.56	4.62
Tetrachloroethene	0.20	U	5.00	5.00	5.00	5.10
Toluene	0.17	U	5.00	5.00	4.99	5.15
1,1,1-Trichloroethane	0.16	U	5.00	5.00	5.16	5.21
Trichloroethene	0.16	U	5.00	5.00	5.02	5.09

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Method Blank - Batch: 280-8905

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-8905/21
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/26/2010 1432
Date Prepared: 03/26/2010 1432

Analysis Batch: 280-8905
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_R1
Lab File ID: R7610.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.627	J	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Method Blank - Batch: 280-8905

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-8905/21
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/26/2010 1432
Date Prepared: 03/26/2010 1432

Analysis Batch: 280-8905
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_R1
Lab File ID: R7610.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	97	70 - 127
Toluene-d8 (Surr)	105	80 - 125
4-Bromofluorobenzene (Surr)	97	78 - 118
Dibromofluoromethane (Surr)	97	77 - 119

Quality Control Results

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Lab Control Sample - Batch: 280-8905

Method: 8260B
Preparation: 5030B

Lab Sample ID:	LCS 280-8905/20	Analysis Batch:	280-8905	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7609.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1406			Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1406				

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Benzene	5.00	4.58	92	77 - 120	
Bromodichloromethane	5.00	4.36	87	78 - 120	
Carbon tetrachloride	5.00	4.58	92	80 - 120	
Chlorobenzene	5.00	4.52	90	78 - 120	
Chloroform	5.00	4.49	90	78 - 120	
1,3-Dichlorobenzene	5.00	4.56	91	75 - 120	
1,1-Dichloroethane	5.00	4.53	91	77 - 120	
trans-1,2-Dichloroethene	5.00	4.69	94	80 - 120	
1,1-Dichloroethene	5.00	4.88	98	68 - 133	
1,2-Dichloropropane	5.00	4.51	90	76 - 120	
Ethylbenzene	5.00	4.61	92	78 - 120	
Methylene Chloride	5.00	4.98	100	71 - 120	
Tetrachloroethene	5.00	4.62	92	77 - 120	
Toluene	5.00	4.60	92	73 - 120	
1,1,1-Trichloroethane	5.00	4.61	92	78 - 120	
Trichloroethene	5.00	4.49	90	78 - 122	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	95	70 - 127
Toluene-d8 (Surr)	101	80 - 125
4-Bromofluorobenzene (Surr)	97	78 - 118
Dibromofluoromethane (Surr)	94	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-8905

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1489-26DL	Analysis Batch:	280-8905	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7614.D
Dilution:	1.0	Run Type:	DL	Initial Weight/Volume:	0.5 mL
Date Analyzed:	03/26/2010 1558			Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1558				

MSD Lab Sample ID:	280-1489-26DL	Analysis Batch:	280-8905	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7615.D
Dilution:	1.0	Run Type:	DL	Initial Weight/Volume:	0.5 mL
Date Analyzed:	03/26/2010 1618			Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1618				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	82	89	77 - 120	8	20		
Bromodichloromethane	82	88	78 - 120	7	20		
Carbon tetrachloride	88	96	80 - 120	9	21		
Chlorobenzene	79	89	78 - 120	12	20		
Chloroform	83	90	78 - 120	8	20		
1,3-Dichlorobenzene	79	87	75 - 120	10	20		
1,1-Dichloroethane	81	90	77 - 120	10	21		
trans-1,2-Dichloroethene	82	93	80 - 120	11	24		
1,1-Dichloroethene	90	101	68 - 133	11	20		
1,2-Dichloropropane	82	89	76 - 120	8	20		
Ethylbenzene	81	92	78 - 120	13	26		
Methylene Chloride	83	88	71 - 120	5	20		
Tetrachloroethene	82	92	77 - 120	12	20		
Toluene	82	90	73 - 120	10	20		
1,1,1-Trichloroethane	87	94	78 - 120	8	20		
Trichloroethene	78	87	78 - 122	12	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101	99	70 - 127
Toluene-d8 (Surr)	100	101	80 - 125
4-Bromofluorobenzene (Surr)	96	97	78 - 118
Dibromofluoromethane (Surr)	97	97	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-8905

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1489-26DL	Units: ug/L	MSD Lab Sample ID:	280-1489-26DL
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/26/2010 1558		Date Analyzed:	03/26/2010 1618
Date Prepared:	03/26/2010 1558		Date Prepared:	03/26/2010 1618
Run Type:	DL		Run Type:	DL

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	6.4	U	200	200	164	177
Bromodichloromethane	6.8	U	200	200	164	177
Carbon tetrachloride	7.6	U	200	200	176	192
Chlorobenzene	6.8	U	200	200	159	179
Chloroform	6.4	U	200	200	167	180
1,3-Dichlorobenzene	5.2	U	200	200	158	175
1,1-Dichloroethane	8.8	U	200	200	162	180
trans-1,2-Dichloroethene	21	J	200	200	185	206
1,1-Dichloroethene	9.2	U	200	200	181	201
1,2-Dichloropropane	7.2	U	200	200	164	178
Ethylbenzene	6.4	U	200	200	161	184
Methylene Chloride	27	J	200	200	194	204
Tetrachloroethene	8.0	U	200	200	164	185
Toluene	6.8	U	200	200	163	180
1,1,1-Trichloroethane	6.4	U	200	200	173	188
Trichloroethene	6.4	U	200	200	156	175

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Method Blank - Batch: 280-9091

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-9091/5
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/29/2010 1825
Date Prepared: 03/29/2010 1825

Analysis Batch: 280-9091
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_G2
Lab File ID: g2_34057.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Method Blank - Batch: 280-9091

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-9091/5	Analysis Batch:	280-9091	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_34057.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1825			Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1825				

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104	70 - 127
Toluene-d8 (Surr)	108	80 - 125
4-Bromofluorobenzene (Surr)	103	78 - 118
Dibromofluoromethane (Surr)	101	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Lab Control Sample/
Lab Control Sample Duplicate Recovery Report - Batch: 280-9091

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-9091/3	Analysis Batch:	280-9091	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_34054.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1722			Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1722				

LCSD Lab Sample ID:	LCSD 280-9091/4	Analysis Batch:	280-9091	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_34055.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1743			Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1743				

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	98	110	77 - 120	12	20		
Bromodichloromethane	92	106	78 - 120	15	20		
Carbon tetrachloride	101	115	80 - 120	13	21		
Chlorobenzene	93	109	78 - 120	15	20		
Chloroform	95	107	78 - 120	12	20		
1,3-Dichlorobenzene	91	104	75 - 120	14	20		
1,1-Dichloroethane	96	107	77 - 120	12	21		
trans-1,2-Dichloroethene	99	110	80 - 120	11	24		
1,1-Dichloroethene	111	122	68 - 133	10	20		
1,2-Dichloropropane	93	109	76 - 120	16	20		
Ethylbenzene	95	110	78 - 120	15	26		
Methylene Chloride	97	110	71 - 120	12	20		
Tetrachloroethene	85	100	77 - 120	17	20		
Toluene	99	119	73 - 120	18	20		
1,1,1-Trichloroethane	103	115	78 - 120	10	20		
Trichloroethene	94	108	78 - 122	14	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	90	105	70 - 127
Toluene-d8 (Surr)	93	112	80 - 125
4-Bromofluorobenzene (Surr)	94	107	78 - 118
Dibromofluoromethane (Surr)	89	102	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Laboratory Control/
Laboratory Duplicate Data Report - Batch: 280-9091

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-9091/3	Units:	ug/L	LCSD Lab Sample ID:	LCSD 280-9091/4
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/29/2010 1722			Date Analyzed:	03/29/2010 1743
Date Prepared:	03/29/2010 1722			Date Prepared:	03/29/2010 1743

Analyte	LCS Spike Amount	LCSD Spike Amount	LCS Result/Qual	LCSD Result/Qual
Benzene	5.00	5.00	4.88	5.49
Bromodichloromethane	5.00	5.00	4.58	5.32
Carbon tetrachloride	5.00	5.00	5.05	5.75
Chlorobenzene	5.00	5.00	4.66	5.43
Chloroform	5.00	5.00	4.76	5.35
1,3-Dichlorobenzene	5.00	5.00	4.54	5.20
1,1-Dichloroethane	5.00	5.00	4.78	5.36
trans-1,2-Dichloroethene	5.00	5.00	4.93	5.52
1,1-Dichloroethene	5.00	5.00	5.53	6.08
1,2-Dichloropropane	5.00	5.00	4.64	5.46
Ethylbenzene	5.00	5.00	4.73	5.52
Methylene Chloride	5.00	5.00	4.87	5.49
Tetrachloroethene	5.00	5.00	4.23	4.99
Toluene	5.00	5.00	4.96	5.94
1,1,1-Trichloroethane	5.00	5.00	5.17	5.74
Trichloroethene	5.00	5.00	4.71	5.41

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-9091

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1595-F-3 MS	Analysis Batch:	280-9091	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_34066.D
Dilution:	1.0			Initial Weight/Volume:	10 mL
Date Analyzed:	03/29/2010 2132			Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 2132				

MSD Lab Sample ID:	280-1595-F-3 MSD	Analysis Batch:	280-9091	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_34067.D
Dilution:	1.0			Initial Weight/Volume:	10 mL
Date Analyzed:	03/29/2010 2153			Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 2153				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	116	97	77 - 120	18	20		
Bromodichloromethane	105	88	78 - 120	17	20		
Carbon tetrachloride	123	103	80 - 120	18	21	F	
Chlorobenzene	109	92	78 - 120	17	20		
Chloroform	115	98	78 - 120	16	20		
1,3-Dichlorobenzene	106	89	75 - 120	18	20		
1,1-Dichloroethane	119	100	77 - 120	17	21		
trans-1,2-Dichloroethene	116	94	80 - 120	21	24		
1,1-Dichloroethene	126	109	68 - 133	15	20		
1,2-Dichloropropane	109	91	76 - 120	17	20		
Ethylbenzene	115	95	78 - 120	20	26		
Methylene Chloride	110	94	71 - 120	15	20		
Tetrachloroethene	105	85	77 - 120	20	20		
Toluene	121	98	73 - 120	21	20	F	F
1,1,1-Trichloroethane	127	107	78 - 120	17	20	F	
Trichloroethene	225	41	78 - 122	22	20	4	4 F

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	107	92	70 - 127
Toluene-d8 (Surr)	116	96	80 - 125
4-Bromofluorobenzene (Surr)	108	93	78 - 118
Dibromofluoromethane (Surr)	101	87	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-9091

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1595-F-3 MS	Units: ug/L	MSD Lab Sample ID: 280-1595-F-3 MSD
Client Matrix:	Water		Client Matrix: Water
Dilution:	1.0		Dilution: 1.0
Date Analyzed:	03/29/2010 2132		Date Analyzed: 03/29/2010 2153
Date Prepared:	03/29/2010 2132		Date Prepared: 03/29/2010 2153

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual		MSD Result/Qual	
Benzene	0.32	U	10.0	10.0	11.6		9.68	
Bromodichloromethane	0.34	U	10.0	10.0	10.5		8.82	
Carbon tetrachloride	0.38	U	10.0	10.0	12.3	F	10.3	
Chlorobenzene	0.34	U	10.0	10.0	10.9		9.20	
Chloroform	0.32	U	10.0	10.0	11.5		9.76	
1,3-Dichlorobenzene	0.26	U	10.0	10.0	10.6		8.85	
1,1-Dichloroethane	0.44	U	10.0	10.0	11.9		10.0	
trans-1,2-Dichloroethene	0.32	J	10.0	10.0	11.9		9.67	
1,1-Dichloroethene	0.46	U	10.0	10.0	12.6		10.9	
1,2-Dichloropropane	0.36	U	10.0	10.0	10.9		9.14	
Ethylbenzene	0.32	U	10.0	10.0	11.5		9.46	
Methylene Chloride	0.64	U	10.0	10.0	11.0		9.44	
Tetrachloroethene	0.40	U	10.0	10.0	10.5		8.52	
Toluene	0.34	U	10.0	10.0	12.1	F	9.84	F
1,1,1-Trichloroethane	0.32	U	10.0	10.0	12.7	F	10.7	
Trichloroethene	72		10.0	10.0	94.7	4	76.2	4 F

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Method Blank - Batch: 280-8660

Method: 8260B SIM
Preparation: 5030B

Lab Sample ID: MB 280-8660/4
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/25/2010 1059
Date Prepared: 03/25/2010 1059

Analysis Batch: 280-8660
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_G2
Lab File ID: g2_33992.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	113	70 - 127

Lab Control Sample - Batch: 280-8660

Method: 8260B SIM
Preparation: 5030B

Lab Sample ID: LCS 280-8660/3
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/25/2010 0940
Date Prepared: 03/25/2010 0940

Analysis Batch: 280-8660
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_G2
Lab File ID: g2_33990.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
1,4-Dioxane	5.00	5.62	112	25 - 141	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	109	70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-8660

Method: 8260B SIM
Preparation: 5030B

MS Lab Sample ID:	280-1555-D-30 MS	Analysis Batch: 280-8660	Instrument ID: MSV_G2
Client Matrix:	Water	Prep Batch: N/A	Lab File ID: g2_33994.D
Dilution:	1.0		Initial Weight/Volume: 20 mL
Date Analyzed:	03/25/2010 1150		Final Weight/Volume: 20 mL
Date Prepared:	03/25/2010 1150		

MSD Lab Sample ID:	280-1555-D-30 MSD	Analysis Batch: 280-8660	Instrument ID: MSV_G2
Client Matrix:	Water	Prep Batch: N/A	Lab File ID: g2_33995.D
Dilution:	1.0		Initial Weight/Volume: 20 mL
Date Analyzed:	03/25/2010 1215		Final Weight/Volume: 20 mL
Date Prepared:	03/25/2010 1215		

Analyte	% Rec. MS	% Rec. MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
1,4-Dioxane	78	85	25 - 141	8	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	88	93	70 - 127

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-8660

Method: 8260B SIM
Preparation: 5030B

MS Lab Sample ID:	280-1555-D-30 MS	Units: ug/L	MSD Lab Sample ID: 280-1555-D-30 MSD
Client Matrix:	Water		Client Matrix: Water
Dilution:	1.0		Dilution: 1.0
Date Analyzed:	03/25/2010 1150		Date Analyzed: 03/25/2010 1215
Date Prepared:	03/25/2010 1150		Date Prepared: 03/25/2010 1215

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
1,4-Dioxane	0.75	U	5.00	5.00	3.92	4.27

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Method Blank - Batch: 280-7923

Method: 6010B
Preparation: 3010A

Lab Sample ID: MB 280-7923/1-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/22/2010 1416
Date Prepared: 03/19/2010 1530

Analysis Batch: 280-8196
Prep Batch: 280-7923
Units: ug/L

Instrument ID: MT_025
Lab File ID: N/A
Initial Weight/Volume: 50 mL
Final Weight/Volume: 50 mL

Analyte	Result	Qual	MDL	RL
Aluminum	18	U	18	100
Iron	22	U	22	100

Lab Control Sample - Batch: 280-7923

Method: 6010B
Preparation: 3010A

Lab Sample ID: LCS 280-7923/2-A
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/22/2010 1418
Date Prepared: 03/19/2010 1530

Analysis Batch: 280-8196
Prep Batch: 280-7923
Units: ug/L

Instrument ID: MT_025
Lab File ID: N/A
Initial Weight/Volume: 50 mL
Final Weight/Volume: 50 mL

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Aluminum	2000	2070	104	87 - 111	
Iron	1000	965	97	89 - 115	

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-7923

Method: 6010B
Preparation: 3010A

MS Lab Sample ID:	280-1489-7	Analysis Batch:	280-8196	Instrument ID:	MT_025
Client Matrix:	Water	Prep Batch:	280-7923	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/22/2010 1426			Final Weight/Volume:	50 mL
Date Prepared:	03/19/2010 1530				

MSD Lab Sample ID:	280-1489-7	Analysis Batch:	280-8196	Instrument ID:	MT_025
Client Matrix:	Water	Prep Batch:	280-7923	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/22/2010 1428			Final Weight/Volume:	50 mL
Date Prepared:	03/19/2010 1530				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Aluminum	103	104	83 - 119	1	25		
Iron	86	90	52 - 155	1	25		

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-7923

Method: 6010B
Preparation: 3010A

MS Lab Sample ID:	280-1489-7	Units: ug/L	MSD Lab Sample ID:	280-1489-7
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/22/2010 1426		Date Analyzed:	03/22/2010 1428
Date Prepared:	03/19/2010 1530		Date Prepared:	03/19/2010 1530

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Aluminum	76	J	2000	2000	2130	2160
Iron	2000		1000	1000	2820	2860

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:280-8529					
LCS 280-8529/4	Lab Control Sample	T	Water	8260B	
LCSD 280-8529/5	Lab Control Sample Duplicate	T	Water	8260B	
MB 280-8529/6	Method Blank	T	Water	8260B	
280-1397-F-1 MS	Matrix Spike	T	Water	8260B	
280-1397-F-1 MSD	Matrix Spike Duplicate	T	Water	8260B	
280-1489-1	PIN99-2875	T	Water	8260B	
280-1489-2	PIN12-0515	T	Water	8260B	
280-1489-3	PIN12-0516	T	Water	8260B	
280-1489-4	PIN12-0517	T	Water	8260B	
280-1489-5	PIN12-0518	T	Water	8260B	
280-1489-6	PIN12-0524	T	Water	8260B	
280-1489-6DL	PIN12-0524	T	Water	8260B	
280-1489-7	PIN15-0537	T	Water	8260B	
280-1489-8	PIN12-0564-1	T	Water	8260B	
280-1489-9	PIN12-0564-2	T	Water	8260B	
Analysis Batch:280-8660					
LCS 280-8660/3	Lab Control Sample	T	Water	8260B SIM	
MB 280-8660/4	Method Blank	T	Water	8260B SIM	
280-1489-8	PIN12-0564-1	T	Water	8260B SIM	
280-1489-9	PIN12-0564-2	T	Water	8260B SIM	
280-1489-10	PIN12-0564-3	T	Water	8260B SIM	
280-1489-11	PIN12-0565-1	T	Water	8260B SIM	
280-1489-12	PIN12-0565-2	T	Water	8260B SIM	
280-1489-13	PIN12-0565-3	T	Water	8260B SIM	
280-1489-14	PIN12-0566-1	T	Water	8260B SIM	
280-1489-15	PIN12-0566-2	T	Water	8260B SIM	
280-1489-16	PIN12-0566-3	T	Water	8260B SIM	
280-1489-17	PIN12-0569-1	T	Water	8260B SIM	
280-1555-D-30 MS	Matrix Spike	T	Water	8260B SIM	
280-1555-D-30 MSD	Matrix Spike Duplicate	T	Water	8260B SIM	

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:280-8666					
LCS 280-8666/4	Lab Control Sample	T	Water	8260B	
LCSD 280-8666/5	Lab Control Sample Duplicate	T	Water	8260B	
MB 280-8666/6	Method Blank	T	Water	8260B	
280-1489-7DL	PIN15-0537	T	Water	8260B	
280-1489-7MSDL	Matrix Spike	T	Water	8260B	
280-1489-7MSDDL	Matrix Spike Duplicate	T	Water	8260B	
280-1489-10	PIN12-0564-3	T	Water	8260B	
280-1489-11	PIN12-0565-1	T	Water	8260B	
280-1489-12	PIN12-0565-2	T	Water	8260B	
280-1489-13	PIN12-0565-3	T	Water	8260B	
280-1489-19	PIN20-2868	T	Water	8260B	
280-1489-20	PIN12-2869	T	Water	8260B	
280-1489-21	PIN12-2870	T	Water	8260B	
280-1489-23	PIN99-2886	T	Water	8260B	
280-1489-24	PIN99-2887	T	Water	8260B	
280-1489-25	PIN99-2889	T	Water	8260B	
280-1489-27	PIN20-M065	T	Water	8260B	
280-1489-28	PIN20-M066	T	Water	8260B	
Analysis Batch:280-8776					
LCS 280-8776/5	Lab Control Sample	T	Water	8260B	
LCSD 280-8776/6	Lab Control Sample Duplicate	T	Water	8260B	
MB 280-8776/7	Method Blank	T	Water	8260B	
280-1489-14	PIN12-0566-1	T	Water	8260B	
280-1489-15	PIN12-0566-2	T	Water	8260B	
280-1489-16	PIN12-0566-3	T	Water	8260B	
280-1489-17	PIN12-0569-1	T	Water	8260B	
280-1489-22	PIN12-2871	T	Water	8260B	
280-1489-32	PIN12-S69B	T	Water	8260B	
280-1489-33	PIN12-S69C	T	Water	8260B	
280-1489-34	PIN12-S69D	T	Water	8260B	
280-1489-34MS	Matrix Spike	T	Water	8260B	
280-1489-34MSD	Matrix Spike Duplicate	T	Water	8260B	
280-1489-35	PIN12-S70D	T	Water	8260B	
280-1489-36	PIN15-0593	T	Water	8260B	

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:280-8905					
LCS 280-8905/20	Lab Control Sample	T	Water	8260B	
MB 280-8905/21	Method Blank	T	Water	8260B	
280-1489-26	PIN20-M001	T	Water	8260B	
280-1489-26MSDL	Matrix Spike	T	Water	8260B	
280-1489-26MSDDL	Matrix Spike Duplicate	T	Water	8260B	
280-1489-29	PIN20-M067	T	Water	8260B	
280-1489-30	PIN20-M068	T	Water	8260B	
280-1489-31	PIN20-M069	T	Water	8260B	
280-1489-37	PIN12-0525	T	Water	8260B	
Analysis Batch:280-9091					
LCS 280-9091/3	Lab Control Sample	T	Water	8260B	
LCSD 280-9091/4	Lab Control Sample Duplicate	T	Water	8260B	
MB 280-9091/5	Method Blank	T	Water	8260B	
280-1489-18	PIN15-0594	T	Water	8260B	
280-1595-F-3 MS	Matrix Spike	T	Water	8260B	
280-1595-F-3 MSD	Matrix Spike Duplicate	T	Water	8260B	

Report Basis
T = Total

Metals					
Prep Batch: 280-7923					
LCS 280-7923/2-A	Lab Control Sample	T	Water	3010A	
MB 280-7923/1-A	Method Blank	T	Water	3010A	
280-1489-7	PIN15-0537	T	Water	3010A	
280-1489-7MS	Matrix Spike	T	Water	3010A	
280-1489-7MSD	Matrix Spike Duplicate	T	Water	3010A	
280-1489-18	PIN15-0594	T	Water	3010A	
280-1489-36	PIN15-0593	T	Water	3010A	
Analysis Batch:280-8196					
LCS 280-7923/2-A	Lab Control Sample	T	Water	6010B	280-7923
MB 280-7923/1-A	Method Blank	T	Water	6010B	280-7923
280-1489-7	PIN15-0537	T	Water	6010B	280-7923
280-1489-7MS	Matrix Spike	T	Water	6010B	280-7923
280-1489-7MSD	Matrix Spike Duplicate	T	Water	6010B	280-7923
280-1489-18	PIN15-0594	T	Water	6010B	280-7923
280-1489-36	PIN15-0593	T	Water	6010B	280-7923

Report Basis
T = Total

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1489-1 **Client ID:** PIN99-2875

Sample Date/Time: 03/13/2010 08:00 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-1		280-8529		03/24/2010 15:29	1	TAL DEN	TDJ
A:8260B	280-1489-C-1		280-8529		03/24/2010 15:29	1	TAL DEN	TDJ

Lab ID: 280-1489-2 **Client ID:** PIN12-0515

Sample Date/Time: 03/13/2010 08:10 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-2		280-8529		03/24/2010 15:50	1	TAL DEN	TDJ
A:8260B	280-1489-C-2		280-8529		03/24/2010 15:50	1	TAL DEN	TDJ

Lab ID: 280-1489-3 **Client ID:** PIN12-0516

Sample Date/Time: 03/13/2010 08:50 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-3		280-8529		03/24/2010 16:10	1	TAL DEN	TDJ
A:8260B	280-1489-C-3		280-8529		03/24/2010 16:10	1	TAL DEN	TDJ

Lab ID: 280-1489-4 **Client ID:** PIN12-0517

Sample Date/Time: 03/13/2010 10:25 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-4		280-8529		03/24/2010 16:31	1	TAL DEN	TDJ
A:8260B	280-1489-C-4		280-8529		03/24/2010 16:31	1	TAL DEN	TDJ

Lab ID: 280-1489-5 **Client ID:** PIN12-0518

Sample Date/Time: 03/13/2010 09:35 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-5		280-8529		03/24/2010 16:51	1	TAL DEN	TDJ
A:8260B	280-1489-C-5		280-8529		03/24/2010 16:51	1	TAL DEN	TDJ

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1489-6 **Client ID:** PIN12-0524

Sample Date/Time: 03/13/2010 11:05 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-B-6		280-8529		03/24/2010 17:12	1	TAL DEN	TDJ
A:8260B	280-1489-B-6		280-8529		03/24/2010 17:12	1	TAL DEN	TDJ
P:5030B	280-1489-B-6	DL	280-8529		03/24/2010 17:32	1	TAL DEN	TDJ
A:8260B	280-1489-B-6	DL	280-8529		03/24/2010 17:32	1	TAL DEN	TDJ

Lab ID: 280-1489-7 **Client ID:** PIN15-0537

Sample Date/Time: 03/15/2010 09:10 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-7		280-8529		03/24/2010 17:53	1	TAL DEN	TDJ
A:8260B	280-1489-C-7		280-8529		03/24/2010 17:53	1	TAL DEN	TDJ
P:5030B	280-1489-A-7	DL	280-8666		03/25/2010 15:25	1	TAL DEN	SAT
A:8260B	280-1489-A-7	DL	280-8666		03/25/2010 15:25	1	TAL DEN	SAT
P:3010A	280-1489-D-7-A		280-8196	280-7923	03/19/2010 15:30	1	TAL DEN	CGG
A:6010B	280-1489-D-7-A		280-8196	280-7923	03/22/2010 14:21	1	TAL DEN	JKH

Lab ID: 280-1489-7 MS **Client ID:** PIN15-0537

Sample Date/Time: 03/15/2010 09:10 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-A-7 MS	DL	280-8666		03/25/2010 15:46	1	TAL DEN	SAT
A:8260B	280-1489-A-7 MS	DL	280-8666		03/25/2010 15:46	1	TAL DEN	SAT
P:3010A	280-1489-D-7-B MS		280-8196	280-7923	03/19/2010 15:30	1	TAL DEN	CGG
A:6010B	280-1489-D-7-B MS		280-8196	280-7923	03/22/2010 14:26	1	TAL DEN	JKH

Lab ID: 280-1489-7 MSD **Client ID:** PIN15-0537

Sample Date/Time: 03/15/2010 09:10 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-A-7 MSD	DL	280-8666		03/25/2010 16:08	1	TAL DEN	SAT
A:8260B	280-1489-A-7 MSD	DL	280-8666		03/25/2010 16:08	1	TAL DEN	SAT
P:3010A	280-1489-D-7-C MSD		280-8196	280-7923	03/19/2010 15:30	1	TAL DEN	CGG
A:6010B	280-1489-D-7-C MSD		280-8196	280-7923	03/22/2010 14:28	1	TAL DEN	JKH

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1489-8 **Client ID:** PIN12-0564-1

Sample Date/Time: 03/13/2010 14:16 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-8		280-8529		03/24/2010 18:14	1	TAL DEN	TDJ
A:8260B	280-1489-C-8		280-8529		03/24/2010 18:14	1	TAL DEN	TDJ
P:5030B	280-1489-B-8		280-8660		03/25/2010 14:47	1	TAL DEN	AEW
A:8260B SIM	280-1489-B-8		280-8660		03/25/2010 14:47	1	TAL DEN	AEW

Lab ID: 280-1489-9 **Client ID:** PIN12-0564-2

Sample Date/Time: 03/13/2010 14:49 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-D-9		280-8529		03/24/2010 18:34	1	TAL DEN	TDJ
A:8260B	280-1489-D-9		280-8529		03/24/2010 18:34	1	TAL DEN	TDJ
P:5030B	280-1489-A-9		280-8660		03/25/2010 15:11	1	TAL DEN	AEW
A:8260B SIM	280-1489-A-9		280-8660		03/25/2010 15:11	1	TAL DEN	AEW

Lab ID: 280-1489-10 **Client ID:** PIN12-0564-3

Sample Date/Time: 03/13/2010 15:54 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-A-10		280-8666		03/25/2010 16:29	1	TAL DEN	SAT
A:8260B	280-1489-A-10		280-8666		03/25/2010 16:29	1	TAL DEN	SAT
P:5030B	280-1489-D-10		280-8660		03/25/2010 15:36	1	TAL DEN	AEW
A:8260B SIM	280-1489-D-10		280-8660		03/25/2010 15:36	1	TAL DEN	AEW

Lab ID: 280-1489-11 **Client ID:** PIN12-0565-1

Sample Date/Time: 03/13/2010 09:23 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-B-11		280-8666		03/25/2010 16:51	1	TAL DEN	SAT
A:8260B	280-1489-B-11		280-8666		03/25/2010 16:51	1	TAL DEN	SAT
P:5030B	280-1489-C-11		280-8660		03/25/2010 16:01	1	TAL DEN	AEW
A:8260B SIM	280-1489-C-11		280-8660		03/25/2010 16:01	1	TAL DEN	AEW

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1489-12 **Client ID:** PIN12-0565-2

Sample Date/Time: 03/13/2010 10:38 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-D-12		280-8666		03/25/2010 17:13	1	TAL DEN	SAT
A:8260B	280-1489-D-12		280-8666		03/25/2010 17:13	1	TAL DEN	SAT
P:5030B	280-1489-B-12		280-8660		03/25/2010 16:27	1	TAL DEN	AEW
A:8260B SIM	280-1489-B-12		280-8660		03/25/2010 16:27	1	TAL DEN	AEW

Lab ID: 280-1489-13 **Client ID:** PIN12-0565-3

Sample Date/Time: 03/13/2010 11:13 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-13		280-8666		03/25/2010 17:34	1	TAL DEN	SAT
A:8260B	280-1489-C-13		280-8666		03/25/2010 17:34	1	TAL DEN	SAT
P:5030B	280-1489-D-13		280-8660		03/25/2010 16:52	1	TAL DEN	AEW
A:8260B SIM	280-1489-D-13		280-8660		03/25/2010 16:52	1	TAL DEN	AEW

Lab ID: 280-1489-14 **Client ID:** PIN12-0566-1

Sample Date/Time: 03/15/2010 09:21 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-D-14		280-8776		03/25/2010 23:36	1	TAL DEN	DPI
A:8260B	280-1489-D-14		280-8776		03/25/2010 23:36	1	TAL DEN	DPI
P:5030B	280-1489-B-14		280-8660		03/25/2010 17:18	1	TAL DEN	AEW
A:8260B SIM	280-1489-B-14		280-8660		03/25/2010 17:18	1	TAL DEN	AEW

Lab ID: 280-1489-15 **Client ID:** PIN12-0566-2

Sample Date/Time: 03/15/2010 09:49 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-D-15		280-8776		03/25/2010 23:57	1	TAL DEN	DPI
A:8260B	280-1489-D-15		280-8776		03/25/2010 23:57	1	TAL DEN	DPI
P:5030B	280-1489-B-15		280-8660		03/25/2010 17:43	1	TAL DEN	AEW
A:8260B SIM	280-1489-B-15		280-8660		03/25/2010 17:43	1	TAL DEN	AEW

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1489-16 **Client ID:** PIN12-0566-3

Sample Date/Time: 03/15/2010 10:45 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-16		280-8776		03/26/2010 00:17	1	TAL DEN	DPI
A:8260B	280-1489-C-16		280-8776		03/26/2010 00:17	1	TAL DEN	DPI
P:5030B	280-1489-D-16		280-8660		03/25/2010 18:09	1	TAL DEN	AEW
A:8260B SIM	280-1489-D-16		280-8660		03/25/2010 18:09	1	TAL DEN	AEW

Lab ID: 280-1489-17 **Client ID:** PIN12-0569-1

Sample Date/Time: 03/15/2010 11:28 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-B-17		280-8776		03/26/2010 00:38	1	TAL DEN	DPI
A:8260B	280-1489-B-17		280-8776		03/26/2010 00:38	1	TAL DEN	DPI
P:5030B	280-1489-D-17		280-8660		03/25/2010 18:34	1	TAL DEN	AEW
A:8260B SIM	280-1489-D-17		280-8660		03/25/2010 18:34	1	TAL DEN	AEW

Lab ID: 280-1489-18 **Client ID:** PIN15-0594

Sample Date/Time: 03/15/2010 11:05 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-18		280-9091		03/29/2010 18:46	1	TAL DEN	DPI
A:8260B	280-1489-C-18		280-9091		03/29/2010 18:46	1	TAL DEN	DPI
P:5030B	280-1489-C-18		280-9091		03/29/2010 19:07	1	TAL DEN	DPI
A:8260B	280-1489-C-18		280-9091		03/29/2010 19:07	1	TAL DEN	DPI
P:3010A	280-1489-A-18-A		280-8196	280-7923	03/19/2010 15:30	1	TAL DEN	CGG
A:6010B	280-1489-A-18-A		280-8196	280-7923	03/22/2010 14:30	1	TAL DEN	JKH

Lab ID: 280-1489-19 **Client ID:** PIN20-2868

Sample Date/Time: 03/13/2010 17:01 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-19		280-8666		03/25/2010 18:39	1	TAL DEN	SAT
A:8260B	280-1489-C-19		280-8666		03/25/2010 18:39	1	TAL DEN	SAT

Lab ID: 280-1489-20 **Client ID:** PIN12-2869

Sample Date/Time: 03/13/2010 12:00 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-20		280-8666		03/25/2010 19:00	1	TAL DEN	SAT
A:8260B	280-1489-C-20		280-8666		03/25/2010 19:00	1	TAL DEN	SAT

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1489-21 **Client ID:** PIN12-2870

Sample Date/Time: 03/13/2010 12:00 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-B-21		280-8666		03/25/2010 19:22	1	TAL DEN	SAT
A:8260B	280-1489-B-21		280-8666		03/25/2010 19:22	1	TAL DEN	SAT

Lab ID: 280-1489-22 **Client ID:** PIN12-2871

Sample Date/Time: 03/15/2010 12:00 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-B-22		280-8776		03/26/2010 01:19	1	TAL DEN	DPI
A:8260B	280-1489-B-22		280-8776		03/26/2010 01:19	1	TAL DEN	DPI

Lab ID: 280-1489-23 **Client ID:** PIN99-2886

Sample Date/Time: 03/13/2010 08:00 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-23		280-8666		03/25/2010 19:43	1	TAL DEN	SAT
A:8260B	280-1489-C-23		280-8666		03/25/2010 19:43	1	TAL DEN	SAT

Lab ID: 280-1489-24 **Client ID:** PIN99-2887

Sample Date/Time: 03/13/2010 08:00 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-A-24		280-8666		03/25/2010 20:05	1	TAL DEN	SAT
A:8260B	280-1489-A-24		280-8666		03/25/2010 20:05	1	TAL DEN	SAT

Lab ID: 280-1489-25 **Client ID:** PIN99-2889

Sample Date/Time: 03/13/2010 08:00 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-B-25		280-8666		03/25/2010 20:27	1	TAL DEN	SAT
A:8260B	280-1489-B-25		280-8666		03/25/2010 20:27	1	TAL DEN	SAT

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1489-26 **Client ID:** PIN20-M001

Sample Date/Time: 03/13/2010 14:35 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-A-26		280-8905		03/26/2010 15:16	1	TAL DEN	SAT
A:8260B	280-1489-A-26		280-8905		03/26/2010 15:16	1	TAL DEN	SAT
P:5030B	280-1489-A-26		280-8905		03/26/2010 15:37	1	TAL DEN	SAT
A:8260B	280-1489-A-26		280-8905		03/26/2010 15:37	1	TAL DEN	SAT

Lab ID: 280-1489-26 MS **Client ID:** PIN20-M001

Sample Date/Time: 03/13/2010 14:35 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-A-26 MS	DL	280-8905		03/26/2010 15:58	1	TAL DEN	SAT
A:8260B	280-1489-A-26 MS	DL	280-8905		03/26/2010 15:58	1	TAL DEN	SAT

Lab ID: 280-1489-26 MSD **Client ID:** PIN20-M001

Sample Date/Time: 03/13/2010 14:35 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-A-26 MSD	DL	280-8905		03/26/2010 16:18	1	TAL DEN	SAT
A:8260B	280-1489-A-26 MSD	DL	280-8905		03/26/2010 16:18	1	TAL DEN	SAT

Lab ID: 280-1489-27 **Client ID:** PIN20-M065

Sample Date/Time: 03/13/2010 10:55 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-27		280-8666		03/25/2010 20:48	1	TAL DEN	SAT
A:8260B	280-1489-C-27		280-8666		03/25/2010 20:48	1	TAL DEN	SAT

Lab ID: 280-1489-28 **Client ID:** PIN20-M066

Sample Date/Time: 03/13/2010 09:45 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-A-28		280-8666		03/25/2010 21:10	1	TAL DEN	SAT
A:8260B	280-1489-A-28		280-8666		03/25/2010 21:10	1	TAL DEN	SAT

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1489-29 **Client ID:** PIN20-M067

Sample Date/Time: 03/13/2010 11:45 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-29		280-8905		03/26/2010 14:56	1	TAL DEN	SAT
A:8260B	280-1489-C-29		280-8905		03/26/2010 14:56	1	TAL DEN	SAT

Lab ID: 280-1489-30 **Client ID:** PIN20-M068

Sample Date/Time: 03/13/2010 15:00 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-B-30		280-8905		03/26/2010 16:39	1	TAL DEN	SAT
A:8260B	280-1489-B-30		280-8905		03/26/2010 16:39	1	TAL DEN	SAT

Lab ID: 280-1489-31 **Client ID:** PIN20-M069

Sample Date/Time: 03/13/2010 15:45 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-B-31		280-8905		03/26/2010 16:59	1	TAL DEN	SAT
A:8260B	280-1489-B-31		280-8905		03/26/2010 16:59	1	TAL DEN	SAT

Lab ID: 280-1489-32 **Client ID:** PIN12-S69B

Sample Date/Time: 03/15/2010 10:30 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-32		280-8776		03/26/2010 01:39	1	TAL DEN	DPI
A:8260B	280-1489-C-32		280-8776		03/26/2010 01:39	1	TAL DEN	DPI

Lab ID: 280-1489-33 **Client ID:** PIN12-S69C

Sample Date/Time: 03/15/2010 09:35 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-33		280-8776		03/26/2010 02:00	1	TAL DEN	DPI
A:8260B	280-1489-C-33		280-8776		03/26/2010 02:00	1	TAL DEN	DPI

Lab ID: 280-1489-34 **Client ID:** PIN12-S69D

Sample Date/Time: 03/15/2010 08:55 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-34		280-8776		03/25/2010 22:14	1	TAL DEN	DPI
A:8260B	280-1489-C-34		280-8776		03/25/2010 22:14	1	TAL DEN	DPI

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1489-34 MS **Client ID:** PIN12-S69D

Sample Date/Time: 03/15/2010 08:55 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-B-34 MS		280-8776		03/25/2010 22:35	1	TAL DEN	DPI
A:8260B	280-1489-B-34 MS		280-8776		03/25/2010 22:35	1	TAL DEN	DPI

Lab ID: 280-1489-34 MSD **Client ID:** PIN12-S69D

Sample Date/Time: 03/15/2010 08:55 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-B-34 MSD		280-8776		03/25/2010 22:55	1	TAL DEN	DPI
A:8260B	280-1489-B-34 MSD		280-8776		03/25/2010 22:55	1	TAL DEN	DPI

Lab ID: 280-1489-35 **Client ID:** PIN12-S70D

Sample Date/Time: 03/15/2010 11:30 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-35		280-8776		03/26/2010 02:20	1	TAL DEN	DPI
A:8260B	280-1489-C-35		280-8776		03/26/2010 02:20	1	TAL DEN	DPI

Lab ID: 280-1489-36 **Client ID:** PIN15-0593

Sample Date/Time: 03/15/2010 12:50 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-B-36		280-8776		03/26/2010 05:45	1	TAL DEN	DPI
A:8260B	280-1489-B-36		280-8776		03/26/2010 05:45	1	TAL DEN	DPI
P:3010A	280-1489-D-36-A		280-8196	280-7923	03/19/2010 15:30	1	TAL DEN	CGG
A:6010B	280-1489-D-36-A		280-8196	280-7923	03/22/2010 14:33	1	TAL DEN	JKH

Lab ID: 280-1489-37 **Client ID:** PIN12-0525

Sample Date/Time: 03/13/2010 11:35 Received Date/Time: 03/18/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1489-C-37		280-8905		03/26/2010 17:20	1	TAL DEN	SAT
A:8260B	280-1489-C-37		280-8905		03/26/2010 17:20	1	TAL DEN	SAT

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
SDG: 10022892

Laboratory Chronicle

Lab ID: **MB** **Client ID:** **N/A**

Sample Date/Time: N/A Received Date/Time: N/A

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	MB 280-8529/6		280-8529		03/24/2010 10:32	1	TAL DEN	TDJ
A:8260B	MB 280-8529/6		280-8529		03/24/2010 10:32	1	TAL DEN	TDJ
P:5030B	MB 280-8666/6		280-8666		03/25/2010 15:03	1	TAL DEN	SAT
A:8260B	MB 280-8666/6		280-8666		03/25/2010 15:03	1	TAL DEN	SAT
P:5030B	MB 280-8776/7		280-8776		03/25/2010 21:54	1	TAL DEN	DPI
A:8260B	MB 280-8776/7		280-8776		03/25/2010 21:54	1	TAL DEN	DPI
P:5030B	MB 280-8905/21		280-8905		03/26/2010 14:32	1	TAL DEN	SAT
A:8260B	MB 280-8905/21		280-8905		03/26/2010 14:32	1	TAL DEN	SAT
P:5030B	MB 280-9091/5		280-9091		03/29/2010 18:25	1	TAL DEN	DPI
A:8260B	MB 280-9091/5		280-9091		03/29/2010 18:25	1	TAL DEN	DPI
P:5030B	MB 280-8660/4		280-8660		03/25/2010 10:59	1	TAL DEN	AEW
A:8260B SIM	MB 280-8660/4		280-8660		03/25/2010 10:59	1	TAL DEN	AEW
P:3010A	MB 280-7923/1-A		280-8196	280-7923	03/19/2010 15:30	1	TAL DEN	CGG
A:6010B	MB 280-7923/1-A		280-8196	280-7923	03/22/2010 14:16	1	TAL DEN	JKH

Lab ID: **LCS** **Client ID:** **N/A**

Sample Date/Time: N/A Received Date/Time: N/A

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	LCS 280-8529/4		280-8529		03/24/2010 09:30	1	TAL DEN	TDJ
A:8260B	LCS 280-8529/4		280-8529		03/24/2010 09:30	1	TAL DEN	TDJ
P:5030B	LCS 280-8666/4		280-8666		03/25/2010 13:37	1	TAL DEN	SAT
A:8260B	LCS 280-8666/4		280-8666		03/25/2010 13:37	1	TAL DEN	SAT
P:5030B	LCS 280-8776/5		280-8776		03/25/2010 20:52	1	TAL DEN	DPI
A:8260B	LCS 280-8776/5		280-8776		03/25/2010 20:52	1	TAL DEN	DPI
P:5030B	LCS 280-8905/20		280-8905		03/26/2010 14:06	1	TAL DEN	SAT
A:8260B	LCS 280-8905/20		280-8905		03/26/2010 14:06	1	TAL DEN	SAT
P:5030B	LCS 280-9091/3		280-9091		03/29/2010 17:22	1	TAL DEN	DPI
A:8260B	LCS 280-9091/3		280-9091		03/29/2010 17:22	1	TAL DEN	DPI
P:5030B	LCS 280-8660/3		280-8660		03/25/2010 09:40	1	TAL DEN	AEW
A:8260B SIM	LCS 280-8660/3		280-8660		03/25/2010 09:40	1	TAL DEN	AEW
P:3010A	LCS 280-7923/2-A		280-8196	280-7923	03/19/2010 15:30	1	TAL DEN	CGG
A:6010B	LCS 280-7923/2-A		280-8196	280-7923	03/22/2010 14:18	1	TAL DEN	JKH

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
SDG: 10022892

Laboratory Chronicle

Lab ID: LCSD **Client ID:** N/A

Sample Date/Time: N/A Received Date/Time: N/A

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	LCSD 280-8529/5		280-8529		03/24/2010 09:51	1	TAL DEN	TDJ
A:8260B	LCSD 280-8529/5		280-8529		03/24/2010 09:51	1	TAL DEN	TDJ
P:5030B	LCSD 280-8666/5		280-8666		03/25/2010 14:20	1	TAL DEN	SAT
A:8260B	LCSD 280-8666/5		280-8666		03/25/2010 14:20	1	TAL DEN	SAT
P:5030B	LCSD 280-8776/6		280-8776		03/25/2010 21:13	1	TAL DEN	DPI
A:8260B	LCSD 280-8776/6		280-8776		03/25/2010 21:13	1	TAL DEN	DPI
P:5030B	LCSD 280-9091/4		280-9091		03/29/2010 17:43	1	TAL DEN	DPI
A:8260B	LCSD 280-9091/4		280-9091		03/29/2010 17:43	1	TAL DEN	DPI

Lab ID: MS **Client ID:** N/A

Sample Date/Time: 03/15/2010 14:00 Received Date/Time: 03/16/2010 14:52

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1397-F-1 MS		280-8529		03/24/2010 14:27	1	TAL DEN	TDJ
A:8260B	280-1397-F-1 MS		280-8529		03/24/2010 14:27	1	TAL DEN	TDJ
P:5030B	280-1595-F-3 MS		280-9091		03/29/2010 21:32	1	TAL DEN	DPI
A:8260B	280-1595-F-3 MS		280-9091		03/29/2010 21:32	1	TAL DEN	DPI
P:5030B	280-1555-D-30 MS		280-8660		03/25/2010 11:50	1	TAL DEN	AEW
A:8260B SIM	280-1555-D-30 MS		280-8660		03/25/2010 11:50	1	TAL DEN	AEW

Lab ID: MSD **Client ID:** N/A

Sample Date/Time: 03/15/2010 14:00 Received Date/Time: 03/16/2010 14:52

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1397-F-1 MSD		280-8529		03/24/2010 14:48	1	TAL DEN	TDJ
A:8260B	280-1397-F-1 MSD		280-8529		03/24/2010 14:48	1	TAL DEN	TDJ
P:5030B	280-1595-F-3 MSD		280-9091		03/29/2010 21:53	1	TAL DEN	DPI
A:8260B	280-1595-F-3 MSD		280-9091		03/29/2010 21:53	1	TAL DEN	DPI
P:5030B	280-1555-D-30 MSD		280-8660		03/25/2010 12:15	1	TAL DEN	AEW
A:8260B SIM	280-1555-D-30 MSD		280-8660		03/25/2010 12:15	1	TAL DEN	AEW

Lab References:
TAL DEN = TestAmerica Denver

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1489-1

SDG No.: 10022892

Instrument ID: MSV_R1 Analysis Batch Number: 8905

Lab Sample ID: IC 280-8905/2 Client Sample ID:

Date Analyzed: 03/26/10 07:45 Lab File ID: R7591.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
trans-1,3-Dichloropropene	9.34	Analyte not Identified by the Data System	tinkhams	03/26/10 11:59

Lab Sample ID: IC 280-8905/4 Client Sample ID:

Date Analyzed: 03/26/10 08:27 Lab File ID: R7593.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Ethanol	5.33	Analyte not Identified by the Data System	tinkhams	03/26/10 11:58

Lab Sample ID: IC 280-8905/10 Client Sample ID:

Date Analyzed: 03/26/10 10:39 Lab File ID: R7599.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Chloroethyl vinyl ether	8.79	Analyte not Identified by the Data System	waterlah	03/30/10 08:37

Lab Sample ID: 280-1489-26 Client Sample ID: PIN20-M001

Date Analyzed: 03/26/10 15:16 Lab File ID: R7612.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Carbon disulfide	6.10	Split Peak	rhoadesw	03/31/10 18:46

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1489-1

SDG No.: 10022892

Instrument ID: MSV_R1 Analysis Batch Number: 8905

Lab Sample ID: IC 280-8905/2 Client Sample ID:

Date Analyzed: 03/26/10 07:45 Lab File ID: R7591.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
trans-1,3-Dichloropropene	9.34	Analyte not Identified by the Data System	tinkhams	03/26/10 11:59

Lab Sample ID: IC 280-8905/4 Client Sample ID:

Date Analyzed: 03/26/10 08:27 Lab File ID: R7593.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Ethanol	5.33	Analyte not Identified by the Data System	tinkhams	03/26/10 11:58

Lab Sample ID: IC 280-8905/10 Client Sample ID:

Date Analyzed: 03/26/10 10:39 Lab File ID: R7599.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Chloroethyl vinyl ether	8.79	Analyte not Identified by the Data System	waterlah	03/30/10 08:37

Lab Sample ID: 280-1489-26 Client Sample ID: PIN20-M001

Date Analyzed: 03/26/10 15:16 Lab File ID: R7612.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Carbon disulfide	6.10	Split Peak	rhoadesw	03/31/10 18:46

WK 4/7/10

8260B

Shipping and Receiving Documents

2.7
3/18/10

Stoller
Legacy Management Team

Chain of Custody / Sample Submittal Form

RIN: 10022892
COC: 10022892.3.1
Sampler(s): baer, lombardi, moe, atkinson, caballero, walters

Project: Pinellas Monitoring
Purchase Order: 3864
Cost Number: 1-502-1-06-509-4-02
Turnaround (Days): 28
Matrix: WA - Water

Laboratory: TestAmerica Denver
Address: 4955 Yarrow Street
Arvada, Colorado 80002
Phone: 303 736 0100

Ship #	Ticket	Sample Date	Time	Site	Location	Container	# Cont.	Preservation	Matrix	Comp.	Grab	Filtered	QC	Analysis
3	IDW 097	03/13/2010	8:00	PIN99	PIN99-2875	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 099	03/13/2010	8:10	PIN12	PIN12-0515	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 100	03/13/2010	8:50	PIN12	PIN12-0516	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 101	03/13/2010	10:25	PIN12	PIN12-0517	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 102	03/13/2010	9:35	PIN12	PIN12-0518	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 104	03/13/2010	11:05	PIN12	PIN12-0524	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 105	03/13/2010	11:35	PIN12	PIN12-0525	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 210	03/15/2010	9:10	PIN15	PIN15-0537	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 210	03/15/2010	9:10	PIN15	PIN15-0537	HDPE 250 mL	1	HNO3	WA			N		Al,Fe
3	IDW 163	03/13/2010	14:16	PIN12	PIN12-0564-1	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
3	IDW 164	03/13/2010	14:49	PIN12	PIN12-0564-2	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
3	IDW 165	03/13/2010	15:54	PIN12	PIN12-0564-3	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
3	IDW 166	03/13/2010	9:23	PIN12	PIN12-0565-1	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
3	IDW 167	03/13/2010	10:38	PIN12	PIN12-0565-2	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
3	IDW 168	03/13/2010	11:13	PIN12	PIN12-0565-3	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
3	IDW 169	03/15/2010	9:21	PIN12	PIN12-0566-1	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane

Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time
Amp.	3-16-10	1705	Amanda Harrison	3-17-10	1700			
Received by (signature)	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**
Elmus Elbig	3-16-10	1705		3-18-10	0800			

1.2°C & 1.4°C CU-07

Stoller
Legacy Management Team

Chain of Custody / Sample Submittal Form

RIN: 10022892
COC: 10022892.3.2
Sampler(s): baer, lombardi, moe, atkinson, caballero, walters

Project: Pinellas Monitoring
Purchase Order: 3864 **Cost Number:** 1-502-1-06-509-4-02
Turnaround (Days): 28 **Matrix:** WA - Water

Laboratory: TestAmerica Denver
Address: 4955 Yarrow Street
Arvada, Colorado 80002
Phone: 303 736 0100

Ship #	Ticket	Sample Date	Time	Site	Location	Container	# Cont.	Preservation	Matrix	Comp.	Grab	Filtered	QC	Analysis
3	IDW 170	03/15/2010	9:49	PIN12	PIN12-0566-2	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
3	IDW 171	03/15/2010	10:45	PIN12	PIN12-0566-3	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
3	IDW 178	03/15/2010	11:28	PIN12	PIN12-0569-1	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
3	IDW 214	03/15/2010	11:05	PIN15	PIN15-0594	Glass 40 mL	2	4 C, HCl	WA			N		VOA limited volume! Edible Oil Substrate present; May interfere with analysis
3	IDW 214	03/15/2010	11:05	PIN15	PIN15-0594	HDPE 250 mL	1	HNO3	WA			N		Al,Fe Edible Oil Substrate present; May interfere with analysis
3	IDW 236	03/13/2010	17:01	PIN20	PIN20-2868	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 237	03/13/2010	12:00	PIN12	PIN12-2869	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 238	03/13/2010	12:00	PIN12	PIN12-2870	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 239	03/15/2010	12:00	PIN12	PIN12-2871	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 254	03/13/2010	8:00	PIN99	PIN99-2886	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 255	03/13/2010	8:00	PIN99	PIN99-2887	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 257	03/13/2010	8:00	PIN99	PIN99-2889	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 218	03/13/2010	14:35	PIN20	PIN20-M001	Glass 40 mL	3	4 C, HCl	WA			N		VOA Edible Oil Substrate present; May interfere with analysis
3	IDW 223	03/13/2010	10:55	PIN20	PIN20-M065	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 224	03/13/2010	9:45	PIN20	PIN20-M066	Glass 40 mL	3	4 C, HCl	WA			N		VOA
3	IDW 225	03/13/2010	11:45	PIN20	PIN20-M067	Glass 40 mL	3	4 C, HCl	WA			N		VOA

Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time
[signature]	3-16-10	1705	[signature]	3/17/10	1700			
Received by (signature)	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**
[signature]	3-16-10	1705	[signature]	3-18-10	0800			

Stoller
Legacy Management Team

Chain of Custody / Sample Submittal Form

RIN: 10022892
COC: 10022892.3.3
Sampler(s): baer, lombardi, moe, atkinson, caballero, walters

Project: Pinellas Monitoring
Purchase Order: 3864
Cost Number: 1-502-1-06-509-4-02
Turnaround (Days): 28
Matrix: WA - Water

Laboratory: TestAmerica Denver
Address: 4955 Yarrow Street
Arvada, Colorado 80002
Phone: 303 736 0100

Ship #	Ticket	Sample Date	Time	Site	Location	Container	# Cont.	Preservation	Matrix	Comp.	Grab	Filtered	QC	Analysis	
3	IDW 226	03/13/2010	15:00	PIN20	PIN20-M068	Glass 40 mL	3	4 C, HCl	WA			N		VOA	
3	IDW 227	03/13/2010	15:45	PIN20	PIN20-M069	Glass 40 mL	3	4 C, HCl	WA			N		VOA	
3	IDW 191	03/15/2010	10:30	PIN12	PIN12-S69B	Glass 40 mL	3	4 C, HCl	WA			N		VOA	
3	IDW 192	03/15/2010	9:35	PIN12	PIN12-S69C	Glass 40 mL	3	4 C, HCl	WA			N		VOA	
3	IDW 193	03/15/2010	8:55	PIN12	PIN12-S69D	Glass 40 mL	3	4 C, HCl	WA			N		VOA	
3	IDW 196	03/15/2010	11:30	PIN12	PIN12-S70D	Glass 40 mL	3	4 C, HCl	WA			N		VOA	
3	IDW 213	03/15/2010	12:50	PIN15	PIN15-0593	Glass 40 mL	3	4 C, HCl	WA			N		VOA	Edible Oil Substrate present; May interfere with analysis
3	IDW 213	03/15/2010	12:50	PIN15	PIN15-0593	HDPE 250 mL	1	HNO3	WA			N		Al,Fe	Edible Oil Substrate present; May interfere with analysis

Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time
	3-16-10	1705		3/17/10	1700			
Received by (signature)	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**
	3-16-10	1705		3-18-10	0800			

Login Sample Receipt Check List

Client: S.M. Stoller Corporation

Job Number: 280-1489-1
SDG Number: 10022892

Login Number: 1489
Creator: Harrington, Nicholas
List Number: 1

List Source: TestAmerica Denver

Question	T / F/ NA	Comment
Radioactivity either was not measured or, if measured, is at or below background	True	
The cooler's custody seal, if present, is intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the sample IDs on the containers and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
VOA sample vials do not have headspace or bubble is <6mm (1/4") in diameter.	True	
If necessary, staff have been informed of any short hold time or quick TAT needs	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Is the Field Sampler's name present on COC?	True	
Sample Preservation Verified	False	METALS VOLUME FOR PIN15-0593 NEUTRAL

12
52

FedEx Express **US Airbill**

FedEx Tracking Number 8712 8177 1820

Form ID No.

1 From This portion can be removed for Recipient's records.

Date 3/17/10 FedEx Tracking Number 871281771820

Sender's Name CUSTODY Phone 813 883 7427

Company TESTAMERICA TAMPA

Address 6712 BENJAMIN RD STE 100 Dept./Floor/Suite/Room

City TAMPA State FL ZIP 33634-4403

2 Your Internal Billing Reference

3 To

Recipient's Name Sample Recv. Phone 303 736 0100

Company TEST AMERICA

HOLD Weekday Print FedEx location address below. NOT available for FedEx First Overnight.

HOLD Saturday Print FedEx location address below. Available ONLY for FedEx Priority Overnight and FedEx 2Day to select locations.

Address Dept./Floor/Suite/Room

We cannot deliver to P.O. boxes or P.O. ZIP codes.

Address 4956 YARROW ST

Print FedEx location address here if HOLD option is selected.

City ARVADA State CO ZIP 80002

0414646911

8712 8177 1820

RECIPIENT: PEEL HERE

fedex.com 1.800.GoFedEx 1.800.463.3339

4a Express Pack

FedEx Priority Ov Next business morning.* shipments will be deliver unless SATURDAY Delive

FedEx 2Day Second business day.* T shipments will be deliver unless SATURDAY Delive

4b Express Freig

FedEx 1Day Freig Next business day.** Frid be delivered on Monday Delivery is selected.

FedEx 2Day Freig Second business day.** on Monday unless SATU

5 Packaging

FedEx Envelope*

Large Pak, and FedEx Sturdy Pak.

6 Special Handling and Delivery Signature Options

SATURDAY Delivery NOT available for FedEx Standard Overnight, FedEx First Overnight, FedEx Express Saver, or FedEx 3Day Freight.

No Signature Required Package may be left without obtaining a signature for delivery.

Direct Signature Someone at recipient's address may sign for delivery. *Fee applies.*

Indirect Signature If no one is available at recipient's address, someone at a neighboring address may sign for delivery. For residential deliveries only. *Fee applies.*

Does this shipment contain dangerous goods?

One box must be checked.

No

Yes As per attached Shipper's Declaration.

Yes Shipper's Declaration not required.

Dry Ice Dry Ice, 9, UN 1845 ____ x ____ kg

Cargo Aircraft Only

Dangerous goods (including dry ice) cannot be shipped in FedEx packaging or placed in a FedEx Express Drop Box.

7 Payment *Bill to:*

Enter FedEx Acct. No. or Credit Card No. below.

Obtain Recip. Acct. No.

Sender Acct. No. in Section 1 will be billed.

Recipient

Third Party

Credit Card

Cash/Check

Total Packages 1 Total Weight 50 LB Credit Card Auth.

†Our liability is limited to $100 unless you declare a higher value. See the current FedEx Service Guide for details.

553

Rev. Date 2/08•Part #158279•©1994–2008 FedEx•PRINTED IN U.S.A.•SRS

FedEx TRK# 0215 8712 8177 1820

THU - 18
PRIORITY O

XH WHHA

Emp# 631384 17MAR10 TPFA

ANALYTICAL REPORT

Job Number: 280-1555-1

SDG Number: 10022892

Job Description: PINELLAS MONITORING

For:
S.M. Stoller Corporation
2597 B 3/4 Road
Grand Junction, CO 81503

Attention: Mr. Steve Donivan

Approved for release.
Kae E Yoder
Project Manager II
4/16/2010 11:06 AM

Kae E Yoder
Project Manager II
kae.yoder@testamericainc.com
04/16/2010

The test results in this report relate only to the samples in this report and meet all requirements of NELAC, with any exceptions noted. Pursuant to NELAP, this report shall not be reproduced except in full, without the written approval of the laboratory. All questions regarding this report should be directed to the TestAmerica Denver Project Manager.

The Lab Certification ID# is E87667.

Reporting limits are adjusted for sample size used, dilutions and moisture content if applicable.

Table of Contents

Cover Title Page 1

Data Summaries 5

Report Narrative 5

Data Qualifiers 6

Executive Summary 7

Method Summary 15

Method / Analyst Summary 16

Sample Summary 17

Sample Datasheets 19

Surrogate Summary 146

QC Data Summary 150

QC Association Summary 178

Lab Chronicle 181

Manual Integration Summary 195

Organic Sample Data 199

GC/MS VOA 199

Method 8260B 199

Method 8260B QC Summary 200

Method 8260B Sample Data 232

Standards Data 740

Method 8260B ICAL Data 740

Method 8260B CCAL Data 870

Raw QC Data 952

Method 8260B Tune Data 952

Method 8260B Blank Data 980

Method 8260B LCS/LCSD Data 1008

Table of Contents

Method 8260B MS/MSD Data 1032

Method 8260B Run Logs 1064

Method 8260B SIM 1193

Method 8260B SIM QC Summary 1194

Method 8260B SIM Sample Data 1208

Standards Data 1256

Method 8260B SIM ICAL Data 1256

Method 8260B SIM CCAL Data 1272

Raw QC Data 1281

Method 8260B SIM Tune Data 1281

Method 8260B SIM Blank Data 1297

Method 8260B SIM LCS/LCSD Data 1303

Method 8260B SIM MS/MSD Data 1309

Method 8260B SIM Run Logs 1321

Inorganic Sample Data 1325

Metals Data 1325

Met Cover Page 1326

Met Sample Data 1327

Met QC Data 1329

Met ICV/CCV 1329

Met CRQL 1331

Met Blanks 1332

Met ICSA/ICSAB 1334

Met MS/MSD/PDS 1336

Met LCS/LCSD 1338

Met Serial Dilution 1339

Table of Contents

Met MDL 1340
Met IECF 1341
Met Linear Ranges 1347
Met Preparation Log 1348
Met Analysis Run Log 1349
Met Raw Data 1353
Met Prep Data 1734
Shipping and Receiving Documents 1735
Client Chain of Custody 1736
Sample Receipt Checklist 1740

CASE NARRATIVE

Client: S.M. Stoller Corporation

Project: PINELLAS MONITORING - 10022892

Report Number: 280-1555-1

With exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. All laboratory quality control samples analyzed in conjunction with the samples in this project were within established control limits, with any exceptions noted. Calculations are performed before rounding to avoid round-off errors in calculated results.

This report includes reporting limits (RLs) less than TestAmerica Denver's practical quantitation limits. These reporting limits are being used specifically at the client's request to meet the needs of this project. Please note that data are not normally reported to these levels without qualification, since they are inherently less reliable and potentially less defensible than required by the current NELAC standards.

Results between the method detection limit (MDL) and reporting limit (RL) are flagged with a "J" qualifier to indicate an estimated value. These results are statistically less reliable than results greater than or equal to the RL and should be considered a qualitative value.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

RECEIPT

The samples were received on 3/19/2010; the samples arrived in good condition, properly preserved and on ice. The temperatures of the coolers at receipt were 0.1 C and 0.4 C.

GC/MS VOLATILES - SW846 8260B

Due to analytes present above the linear calibration curve, a reduced aliquot size had to be used for the preparation/analysis of samples PIN12-0554C (IDW 144), PIN12-0569-2 (IDW 179), PIN12-0569-3 (IDW 180), PIN12-2872 (IDW 240), PIN12-S67B (IDW 185), PIN12-S67C (IDW 186) and PIN12-S68D (IDW 190). To provide the lowest possible detection limits, multiple runs are reported.

Methylene chloride and Acetone, common laboratory contaminants, were detected in the method blanks associated with batches 280-8984 and 280-9015, respectively, at levels that were above the method detection limit but below the reporting limit. The values should be considered estimates, and have been flagged “J”. If the associated samples reported a result above the MDL and/or RL, the result has been “B” flagged.

Styrene was detected in the method blanks associated with batches 280-9023 and 280-9155 at levels that were above the method detection limit but below the reporting limit. The values should be considered estimates, and have been flagged “J”. If the associated samples reported a result above the MDL and/or RL, the result has been “B” flagged.

Carbon tetrachloride and Tetrachloroethene failed the recovery criteria low for the MS associated with batch 280-9015. The LCS and LCSD were within control limits.

1,1-Dichloroethene failed the recovery criteria high for the MS performed on sample PIN12-0529 (IDW 109) in batch 280-9155. The LCS was within control limits.

The Continuing Calibration Verification (CCV) associated with samples in batch 280-8984 exceeded the control criteria of +/-35%, biased high, for 2,2-Dichloropropane (+38.6%). No detectable concentrations of 2,2-Dichloropropane are present in the associated samples; therefore, data are reported as is.

No other anomalies were encountered.

GC/MS VOLATILES - SW846 8260B-SIM

No anomalies were encountered.

TOTAL METALS - SW846 6010B

The Linear Range date listed on the CLP forms for ICP instrument 26 is incorrect. The correct Linear range date is 1/25/2010.

No other anomalies were encountered.

DATA REPORTING QUALIFIERS

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Section	Qualifier	Description
GC/MS VOA		
	B	Compound was found in the blank and sample.
	U	Indicates the analyte was analyzed for but not detected.
	F	MS or MSD exceeds the control limits
	J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
Metals		
	U	Indicates the analyte was analyzed for but not detected.
	J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1555-1	**PIN12-0513**					
Acetone		4.0	J B	10	ug/L	8260B
Chloroethane		0.70	J	1.0	ug/L	8260B
1,1-Dichloroethane		3.4		1.0	ug/L	8260B
cis-1,2-Dichloroethene		0.17	J	1.0	ug/L	8260B
Methylene Chloride		0.35	J	1.0	ug/L	8260B
Vinyl chloride		0.64	J	1.0	ug/L	8260B
280-1555-2	**PIN12-0526**					
Acetone		5.5	J B	10	ug/L	8260B
cis-1,2-Dichloroethene		0.30	J	1.0	ug/L	8260B
trans-1,2-Dichloroethene		1.2		1.0	ug/L	8260B
Methylene Chloride		0.37	J	1.0	ug/L	8260B
Vinyl chloride		3.2		1.0	ug/L	8260B
280-1555-3	**PIN12-0529**					
Methylene Chloride		0.66	J	1.0	ug/L	8260B
Styrene		0.43	J B	1.0	ug/L	8260B
280-1555-4	**PIN12-0537**					
Acetone		3.4	J B	10	ug/L	8260B
280-1555-5	**PIN12-0538**					
Acetone		4.0	J B	10	ug/L	8260B
Methylene Chloride		0.35	J	1.0	ug/L	8260B
280-1555-6	**PIN12-0539**					
Methylene Chloride		0.58	J	1.0	ug/L	8260B
Styrene		0.42	J B	1.0	ug/L	8260B
280-1555-7	**PIN12-0541**					
Acetone		3.8	J B	10	ug/L	8260B
1,1-Dichloroethane		0.59	J	1.0	ug/L	8260B
cis-1,2-Dichloroethene		0.60	J	1.0	ug/L	8260B
1,1-Dichloroethene		0.24	J	1.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1555-8	**PIN12-0542**					
1,1-Dichloroethane		2.5		1.0	ug/L	8260B
cis-1,2-Dichloroethene		1.8		1.0	ug/L	8260B
1,1-Dichloroethene		0.56	J	1.0	ug/L	8260B
280-1555-9	**PIN12-0547**					
Methylene Chloride		0.81	J	1.0	ug/L	8260B
Styrene		0.42	J B	1.0	ug/L	8260B
280-1555-10	**PIN12-0549**					
Acetone		5.4	J B	10	ug/L	8260B
280-1555-11	**PIN12-0552-1**					
Acetone		8.2	J B	10	ug/L	8260B
cis-1,2-Dichloroethene		14		1.0	ug/L	8260B
trans-1,2-Dichloroethene		0.15	J	1.0	ug/L	8260B
1,1-Dichloroethene		0.52	J	1.0	ug/L	8260B
Toluene		0.47	J	1.0	ug/L	8260B
Vinyl chloride		3.9		1.0	ug/L	8260B
280-1555-12	**PIN12-0552-2**					
Acetone		9.9	J B	10	ug/L	8260B
cis-1,2-Dichloroethene		24		1.0	ug/L	8260B
trans-1,2-Dichloroethene		0.29	J	1.0	ug/L	8260B
1,1-Dichloroethene		1.0		1.0	ug/L	8260B
Toluene		0.40	J	1.0	ug/L	8260B
Vinyl chloride		6.8		1.0	ug/L	8260B
280-1555-13	**PIN12-0552-3**					
Acetone		10	J B	10	ug/L	8260B
1,1-Dichloroethane		0.25	J	1.0	ug/L	8260B
cis-1,2-Dichloroethene		48		1.0	ug/L	8260B
trans-1,2-Dichloroethene		0.70	J	1.0	ug/L	8260B
1,1-Dichloroethene		2.0		1.0	ug/L	8260B
Toluene		0.34	J	1.0	ug/L	8260B
Vinyl chloride		30		1.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1555-14	**PIN12-0554B**					
Acetone		2.3	J B	10	ug/L	8260B
280-1555-15	**PIN12-0554C**					
Acetone		4.0	J B	10	ug/L	8260B
1,1-Dichloroethane		65		2.0	ug/L	8260B
cis-1,2-Dichloroethene		16		1.0	ug/L	8260B
trans-1,2-Dichloroethene		4.6		1.0	ug/L	8260B
1,1-Dichloroethene		1.5		1.0	ug/L	8260B
Vinyl chloride		82		2.0	ug/L	8260B
280-1555-16	**PIN12-0555A**					
Methylene Chloride		0.34	J	1.0	ug/L	8260B
Styrene		0.39	J B	1.0	ug/L	8260B
280-1555-17	**PIN12-0555B**					
Acetone		9.4	J	10	ug/L	8260B
Styrene		0.41	J B	1.0	ug/L	8260B
280-1555-18	**PIN12-0555C**					
cis-1,2-Dichloroethene		1.2		1.0	ug/L	8260B
trans-1,2-Dichloroethene		0.61	J	1.0	ug/L	8260B
Methylene Chloride		0.39	J	1.0	ug/L	8260B
Styrene		0.43	J B	1.0	ug/L	8260B
280-1555-19	**PIN12-0562-1**					
Methylene Chloride		0.36	J	1.0	ug/L	8260B
Styrene		0.40	J B	1.0	ug/L	8260B
Toluene		0.36	J	1.0	ug/L	8260B
280-1555-20	**PIN12-0562-2**					
Chloroform		0.39	J	1.0	ug/L	8260B
Styrene		0.41	J B	1.0	ug/L	8260B
Toluene		0.19	J	1.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1555-21	**PIN12-0562-3**					
Chloroethane		0.62	J	1.0	ug/L	8260B
Styrene		0.49	J B	1.0	ug/L	8260B
Toluene		0.18	J	1.0	ug/L	8260B
280-1555-22	**PIN12-0563-1**					
Styrene		0.43	J B	1.0	ug/L	8260B
Toluene		0.28	J	1.0	ug/L	8260B
280-1555-23	**PIN12-0563-2**					
cis-1,2-Dichloroethene		0.29	J	1.0	ug/L	8260B
trans-1,3-Dichloropropene		0.94	J	1.0	ug/L	8260B
Styrene		0.49	J B	1.0	ug/L	8260B
Toluene		0.26	J	1.0	ug/L	8260B
280-1555-24	**PIN12-0563-3**					
Methylene Chloride		0.33	J	1.0	ug/L	8260B
Styrene		0.42	J B	1.0	ug/L	8260B
Toluene		0.27	J	1.0	ug/L	8260B
Vinyl chloride		0.54	J	1.0	ug/L	8260B
280-1555-25	**PIN12-0567-1**					
Styrene		0.42	J B	1.0	ug/L	8260B
280-1555-26	**PIN12-0567-2**					
Styrene		0.41	J B	1.0	ug/L	8260B
280-1555-27	**PIN12-0567-3**					
Styrene		0.45	J B	1.0	ug/L	8260B
Toluene		0.26	J	1.0	ug/L	8260B
280-1555-28	**PIN12-0568-1**					
Acetone		7.6	J	10	ug/L	8260B
Methylene Chloride		0.33	J B	1.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1555-29	**PIN12-0568-2**					
Acetone		7.1	J	10	ug/L	8260B
Methylene Chloride		0.36	J B	1.0	ug/L	8260B
280-1555-30	**PIN12-0568-3**					
Styrene		0.49	J B	1.0	ug/L	8260B
280-1555-31	**PIN12-0569-2**					
Acetone		13		10	ug/L	8260B
cis-1,2-Dichloroethene		170		10	ug/L	8260B
trans-1,2-Dichloroethene		1.7		1.0	ug/L	8260B
1,1-Dichloroethene		10		1.0	ug/L	8260B
Methylene Chloride		0.33	J B	1.0	ug/L	8260B
Vinyl chloride		65		10	ug/L	8260B
280-1555-32	**PIN12-0569-3**					
Acetone		17	J	50	ug/L	8260B
cis-1,2-Dichloroethene		1200		50	ug/L	8260B
trans-1,2-Dichloroethene		8.4		5.0	ug/L	8260B
1,1-Dichloroethene		56		5.0	ug/L	8260B
Methylene Chloride		1.7	J B	5.0	ug/L	8260B
Vinyl chloride		350		50	ug/L	8260B
280-1555-33	**PIN12-2872**					
cis-1,2-Dichloroethene		180		10	ug/L	8260B
trans-1,2-Dichloroethene		1.8		1.0	ug/L	8260B
1,1-Dichloroethene		11		1.0	ug/L	8260B
Vinyl chloride		69		10	ug/L	8260B
280-1555-34	**PIN99-2876**					
Acetone		3.5	J	10	ug/L	8260B
Methylene Chloride		1.5	B	1.0	ug/L	8260B
280-1555-35	**PIN99-2882**					
Acetone		3.0	J B	10	ug/L	8260B
Methylene Chloride		1.6		1.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1555-36	**PIN99-2890**					
Methylene Chloride		1.4		1.0	ug/L	8260B
Styrene		0.41	J B	1.0	ug/L	8260B
280-1555-37	**PIN99-2901**					
Methylene Chloride		1.2		1.0	ug/L	8260B
Styrene		0.48	J B	1.0	ug/L	8260B
Toluene		0.81	J	1.0	ug/L	8260B
280-1555-38	**PIN99-2902**					
Methylene Chloride		1.3		1.0	ug/L	8260B
Styrene		0.41	J B	1.0	ug/L	8260B
Toluene		0.26	J	1.0	ug/L	8260B
Aluminum		27	J	100	ug/L	6010B
280-1555-39	**PIN12-S67B**					
1,1-Dichloroethane		52		2.0	ug/L	8260B
cis-1,2-Dichloroethene		20		2.0	ug/L	8260B
trans-1,2-Dichloroethene		4.5		2.0	ug/L	8260B
Methylene Chloride		1.3	J	2.0	ug/L	8260B
Styrene		0.79	J B	2.0	ug/L	8260B
Vinyl chloride		300		20	ug/L	8260B
280-1555-40	**PIN12-S67C**					
1,1-Dichloroethane		13		1.0	ug/L	8260B
cis-1,2-Dichloroethene		130		10	ug/L	8260B
trans-1,2-Dichloroethene		26		1.0	ug/L	8260B
1,1-Dichloroethene		1.9		1.0	ug/L	8260B
Methylene Chloride		0.39	J	1.0	ug/L	8260B
Vinyl chloride		140		10	ug/L	8260B
280-1555-41	**PIN12-S67D**					
1,1-Dichloroethane		1.6		1.0	ug/L	8260B
cis-1,2-Dichloroethene		17		1.0	ug/L	8260B
trans-1,2-Dichloroethene		5.6		1.0	ug/L	8260B
Methylene Chloride		0.36	J	1.0	ug/L	8260B
Vinyl chloride		27		1.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1555-42	**PIN12-S68B**					
Styrene		0.43	J B	1.0	ug/L	8260B
280-1555-43	**PIN12-S68C**					
Acetone		4.9	J	10	ug/L	8260B
1,1-Dichloroethane		4.2		1.0	ug/L	8260B
cis-1,2-Dichloroethene		2.8		1.0	ug/L	8260B
Methylene Chloride		0.38	J B	1.0	ug/L	8260B
Vinyl chloride		4.5		1.0	ug/L	8260B
280-1555-44	**PIN12-S68D**					
1,1-Dichloroethane		3.0		1.0	ug/L	8260B
cis-1,2-Dichloroethene		54		2.0	ug/L	8260B
trans-1,2-Dichloroethene		0.94	J	1.0	ug/L	8260B
Vinyl chloride		43		1.0	ug/L	8260B
280-1555-45	**PIN12-S70B**					
cis-1,2-Dichloroethene		16		1.0	ug/L	8260B
trans-1,2-Dichloroethene		0.50	J	1.0	ug/L	8260B
Methylene Chloride		0.34	J B	1.0	ug/L	8260B
Vinyl chloride		11		1.0	ug/L	8260B
280-1555-46	**PIN12-S70C**					
Acetone		6.5	J B	10	ug/L	8260B
1,1-Dichloroethane		3.5		1.0	ug/L	8260B
cis-1,2-Dichloroethene		18		1.0	ug/L	8260B
trans-1,2-Dichloroethene		5.2		1.0	ug/L	8260B
1,1-Dichloroethene		0.39	J	1.0	ug/L	8260B
Methylene Chloride		0.40	J	1.0	ug/L	8260B
Vinyl chloride		11		1.0	ug/L	8260B
280-1555-47	**PIN12-S72B**					
Styrene		0.43	J B	1.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1555-48	**PIN12-S72C**					
1,1-Dichloroethane		2.8		1.0	ug/L	8260B
cis-1,2-Dichloroethene		2.2		1.0	ug/L	8260B
1,1-Dichloroethene		0.91	J	1.0	ug/L	8260B
Vinyl chloride		1.3		1.0	ug/L	8260B
280-1555-49	**PIN12-S72D**					
Styrene		0.42	J B	1.0	ug/L	8260B
280-1555-50	**PIN12-S73C**					
cis-1,2-Dichloroethene		0.21	J	1.0	ug/L	8260B
trans-1,2-Dichloroethene		1.2		1.0	ug/L	8260B
Vinyl chloride		5.4		1.0	ug/L	8260B
280-1555-51	**PIN12-S73D**					
Styrene		0.45	J B	1.0	ug/L	8260B
Toluene		0.23	J	1.0	ug/L	8260B

METHOD SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Description	Lab Location	Method	Preparation Method
Matrix: Water			
Volatile Organic Compounds (GC/MS)	TAL DEN	SW846 8260B	
Purge and Trap	TAL DEN		SW846 5030B
Volatile Organic Compounds (GC/MS-SIM)	TAL DEN	SW846 8260B SIM	
Purge and Trap	TAL DEN		SW846 5030B
Metals (ICP)	TAL DEN	SW846 6010B	
Preparation, Total Metals	TAL DEN		SW846 3010A

Lab References:

TAL DEN = TestAmerica Denver

Method References:

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

METHOD / ANALYST SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Method	Analyst	Analyst ID
SW846 8260B	Ilczyszyn, Dennis P	DPI
SW846 8260B	Tinkham, Sarah A	SAT
SW846 8260B	Waterland, Hayley E	HEW
SW846 8260B SIM	Wolfe, Ashley E	AEW
SW846 6010B	Wells, David	DW

SAMPLE SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
280-1555-1	PIN12-0513	Water	03/16/2010 0925	03/19/2010 0900
280-1555-2	PIN12-0526	Water	03/16/2010 1025	03/19/2010 0900
280-1555-3	PIN12-0529	Water	03/17/2010 1057	03/19/2010 0900
280-1555-4	PIN12-0537	Water	03/16/2010 1540	03/19/2010 0900
280-1555-5	PIN12-0538	Water	03/16/2010 1607	03/19/2010 0900
280-1555-6	PIN12-0539	Water	03/17/2010 1102	03/19/2010 0900
280-1555-7	PIN12-0541	Water	03/16/2010 1345	03/19/2010 0900
280-1555-8	PIN12-0542	Water	03/16/2010 1413	03/19/2010 0900
280-1555-9	PIN12-0547	Water	03/17/2010 1012	03/19/2010 0900
280-1555-10	PIN12-0549	Water	03/16/2010 1500	03/19/2010 0900
280-1555-11	PIN12-0552-1	Water	03/16/2010 1340	03/19/2010 0900
280-1555-12	PIN12-0552-2	Water	03/16/2010 1438	03/19/2010 0900
280-1555-13	PIN12-0552-3	Water	03/16/2010 1509	03/19/2010 0900
280-1555-14	PIN12-0554B	Water	03/16/2010 1045	03/19/2010 0900
280-1555-15	PIN12-0554C	Water	03/16/2010 1118	03/19/2010 0900
280-1555-16	PIN12-0555A	Water	03/16/2010 0955	03/19/2010 0900
280-1555-17	PIN12-0555B	Water	03/16/2010 0925	03/19/2010 0900
280-1555-18	PIN12-0555C	Water	03/16/2010 0845	03/19/2010 0900
280-1555-19	PIN12-0562-1	Water	03/16/2010 1120	03/19/2010 0900
280-1555-20	PIN12-0562-2	Water	03/16/2010 1556	03/19/2010 0900
280-1555-21	PIN12-0562-3	Water	03/16/2010 1632	03/19/2010 0900
280-1555-22	PIN12-0563-1	Water	03/17/2010 0903	03/19/2010 0900
280-1555-23	PIN12-0563-2	Water	03/17/2010 0931	03/19/2010 0900
280-1555-24	PIN12-0563-3	Water	03/17/2010 0956	03/19/2010 0900
280-1555-25	PIN12-0567-1	Water	03/16/2010 0921	03/19/2010 0900
280-1555-26	PIN12-0567-2	Water	03/16/2010 0947	03/19/2010 0900
280-1555-27	PIN12-0567-3	Water	03/16/2010 1039	03/19/2010 0900
280-1555-28	PIN12-0568-1	Water	03/15/2010 1509	03/19/2010 0900
280-1555-29	PIN12-0568-2	Water	03/15/2010 1544	03/19/2010 0900
280-1555-30	PIN12-0568-3	Water	03/16/2010 0847	03/19/2010 0900
280-1555-31	PIN12-0569-2	Water	03/15/2010 1342	03/19/2010 0900
280-1555-32	PIN12-0569-3	Water	03/15/2010 1420	03/19/2010 0900
280-1555-33	PIN12-2872	Water	03/15/2010 1200	03/19/2010 0900
280-1555-34	PIN99-2876	Water	03/15/2010 0800	03/19/2010 0900
280-1555-35	PIN99-2882	Water	03/15/2010 1200	03/19/2010 0900
280-1555-36	PIN99-2890	Water	03/16/2010 0812	03/19/2010 0900
280-1555-37	PIN99-2901	Water	03/16/2010 0730	03/19/2010 0900
280-1555-37MS	PIN99-2901	Water	03/16/2010 0730	03/19/2010 0900
280-1555-37MSD	PIN99-2901	Water	03/16/2010 0730	03/19/2010 0900
280-1555-38	PIN99-2902	Water	03/17/2010 0730	03/19/2010 0900
280-1555-39	PIN12-S67B	Water	03/17/2010 1055	03/19/2010 0900
280-1555-40	PIN12-S67C	Water	03/17/2010 1010	03/19/2010 0900
280-1555-41	PIN12-S67D	Water	03/17/2010 0905	03/19/2010 0900
280-1555-42	PIN12-S68B	Water	03/16/2010 0840	03/19/2010 0900
280-1555-43	PIN12-S68C	Water	03/15/2010 1630	03/19/2010 0900
280-1555-44	PIN12-S68D	Water	03/15/2010 1550	03/19/2010 0900

SAMPLE SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
280-1555-45	PIN12-S70B	Water	03/15/2010 1445	03/19/2010 0900
280-1555-46	PIN12-S70C	Water	03/15/2010 1345	03/19/2010 0900
280-1555-47	PIN12-S72B	Water	03/16/2010 1430	03/19/2010 0900
280-1555-48	PIN12-S72C	Water	03/16/2010 1345	03/19/2010 0900
280-1555-49	PIN12-S72D	Water	03/16/2010 1105	03/19/2010 0900
280-1555-50	PIN12-S73C	Water	03/16/2010 1610	03/19/2010 0900
280-1555-51	PIN12-S73D	Water	03/16/2010 1520	03/19/2010 0900

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0513**

Lab Sample ID: 280-1555-1
Client Matrix: Water

Date Sampled: 03/16/2010 0925
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7692.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1439		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1439			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	4.0	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.70	J	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	3.4		0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.17	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.35	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0513**

Lab Sample ID: 280-1555-1
Client Matrix: Water

Date Sampled: 03/16/2010 0925
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7692.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1439		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1439			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.64	J	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	108		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	95		78 - 118
Dibromofluoromethane (Surr)	97		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0526**

Lab Sample ID: 280-1555-2
Client Matrix: Water

Date Sampled: 03/16/2010 1025
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7693.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1500		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1500			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	5.5	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.30	J	0.15	1.0
trans-1,2-Dichloroethene	1.2		0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.37	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0526**

Lab Sample ID: 280-1555-2
Client Matrix: Water

Date Sampled: 03/16/2010 1025
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7693.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1500		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1500			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	3.2		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	111		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	96		78 - 118
Dibromofluoromethane (Surr)	99		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0529**

Lab Sample ID: 280-1555-3
Client Matrix: Water

Date Sampled: 03/17/2010 1057
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5053.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1112		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1112			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.66	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0529**

Lab Sample ID: 280-1555-3
Client Matrix: Water

Date Sampled: 03/17/2010 1057
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5053.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1112		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1112			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.43	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	91		70 - 127
Toluene-d8 (Surr)	99		80 - 125
4-Bromofluorobenzene (Surr)	93		78 - 118
Dibromofluoromethane (Surr)	93		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0537**

Lab Sample ID: 280-1555-4
Client Matrix: Water

Date Sampled: 03/16/2010 1540
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7706.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1927		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1927			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	3.4	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0537**

Lab Sample ID: 280-1555-4
Client Matrix: Water

Date Sampled: 03/16/2010 1540
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7706.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1927		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1927			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	119		70 - 127
Toluene-d8 (Surr)	102		80 - 125
4-Bromofluorobenzene (Surr)	98		78 - 118
Dibromofluoromethane (Surr)	101		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0538**

Lab Sample ID: 280-1555-5
Client Matrix: Water

Date Sampled: 03/16/2010 1607
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7696.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1601		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1601			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	4.0	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.35	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0538**

Lab Sample ID: 280-1555-5
Client Matrix: Water

Date Sampled: 03/16/2010 1607
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7696.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1601		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1601			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	110		70 - 127
Toluene-d8 (Surr)	99		80 - 125
4-Bromofluorobenzene (Surr)	93		78 - 118
Dibromofluoromethane (Surr)	98		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0539**

Lab Sample ID: 280-1555-6
Client Matrix: Water

Date Sampled: 03/17/2010 1102
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5057.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1234		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1234			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.58	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0539**

Lab Sample ID: 280-1555-6
Client Matrix: Water

Date Sampled: 03/17/2010 1102
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5057.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1234		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1234			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.42	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	100		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	98		78 - 118
Dibromofluoromethane (Surr)	98		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0541**

Lab Sample ID: 280-1555-7
Client Matrix: Water

Date Sampled: 03/16/2010 1345
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7697.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1622		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1622			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	3.8	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.59	J	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.60	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.24	J	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0541**

Lab Sample ID: 280-1555-7
Client Matrix: Water

Date Sampled: 03/16/2010 1345
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7697.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1622		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1622			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	109		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	97		78 - 118
Dibromofluoromethane (Surr)	100		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0542**

Lab Sample ID: 280-1555-8
Client Matrix: Water

Date Sampled: 03/16/2010 1413
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7698.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1642		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1642			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	2.5		0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	1.8		0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.56	J	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0542**

Lab Sample ID: 280-1555-8
Client Matrix: Water

Date Sampled: 03/16/2010 1413
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7698.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1642		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1642			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	115		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	99		78 - 118
Dibromofluoromethane (Surr)	100		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0547**

Lab Sample ID: 280-1555-9
Client Matrix: Water

Date Sampled: 03/17/2010 1012
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5058.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1255		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1255			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.81	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0547**

Lab Sample ID: 280-1555-9
Client Matrix: Water

Date Sampled: 03/17/2010 1012
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5058.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1255		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1255			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.42	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	98		70 - 127
Toluene-d8 (Surr)	97		80 - 125
4-Bromofluorobenzene (Surr)	98		78 - 118
Dibromofluoromethane (Surr)	100		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0549**

Lab Sample ID: 280-1555-10
Client Matrix: Water

Date Sampled: 03/16/2010 1500
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7699.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1703		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1703			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	5.4	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0549**

Lab Sample ID: 280-1555-10
Client Matrix: Water

Date Sampled: 03/16/2010 1500
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7699.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1703		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1703			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	115		70 - 127
Toluene-d8 (Surr)	99		80 - 125
4-Bromofluorobenzene (Surr)	94		78 - 118
Dibromofluoromethane (Surr)	100		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0552-1**

Lab Sample ID: 280-1555-11
Client Matrix: Water

Date Sampled: 03/16/2010 1340
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7700.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1724		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1724			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	8.2	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	14		0.15	1.0
trans-1,2-Dichloroethene	0.15	J	0.15	1.0
1,1-Dichloroethene	0.52	J	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0552-1**

Lab Sample ID: 280-1555-11
Client Matrix: Water

Date Sampled: 03/16/2010 1340
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7700.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1724		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1724			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.47	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	3.9		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	110		70 - 127
Toluene-d8 (Surr)	102		80 - 125
4-Bromofluorobenzene (Surr)	96		78 - 118
Dibromofluoromethane (Surr)	98		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0552-2**

Lab Sample ID: 280-1555-12
Client Matrix: Water

Date Sampled: 03/16/2010 1438
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7701.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1744		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1744			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	9.9	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	24		0.15	1.0
trans-1,2-Dichloroethene	0.29	J	0.15	1.0
1,1-Dichloroethene	1.0		0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0552-2**

Lab Sample ID: 280-1555-12
Client Matrix: Water

Date Sampled: 03/16/2010 1438
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7701.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1744		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1744			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.40	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	6.8		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	111		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	98		78 - 118
Dibromofluoromethane (Surr)	99		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0552-3**

Lab Sample ID: 280-1555-13
Client Matrix: Water

Date Sampled: 03/16/2010 1509
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7702.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1805		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1805			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	10	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.25	J	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	48		0.15	1.0
trans-1,2-Dichloroethene	0.70	J	0.15	1.0
1,1-Dichloroethene	2.0		0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0552-3**

Lab Sample ID: 280-1555-13
Client Matrix: Water

Date Sampled: 03/16/2010 1509
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7702.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1805		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1805			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.34	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	30		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	113		70 - 127
Toluene-d8 (Surr)	103		80 - 125
4-Bromofluorobenzene (Surr)	97		78 - 118
Dibromofluoromethane (Surr)	99		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0554B**

Lab Sample ID: 280-1555-14
Client Matrix: Water

Date Sampled: 03/16/2010 1045
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7703.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1826		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1826			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	2.3	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0554B**

Lab Sample ID: 280-1555-14
Client Matrix: Water

Date Sampled: 03/16/2010 1045
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7703.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1826		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1826			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	112		70 - 127
Toluene-d8 (Surr)	102		80 - 125
4-Bromofluorobenzene (Surr)	96		78 - 118
Dibromofluoromethane (Surr)	99		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0554C**

Lab Sample ID: 280-1555-15
Client Matrix: Water

Date Sampled: 03/16/2010 1118
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7704.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1846		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1846			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	4.0	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	16		0.15	1.0
trans-1,2-Dichloroethene	4.6		0.15	1.0
1,1-Dichloroethene	1.5		0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0554C**

Lab Sample ID: 280-1555-15
Client Matrix: Water

Date Sampled: 03/16/2010 1118
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7704.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1846		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1846			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	112		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	95		78 - 118
Dibromofluoromethane (Surr)	100		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0554C**

Lab Sample ID: 280-1555-15
Client Matrix: Water

Date Sampled: 03/16/2010 1118
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155		Instrument ID:	MSV_R2
Preparation:	5030B			Lab File ID:	RR5052.D
Dilution:	1.0			Initial Weight/Volume:	10 mL
Date Analyzed:	03/30/2010 1052	Run Type:	DL	Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1052				

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,1-Dichloroethane	65		0.44	2.0
Vinyl chloride	82		0.80	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	97		70 - 127
Toluene-d8 (Surr)	100		80 - 125
4-Bromofluorobenzene (Surr)	93		78 - 118
Dibromofluoromethane (Surr)	93		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0555A**

Lab Sample ID: 280-1555-16
Client Matrix: Water

Date Sampled: 03/16/2010 0955
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5013.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1259		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1259			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.34	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0555A**

Lab Sample ID: 280-1555-16
Client Matrix: Water

Date Sampled: 03/16/2010 0955
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5013.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1259		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1259			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.39	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	89		70 - 127
Toluene-d8 (Surr)	98		80 - 125
4-Bromofluorobenzene (Surr)	108		78 - 118
Dibromofluoromethane (Surr)	91		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0555B**

Lab Sample ID: 280-1555-17
Client Matrix: Water

Date Sampled: 03/16/2010 0925
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5018.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1441		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1441			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	9.4	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0555B**

Lab Sample ID: 280-1555-17
Client Matrix: Water

Date Sampled: 03/16/2010 0925
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5018.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1441		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1441			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.41	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	89		70 - 127
Toluene-d8 (Surr)	103		80 - 125
4-Bromofluorobenzene (Surr)	114		78 - 118
Dibromofluoromethane (Surr)	89		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0555C**

Lab Sample ID: 280-1555-18
Client Matrix: Water

Date Sampled: 03/16/2010 0845
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5036.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 2047		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 2047			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	1.2		0.15	1.0
trans-1,2-Dichloroethene	0.61	J	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.39	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0555C**

Lab Sample ID: 280-1555-18
Client Matrix: Water

Date Sampled: 03/16/2010 0845
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5036.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 2047		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 2047			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.43	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	107		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	96		78 - 118
Dibromofluoromethane (Surr)	107		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0562-1**

Lab Sample ID: 280-1555-19
Client Matrix: Water

Date Sampled: 03/16/2010 1120
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5021.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1541		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1541			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.36	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0562-1**

Lab Sample ID: 280-1555-19
Client Matrix: Water

Date Sampled: 03/16/2010 1120
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5021.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1541		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1541			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.40	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.36	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	108		70 - 127
Toluene-d8 (Surr)	103		80 - 125
4-Bromofluorobenzene (Surr)	100		78 - 118
Dibromofluoromethane (Surr)	99		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0562-2**

Lab Sample ID: 280-1555-20
Client Matrix: Water

Date Sampled: 03/16/2010 1556
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5022.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1602		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1602			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.39	J	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0562-2**

Lab Sample ID: 280-1555-20
Client Matrix: Water

Date Sampled: 03/16/2010 1556
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5022.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1602		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1602			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.41	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.19	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	107		70 - 127
Toluene-d8 (Surr)	106		80 - 125
4-Bromofluorobenzene (Surr)	100		78 - 118
Dibromofluoromethane (Surr)	100		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0562-3**

Lab Sample ID: 280-1555-21
Client Matrix: Water

Date Sampled: 03/16/2010 1632
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5023.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1622		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1622			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.62	J	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0562-3**

Lab Sample ID: 280-1555-21
Client Matrix: Water

Date Sampled: 03/16/2010 1632
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5023.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1622		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1622			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.49	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.18	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	102		70 - 127
Toluene-d8 (Surr)	98		80 - 125
4-Bromofluorobenzene (Surr)	104		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0563-1**

Lab Sample ID: 280-1555-22
Client Matrix: Water

Date Sampled: 03/17/2010 0903
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5059.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1318		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1318			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0563-1**

Lab Sample ID: 280-1555-22
Client Matrix: Water

Date Sampled: 03/17/2010 0903
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5059.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1318		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1318			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.43	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.28	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	96		78 - 118
Dibromofluoromethane (Surr)	96		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0563-2**

Lab Sample ID: 280-1555-23
Client Matrix: Water

Date Sampled: 03/17/2010 0931
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5060.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1338		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1338			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.29	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.94	J	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0563-2**

Lab Sample ID: 280-1555-23
Client Matrix: Water

Date Sampled: 03/17/2010 0931
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5060.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1338		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1338			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.49	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.26	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	102		70 - 127
Toluene-d8 (Surr)	98		80 - 125
4-Bromofluorobenzene (Surr)	100		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0563-3**

Lab Sample ID: 280-1555-24
Client Matrix: Water

Date Sampled: 03/17/2010 0956
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5061.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1359		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1359			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.33	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0563-3**

Lab Sample ID: 280-1555-24
Client Matrix: Water

Date Sampled: 03/17/2010 0956
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5061.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1359		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1359			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.42	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.27	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.54	J	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	109		70 - 127
Toluene-d8 (Surr)	99		80 - 125
4-Bromofluorobenzene (Surr)	94		78 - 118
Dibromofluoromethane (Surr)	107		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0567-1**

Lab Sample ID: 280-1555-25
Client Matrix: Water

Date Sampled: 03/16/2010 0921
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5024.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1642		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1642			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0567-1**

Lab Sample ID: 280-1555-25
Client Matrix: Water

Date Sampled: 03/16/2010 0921
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5024.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1642		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1642			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.42	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	110		70 - 127
Toluene-d8 (Surr)	102		80 - 125
4-Bromofluorobenzene (Surr)	98		78 - 118
Dibromofluoromethane (Surr)	106		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0567-2**

Lab Sample ID: 280-1555-26
Client Matrix: Water

Date Sampled: 03/16/2010 0947
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5025.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1703		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1703			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0567-2**

Lab Sample ID: 280-1555-26
Client Matrix: Water

Date Sampled: 03/16/2010 0947
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5025.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1703		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1703			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.41	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	113		70 - 127
Toluene-d8 (Surr)	104		80 - 125
4-Bromofluorobenzene (Surr)	100		78 - 118
Dibromofluoromethane (Surr)	108		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0567-3**

Lab Sample ID: 280-1555-27
Client Matrix: Water

Date Sampled: 03/16/2010 1039
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5026.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1723		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1723			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0567-3**

Lab Sample ID: 280-1555-27
Client Matrix: Water

Date Sampled: 03/16/2010 1039
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5026.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1723		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1723			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.45	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.26	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	117		70 - 127
Toluene-d8 (Surr)	100		80 - 125
4-Bromofluorobenzene (Surr)	102		78 - 118
Dibromofluoromethane (Surr)	118		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0568-1**

Lab Sample ID: 280-1555-28
Client Matrix: Water

Date Sampled: 03/15/2010 1509
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7644.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0256		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0256			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	7.6	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.33	J B	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0568-1**

Lab Sample ID: 280-1555-28
Client Matrix: Water

Date Sampled: 03/15/2010 1509
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7644.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0256		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0256			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104		70 - 127
Toluene-d8 (Surr)	103		80 - 125
4-Bromofluorobenzene (Surr)	101		78 - 118
Dibromofluoromethane (Surr)	98		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0568-2**

Lab Sample ID: 280-1555-29
Client Matrix: Water

Date Sampled: 03/15/2010 1544
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7645.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0316		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0316			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	7.1	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.36	J B	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0568-2**

Lab Sample ID: 280-1555-29
Client Matrix: Water

Date Sampled: 03/15/2010 1544
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7645.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0316		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0316			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	112		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	100		78 - 118
Dibromofluoromethane (Surr)	100		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0568-3**

Lab Sample ID: 280-1555-30
Client Matrix: Water

Date Sampled: 03/16/2010 0847
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5027.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1744		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1744			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0568-3**

Lab Sample ID: 280-1555-30
Client Matrix: Water

Date Sampled: 03/16/2010 0847
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5027.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1744		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1744			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.49	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	102		70 - 127
Toluene-d8 (Surr)	100		80 - 125
4-Bromofluorobenzene (Surr)	105		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0569-2**

Lab Sample ID: 280-1555-31
Client Matrix: Water

Date Sampled: 03/15/2010 1342
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7649.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0438		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0438			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	13		1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
trans-1,2-Dichloroethene	1.7		0.15	1.0
1,1-Dichloroethene	10		0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.33	J B	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0569-2**

Lab Sample ID: 280-1555-31
Client Matrix: Water

Date Sampled: 03/15/2010 1342
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7649.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0438		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0438			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	113		70 - 127
Toluene-d8 (Surr)	103		80 - 125
4-Bromofluorobenzene (Surr)	101		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0569-2**

Lab Sample ID: 280-1555-31
Client Matrix: Water

Date Sampled: 03/15/2010 1342
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984		Instrument ID:	MSV_R1
Preparation:	5030B			Lab File ID:	R7650.D
Dilution:	1.0			Initial Weight/Volume:	2 mL
Date Analyzed:	03/27/2010 0459	Run Type:	DL	Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0459				

Analyte	Result (ug/L)	Qualifier	MDL	RL
cis-1,2-Dichloroethene	170		1.5	10
Vinyl chloride	65		4.0	10

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	108		70 - 127
Toluene-d8 (Surr)	99		80 - 125
4-Bromofluorobenzene (Surr)	99		78 - 118
Dibromofluoromethane (Surr)	102		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0569-3**

Lab Sample ID: 280-1555-32
Client Matrix: Water

Date Sampled: 03/15/2010 1420
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7651.D
Dilution:	1.0		Initial Weight/Volume:	4 mL
Date Analyzed:	03/27/2010 0520		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0520			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	17	J	9.5	50
Benzene	0.80	U	0.80	5.0
Bromobenzene	0.85	U	0.85	5.0
Bromochloromethane	0.50	U	0.50	5.0
Bromodichloromethane	0.85	U	0.85	5.0
Bromoform	0.95	U	0.95	5.0
Bromomethane	1.0	U	1.0	5.0
2-Butanone (MEK)	10	U	10	25
n-Butylbenzene	0.70	U	0.70	5.0
sec-Butylbenzene	0.85	U	0.85	5.0
tert-Butylbenzene	0.80	U	0.80	5.0
Carbon disulfide	2.2	U	2.2	5.0
Carbon tetrachloride	0.95	U	0.95	5.0
Chlorobenzene	0.85	U	0.85	5.0
Dibromochloromethane	0.85	U	0.85	5.0
Chloroethane	2.0	U	2.0	5.0
Chloroform	0.80	U	0.80	5.0
Chloromethane	1.5	U	1.5	5.0
2-Chlorotoluene	0.85	U	0.85	5.0
4-Chlorotoluene	1.0	U	1.0	5.0
1,2-Dibromo-3-Chloropropane	2.4	U	2.4	5.0
Dibromomethane	0.85	U	0.85	5.0
1,2-Dichlorobenzene	0.75	U	0.75	5.0
1,3-Dichlorobenzene	0.65	U	0.65	5.0
1,4-Dichlorobenzene	0.80	U	0.80	5.0
Dichlorodifluoromethane	1.6	U	1.6	5.0
1,1-Dichloroethane	1.1	U	1.1	5.0
1,2-Dichloroethane	0.65	U	0.65	5.0
trans-1,2-Dichloroethene	8.4		0.75	5.0
1,1-Dichloroethene	56		1.2	5.0
1,2-Dichloropropane	0.90	U	0.90	5.0
1,3-Dichloropropane	1.1	U	1.1	5.0
2,2-Dichloropropane	0.90	U	0.90	5.0
cis-1,3-Dichloropropene	0.80	U	0.80	5.0
trans-1,3-Dichloropropene	0.95	U	0.95	5.0
1,1-Dichloropropene	0.95	U	0.95	5.0
Ethylbenzene	0.80	U	0.80	5.0
Hexachlorobutadiene	0.60	U	0.60	5.0
2-Hexanone	8.5	U	8.5	25
Isopropylbenzene	0.95	U	0.95	5.0
4-Isopropyltoluene	1.0	U	1.0	5.0
Methylene Chloride	1.7	J B	1.6	5.0
4-Methyl-2-pentanone	4.9	U	4.9	25
Naphthalene	1.1	U	1.1	5.0
n-Propylbenzene	0.80	U	0.80	5.0
Styrene	0.85	U	0.85	5.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0569-3**

Lab Sample ID: 280-1555-32
Client Matrix: Water

Date Sampled: 03/15/2010 1420
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7651.D
Dilution:	1.0		Initial Weight/Volume:	4 mL
Date Analyzed:	03/27/2010 0520		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0520			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,1,1,2-Tetrachloroethane	1.0	U	1.0	5.0
1,1,2,2-Tetrachloroethane	1.0	U	1.0	5.0
Tetrachloroethene	1.0	U	1.0	5.0
Toluene	0.85	U	0.85	5.0
1,2,3-Trichlorobenzene	1.0	U	1.0	5.0
1,2,4-Trichlorobenzene	1.0	U	1.0	5.0
1,1,1-Trichloroethane	0.80	U	0.80	5.0
1,1,2-Trichloroethane	1.4	U	1.4	5.0
Trichloroethene	0.80	U	0.80	5.0
Trichlorofluoromethane	1.4	U	1.4	5.0
1,2,3-Trichloropropane	1.6	U	1.6	5.0
1,2,4-Trimethylbenzene	0.75	U	0.75	5.0
1,3,5-Trimethylbenzene	0.80	U	0.80	5.0
Xylenes, Total	0.95	U	0.95	5.0
1,2-Dibromoethane	0.90	U	0.90	5.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	112		70 - 127
Toluene-d8 (Surr)	99		80 - 125
4-Bromofluorobenzene (Surr)	101		78 - 118
Dibromofluoromethane (Surr)	102		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0569-3**

Lab Sample ID: 280-1555-32
Client Matrix: Water

Date Sampled: 03/15/2010 1420
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984		Instrument ID:	MSV_R1
Preparation:	5030B			Lab File ID:	R7652.D
Dilution:	1.0			Initial Weight/Volume:	0.4 mL
Date Analyzed:	03/27/2010 0541	Run Type:	DL	Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0541				

Analyte	Result (ug/L)	Qualifier	MDL	RL
cis-1,2-Dichloroethene	1200		7.5	50
Vinyl chloride	350		20	50

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	107		70 - 127
Toluene-d8 (Surr)	100		80 - 125
4-Bromofluorobenzene (Surr)	97		78 - 118
Dibromofluoromethane (Surr)	100		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-2872**

Lab Sample ID: 280-1555-33
Client Matrix: Water

Date Sampled: 03/15/2010 1200
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7653.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0601		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0601			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
trans-1,2-Dichloroethene	1.8		0.15	1.0
1,1-Dichloroethene	11		0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-2872**

Lab Sample ID: 280-1555-33
Client Matrix: Water

Date Sampled: 03/15/2010 1200
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7653.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0601		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0601			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	111		70 - 127
Toluene-d8 (Surr)	99		80 - 125
4-Bromofluorobenzene (Surr)	97		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-2872**

Lab Sample ID: 280-1555-33
Client Matrix: Water

Date Sampled: 03/15/2010 1200
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984		Instrument ID:	MSV_R1
Preparation:	5030B			Lab File ID:	R7654.D
Dilution:	1.0			Initial Weight/Volume:	2 mL
Date Analyzed:	03/27/2010 0622	Run Type:	DL	Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0622				

Analyte	Result (ug/L)	Qualifier	MDL	RL
cis-1,2-Dichloroethene	180		1.5	10
Vinyl chloride	69		4.0	10

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	108		70 - 127
Toluene-d8 (Surr)	98		80 - 125
4-Bromofluorobenzene (Surr)	99		78 - 118
Dibromofluoromethane (Surr)	101		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2876**

Lab Sample ID: 280-1555-34
Client Matrix: Water

Date Sampled: 03/15/2010 0800
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7646.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0336		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0336			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	3.5	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	1.5	B	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2876**

Lab Sample ID: 280-1555-34
Client Matrix: Water

Date Sampled: 03/15/2010 0800
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7646.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0336		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0336			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	107		70 - 127
Toluene-d8 (Surr)	102		80 - 125
4-Bromofluorobenzene (Surr)	103		78 - 118
Dibromofluoromethane (Surr)	99		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2882**

Lab Sample ID: 280-1555-35
Client Matrix: Water

Date Sampled: 03/15/2010 1200
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7690.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1358		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1358			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	3.0	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	1.6		0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2882**

Lab Sample ID: 280-1555-35
Client Matrix: Water

Date Sampled: 03/15/2010 1200
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method: 8260B
Preparation: 5030B
Dilution: 1.0
Date Analyzed: 03/29/2010 1358
Date Prepared: 03/29/2010 1358

Analysis Batch: 280-9015

Instrument ID: MSV_R1
Lab File ID: R7690.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	95		78 - 118
Dibromofluoromethane (Surr)	95		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2890**

Lab Sample ID: 280-1555-36
Client Matrix: Water

Date Sampled: 03/16/2010 0812
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5028.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1804		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1804			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	1.4		0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2890**

Lab Sample ID: 280-1555-36
Client Matrix: Water

Date Sampled: 03/16/2010 0812
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5028.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1804		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1804			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.41	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104		70 - 127
Toluene-d8 (Surr)	108		80 - 125
4-Bromofluorobenzene (Surr)	103		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2901**

Lab Sample ID: 280-1555-37
Client Matrix: Water

Date Sampled: 03/16/2010 0730
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5029.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1825		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1825			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	1.2		0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2901**

Lab Sample ID: 280-1555-37
Client Matrix: Water

Date Sampled: 03/16/2010 0730
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5029.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1825		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1825			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.48	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.81	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103		70 - 127
Toluene-d8 (Surr)	107		80 - 125
4-Bromofluorobenzene (Surr)	106		78 - 118
Dibromofluoromethane (Surr)	102		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2902**

Lab Sample ID: 280-1555-38
Client Matrix: Water

Date Sampled: 03/17/2010 0730
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5062.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1420		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1420			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	1.3		0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2902**

Lab Sample ID: 280-1555-38
Client Matrix: Water

Date Sampled: 03/17/2010 0730
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5062.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1420		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1420			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.41	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.26	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	107		70 - 127
Toluene-d8 (Surr)	114		80 - 125
4-Bromofluorobenzene (Surr)	105		78 - 118
Dibromofluoromethane (Surr)	108		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S67B**

Lab Sample ID: 280-1555-39
Client Matrix: Water

Date Sampled: 03/17/2010 1055
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5063.D
Dilution:	1.0		Initial Weight/Volume:	10 mL
Date Analyzed:	03/30/2010 1440		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1440			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	3.8	U	3.8	20
Benzene	0.32	U	0.32	2.0
Bromobenzene	0.34	U	0.34	2.0
Bromochloromethane	0.20	U	0.20	2.0
Bromodichloromethane	0.34	U	0.34	2.0
Bromoform	0.38	U	0.38	2.0
Bromomethane	0.42	U	0.42	2.0
2-Butanone (MEK)	4.0	U	4.0	10
n-Butylbenzene	0.28	U	0.28	2.0
sec-Butylbenzene	0.34	U	0.34	2.0
tert-Butylbenzene	0.32	U	0.32	2.0
Carbon disulfide	0.90	U	0.90	2.0
Carbon tetrachloride	0.38	U	0.38	2.0
Chlorobenzene	0.34	U	0.34	2.0
Dibromochloromethane	0.34	U	0.34	2.0
Chloroethane	0.82	U	0.82	2.0
Chloroform	0.32	U	0.32	2.0
Chloromethane	0.60	U	0.60	2.0
2-Chlorotoluene	0.34	U	0.34	2.0
4-Chlorotoluene	0.42	U	0.42	2.0
1,2-Dibromo-3-Chloropropane	0.94	U	0.94	2.0
Dibromomethane	0.34	U	0.34	2.0
1,2-Dichlorobenzene	0.30	U	0.30	2.0
1,3-Dichlorobenzene	0.26	U	0.26	2.0
1,4-Dichlorobenzene	0.32	U	0.32	2.0
Dichlorodifluoromethane	0.62	U	0.62	2.0
1,1-Dichloroethane	52		0.44	2.0
1,2-Dichloroethane	0.26	U	0.26	2.0
cis-1,2-Dichloroethene	20		0.30	2.0
trans-1,2-Dichloroethene	4.5		0.30	2.0
1,1-Dichloroethene	0.46	U	0.46	2.0
1,2-Dichloropropane	0.36	U	0.36	2.0
1,3-Dichloropropane	0.44	U	0.44	2.0
2,2-Dichloropropane	0.36	U	0.36	2.0
cis-1,3-Dichloropropene	0.32	U	0.32	2.0
trans-1,3-Dichloropropene	0.38	U	0.38	2.0
1,1-Dichloropropene	0.38	U	0.38	2.0
Ethylbenzene	0.32	U	0.32	2.0
Hexachlorobutadiene	0.24	U	0.24	2.0
2-Hexanone	3.4	U	3.4	10
Isopropylbenzene	0.38	U	0.38	2.0
4-Isopropyltoluene	0.40	U	0.40	2.0
Methylene Chloride	1.3	J	0.64	2.0
4-Methyl-2-pentanone	2.0	U	2.0	10
Naphthalene	0.44	U	0.44	2.0
n-Propylbenzene	0.32	U	0.32	2.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S67B**

Lab Sample ID: 280-1555-39
Client Matrix: Water

Date Sampled: 03/17/2010 1055
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5063.D
Dilution:	1.0		Initial Weight/Volume:	10 mL
Date Analyzed:	03/30/2010 1440		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1440			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.79	J B	0.34	2.0
1,1,1,2-Tetrachloroethane	0.42	U	0.42	2.0
1,1,2,2-Tetrachloroethane	0.42	U	0.42	2.0
Tetrachloroethene	0.40	U	0.40	2.0
Toluene	0.34	U	0.34	2.0
1,2,3-Trichlorobenzene	0.42	U	0.42	2.0
1,2,4-Trichlorobenzene	0.42	U	0.42	2.0
1,1,1-Trichloroethane	0.32	U	0.32	2.0
1,1,2-Trichloroethane	0.54	U	0.54	2.0
Trichloroethene	0.32	U	0.32	2.0
Trichlorofluoromethane	0.58	U	0.58	2.0
1,2,3-Trichloropropane	0.66	U	0.66	2.0
1,2,4-Trimethylbenzene	0.30	U	0.30	2.0
1,3,5-Trimethylbenzene	0.32	U	0.32	2.0
Xylenes, Total	0.38	U	0.38	2.0
1,2-Dibromoethane	0.36	U	0.36	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103		70 - 127
Toluene-d8 (Surr)	105		80 - 125
4-Bromofluorobenzene (Surr)	95		78 - 118
Dibromofluoromethane (Surr)	101		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S67B**

Lab Sample ID: 280-1555-39
Client Matrix: Water

Date Sampled: 03/17/2010 1055
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155		Instrument ID:	MSV_R2
Preparation:	5030B			Lab File ID:	RR5064.D
Dilution:	1.0			Initial Weight/Volume:	1 mL
Date Analyzed:	03/30/2010 1501	Run Type:	DL	Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1501				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Vinyl chloride	300		8.0	20

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	99		70 - 127
Toluene-d8 (Surr)	103		80 - 125
4-Bromofluorobenzene (Surr)	96		78 - 118
Dibromofluoromethane (Surr)	99		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S67C**

Lab Sample ID: 280-1555-40
Client Matrix: Water

Date Sampled: 03/17/2010 1010
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5065.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1522		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1522			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	13		0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
trans-1,2-Dichloroethene	26		0.15	1.0
1,1-Dichloroethene	1.9		0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.39	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S67C**

Lab Sample ID: 280-1555-40
Client Matrix: Water

Date Sampled: 03/17/2010 1010
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5065.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1522		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1522			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	110		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	98		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S67C**

Lab Sample ID: 280-1555-40
Client Matrix: Water

Date Sampled: 03/17/2010 1010
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155		Instrument ID:	MSV_R2
Preparation:	5030B			Lab File ID:	RR5066.D
Dilution:	1.0			Initial Weight/Volume:	2 mL
Date Analyzed:	03/30/2010 1543	Run Type:	DL	Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1543				

Analyte	Result (ug/L)	Qualifier	MDL	RL
cis-1,2-Dichloroethene	130		1.5	10
Vinyl chloride	140		4.0	10

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	102		70 - 127
Toluene-d8 (Surr)	106		80 - 125
4-Bromofluorobenzene (Surr)	102		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S67D**

Lab Sample ID: 280-1555-41
Client Matrix: Water

Date Sampled: 03/17/2010 0905
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5067.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1603		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1603			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	1.6		0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	17		0.15	1.0
trans-1,2-Dichloroethene	5.6		0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.36	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S67D**

Lab Sample ID: 280-1555-41
Client Matrix: Water

Date Sampled: 03/17/2010 0905
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5067.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1603		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1603			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	27		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	106		70 - 127
Toluene-d8 (Surr)	97		80 - 125
4-Bromofluorobenzene (Surr)	98		78 - 118
Dibromofluoromethane (Surr)	101		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S68B**

Lab Sample ID: 280-1555-42
Client Matrix: Water

Date Sampled: 03/16/2010 0840
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5030.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1845		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1845			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S68B**

Lab Sample ID: 280-1555-42
Client Matrix: Water

Date Sampled: 03/16/2010 0840
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5030.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1845		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1845			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.43	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	110		70 - 127
Toluene-d8 (Surr)	103		80 - 125
4-Bromofluorobenzene (Surr)	103		78 - 118
Dibromofluoromethane (Surr)	106		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S68C**

Lab Sample ID: 280-1555-43
Client Matrix: Water

Date Sampled: 03/15/2010 1630
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7647.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0357		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0357			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	4.9	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	4.2		0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	2.8		0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.38	J B	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S68C**

Lab Sample ID: 280-1555-43
Client Matrix: Water

Date Sampled: 03/15/2010 1630
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7647.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0357		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0357			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	4.5		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	109		70 - 127
Toluene-d8 (Surr)	98		80 - 125
4-Bromofluorobenzene (Surr)	95		78 - 118
Dibromofluoromethane (Surr)	100		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S68D**

Lab Sample ID: 280-1555-44
Client Matrix: Water

Date Sampled: 03/15/2010 1550
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7655.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0643		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0643			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	3.0		0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
trans-1,2-Dichloroethene	0.94	J	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S68D**

Lab Sample ID: 280-1555-44
Client Matrix: Water

Date Sampled: 03/15/2010 1550
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7655.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0643		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0643			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	43		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	115		70 - 127
Toluene-d8 (Surr)	99		80 - 125
4-Bromofluorobenzene (Surr)	96		78 - 118
Dibromofluoromethane (Surr)	105		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S68D**

Lab Sample ID: 280-1555-44
Client Matrix: Water

Date Sampled: 03/15/2010 1550
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023		Instrument ID:	MSV_R2
Preparation:	5030B			Lab File ID:	RR5020.D
Dilution:	1.0			Initial Weight/Volume:	10 mL
Date Analyzed:	03/29/2010 1521	Run Type:	DL	Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1521				

Analyte	Result (ug/L)	Qualifier	MDL	RL
cis-1,2-Dichloroethene	54		0.30	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	106		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	99		78 - 118
Dibromofluoromethane (Surr)	100		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S70B**

Lab Sample ID: 280-1555-45
Client Matrix: Water

Date Sampled: 03/15/2010 1445
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7648.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0418		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0418			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	16		0.15	1.0
trans-1,2-Dichloroethene	0.50	J	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.34	J B	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S70B**

Lab Sample ID: 280-1555-45
Client Matrix: Water

Date Sampled: 03/15/2010 1445
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-8984	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7648.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/27/2010 0418		Final Weight/Volume:	20 mL
Date Prepared:	03/27/2010 0418			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	11		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	112		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	100		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S70C**

Lab Sample ID: 280-1555-46
Client Matrix: Water

Date Sampled: 03/15/2010 1345
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7691.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1419		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1419			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	6.5	J B	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	3.5		0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	18		0.15	1.0
trans-1,2-Dichloroethene	5.2		0.15	1.0
1,1-Dichloroethene	0.39	J	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.40	J	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S70C**

Lab Sample ID: 280-1555-46
Client Matrix: Water

Date Sampled: 03/15/2010 1345
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9015	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7691.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1419		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1419			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	11		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	110		70 - 127
Toluene-d8 (Surr)	101		80 - 125
4-Bromofluorobenzene (Surr)	96		78 - 118
Dibromofluoromethane (Surr)	98		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S72B**

Lab Sample ID: 280-1555-47
Client Matrix: Water

Date Sampled: 03/16/2010 1430
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5031.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1905		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1905			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S72B**

Lab Sample ID: 280-1555-47
Client Matrix: Water

Date Sampled: 03/16/2010 1430
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5031.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1905		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1905			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.43	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103		70 - 127
Toluene-d8 (Surr)	94		80 - 125
4-Bromofluorobenzene (Surr)	109		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S72C**

Lab Sample ID: 280-1555-48
Client Matrix: Water

Date Sampled: 03/16/2010 1345
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5014.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1319		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1319			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	2.8		0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	2.2		0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.91	J	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S72C**

Lab Sample ID: 280-1555-48
Client Matrix: Water

Date Sampled: 03/16/2010 1345
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5014.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1319		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1319			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	1.3		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	88		70 - 127
Toluene-d8 (Surr)	98		80 - 125
4-Bromofluorobenzene (Surr)	107		78 - 118
Dibromofluoromethane (Surr)	90		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S72D**

Lab Sample ID: 280-1555-49
Client Matrix: Water

Date Sampled: 03/16/2010 1105
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5032.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1925		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1925			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S72D**

Lab Sample ID: 280-1555-49
Client Matrix: Water

Date Sampled: 03/16/2010 1105
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5032.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1925		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1925			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.42	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	87		70 - 127
Toluene-d8 (Surr)	89		80 - 125
4-Bromofluorobenzene (Surr)	92		78 - 118
Dibromofluoromethane (Surr)	96		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S73C**

Lab Sample ID: 280-1555-50
Client Matrix: Water

Date Sampled: 03/16/2010 1610
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5033.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1946		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1946			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.21	J	0.15	1.0
trans-1,2-Dichloroethene	1.2		0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S73C**

Lab Sample ID: 280-1555-50
Client Matrix: Water

Date Sampled: 03/16/2010 1610
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5033.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1946		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1946			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	5.4		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	119		70 - 127
Toluene-d8 (Surr)	98		80 - 125
4-Bromofluorobenzene (Surr)	100		78 - 118
Dibromofluoromethane (Surr)	119		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S73D**

Lab Sample ID: 280-1555-51
Client Matrix: Water

Date Sampled: 03/16/2010 1520
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5034.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 2006		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 2006			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S73D**

Lab Sample ID: 280-1555-51
Client Matrix: Water

Date Sampled: 03/16/2010 1520
Date Received: 03/19/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9023	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5034.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 2006		Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 2006			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.45	J B	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.23	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	95		70 - 127
Toluene-d8 (Surr)	97		80 - 125
4-Bromofluorobenzene (Surr)	88		78 - 118
Dibromofluoromethane (Surr)	105		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0529**

Lab Sample ID: 280-1555-3
Client Matrix: Water

Date Sampled: 03/17/2010 1057
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8507	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33965.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/23/2010 1438		Final Weight/Volume:	20 mL
Date Prepared:	03/23/2010 1438			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0539**

Lab Sample ID: 280-1555-6
Client Matrix: Water

Date Sampled: 03/17/2010 1102
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8507	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33968.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/23/2010 1554		Final Weight/Volume:	20 mL
Date Prepared:	03/23/2010 1554			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	93		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0552-1**

Lab Sample ID: 280-1555-11
Client Matrix: Water

Date Sampled: 03/16/2010 1340
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8507	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33969.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/23/2010 1619		Final Weight/Volume:	20 mL
Date Prepared:	03/23/2010 1619			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	92		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0552-2**

Lab Sample ID: 280-1555-12
Client Matrix: Water

Date Sampled: 03/16/2010 1438
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8507	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33970.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/23/2010 1645		Final Weight/Volume:	20 mL
Date Prepared:	03/23/2010 1645			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	98		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0552-3**

Lab Sample ID: 280-1555-13
Client Matrix: Water

Date Sampled: 03/16/2010 1509
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8507	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33971.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/23/2010 1710		Final Weight/Volume:	20 mL
Date Prepared:	03/23/2010 1710			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	94		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0567-1**

Lab Sample ID: 280-1555-25
Client Matrix: Water

Date Sampled: 03/16/2010 0921
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8507	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33972.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/23/2010 1736		Final Weight/Volume:	20 mL
Date Prepared:	03/23/2010 1736			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	98		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0567-2**

Lab Sample ID: 280-1555-26
Client Matrix: Water

Date Sampled: 03/16/2010 0947
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8507	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33973.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/23/2010 1801		Final Weight/Volume:	20 mL
Date Prepared:	03/23/2010 1801			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	91		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0567-3**

Lab Sample ID: 280-1555-27
Client Matrix: Water

Date Sampled: 03/16/2010 1039
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8507	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33974.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/23/2010 1826		Final Weight/Volume:	20 mL
Date Prepared:	03/23/2010 1826			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	98		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0568-1**

Lab Sample ID: 280-1555-28
Client Matrix: Water

Date Sampled: 03/15/2010 1509
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8507	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33975.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/23/2010 1852		Final Weight/Volume:	20 mL
Date Prepared:	03/23/2010 1852			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	93		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0568-2**

Lab Sample ID: 280-1555-29
Client Matrix: Water

Date Sampled: 03/15/2010 1544
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8507	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33976.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/23/2010 1917		Final Weight/Volume:	20 mL
Date Prepared:	03/23/2010 1917			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	100		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0568-3**

Lab Sample ID: 280-1555-30
Client Matrix: Water

Date Sampled: 03/16/2010 0847
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33993.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1124		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1124			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	91		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0569-2**

Lab Sample ID: 280-1555-31
Client Matrix: Water

Date Sampled: 03/15/2010 1342
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33996.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1240		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1240			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	83		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0569-3**

Lab Sample ID: 280-1555-32
Client Matrix: Water

Date Sampled: 03/15/2010 1420
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33997.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1305		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1305			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	79		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN12-2872**

Lab Sample ID: 280-1555-33
Client Matrix: Water

Date Sampled: 03/15/2010 1200
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33998.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1331		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1331			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	91		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2901**

Lab Sample ID: 280-1555-37
Client Matrix: Water

Date Sampled: 03/16/2010 0730
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_33999.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1356		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1356			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	92		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2902**

Lab Sample ID: 280-1555-38
Client Matrix: Water

Date Sampled: 03/17/2010 0730
Date Received: 03/19/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8660	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34000.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1422		Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1422			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	99		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2901**

Lab Sample ID: 280-1555-37
Client Matrix: Water

Date Sampled: 03/16/2010 0730
Date Received: 03/19/2010 0900

6010B Metals (ICP)

Method:	6010B	Analysis Batch: 280-8382		Instrument ID:	MT_026
Preparation:	3010A	Prep Batch: 280-8172		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/23/2010 2011			Final Weight/Volume:	50 mL
Date Prepared:	03/23/2010 0730				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Aluminum	18	U	18	100
Iron	22	U	22	100

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2902**

Lab Sample ID: 280-1555-38
Client Matrix: Water

Date Sampled: 03/17/2010 0730
Date Received: 03/19/2010 0900

6010B Metals (ICP)

Method:	6010B	Analysis Batch: 280-8382		Instrument ID:	MT_026
Preparation:	3010A	Prep Batch: 280-8172		Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/23/2010 2020			Final Weight/Volume:	50 mL
Date Prepared:	03/23/2010 0730				

Analyte	Result (ug/L)	Qualifier	MDL	RL
Aluminum	27	J	18	100
Iron	22	U	22	100

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B Volatile Organic Compounds (GC/MS)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DBFM %Rec	DCA %Rec	TOL %Rec	BFB %Rec
280-1555-1	PIN12-0513	97	108	101	95
280-1555-2	PIN12-0526	99	111	101	96
280-1555-3	PIN12-0529	93	91	99	93
280-1555-4	PIN12-0537	101	119	102	98
280-1555-5	PIN12-0538	98	110	99	93
280-1555-6	PIN12-0539	98	100	101	98
280-1555-7	PIN12-0541	100	109	101	97
280-1555-8	PIN12-0542	100	115	101	99
280-1555-9	PIN12-0547	100	98	97	98
280-1555-10	PIN12-0549	100	115	99	94
280-1555-11	PIN12-0552-1	98	110	102	96
280-1555-12	PIN12-0552-2	99	111	101	98
280-1555-13	PIN12-0552-3	99	113	103	97
280-1555-14	PIN12-0554B	99	112	102	96
280-1555-15	PIN12-0554C	100	112	101	95
280-1555-15 DL	PIN12-0554C DL	93	97	100	93
280-1555-16	PIN12-0555A	91	89	98	108
280-1555-17	PIN12-0555B	89	89	103	114
280-1555-18	PIN12-0555C	107	107	101	96
280-1555-19	PIN12-0562-1	99	108	103	100
280-1555-20	PIN12-0562-2	100	107	106	100
280-1555-21	PIN12-0562-3	104	102	98	104
280-1555-22	PIN12-0563-1	96	103	96	96
280-1555-23	PIN12-0563-2	103	102	98	100
280-1555-24	PIN12-0563-3	107	109	99	94
280-1555-25	PIN12-0567-1	106	110	102	98
280-1555-26	PIN12-0567-2	108	113	104	100
280-1555-27	PIN12-0567-3	118	117	100	102
280-1555-28	PIN12-0568-1	98	104	103	101

Surrogate	Acceptance Limits
DBFM = Dibromofluoromethane (Surr)	77-119
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127
TOL = Toluene-d8 (Surr)	80-125
BFB = 4-Bromofluorobenzene (Surr)	78-118

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B Volatile Organic Compounds (GC/MS)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DBFM %Rec	DCA %Rec	TOL %Rec	BFB %Rec
280-1555-29	PIN12-0568-2	100	112	101	100
280-1555-30	PIN12-0568-3	103	102	100	105
280-1555-31	PIN12-0569-2	103	113	103	101
280-1555-31 DL	PIN12-0569-2 DL	102	108	99	99
280-1555-32	PIN12-0569-3	102	112	99	101
280-1555-32 DL	PIN12-0569-3 DL	100	107	100	97
280-1555-33	PIN12-2872	104	111	99	97
280-1555-33 DL	PIN12-2872 DL	101	108	98	99
280-1555-34	PIN99-2876	99	107	102	103
280-1555-35	PIN99-2882	95	101	101	95
280-1555-36	PIN99-2890	104	104	108	103
280-1555-37	PIN99-2901	102	103	107	106
280-1555-38	PIN99-2902	108	107	114	105
280-1555-39	PIN12-S67B	101	103	105	95
280-1555-39 DL	PIN12-S67B DL	99	99	103	96
280-1555-40	PIN12-S67C	103	110	101	98
280-1555-40 DL	PIN12-S67C DL	104	102	106	102
280-1555-41	PIN12-S67D	101	106	97	98
280-1555-42	PIN12-S68B	106	110	103	103
280-1555-43	PIN12-S68C	100	109	98	95
280-1555-44	PIN12-S68D	105	115	99	96
280-1555-44 DL	PIN12-S68D DL	100	106	101	99
280-1555-45	PIN12-S70B	103	112	101	100
280-1555-46	PIN12-S70C	98	110	101	96
280-1555-47	PIN12-S72B	103	103	94	109
280-1555-48	PIN12-S72C	90	88	98	107
280-1555-49	PIN12-S72D	96	87	89	92
280-1555-50	PIN12-S73C	119	119	98	100
280-1555-51	PIN12-S73D	105	95	97	88
MB 280-8984/6		101	108	102	101

Surrogate	Acceptance Limits
DBFM = Dibromofluoromethane (Surr)	77-119
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127
TOL = Toluene-d8 (Surr)	80-125
BFB = 4-Bromofluorobenzene (Surr)	78-118

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B Volatile Organic Compounds (GC/MS)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DBFM %Rec	DCA %Rec	TOL %Rec	BFB %Rec
MB 280-9015/6		96	99	102	99
MB 280-9023/5		91	89	99	107
MB 280-9155/5		101	99	104	102
LCS 280-8984/4		98	100	100	96
LCS 280-9015/4		95	96	106	99
LCS 280-9023/4		92	92	107	114
LCS 280-9155/4		99	97	108	104
LCSD 280-8984/5		97	102	100	97
LCSD 280-9015/5		93	95	103	97
280-1555-3 MS	PIN12-0529 MS	96	97	100	99
280-1555-48 MS	PIN12-S72C MS	91	95	100	110
280-1737-E-2 MS		99	101	100	96
280-1536-H-2 MS		99	103	100	96
280-1555-3 MSD	PIN12-0529 MSD	96	95	102	97
280-1555-48 MSD	PIN12-S72C MSD	94	96	96	114
280-1737-E-2 MSD		96	100	102	97
280-1536-F-2 MSD		98	103	101	96

Surrogate	Acceptance Limits
DBFM = Dibromofluoromethane (Surr)	77-119
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127
TOL = Toluene-d8 (Surr)	80-125
BFB = 4-Bromofluorobenzene (Surr)	78-118

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DCA %Rec
280-1555-3	PIN12-0529	103
280-1555-6	PIN12-0539	93
280-1555-11	PIN12-0552-1	92
280-1555-12	PIN12-0552-2	98
280-1555-13	PIN12-0552-3	94
280-1555-25	PIN12-0567-1	98
280-1555-26	PIN12-0567-2	91
280-1555-27	PIN12-0567-3	98
280-1555-28	PIN12-0568-1	93
280-1555-29	PIN12-0568-2	100
280-1555-30	PIN12-0568-3	91
280-1555-31	PIN12-0569-2	83
280-1555-32	PIN12-0569-3	79
280-1555-33	PIN12-2872	91
280-1555-37	PIN99-2901	92
280-1555-38	PIN99-2902	99
MB 280-8507/4		110
MB 280-8660/4		113
LCS 280-8507/3		107
LCS 280-8660/3		109
280-1555-3 MS	PIN12-0529 MS	98
280-1555-30 MS	PIN12-0568-3 MS	88
280-1555-3 MSD	PIN12-0529 MSD	98
280-1555-30 MSD	PIN12-0568-3 MSD	93

Surrogate	Acceptance Limits
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Method Blank - Batch: 280-8984

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-8984/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/26/2010 2105
Date Prepared: 03/26/2010 2105

Analysis Batch: 280-8984
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_R1
Lab File ID: R7627.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.443	J	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Method Blank - Batch: 280-8984

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-8984/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/26/2010 2105
Date Prepared: 03/26/2010 2105

Analysis Batch: 280-8984
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_R1
Lab File ID: R7627.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	108	70 - 127
Toluene-d8 (Surr)	102	80 - 125
4-Bromofluorobenzene (Surr)	101	78 - 118
Dibromofluoromethane (Surr)	101	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Control Sample/
Lab Control Sample Duplicate Recovery Report - Batch: 280-8984

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-8984/4	Analysis Batch:	280-8984	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7624.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 2002			Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 2002				

LCSD Lab Sample ID:	LCSD 280-8984/5	Analysis Batch:	280-8984	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7625.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 2023			Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 2023				

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	91	92	77 - 120	1	20		
Bromodichloromethane	86	89	78 - 120	3	20		
Carbon tetrachloride	106	106	80 - 120	0	21		
Chlorobenzene	88	90	78 - 120	2	20		
Chloroform	92	93	78 - 120	1	20		
1,3-Dichlorobenzene	87	91	75 - 120	4	20		
1,1-Dichloroethane	93	92	77 - 120	1	21		
trans-1,2-Dichloroethene	100	99	80 - 120	1	24		
1,1-Dichloroethene	109	110	68 - 133	1	20		
1,2-Dichloropropane	88	90	76 - 120	3	20		
Ethylbenzene	94	95	78 - 120	1	26		
Methylene Chloride	96	98	71 - 120	2	20		
Tetrachloroethene	102	103	77 - 120	1	20		
Toluene	94	94	73 - 120	1	20		
1,1,1-Trichloroethane	104	103	78 - 120	1	20		
Trichloroethene	96	95	78 - 122	1	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	100	102	70 - 127
Toluene-d8 (Surr)	100	100	80 - 125
4-Bromofluorobenzene (Surr)	96	97	78 - 118
Dibromofluoromethane (Surr)	98	97	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Laboratory Control/
Laboratory Duplicate Data Report - Batch: 280-8984

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-8984/4	Units:	ug/L	LCSD Lab Sample ID:	LCSD 280-8984/5
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/26/2010 2002			Date Analyzed:	03/26/2010 2023
Date Prepared:	03/26/2010 2002			Date Prepared:	03/26/2010 2023

Analyte	LCS Spike Amount	LCSD Spike Amount	LCS Result/Qual	LCSD Result/Qual
Benzene	5.00	5.00	4.54	4.60
Bromodichloromethane	5.00	5.00	4.32	4.44
Carbon tetrachloride	5.00	5.00	5.32	5.32
Chlorobenzene	5.00	5.00	4.41	4.51
Chloroform	5.00	5.00	4.60	4.65
1,3-Dichlorobenzene	5.00	5.00	4.35	4.54
1,1-Dichloroethane	5.00	5.00	4.67	4.62
trans-1,2-Dichloroethene	5.00	5.00	5.00	4.93
1,1-Dichloroethene	5.00	5.00	5.45	5.51
1,2-Dichloropropane	5.00	5.00	4.38	4.50
Ethylbenzene	5.00	5.00	4.71	4.74
Methylene Chloride	5.00	5.00	4.78	4.88
Tetrachloroethene	5.00	5.00	5.10	5.15
Toluene	5.00	5.00	4.69	4.72
1,1,1-Trichloroethane	5.00	5.00	5.21	5.17
Trichloroethene	5.00	5.00	4.78	4.73

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-8984

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1737-E-2 MS	Analysis Batch:	280-8984	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7629.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 2146			Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 2146				

MSD Lab Sample ID:	280-1737-E-2 MSD	Analysis Batch:	280-8984	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7630.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 2207			Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 2207				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	94	92	77 - 120	2	20		
Bromodichloromethane	88	87	78 - 120	1	20		
Carbon tetrachloride	111	103	80 - 120	7	21		
Chlorobenzene	90	90	78 - 120	0	20		
Chloroform	86	81	78 - 120	2	20		
1,3-Dichlorobenzene	88	88	75 - 120	1	20		
1,1-Dichloroethane	95	92	77 - 120	3	21		
trans-1,2-Dichloroethene	100	99	80 - 120	1	24		
1,1-Dichloroethene	113	108	68 - 133	5	20		
1,2-Dichloropropane	90	90	76 - 120	1	20		
Ethylbenzene	96	96	78 - 120	0	26		
Methylene Chloride	93	90	71 - 120	3	20		
Tetrachloroethene	104	103	77 - 120	1	20		
Toluene	94	92	73 - 120	1	20		
1,1,1-Trichloroethane	108	103	78 - 120	4	20		
Trichloroethene	96	93	78 - 122	3	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101	100	70 - 127
Toluene-d8 (Surr)	100	102	80 - 125
4-Bromofluorobenzene (Surr)	96	97	78 - 118
Dibromofluoromethane (Surr)	99	96	77 - 119

Quality Control Results

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-8984

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1737-E-2 MS	Units: ug/L	MSD Lab Sample ID:	280-1737-E-2 MSD
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/26/2010 2146		Date Analyzed:	03/26/2010 2207
Date Prepared:	03/26/2010 2146		Date Prepared:	03/26/2010 2207

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	0.16	U	5.00	5.00	4.71	4.60
Bromodichloromethane	3.0		5.00	5.00	7.44	7.36
Carbon tetrachloride	0.19	U	5.00	5.00	5.54	5.16
Chlorobenzene	0.17	U	5.00	5.00	4.51	4.52
Chloroform	8.0		5.00	5.00	12.3	12.0
1,3-Dichlorobenzene	0.13	U	5.00	5.00	4.42	4.39
1,1-Dichloroethane	0.22	U	5.00	5.00	4.74	4.61
trans-1,2-Dichloroethene	0.15	U	5.00	5.00	5.02	4.94
1,1-Dichloroethene	0.23	U	5.00	5.00	5.65	5.39
1,2-Dichloropropane	0.18	U	5.00	5.00	4.52	4.49
Ethylbenzene	0.16	U	5.00	5.00	4.79	4.79
Methylene Chloride	0.37	J	5.00	5.00	5.00	4.88
Tetrachloroethene	0.20	U	5.00	5.00	5.21	5.14
Toluene	0.64	J	5.00	5.00	5.32	5.26
1,1,1-Trichloroethane	0.16	U	5.00	5.00	5.38	5.14
Trichloroethene	0.16	U	5.00	5.00	4.80	4.63

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Method Blank - Batch: 280-9015

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-9015/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/29/2010 1035
Date Prepared: 03/29/2010 1035

Analysis Batch: 280-9015
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_R1
Lab File ID: R7681.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
Acetone	2.64	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Method Blank - Batch: 280-9015

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-9015/6
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/29/2010 1035
Date Prepared: 03/29/2010 1035

Analysis Batch: 280-9015
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_R1
Lab File ID: R7681.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	99	70 - 127
Toluene-d8 (Surr)	102	80 - 125
4-Bromofluorobenzene (Surr)	99	78 - 118
Dibromofluoromethane (Surr)	96	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Control Sample/
Lab Control Sample Duplicate Recovery Report - Batch: 280-9015

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-9015/4	Analysis Batch:	280-9015	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7679.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 0954			Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 0954				

LCSD Lab Sample ID:	LCSD 280-9015/5	Analysis Batch:	280-9015	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7680.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1015			Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1015				

Analyte	% Rec. LCS	% Rec. LCSD	Limit	RPD	RPD Limit	LCS Qual	LCSD Qual
Benzene	87	86	77 - 120	2	20		
Bromodichloromethane	81	81	78 - 120	1	20		
Carbon tetrachloride	86	83	80 - 120	3	21		
Chlorobenzene	87	86	78 - 120	1	20		
Chloroform	87	86	78 - 120	1	20		
1,3-Dichlorobenzene	89	87	75 - 120	2	20		
1,1-Dichloroethane	89	87	77 - 120	1	21		
trans-1,2-Dichloroethene	89	88	80 - 120	1	24		
1,1-Dichloroethene	98	95	68 - 133	3	20		
1,2-Dichloropropane	87	86	76 - 120	1	20		
Ethylbenzene	90	88	78 - 120	2	26		
Methylene Chloride	93	89	71 - 120	3	20		
Tetrachloroethene	91	85	77 - 120	7	20		
Toluene	85	84	73 - 120	1	20		
1,1,1-Trichloroethane	85	84	78 - 120	1	20		
Trichloroethene	86	83	78 - 122	3	20		

Surrogate	LCS % Rec	LCSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	96	95	70 - 127
Toluene-d8 (Surr)	106	103	80 - 125
4-Bromofluorobenzene (Surr)	99	97	78 - 118
Dibromofluoromethane (Surr)	95	93	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Laboratory Control/
Laboratory Duplicate Data Report - Batch: 280-9015

Method: 8260B
Preparation: 5030B

LCS Lab Sample ID:	LCS 280-9015/4	Units:	ug/L	LCSD Lab Sample ID:	LCSD 280-9015/5
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/29/2010 0954			Date Analyzed:	03/29/2010 1015
Date Prepared:	03/29/2010 0954			Date Prepared:	03/29/2010 1015

Analyte	LCS Spike Amount	LCSD Spike Amount	LCS Result/Qual	LCSD Result/Qual
Benzene	5.00	5.00	4.34	4.28
Bromodichloromethane	5.00	5.00	4.06	4.03
Carbon tetrachloride	5.00	5.00	4.29	4.16
Chlorobenzene	5.00	5.00	4.36	4.32
Chloroform	5.00	5.00	4.37	4.31
1,3-Dichlorobenzene	5.00	5.00	4.44	4.33
1,1-Dichloroethane	5.00	5.00	4.43	4.37
trans-1,2-Dichloroethene	5.00	5.00	4.43	4.38
1,1-Dichloroethene	5.00	5.00	4.88	4.74
1,2-Dichloropropane	5.00	5.00	4.33	4.29
Ethylbenzene	5.00	5.00	4.48	4.41
Methylene Chloride	5.00	5.00	4.63	4.47
Tetrachloroethene	5.00	5.00	4.56	4.27
Toluene	5.00	5.00	4.24	4.18
1,1,1-Trichloroethane	5.00	5.00	4.27	4.21
Trichloroethene	5.00	5.00	4.29	4.15

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-9015

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1536-H-2 MS	Analysis Batch:	280-9015	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7686.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1236			Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1236				

MSD Lab Sample ID:	280-1536-F-2 MSD	Analysis Batch:	280-9015	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7687.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1257			Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1257				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	85	85	77 - 120	1	20		
Bromodichloromethane	84	84	78 - 120	0	20		
Carbon tetrachloride	78	83	80 - 120	6	21	F	
Chlorobenzene	83	86	78 - 120	3	20		
Chloroform	90	89	78 - 120	0	20		
1,3-Dichlorobenzene	82	84	75 - 120	3	20		
1,1-Dichloroethane	87	88	77 - 120	1	21		
trans-1,2-Dichloroethene	85	88	80 - 120	3	24		
1,1-Dichloroethene	88	92	68 - 133	5	20		
1,2-Dichloropropane	86	88	76 - 120	2	20		
Ethylbenzene	81	86	78 - 120	5	26		
Methylene Chloride	91	90	71 - 120	1	20		
Tetrachloroethene	76	81	77 - 120	7	20	F	
Toluene	83	85	73 - 120	3	20		
1,1,1-Trichloroethane	81	84	78 - 120	3	20		
Trichloroethene	81	82	78 - 122	1	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103	103	70 - 127
Toluene-d8 (Surr)	100	101	80 - 125
4-Bromofluorobenzene (Surr)	96	96	78 - 118
Dibromofluoromethane (Surr)	99	98	77 - 119

Quality Control Results

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-9015

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1536-H-2 MS	Units: ug/L	MSD Lab Sample ID:	280-1536-F-2 MSD
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/29/2010 1236		Date Analyzed:	03/29/2010 1257
Date Prepared:	03/29/2010 1236		Date Prepared:	03/29/2010 1257

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual		MSD Result/Qual
Benzene	0.16	U	5.00	5.00	4.23		4.27
Bromodichloromethane	0.17	U	5.00	5.00	4.21		4.22
Carbon tetrachloride	0.19	U	5.00	5.00	3.91	F	4.16
Chlorobenzene	0.17	U	5.00	5.00	4.15		4.29
Chloroform	0.16	U	5.00	5.00	4.50		4.47
1,3-Dichlorobenzene	0.13	U	5.00	5.00	4.08		4.21
1,1-Dichloroethane	0.22	U	5.00	5.00	4.36		4.41
trans-1,2-Dichloroethene	0.15	U	5.00	5.00	4.24		4.38
1,1-Dichloroethene	0.23	U	5.00	5.00	4.39		4.62
1,2-Dichloropropane	0.18	U	5.00	5.00	4.32		4.42
Ethylbenzene	0.16	U	5.00	5.00	4.06		4.28
Methylene Chloride	0.44	J	5.00	5.00	4.98		4.93
Tetrachloroethene	0.20	U	5.00	5.00	3.80	F	4.06
Toluene	0.17	U	5.00	5.00	4.13		4.25
1,1,1-Trichloroethane	0.16	U	5.00	5.00	4.05		4.19
Trichloroethene	0.16	U	5.00	5.00	4.05		4.11

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Method Blank - Batch: 280-9023

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-9023/5	Analysis Batch:	280-9023	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR5011.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1212			Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1212				

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Method Blank - Batch: 280-9023

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-9023/5	Analysis Batch:	280-9023	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR5011.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1212			Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1212				

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.393	J	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	89	70 - 127
Toluene-d8 (Surr)	99	80 - 125
4-Bromofluorobenzene (Surr)	107	78 - 118
Dibromofluoromethane (Surr)	91	77 - 119

Quality Control Results

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Control Sample - Batch: 280-9023

Method: 8260B
Preparation: 5030B

Lab Sample ID:	LCS 280-9023/4	Analysis Batch:	280-9023	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR5008.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1111			Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1111				

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Benzene	5.00	5.02	100	77 - 120	
Bromodichloromethane	5.00	4.67	93	78 - 120	
Carbon tetrachloride	5.00	5.12	102	80 - 120	
Chlorobenzene	5.00	5.07	101	78 - 120	
Chloroform	5.00	4.88	98	78 - 120	
1,3-Dichlorobenzene	5.00	5.20	104	75 - 120	
1,1-Dichloroethane	5.00	4.91	98	77 - 120	
trans-1,2-Dichloroethene	5.00	4.99	100	80 - 120	
1,1-Dichloroethene	5.00	5.60	112	68 - 133	
1,2-Dichloropropane	5.00	4.80	96	76 - 120	
Ethylbenzene	5.00	5.37	107	78 - 120	
Methylene Chloride	5.00	4.83	97	71 - 120	
Tetrachloroethene	5.00	5.55	111	77 - 120	
Toluene	5.00	4.89	98	73 - 120	
1,1,1-Trichloroethane	5.00	5.29	106	78 - 120	
Trichloroethene	5.00	4.86	97	78 - 122	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	92	70 - 127
Toluene-d8 (Surr)	107	80 - 125
4-Bromofluorobenzene (Surr)	114	78 - 118
Dibromofluoromethane (Surr)	92	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-9023

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1555-48	Analysis Batch:	280-9023	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR5015.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1340			Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1340				

MSD Lab Sample ID:	280-1555-48	Analysis Batch:	280-9023	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR5016.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/29/2010 1400			Final Weight/Volume:	20 mL
Date Prepared:	03/29/2010 1400				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	88	94	77 - 120	6	20		
Bromodichloromethane	86	91	78 - 120	6	20		
Carbon tetrachloride	89	101	80 - 120	12	21		
Chlorobenzene	88	97	78 - 120	9	20		
Chloroform	89	97	78 - 120	9	20		
1,3-Dichlorobenzene	90	96	75 - 120	6	20		
1,1-Dichloroethane	90	102	77 - 120	8	21		
trans-1,2-Dichloroethene	91	101	80 - 120	10	24		
1,1-Dichloroethene	95	113	68 - 133	15	20		
1,2-Dichloropropane	88	88	76 - 120	0	20		
Ethylbenzene	89	102	78 - 120	14	26		
Methylene Chloride	93	102	71 - 120	10	20		
Tetrachloroethene	87	93	77 - 120	7	20		
Toluene	87	93	73 - 120	6	20		
1,1,1-Trichloroethane	92	107	78 - 120	15	20		
Trichloroethene	85	90	78 - 122	6	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	95	96	70 - 127
Toluene-d8 (Surr)	100	96	80 - 125
4-Bromofluorobenzene (Surr)	110	114	78 - 118
Dibromofluoromethane (Surr)	91	94	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-9023

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1555-48	Units: ug/L	MSD Lab Sample ID:	280-1555-48
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/29/2010 1340		Date Analyzed:	03/29/2010 1400
Date Prepared:	03/29/2010 1340		Date Prepared:	03/29/2010 1400

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	0.16	U	5.00	5.00	4.42	4.70
Bromodichloromethane	0.17	U	5.00	5.00	4.30	4.54
Carbon tetrachloride	0.19	U	5.00	5.00	4.47	5.05
Chlorobenzene	0.17	U	5.00	5.00	4.41	4.83
Chloroform	0.16	U	5.00	5.00	4.44	4.86
1,3-Dichlorobenzene	0.13	U	5.00	5.00	4.51	4.82
1,1-Dichloroethane	2.8		5.00	5.00	7.32	7.93
trans-1,2-Dichloroethene	0.15	U	5.00	5.00	4.54	5.04
1,1-Dichloroethene	0.91	J	5.00	5.00	5.65	6.56
1,2-Dichloropropane	0.18	U	5.00	5.00	4.38	4.39
Ethylbenzene	0.16	U	5.00	5.00	4.45	5.10
Methylene Chloride	0.32	U	5.00	5.00	4.64	5.12
Tetrachloroethene	0.20	U	5.00	5.00	4.34	4.67
Toluene	0.17	U	5.00	5.00	4.37	4.67
1,1,1-Trichloroethane	0.16	U	5.00	5.00	4.58	5.33
Trichloroethene	0.16	U	5.00	5.00	4.23	4.51

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Method Blank - Batch: 280-9155

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-9155/5	Analysis Batch:	280-9155	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR5051.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1031			Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1031				

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Method Blank - Batch: 280-9155

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-9155/5	Analysis Batch:	280-9155	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR5051.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1031			Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1031				

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.428	J	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	99	70 - 127
Toluene-d8 (Surr)	104	80 - 125
4-Bromofluorobenzene (Surr)	102	78 - 118
Dibromofluoromethane (Surr)	101	77 - 119

Quality Control Results

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Lab Control Sample - Batch: 280-9155

Method: 8260B
Preparation: 5030B

Lab Sample ID:	LCS 280-9155/4	Analysis Batch:	280-9155	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR5048.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 0929			Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 0929				

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Benzene	5.00	5.52	110	77 - 120	
Bromodichloromethane	5.00	4.89	98	78 - 120	
Carbon tetrachloride	5.00	5.62	112	80 - 120	
Chlorobenzene	5.00	5.34	107	78 - 120	
Chloroform	5.00	5.07	101	78 - 120	
1,3-Dichlorobenzene	5.00	5.81	116	75 - 120	
1,1-Dichloroethane	5.00	5.39	108	77 - 120	
trans-1,2-Dichloroethene	5.00	5.57	111	80 - 120	
1,1-Dichloroethene	5.00	6.35	127	68 - 133	
1,2-Dichloropropane	5.00	4.92	98	76 - 120	
Ethylbenzene	5.00	5.75	115	78 - 120	
Methylene Chloride	5.00	5.22	104	71 - 120	
Tetrachloroethene	5.00	5.89	118	77 - 120	
Toluene	5.00	4.97	99	73 - 120	
1,1,1-Trichloroethane	5.00	5.17	103	78 - 120	
Trichloroethene	5.00	5.12	102	78 - 122	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	97	70 - 127
Toluene-d8 (Surr)	108	80 - 125
4-Bromofluorobenzene (Surr)	104	78 - 118
Dibromofluoromethane (Surr)	99	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-9155

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1555-3	Analysis Batch:	280-9155	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR5054.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1133			Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1133				

MSD Lab Sample ID:	280-1555-3	Analysis Batch:	280-9155	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR5055.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1153			Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1153				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	108	103	77 - 120	5	20		
Bromodichloromethane	94	95	78 - 120	1	20		
Carbon tetrachloride	111	97	80 - 120	14	21		
Chlorobenzene	103	102	78 - 120	1	20		
Chloroform	105	102	78 - 120	3	20		
1,3-Dichlorobenzene	107	105	75 - 120	2	20		
1,1-Dichloroethane	109	103	77 - 120	5	21		
trans-1,2-Dichloroethene	113	105	80 - 120	8	24		
1,1-Dichloroethene	134	117	68 - 133	13	20	F	
1,2-Dichloropropane	95	97	76 - 120	2	20		
Ethylbenzene	111	107	78 - 120	4	26		
Methylene Chloride	111	104	71 - 120	5	20		
Tetrachloroethene	112	102	77 - 120	9	20		
Toluene	97	99	73 - 120	2	20		
1,1,1-Trichloroethane	107	93	78 - 120	14	20		
Trichloroethene	100	96	78 - 122	4	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	97	95	70 - 127
Toluene-d8 (Surr)	100	102	80 - 125
4-Bromofluorobenzene (Surr)	99	97	78 - 118
Dibromofluoromethane (Surr)	96	96	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-9155

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1555-3	Units: ug/L	MSD Lab Sample ID:	280-1555-3
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/30/2010 1133		Date Analyzed:	03/30/2010 1153
Date Prepared:	03/30/2010 1133		Date Prepared:	03/30/2010 1153

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual		MSD Result/Qual
Benzene	0.16	U	5.00	5.00	5.39		5.14
Bromodichloromethane	0.17	U	5.00	5.00	4.72		4.76
Carbon tetrachloride	0.19	U	5.00	5.00	5.56		4.84
Chlorobenzene	0.17	U	5.00	5.00	5.16		5.11
Chloroform	0.16	U	5.00	5.00	5.24		5.09
1,3-Dichlorobenzene	0.13	U	5.00	5.00	5.36		5.27
1,1-Dichloroethane	0.22	U	5.00	5.00	5.45		5.16
trans-1,2-Dichloroethene	0.15	U	5.00	5.00	5.67		5.25
1,1-Dichloroethene	0.23	U	5.00	5.00	6.68	F	5.87
1,2-Dichloropropane	0.18	U	5.00	5.00	4.77		4.86
Ethylbenzene	0.16	U	5.00	5.00	5.53		5.33
Methylene Chloride	0.66	J	5.00	5.00	6.21		5.88
Tetrachloroethene	0.20	U	5.00	5.00	5.62		5.12
Toluene	0.17	U	5.00	5.00	4.84		4.94
1,1,1-Trichloroethane	0.16	U	5.00	5.00	5.37		4.67
Trichloroethene	0.16	U	5.00	5.00	5.01		4.82

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Method Blank - Batch: 280-8507

Method: 8260B SIM
Preparation: 5030B

Lab Sample ID:	MB 280-8507/4	Analysis Batch:	280-8507	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_33954.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/23/2010 0957			Final Weight/Volume:	20 mL
Date Prepared:	03/23/2010 0957				

Analyte	Result	Qual	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	110	70 - 127

Lab Control Sample - Batch: 280-8507

Method: 8260B SIM
Preparation: 5030B

Lab Sample ID:	LCS 280-8507/3	Analysis Batch:	280-8507	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_33952.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/23/2010 0906			Final Weight/Volume:	20 mL
Date Prepared:	03/23/2010 0906				

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
1,4-Dioxane	5.00	6.84	137	25 - 141	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	107	70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-8507

Method: 8260B SIM
Preparation: 5030B

MS Lab Sample ID:	280-1555-3	Analysis Batch:	280-8507	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_33966.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/23/2010 1503			Final Weight/Volume:	20 mL
Date Prepared:	03/23/2010 1503				

MSD Lab Sample ID:	280-1555-3	Analysis Batch:	280-8507	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_33967.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/23/2010 1528			Final Weight/Volume:	20 mL
Date Prepared:	03/23/2010 1528				

Analyte	% Rec. MS	% Rec. MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
1,4-Dioxane	85	89	25 - 141	4	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	98	98	70 - 127

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-8507

Method: 8260B SIM
Preparation: 5030B

MS Lab Sample ID:	280-1555-3	Units: ug/L	MSD Lab Sample ID:	280-1555-3
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/23/2010 1503		Date Analyzed:	03/23/2010 1528
Date Prepared:	03/23/2010 1503		Date Prepared:	03/23/2010 1528

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
1,4-Dioxane	0.75	U	5.00	5.00	4.26	4.45

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Method Blank - Batch: 280-8660

Method: 8260B SIM
Preparation: 5030B

Lab Sample ID:	MB 280-8660/4	Analysis Batch:	280-8660	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_33992.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1059			Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1059				

Analyte	Result	Qual	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	113	70 - 127

Lab Control Sample - Batch: 280-8660

Method: 8260B SIM
Preparation: 5030B

Lab Sample ID:	LCS 280-8660/3	Analysis Batch:	280-8660	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_33990.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 0940			Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 0940				

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
1,4-Dioxane	5.00	5.62	112	25 - 141	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	109	70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-8660

Method: 8260B SIM
Preparation: 5030B

MS Lab Sample ID:	280-1555-30	Analysis Batch:	280-8660	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_33994.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1150			Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1150				

MSD Lab Sample ID:	280-1555-30	Analysis Batch:	280-8660	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_33995.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/25/2010 1215			Final Weight/Volume:	20 mL
Date Prepared:	03/25/2010 1215				

Analyte	% Rec. MS	% Rec. MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
1,4-Dioxane	78	85	25 - 141	8	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	88	93	70 - 127

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-8660

Method: 8260B SIM
Preparation: 5030B

MS Lab Sample ID:	280-1555-30	Units:	ug/L	MSD Lab Sample ID:	280-1555-30
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/25/2010 1150			Date Analyzed:	03/25/2010 1215
Date Prepared:	03/25/2010 1150			Date Prepared:	03/25/2010 1215

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
1,4-Dioxane	0.75	U	5.00	5.00	3.92	4.27

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Method Blank - Batch: 280-8172

Method: 6010B
Preparation: 3010A

Lab Sample ID:	MB 280-8172/1-A	Analysis Batch:	280-8382	Instrument ID:	MT_026
Client Matrix:	Water	Prep Batch:	280-8172	Lab File ID:	N/A
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	50 mL
Date Analyzed:	03/23/2010 2004			Final Weight/Volume:	50 mL
Date Prepared:	03/23/2010 0730				

Analyte	Result	Qual	MDL	RL
Aluminum	18	U	18	100
Iron	22	U	22	100

Lab Control Sample - Batch: 280-8172

Method: 6010B
Preparation: 3010A

Lab Sample ID:	LCS 280-8172/2-A	Analysis Batch:	280-8382	Instrument ID:	MT_026
Client Matrix:	Water	Prep Batch:	280-8172	Lab File ID:	N/A
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	50 mL
Date Analyzed:	03/23/2010 2006			Final Weight/Volume:	50 mL
Date Prepared:	03/23/2010 0730				

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Aluminum	2000	1940	97	87 - 111	
Iron	1000	945	94	89 - 115	

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-8172

Method: 6010B
Preparation: 3010A

MS Lab Sample ID:	280-1555-37	Analysis Batch:	280-8382	Instrument ID:	MT_026
Client Matrix:	Water	Prep Batch:	280-8172	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/23/2010 2015			Final Weight/Volume:	50 mL
Date Prepared:	03/23/2010 0730				

MSD Lab Sample ID:	280-1555-37	Analysis Batch:	280-8382	Instrument ID:	MT_026
Client Matrix:	Water	Prep Batch:	280-8172	Lab File ID:	N/A
Dilution:	1.0			Initial Weight/Volume:	50 mL
Date Analyzed:	03/23/2010 2018			Final Weight/Volume:	50 mL
Date Prepared:	03/23/2010 0730				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Aluminum	96	97	83 - 119	1	25		
Iron	93	94	52 - 155	1	25		

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-8172

Method: 6010B
Preparation: 3010A

MS Lab Sample ID:	280-1555-37	Units: ug/L	MSD Lab Sample ID:	280-1555-37
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/23/2010 2015		Date Analyzed:	03/23/2010 2018
Date Prepared:	03/23/2010 0730		Date Prepared:	03/23/2010 0730

	Sample		MS Spike	MSD Spike	MS	MSD
Analyte	Result/Qual		Amount	Amount	Result/Qual	Result/Qual
Aluminum	18	U	2000	2000	1920	1940
Iron	22	U	1000	1000	925	936

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:280-8507					
LCS 280-8507/3	Lab Control Sample	T	Water	8260B SIM	
MB 280-8507/4	Method Blank	T	Water	8260B SIM	
280-1555-3	PIN12-0529	T	Water	8260B SIM	
280-1555-3MS	Matrix Spike	T	Water	8260B SIM	
280-1555-3MSD	Matrix Spike Duplicate	T	Water	8260B SIM	
280-1555-6	PIN12-0539	T	Water	8260B SIM	
280-1555-11	PIN12-0552-1	T	Water	8260B SIM	
280-1555-12	PIN12-0552-2	T	Water	8260B SIM	
280-1555-13	PIN12-0552-3	T	Water	8260B SIM	
280-1555-25	PIN12-0567-1	T	Water	8260B SIM	
280-1555-26	PIN12-0567-2	T	Water	8260B SIM	
280-1555-27	PIN12-0567-3	T	Water	8260B SIM	
280-1555-28	PIN12-0568-1	T	Water	8260B SIM	
280-1555-29	PIN12-0568-2	T	Water	8260B SIM	
Analysis Batch:280-8660					
LCS 280-8660/3	Lab Control Sample	T	Water	8260B SIM	
MB 280-8660/4	Method Blank	T	Water	8260B SIM	
280-1555-30	PIN12-0568-3	T	Water	8260B SIM	
280-1555-30MS	Matrix Spike	T	Water	8260B SIM	
280-1555-30MSD	Matrix Spike Duplicate	T	Water	8260B SIM	
280-1555-31	PIN12-0569-2	T	Water	8260B SIM	
280-1555-32	PIN12-0569-3	T	Water	8260B SIM	
280-1555-33	PIN12-2872	T	Water	8260B SIM	
280-1555-37	PIN99-2901	T	Water	8260B SIM	
280-1555-38	PIN99-2902	T	Water	8260B SIM	
Analysis Batch:280-8984					
LCS 280-8984/4	Lab Control Sample	T	Water	8260B	
LCSD 280-8984/5	Lab Control Sample Duplicate	T	Water	8260B	
MB 280-8984/6	Method Blank	T	Water	8260B	
280-1555-28	PIN12-0568-1	T	Water	8260B	
280-1555-29	PIN12-0568-2	T	Water	8260B	
280-1555-31	PIN12-0569-2	T	Water	8260B	
280-1555-31DL	PIN12-0569-2	T	Water	8260B	
280-1555-32	PIN12-0569-3	T	Water	8260B	
280-1555-32DL	PIN12-0569-3	T	Water	8260B	
280-1555-33	PIN12-2872	T	Water	8260B	
280-1555-33DL	PIN12-2872	T	Water	8260B	
280-1555-34	PIN99-2876	T	Water	8260B	
280-1555-43	PIN12-S68C	T	Water	8260B	
280-1555-44	PIN12-S68D	T	Water	8260B	
280-1555-45	PIN12-S70B	T	Water	8260B	
280-1737-E-2 MS	Matrix Spike	T	Water	8260B	
280-1737-E-2 MSD	Matrix Spike Duplicate	T	Water	8260B	

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:280-9015					
LCS 280-9015/4	Lab Control Sample	T	Water	8260B	
LCSD 280-9015/5	Lab Control Sample Duplicate	T	Water	8260B	
MB 280-9015/6	Method Blank	T	Water	8260B	
280-1536-H-2 MS	Matrix Spike	T	Water	8260B	
280-1536-F-2 MSD	Matrix Spike Duplicate	T	Water	8260B	
280-1555-1	PIN12-0513	T	Water	8260B	
280-1555-2	PIN12-0526	T	Water	8260B	
280-1555-4	PIN12-0537	T	Water	8260B	
280-1555-5	PIN12-0538	T	Water	8260B	
280-1555-7	PIN12-0541	T	Water	8260B	
280-1555-8	PIN12-0542	T	Water	8260B	
280-1555-10	PIN12-0549	T	Water	8260B	
280-1555-11	PIN12-0552-1	T	Water	8260B	
280-1555-12	PIN12-0552-2	T	Water	8260B	
280-1555-13	PIN12-0552-3	T	Water	8260B	
280-1555-14	PIN12-0554B	T	Water	8260B	
280-1555-15	PIN12-0554C	T	Water	8260B	
280-1555-35	PIN99-2882	T	Water	8260B	
280-1555-46	PIN12-S70C	T	Water	8260B	
Analysis Batch:280-9023					
LCS 280-9023/4	Lab Control Sample	T	Water	8260B	
MB 280-9023/5	Method Blank	T	Water	8260B	
280-1555-16	PIN12-0555A	T	Water	8260B	
280-1555-17	PIN12-0555B	T	Water	8260B	
280-1555-18	PIN12-0555C	T	Water	8260B	
280-1555-19	PIN12-0562-1	T	Water	8260B	
280-1555-20	PIN12-0562-2	T	Water	8260B	
280-1555-21	PIN12-0562-3	T	Water	8260B	
280-1555-25	PIN12-0567-1	T	Water	8260B	
280-1555-26	PIN12-0567-2	T	Water	8260B	
280-1555-27	PIN12-0567-3	T	Water	8260B	
280-1555-30	PIN12-0568-3	T	Water	8260B	
280-1555-36	PIN99-2890	T	Water	8260B	
280-1555-37	PIN99-2901	T	Water	8260B	
280-1555-42	PIN12-S68B	T	Water	8260B	
280-1555-44DL	PIN12-S68D	T	Water	8260B	
280-1555-47	PIN12-S72B	T	Water	8260B	
280-1555-48	PIN12-S72C	T	Water	8260B	
280-1555-48MS	Matrix Spike	T	Water	8260B	
280-1555-48MSD	Matrix Spike Duplicate	T	Water	8260B	
280-1555-49	PIN12-S72D	T	Water	8260B	
280-1555-50	PIN12-S73C	T	Water	8260B	
280-1555-51	PIN12-S73D	T	Water	8260B	

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:280-9155					
LCS 280-9155/4	Lab Control Sample	T	Water	8260B	
MB 280-9155/5	Method Blank	T	Water	8260B	
280-1555-3	PIN12-0529	T	Water	8260B	
280-1555-3MS	Matrix Spike	T	Water	8260B	
280-1555-3MSD	Matrix Spike Duplicate	T	Water	8260B	
280-1555-6	PIN12-0539	T	Water	8260B	
280-1555-9	PIN12-0547	T	Water	8260B	
280-1555-15DL	PIN12-0554C	T	Water	8260B	
280-1555-22	PIN12-0563-1	T	Water	8260B	
280-1555-23	PIN12-0563-2	T	Water	8260B	
280-1555-24	PIN12-0563-3	T	Water	8260B	
280-1555-38	PIN99-2902	T	Water	8260B	
280-1555-39	PIN12-S67B	T	Water	8260B	
280-1555-39DL	PIN12-S67B	T	Water	8260B	
280-1555-40	PIN12-S67C	T	Water	8260B	
280-1555-40DL	PIN12-S67C	T	Water	8260B	
280-1555-41	PIN12-S67D	T	Water	8260B	

Report Basis
T = Total

Metals					
Prep Batch: 280-8172					
LCS 280-8172/2-A	Lab Control Sample	T	Water	3010A	
MB 280-8172/1-A	Method Blank	T	Water	3010A	
280-1555-37	PIN99-2901	T	Water	3010A	
280-1555-37MS	Matrix Spike	T	Water	3010A	
280-1555-37MSD	Matrix Spike Duplicate	T	Water	3010A	
280-1555-38	PIN99-2902	T	Water	3010A	
Analysis Batch:280-8382					
LCS 280-8172/2-A	Lab Control Sample	T	Water	6010B	280-8172
MB 280-8172/1-A	Method Blank	T	Water	6010B	280-8172
280-1555-37	PIN99-2901	T	Water	6010B	280-8172
280-1555-37MS	Matrix Spike	T	Water	6010B	280-8172
280-1555-37MSD	Matrix Spike Duplicate	T	Water	6010B	280-8172
280-1555-38	PIN99-2902	T	Water	6010B	280-8172

Report Basis
T = Total

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1555-1 **Client ID:** PIN12-0513

Sample Date/Time: 03/16/2010 09:25 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-1		280-9015		03/29/2010 14:39	1	TAL DEN	SAT
A:8260B	280-1555-A-1		280-9015		03/29/2010 14:39	1	TAL DEN	SAT

Lab ID: 280-1555-2 **Client ID:** PIN12-0526

Sample Date/Time: 03/16/2010 10:25 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-2		280-9015		03/29/2010 15:00	1	TAL DEN	SAT
A:8260B	280-1555-A-2		280-9015		03/29/2010 15:00	1	TAL DEN	SAT

Lab ID: 280-1555-3 **Client ID:** PIN12-0529

Sample Date/Time: 03/17/2010 10:57 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-3		280-9155		03/30/2010 11:12	1	TAL DEN	HEW
A:8260B	280-1555-A-3		280-9155		03/30/2010 11:12	1	TAL DEN	HEW
P:5030B	280-1555-B-3		280-8507		03/23/2010 14:38	1	TAL DEN	AEW
A:8260B SIM	280-1555-B-3		280-8507		03/23/2010 14:38	1	TAL DEN	AEW

Lab ID: 280-1555-3 MS **Client ID:** PIN12-0529

Sample Date/Time: 03/17/2010 10:57 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-3 MS		280-9155		03/30/2010 11:33	1	TAL DEN	HEW
A:8260B	280-1555-A-3 MS		280-9155		03/30/2010 11:33	1	TAL DEN	HEW
P:5030B	280-1555-C-3 MS		280-8507		03/23/2010 15:03	1	TAL DEN	AEW
A:8260B SIM	280-1555-C-3 MS		280-8507		03/23/2010 15:03	1	TAL DEN	AEW

Lab ID: 280-1555-3 MSD **Client ID:** PIN12-0529

Sample Date/Time: 03/17/2010 10:57 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-D-3 MSD		280-9155		03/30/2010 11:53	1	TAL DEN	HEW
A:8260B	280-1555-D-3 MSD		280-9155		03/30/2010 11:53	1	TAL DEN	HEW
P:5030B	280-1555-C-3 MSD		280-8507		03/23/2010 15:28	1	TAL DEN	AEW
A:8260B SIM	280-1555-C-3 MSD		280-8507		03/23/2010 15:28	1	TAL DEN	AEW

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1555-4 **Client ID:** PIN12-0537

Sample Date/Time: 03/16/2010 15:40 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-4		280-9015		03/29/2010 19:27	1	TAL DEN	SAT
A:8260B	280-1555-C-4		280-9015		03/29/2010 19:27	1	TAL DEN	SAT

Lab ID: 280-1555-5 **Client ID:** PIN12-0538

Sample Date/Time: 03/16/2010 16:07 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-5		280-9015		03/29/2010 16:01	1	TAL DEN	SAT
A:8260B	280-1555-A-5		280-9015		03/29/2010 16:01	1	TAL DEN	SAT

Lab ID: 280-1555-6 **Client ID:** PIN12-0539

Sample Date/Time: 03/17/2010 11:02 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-6		280-9155		03/30/2010 12:34	1	TAL DEN	HEW
A:8260B	280-1555-B-6		280-9155		03/30/2010 12:34	1	TAL DEN	HEW
P:5030B	280-1555-D-6		280-8507		03/23/2010 15:54	1	TAL DEN	AEW
A:8260B SIM	280-1555-D-6		280-8507		03/23/2010 15:54	1	TAL DEN	AEW

Lab ID: 280-1555-7 **Client ID:** PIN12-0541

Sample Date/Time: 03/16/2010 13:45 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-7		280-9015		03/29/2010 16:22	1	TAL DEN	SAT
A:8260B	280-1555-A-7		280-9015		03/29/2010 16:22	1	TAL DEN	SAT

Lab ID: 280-1555-8 **Client ID:** PIN12-0542

Sample Date/Time: 03/16/2010 14:13 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-8		280-9015		03/29/2010 16:42	1	TAL DEN	SAT
A:8260B	280-1555-B-8		280-9015		03/29/2010 16:42	1	TAL DEN	SAT

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1555-9 **Client ID:** PIN12-0547

Sample Date/Time: 03/17/2010 10:12 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-9		280-9155		03/30/2010 12:55	1	TAL DEN	HEW
A:8260B	280-1555-C-9		280-9155		03/30/2010 12:55	1	TAL DEN	HEW

Lab ID: 280-1555-10 **Client ID:** PIN12-0549

Sample Date/Time: 03/16/2010 15:00 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-10		280-9015		03/29/2010 17:03	1	TAL DEN	SAT
A:8260B	280-1555-A-10		280-9015		03/29/2010 17:03	1	TAL DEN	SAT

Lab ID: 280-1555-11 **Client ID:** PIN12-0552-1

Sample Date/Time: 03/16/2010 13:40 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-11		280-9015		03/29/2010 17:24	1	TAL DEN	SAT
A:8260B	280-1555-B-11		280-9015		03/29/2010 17:24	1	TAL DEN	SAT
P:5030B	280-1555-A-11		280-8507		03/23/2010 16:19	1	TAL DEN	AEW
A:8260B SIM	280-1555-A-11		280-8507		03/23/2010 16:19	1	TAL DEN	AEW

Lab ID: 280-1555-12 **Client ID:** PIN12-0552-2

Sample Date/Time: 03/16/2010 14:38 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-12		280-9015		03/29/2010 17:44	1	TAL DEN	SAT
A:8260B	280-1555-C-12		280-9015		03/29/2010 17:44	1	TAL DEN	SAT
P:5030B	280-1555-A-12		280-8507		03/23/2010 16:45	1	TAL DEN	AEW
A:8260B SIM	280-1555-A-12		280-8507		03/23/2010 16:45	1	TAL DEN	AEW

Lab ID: 280-1555-13 **Client ID:** PIN12-0552-3

Sample Date/Time: 03/16/2010 15:09 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-13		280-9015		03/29/2010 18:05	1	TAL DEN	SAT
A:8260B	280-1555-B-13		280-9015		03/29/2010 18:05	1	TAL DEN	SAT
P:5030B	280-1555-A-13		280-8507		03/23/2010 17:10	1	TAL DEN	AEW
A:8260B SIM	280-1555-A-13		280-8507		03/23/2010 17:10	1	TAL DEN	AEW

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1555-14 **Client ID:** PIN12-0554B

Sample Date/Time: 03/16/2010 10:45 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-14		280-9015		03/29/2010 18:26	1	TAL DEN	SAT
A:8260B	280-1555-A-14		280-9015		03/29/2010 18:26	1	TAL DEN	SAT

Lab ID: 280-1555-15 **Client ID:** PIN12-0554C

Sample Date/Time: 03/16/2010 11:18 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-15		280-9015		03/29/2010 18:46	1	TAL DEN	SAT
A:8260B	280-1555-A-15		280-9015		03/29/2010 18:46	1	TAL DEN	SAT
P:5030B	280-1555-B-15	DL	280-9155		03/30/2010 10:52	1	TAL DEN	HEW
A:8260B	280-1555-B-15	DL	280-9155		03/30/2010 10:52	1	TAL DEN	HEW

Lab ID: 280-1555-16 **Client ID:** PIN12-0555A

Sample Date/Time: 03/16/2010 09:55 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-16		280-9023		03/29/2010 12:59	1	TAL DEN	HEW
A:8260B	280-1555-C-16		280-9023		03/29/2010 12:59	1	TAL DEN	HEW

Lab ID: 280-1555-17 **Client ID:** PIN12-0555B

Sample Date/Time: 03/16/2010 09:25 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-17		280-9023		03/29/2010 14:41	1	TAL DEN	HEW
A:8260B	280-1555-B-17		280-9023		03/29/2010 14:41	1	TAL DEN	HEW

Lab ID: 280-1555-18 **Client ID:** PIN12-0555C

Sample Date/Time: 03/16/2010 08:45 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-18		280-9023		03/29/2010 20:47	1	TAL DEN	HEW
A:8260B	280-1555-B-18		280-9023		03/29/2010 20:47	1	TAL DEN	HEW

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1555-19 **Client ID:** PIN12-0562-1

Sample Date/Time: 03/16/2010 11:20 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-19		280-9023		03/29/2010 15:41	1	TAL DEN	HEW
A:8260B	280-1555-C-19		280-9023		03/29/2010 15:41	1	TAL DEN	HEW

Lab ID: 280-1555-20 **Client ID:** PIN12-0562-2

Sample Date/Time: 03/16/2010 15:56 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-20		280-9023		03/29/2010 16:02	1	TAL DEN	HEW
A:8260B	280-1555-A-20		280-9023		03/29/2010 16:02	1	TAL DEN	HEW

Lab ID: 280-1555-21 **Client ID:** PIN12-0562-3

Sample Date/Time: 03/16/2010 16:32 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-21		280-9023		03/29/2010 16:22	1	TAL DEN	HEW
A:8260B	280-1555-B-21		280-9023		03/29/2010 16:22	1	TAL DEN	HEW

Lab ID: 280-1555-22 **Client ID:** PIN12-0563-1

Sample Date/Time: 03/17/2010 09:03 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-22		280-9155		03/30/2010 13:18	1	TAL DEN	HEW
A:8260B	280-1555-B-22		280-9155		03/30/2010 13:18	1	TAL DEN	HEW

Lab ID: 280-1555-23 **Client ID:** PIN12-0563-2

Sample Date/Time: 03/17/2010 09:31 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-23		280-9155		03/30/2010 13:38	1	TAL DEN	HEW
A:8260B	280-1555-B-23		280-9155		03/30/2010 13:38	1	TAL DEN	HEW

Lab ID: 280-1555-24 **Client ID:** PIN12-0563-3

Sample Date/Time: 03/17/2010 09:56 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-24		280-9155		03/30/2010 13:59	1	TAL DEN	HEW
A:8260B	280-1555-B-24		280-9155		03/30/2010 13:59	1	TAL DEN	HEW

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG: 10022892

Laboratory Chronicle

Lab ID: **280-1555-25** **Client ID:** **PIN12-0567-1**

Sample Date/Time: 03/16/2010 09:21 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-25		280-9023		03/29/2010 16:42	1	TAL DEN	HEW
A:8260B	280-1555-B-25		280-9023		03/29/2010 16:42	1	TAL DEN	HEW
P:5030B	280-1555-A-25		280-8507		03/23/2010 17:36	1	TAL DEN	AEW
A:8260B SIM	280-1555-A-25		280-8507		03/23/2010 17:36	1	TAL DEN	AEW

Lab ID: **280-1555-26** **Client ID:** **PIN12-0567-2**

Sample Date/Time: 03/16/2010 09:47 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-26		280-9023		03/29/2010 17:03	1	TAL DEN	HEW
A:8260B	280-1555-B-26		280-9023		03/29/2010 17:03	1	TAL DEN	HEW
P:5030B	280-1555-C-26		280-8507		03/23/2010 18:01	1	TAL DEN	AEW
A:8260B SIM	280-1555-C-26		280-8507		03/23/2010 18:01	1	TAL DEN	AEW

Lab ID: **280-1555-27** **Client ID:** **PIN12-0567-3**

Sample Date/Time: 03/16/2010 10:39 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-27		280-9023		03/29/2010 17:23	1	TAL DEN	HEW
A:8260B	280-1555-B-27		280-9023		03/29/2010 17:23	1	TAL DEN	HEW
P:5030B	280-1555-A-27		280-8507		03/23/2010 18:26	1	TAL DEN	AEW
A:8260B SIM	280-1555-A-27		280-8507		03/23/2010 18:26	1	TAL DEN	AEW

Lab ID: **280-1555-28** **Client ID:** **PIN12-0568-1**

Sample Date/Time: 03/15/2010 15:09 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-28		280-8984		03/27/2010 02:56	1	TAL DEN	DPI
A:8260B	280-1555-C-28		280-8984		03/27/2010 02:56	1	TAL DEN	DPI
P:5030B	280-1555-A-28		280-8507		03/23/2010 18:52	1	TAL DEN	AEW
A:8260B SIM	280-1555-A-28		280-8507		03/23/2010 18:52	1	TAL DEN	AEW

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1555-29 **Client ID:** PIN12-0568-2

Sample Date/Time: 03/15/2010 15:44 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-29		280-8984		03/27/2010 03:16	1	TAL DEN	DPI
A:8260B	280-1555-B-29		280-8984		03/27/2010 03:16	1	TAL DEN	DPI
P:5030B	280-1555-A-29		280-8507		03/23/2010 19:17	1	TAL DEN	AEW
A:8260B SIM	280-1555-A-29		280-8507		03/23/2010 19:17	1	TAL DEN	AEW

Lab ID: 280-1555-30 **Client ID:** PIN12-0568-3

Sample Date/Time: 03/16/2010 08:47 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-30		280-9023		03/29/2010 17:44	1	TAL DEN	HEW
A:8260B	280-1555-C-30		280-9023		03/29/2010 17:44	1	TAL DEN	HEW
P:5030B	280-1555-A-30		280-8660		03/25/2010 11:24	1	TAL DEN	AEW
A:8260B SIM	280-1555-A-30		280-8660		03/25/2010 11:24	1	TAL DEN	AEW

Lab ID: 280-1555-30 MS **Client ID:** PIN12-0568-3

Sample Date/Time: 03/16/2010 08:47 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-D-30 MS		280-8660		03/25/2010 11:50	1	TAL DEN	AEW
A:8260B SIM	280-1555-D-30 MS		280-8660		03/25/2010 11:50	1	TAL DEN	AEW

Lab ID: 280-1555-30 MSD **Client ID:** PIN12-0568-3

Sample Date/Time: 03/16/2010 08:47 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-D-30 MSD		280-8660		03/25/2010 12:15	1	TAL DEN	AEW
A:8260B SIM	280-1555-D-30 MSD		280-8660		03/25/2010 12:15	1	TAL DEN	AEW

Lab ID: 280-1555-31 **Client ID:** PIN12-0569-2

Sample Date/Time: 03/15/2010 13:42 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-31		280-8984		03/27/2010 04:38	1	TAL DEN	DPI
A:8260B	280-1555-A-31		280-8984		03/27/2010 04:38	1	TAL DEN	DPI
P:5030B	280-1555-A-31	DL	280-8984		03/27/2010 04:59	1	TAL DEN	DPI
A:8260B	280-1555-A-31	DL	280-8984		03/27/2010 04:59	1	TAL DEN	DPI
P:5030B	280-1555-B-31		280-8660		03/25/2010 12:40	1	TAL DEN	AEW
A:8260B SIM	280-1555-B-31		280-8660		03/25/2010 12:40	1	TAL DEN	AEW

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1555-32 **Client ID:** PIN12-0569-3

Sample Date/Time: 03/15/2010 14:20 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-32		280-8984		03/27/2010 05:20	1	TAL DEN	DPI
A:8260B	280-1555-B-32		280-8984		03/27/2010 05:20	1	TAL DEN	DPI
P:5030B	280-1555-B-32	DL	280-8984		03/27/2010 05:41	1	TAL DEN	DPI
A:8260B	280-1555-B-32	DL	280-8984		03/27/2010 05:41	1	TAL DEN	DPI
P:5030B	280-1555-A-32		280-8660		03/25/2010 13:05	1	TAL DEN	AEW
A:8260B SIM	280-1555-A-32		280-8660		03/25/2010 13:05	1	TAL DEN	AEW

Lab ID: 280-1555-33 **Client ID:** PIN12-2872

Sample Date/Time: 03/15/2010 12:00 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-33		280-8984		03/27/2010 06:01	1	TAL DEN	DPI
A:8260B	280-1555-B-33		280-8984		03/27/2010 06:01	1	TAL DEN	DPI
P:5030B	280-1555-B-33	DL	280-8984		03/27/2010 06:22	1	TAL DEN	DPI
A:8260B	280-1555-B-33	DL	280-8984		03/27/2010 06:22	1	TAL DEN	DPI
P:5030B	280-1555-A-33		280-8660		03/25/2010 13:31	1	TAL DEN	AEW
A:8260B SIM	280-1555-A-33		280-8660		03/25/2010 13:31	1	TAL DEN	AEW

Lab ID: 280-1555-34 **Client ID:** PIN99-2876

Sample Date/Time: 03/15/2010 08:00 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-34		280-8984		03/27/2010 03:36	1	TAL DEN	DPI
A:8260B	280-1555-B-34		280-8984		03/27/2010 03:36	1	TAL DEN	DPI

Lab ID: 280-1555-35 **Client ID:** PIN99-2882

Sample Date/Time: 03/15/2010 12:00 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-35		280-9015		03/29/2010 13:58	1	TAL DEN	SAT
A:8260B	280-1555-A-35		280-9015		03/29/2010 13:58	1	TAL DEN	SAT

Lab ID: 280-1555-36 **Client ID:** PIN99-2890

Sample Date/Time: 03/16/2010 08:12 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-36		280-9023		03/29/2010 18:04	1	TAL DEN	HEW
A:8260B	280-1555-C-36		280-9023		03/29/2010 18:04	1	TAL DEN	HEW

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1555-37 **Client ID:** PIN99-2901

Sample Date/Time: 03/16/2010 07:30 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-37		280-9023		03/29/2010 18:25	1	TAL DEN	HEW
A:8260B	280-1555-C-37		280-9023		03/29/2010 18:25	1	TAL DEN	HEW
P:5030B	280-1555-E-37		280-8660		03/25/2010 13:56	1	TAL DEN	AEW
A:8260B SIM	280-1555-E-37		280-8660		03/25/2010 13:56	1	TAL DEN	AEW
P:3010A	280-1555-A-37-A		280-8382	280-8172	03/23/2010 07:30	1	TAL DEN	JW
A:6010B	280-1555-A-37-A		280-8382	280-8172	03/23/2010 20:11	1	TAL DEN	DW

Lab ID: 280-1555-37 **Client ID:** PIN99-2901

Sample Date/Time: 03/16/2010 07:30 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:3010A	280-1555-A-37-B MS		280-8382	280-8172	03/23/2010 07:30	1	TAL DEN	JW
A:6010B	280-1555-A-37-B MS		280-8382	280-8172	03/23/2010 20:15	1	TAL DEN	DW

Lab ID: 280-1555-37 **Client ID:** PIN99-2901

Sample Date/Time: 03/16/2010 07:30 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:3010A	280-1555-A-37-C MSD		280-8382	280-8172	03/23/2010 07:30	1	TAL DEN	JW
A:6010B	280-1555-A-37-C MSD		280-8382	280-8172	03/23/2010 20:18	1	TAL DEN	DW

Lab ID: 280-1555-38 **Client ID:** PIN99-2902

Sample Date/Time: 03/17/2010 07:30 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-D-38		280-9155		03/30/2010 14:20	1	TAL DEN	HEW
A:8260B	280-1555-D-38		280-9155		03/30/2010 14:20	1	TAL DEN	HEW
P:5030B	280-1555-C-38		280-8660		03/25/2010 14:22	1	TAL DEN	AEW
A:8260B SIM	280-1555-C-38		280-8660		03/25/2010 14:22	1	TAL DEN	AEW
P:3010A	280-1555-A-38-A		280-8382	280-8172	03/23/2010 07:30	1	TAL DEN	JW
A:6010B	280-1555-A-38-A		280-8382	280-8172	03/23/2010 20:20	1	TAL DEN	DW

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1555-39 **Client ID:** PIN12-S67B

Sample Date/Time: 03/17/2010 10:55 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-39		280-9155		03/30/2010 14:40	1	TAL DEN	HEW
A:8260B	280-1555-B-39		280-9155		03/30/2010 14:40	1	TAL DEN	HEW
P:5030B	280-1555-B-39	DL	280-9155		03/30/2010 15:01	1	TAL DEN	HEW
A:8260B	280-1555-B-39	DL	280-9155		03/30/2010 15:01	1	TAL DEN	HEW

Lab ID: 280-1555-40 **Client ID:** PIN12-S67C

Sample Date/Time: 03/17/2010 10:10 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-40		280-9155		03/30/2010 15:22	1	TAL DEN	HEW
A:8260B	280-1555-B-40		280-9155		03/30/2010 15:22	1	TAL DEN	HEW
P:5030B	280-1555-B-40	DL	280-9155		03/30/2010 15:43	1	TAL DEN	HEW
A:8260B	280-1555-B-40	DL	280-9155		03/30/2010 15:43	1	TAL DEN	HEW

Lab ID: 280-1555-41 **Client ID:** PIN12-S67D

Sample Date/Time: 03/17/2010 09:05 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-41		280-9155		03/30/2010 16:03	1	TAL DEN	HEW
A:8260B	280-1555-C-41		280-9155		03/30/2010 16:03	1	TAL DEN	HEW

Lab ID: 280-1555-42 **Client ID:** PIN12-S68B

Sample Date/Time: 03/16/2010 08:40 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-42		280-9023		03/29/2010 18:45	1	TAL DEN	HEW
A:8260B	280-1555-C-42		280-9023		03/29/2010 18:45	1	TAL DEN	HEW

Lab ID: 280-1555-43 **Client ID:** PIN12-S68C

Sample Date/Time: 03/15/2010 16:30 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-43		280-8984		03/27/2010 03:57	1	TAL DEN	DPI
A:8260B	280-1555-B-43		280-8984		03/27/2010 03:57	1	TAL DEN	DPI

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1555-44 **Client ID:** PIN12-S68D

Sample Date/Time: 03/15/2010 15:50 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-44		280-8984		03/27/2010 06:43	1	TAL DEN	DPI
A:8260B	280-1555-B-44		280-8984		03/27/2010 06:43	1	TAL DEN	DPI
P:5030B	280-1555-C-44	DL	280-9023		03/29/2010 15:21	1	TAL DEN	HEW
A:8260B	280-1555-C-44	DL	280-9023		03/29/2010 15:21	1	TAL DEN	HEW

Lab ID: 280-1555-45 **Client ID:** PIN12-S70B

Sample Date/Time: 03/15/2010 14:45 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-45		280-8984		03/27/2010 04:18	1	TAL DEN	DPI
A:8260B	280-1555-A-45		280-8984		03/27/2010 04:18	1	TAL DEN	DPI

Lab ID: 280-1555-46 **Client ID:** PIN12-S70C

Sample Date/Time: 03/15/2010 13:45 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-46		280-9015		03/29/2010 14:19	1	TAL DEN	SAT
A:8260B	280-1555-A-46		280-9015		03/29/2010 14:19	1	TAL DEN	SAT

Lab ID: 280-1555-47 **Client ID:** PIN12-S72B

Sample Date/Time: 03/16/2010 14:30 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-47		280-9023		03/29/2010 19:05	1	TAL DEN	HEW
A:8260B	280-1555-B-47		280-9023		03/29/2010 19:05	1	TAL DEN	HEW

Lab ID: 280-1555-48 **Client ID:** PIN12-S72C

Sample Date/Time: 03/16/2010 13:45 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-48		280-9023		03/29/2010 13:19	1	TAL DEN	HEW
A:8260B	280-1555-C-48		280-9023		03/29/2010 13:19	1	TAL DEN	HEW

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG: 10022892

Laboratory Chronicle

Lab ID: **280-1555-48 MS** **Client ID:** **PIN12-S72C**

Sample Date/Time: 03/16/2010 13:45 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-48 MS		280-9023		03/29/2010 13:40	1	TAL DEN	HEW
A:8260B	280-1555-C-48 MS		280-9023		03/29/2010 13:40	1	TAL DEN	HEW

Lab ID: **280-1555-48 MSD** **Client ID:** **PIN12-S72C**

Sample Date/Time: 03/16/2010 13:45 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-48 MSD		280-9023		03/29/2010 14:00	1	TAL DEN	HEW
A:8260B	280-1555-A-48 MSD		280-9023		03/29/2010 14:00	1	TAL DEN	HEW

Lab ID: **280-1555-49** **Client ID:** **PIN12-S72D**

Sample Date/Time: 03/16/2010 11:05 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-B-49		280-9023		03/29/2010 19:25	1	TAL DEN	HEW
A:8260B	280-1555-B-49		280-9023		03/29/2010 19:25	1	TAL DEN	HEW

Lab ID: **280-1555-50** **Client ID:** **PIN12-S73C**

Sample Date/Time: 03/16/2010 16:10 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-50		280-9023		03/29/2010 19:46	1	TAL DEN	HEW
A:8260B	280-1555-C-50		280-9023		03/29/2010 19:46	1	TAL DEN	HEW

Lab ID: **280-1555-51** **Client ID:** **PIN12-S73D**

Sample Date/Time: 03/16/2010 15:20 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-C-51		280-9023		03/29/2010 20:06	1	TAL DEN	HEW
A:8260B	280-1555-C-51		280-9023		03/29/2010 20:06	1	TAL DEN	HEW

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG: 10022892

Laboratory Chronicle

Lab ID: **MB** **Client ID:** **N/A**

Sample Date/Time: N/A Received Date/Time: N/A

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	MB 280-8984/6		280-8984		03/26/2010 21:05	1	TAL DEN	DPI
A:8260B	MB 280-8984/6		280-8984		03/26/2010 21:05	1	TAL DEN	DPI
P:5030B	MB 280-9015/6		280-9015		03/29/2010 10:35	1	TAL DEN	SAT
A:8260B	MB 280-9015/6		280-9015		03/29/2010 10:35	1	TAL DEN	SAT
P:5030B	MB 280-9023/5		280-9023		03/29/2010 12:12	1	TAL DEN	HEW
A:8260B	MB 280-9023/5		280-9023		03/29/2010 12:12	1	TAL DEN	HEW
P:5030B	MB 280-9155/5		280-9155		03/30/2010 10:31	1	TAL DEN	HEW
A:8260B	MB 280-9155/5		280-9155		03/30/2010 10:31	1	TAL DEN	HEW
P:5030B	MB 280-8507/4		280-8507		03/23/2010 09:57	1	TAL DEN	AEW
A:8260B SIM	MB 280-8507/4		280-8507		03/23/2010 09:57	1	TAL DEN	AEW
P:5030B	MB 280-8660/4		280-8660		03/25/2010 10:59	1	TAL DEN	AEW
A:8260B SIM	MB 280-8660/4		280-8660		03/25/2010 10:59	1	TAL DEN	AEW
P:3010A	MB 280-8172/1-A		280-8382	280-8172	03/23/2010 07:30	1	TAL DEN	JW
A:6010B	MB 280-8172/1-A		280-8382	280-8172	03/23/2010 20:04	1	TAL DEN	DW

Lab ID: **LCS** **Client ID:** **N/A**

Sample Date/Time: N/A Received Date/Time: N/A

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	LCS 280-8984/4		280-8984		03/26/2010 20:02	1	TAL DEN	DPI
A:8260B	LCS 280-8984/4		280-8984		03/26/2010 20:02	1	TAL DEN	DPI
P:5030B	LCS 280-9015/4		280-9015		03/29/2010 09:54	1	TAL DEN	SAT
A:8260B	LCS 280-9015/4		280-9015		03/29/2010 09:54	1	TAL DEN	SAT
P:5030B	LCS 280-9023/4		280-9023		03/29/2010 11:11	1	TAL DEN	HEW
A:8260B	LCS 280-9023/4		280-9023		03/29/2010 11:11	1	TAL DEN	HEW
P:5030B	LCS 280-9155/4		280-9155		03/30/2010 09:29	1	TAL DEN	HEW
A:8260B	LCS 280-9155/4		280-9155		03/30/2010 09:29	1	TAL DEN	HEW
P:5030B	LCS 280-8507/3		280-8507		03/23/2010 09:06	1	TAL DEN	AEW
A:8260B SIM	LCS 280-8507/3		280-8507		03/23/2010 09:06	1	TAL DEN	AEW
P:5030B	LCS 280-8660/3		280-8660		03/25/2010 09:40	1	TAL DEN	AEW
A:8260B SIM	LCS 280-8660/3		280-8660		03/25/2010 09:40	1	TAL DEN	AEW
P:3010A	LCS 280-8172/2-A		280-8382	280-8172	03/23/2010 07:30	1	TAL DEN	JW
A:6010B	LCS 280-8172/2-A		280-8382	280-8172	03/23/2010 20:06	1	TAL DEN	DW

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG: 10022892

Laboratory Chronicle

Lab ID: **LCSD** **Client ID:** **N/A**

Sample Date/Time: N/A Received Date/Time: N/A

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	LCSD 280-8984/5		280-8984		03/26/2010 20:23	1	TAL DEN	DPI
A:8260B	LCSD 280-8984/5		280-8984		03/26/2010 20:23	1	TAL DEN	DPI
P:5030B	LCSD 280-9015/5		280-9015		03/29/2010 10:15	1	TAL DEN	SAT
A:8260B	LCSD 280-9015/5		280-9015		03/29/2010 10:15	1	TAL DEN	SAT

Lab ID: **MS** **Client ID:** **N/A**

Sample Date/Time: 03/25/2010 15:30 Received Date/Time: 03/26/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1737-E-2 MS		280-8984		03/26/2010 21:46	1	TAL DEN	DPI
A:8260B	280-1737-E-2 MS		280-8984		03/26/2010 21:46	1	TAL DEN	DPI
P:5030B	280-1536-H-2 MS		280-9015		03/29/2010 12:36	1	TAL DEN	SAT
A:8260B	280-1536-H-2 MS		280-9015		03/29/2010 12:36	1	TAL DEN	SAT

Lab ID: **MSD** **Client ID:** **N/A**

Sample Date/Time: 03/25/2010 15:30 Received Date/Time: 03/26/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1737-E-2 MSD		280-8984		03/26/2010 22:07	1	TAL DEN	DPI
A:8260B	280-1737-E-2 MSD		280-8984		03/26/2010 22:07	1	TAL DEN	DPI
P:5030B	280-1536-F-2 MSD		280-9015		03/29/2010 12:57	1	TAL DEN	SAT
A:8260B	280-1536-F-2 MSD		280-9015		03/29/2010 12:57	1	TAL DEN	SAT

Lab References:
TAL DEN = TestAmerica Denver

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1555-1

SDG No.: 10022892

Instrument ID: MSV_R1 Analysis Batch Number: 8905

Lab Sample ID: IC 280-8905/2 Client Sample ID:

Date Analyzed: 03/26/10 07:45 Lab File ID: R7591.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
trans-1,3-Dichloropropene	9.34	Analyte not Identified by the Data System	tinkhams	03/26/10 11:59

Lab Sample ID: IC 280-8905/4 Client Sample ID:

Date Analyzed: 03/26/10 08:27 Lab File ID: R7593.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Ethanol	5.33	Analyte not Identified by the Data System	tinkhams	03/26/10 11:58

Lab Sample ID: IC 280-8905/10 Client Sample ID:

Date Analyzed: 03/26/10 10:39 Lab File ID: R7599.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Chloroethyl vinyl ether	8.79	Analyte not Identified by the Data System	waterlah	03/30/10 08:37

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1555-1

SDG No.: 10022892

Instrument ID: MSV_R2 Analysis Batch Number: 8473

Lab Sample ID: IC 280-8473/5 Client Sample ID:

Date Analyzed: 03/24/10 09:17 Lab File ID: RR4926.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Acetonitrile	4.57	Split Peak	waterlah	03/24/10 10:00

8260B

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1555-1

SDG No.: 10022892

Instrument ID: MSV_R1 Analysis Batch Number: 8905

Lab Sample ID: IC 280-8905/2 Client Sample ID:

Date Analyzed: 03/26/10 07:45 Lab File ID: R7591.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
trans-1,3-Dichloropropene	9.34	Analyte not Identified by the Data System	tinkhams	03/26/10 11:59

Lab Sample ID: IC 280-8905/4 Client Sample ID:

Date Analyzed: 03/26/10 08:27 Lab File ID: R7593.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Ethanol	5.33	Analyte not Identified by the Data System	tinkhams	03/26/10 11:58

Lab Sample ID: IC 280-8905/10 Client Sample ID:

Date Analyzed: 03/26/10 10:39 Lab File ID: R7599.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Chloroethyl vinyl ether	8.79	Analyte not Identified by the Data System	waterlah	03/30/10 08:37

wp 4/9

8260B

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1555-1

SDG No.: 10022892

Instrument ID: MSV_R2 Analysis Batch Number: 8473

Lab Sample ID: IC 280-8473/5 Client Sample ID:

Date Analyzed: 03/24/10 09:17 Lab File ID: RR4926.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Acetonitrile	4.57	Split Peak	waterlah	03/24/10 10:00

4/9

8260B

Shipping and Receiving Documents

Stoller
Legacy Management Team

Chain of Custody / Sample Submittal Form

RIN: 10022892
COC: 10022892.4.1
Sampler(s): baer, lombardi, caballero, atkinson, walters, moe

Project: Pinellas Monitoring
Purchase Order: 3864
Cost Number: 1-502-1-06-509-4-02
Turnaround (Days): 28
Matrix: WA - Water

Laboratory: TestAmerica Denver
Address: 4955 Yarrow Street
Arvada, Colorado 80002
Phone: 303 736 0100

Ship #	Ticket	Sample Date	Time	Site	Location	Container	# Cont.	Preservation	Matrix	Comp.	Grab	Filtered	QC	Analysis
4	IDW 098	03/16/2010	9:25	PIN12	PIN12-0513	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 106	03/16/2010	10:25	PIN12	PIN12-0526	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 109	03/17/2010	10:57	PIN12	PIN12-0529	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
4	IDW 117	03/16/2010	15:40	PIN12	PIN12-0537	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 118	03/16/2010	16:07	PIN12	PIN12-0538	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 119	03/17/2010	11:02	PIN12	PIN12-0539	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
4	IDW 121	03/16/2010	13:45	PIN12	PIN12-0541	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 122	03/16/2010	14:13	PIN12	PIN12-0542	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 127	03/17/2010	10:12	PIN12	PIN12-0547	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 129	03/16/2010	15:00	PIN12	PIN12-0549	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 136	03/16/2010	13:40	PIN12	PIN12-0552-1	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
4	IDW 137	03/16/2010	14:38	PIN12	PIN12-0552-2	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
4	IDW 138	03/16/2010	15:09	PIN12	PIN12-0552-3	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
4	IDW 143	03/16/2010	10:45	PIN12	PIN12-0554B	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 144	03/16/2010	11:18	PIN12	PIN12-0554C	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 145	03/16/2010	9:55	PIN12	PIN12-0555A	Glass 40 mL	3	4 C, HCl	WA			N		VOA

Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time
[illegible]	3-17-10	1815	[illegible]	3/18/10	1700			
Received by (signature)	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**
[illegible]	3-17-10	1815	[illegible]	3/19/10	0900			

Stoller
Legacy Management Team

0.1, 0.4
89
IRI
3/19/10

Chain of Custody / Sample Submittal Form

RIN: 10022892
COC: 10022892.4.2
Sampler(s): baer, lombardi, caballero, atkinson, walters, moe

Project: Pinellas Monitoring
Purchase Order: 3864
Cost Number: 1-502-1-06-509-4-02
Turnaround (Days): 28
Matrix: WA - Water

Laboratory: TestAmerica Denver
Address: 4955 Yarrow Street
Arvada, Colorado 80002
Phone: 303 736 0100

Ship #	Ticket	Sample Date	Time	Site	Location	Container	# Cont.	Preservation	Matrix	Comp.	Grab	Filtered	QC	Analysis
4	IDW 146	03/16/2010	9:25	PIN12	PIN12-0555B	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 147	03/16/2010	8:45	PIN12	PIN12-0555C	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 157	03/16/2010	11:20	PIN12	PIN12-0562-1	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 158	03/16/2010	15:56	PIN12	PIN12-0562-2	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 159	03/16/2010	16:32	PIN12	PIN12-0562-3	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 160	03/17/2010	9:03	PIN12	PIN12-0563-1	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 161	03/17/2010	9:31	PIN12	PIN12-0563-2	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 162	03/17/2010	9:56	PIN12	PIN12-0563-3	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 172	03/16/2010	9:21	PIN12	PIN12-0567-1	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
4	IDW 173	03/16/2010	9:47	PIN12	PIN12-0567-2	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
4	IDW 174	03/16/2010	10:39	PIN12	PIN12-0567-3	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
4	IDW 175	03/15/2010	15:09	PIN12	PIN12-0568-1	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
4	IDW 176	03/15/2010	15:44	PIN12	PIN12-0568-2	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
4	IDW 177	03/16/2010	8:47	PIN12	PIN12-0568-3	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
4	IDW 179	03/15/2010	13:42	PIN12	PIN12-0569-2	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
4	IDW 180	03/15/2010	14:20	PIN12	PIN12-0569-3	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane

Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time
[signature]	3-17-10	1815	[signature]	3/18/10	1700			
Received by (signature)	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**
[signature]	3-17-10	1815	[signature]	3/19/10	0900			

Stoller
Legacy Management Team

Chain of Custody / Sample Submittal Form

RIN: 10022892
COC: 10022892.4.3
Sampler(s): baer, lombardi, caballero, atkinson, walters, moe

Project: Pinellas Monitoring
Purchase Order: 3864
Cost Number: 1-502-1-06-509-4-02
Turnaround (Days): 28
Matrix: WA - Water

Laboratory: TestAmerica Denver
Address: 4955 Yarrow Street
Arvada, Colorado 80002
Phone: 303 736 0100

Ship #	Ticket	Sample Date	Time	Site	Location	Container	# Cont.	Preservation	Matrix	Comp.	Grab	Filtered	QC	Analysis
4	IDW 240	03/15/2010	12:00	PIN12	PIN12-2872	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
4	IDW 244	03/15/2010	8:00	PIN99	PIN99-2876	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 250	03/15/2010	12:00	PIN99	PIN99-2882	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 258	03/16/2010	8:12	PIN99	PIN99-2890	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IEX 515	03/16/2010	7:30	PIN99	PIN99-2901	250 HDPE #5 / Glass 40 mL	3	4 C, HCl	WA			N		VOA, Dioxane + metals Al, Fe pres. w/ HNO3
4	IEX 516	03/17/2010	7:30	PIN99	PIN99-2902	250 HDPE #5 / Glass 40 mL	3	4 C, HCl	WA			N		VOA, Dioxane + metals, Al, Fe pres. w/ HNO3
4	IDW 185	03/17/2010	10:55	PIN12	PIN12-S67B	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 186	03/17/2010	10:10	PIN12	PIN12-S67C	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 187	03/17/2010	9:05	PIN12	PIN12-S67D	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 188	03/16/2010	8:40	PIN12	PIN12-S68B	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 189	03/15/2010	16:30	PIN12	PIN12-S68C	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 190	03/15/2010	15:50	PIN12	PIN12-S68D	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 194	03/15/2010	14:45	PIN12	PIN12-S70B	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 195	03/15/2010	13:45	PIN12	PIN12-S70C	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 200	03/16/2010	14:30	PIN12	PIN12-S72B	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 201	03/16/2010	13:45	PIN12	PIN12-S72C	Glass 40 mL	3	4 C, HCl	WA			N		VOA

Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time
[illegible]	3-17-10	1815	[illegible]	3/17/10	170			
Received by (signature)	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**
[illegible]	3-17-10	1845	[illegible]	3/19/10	0900			

Stoller
Legacy Management Team

Chain of Custody / Sample Submittal Form

RIN: 10022892
COC: 10022892.4.4
Sampler(s): baer, lombardi, caballero, atkinson, walters, moe

Project: Pinellas Monitoring
Purchase Order: 3864
Cost Number: 1-502-1-06-509-4-02
Turnaround (Days): 28
Matrix: WA - Water

Laboratory: TestAmerica Denver
Address: 4955 Yarrow Street
Arvada, Colorado 80002
Phone: 303 736 0100

Ship #	Ticket	Sample Date	Time	Site	Location	Container	# Cont.	Preservation	Matrix	Comp.	Grab	Filtered	QC	Analysis
4	IDW 202	03/16/2010	11:05	PIN12	PIN12-S72D	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 204	03/16/2010	16:10	PIN12	PIN12-S73C	Glass 40 mL	3	4 C, HCl	WA			N		VOA
4	IDW 205	03/16/2010	15:20	PIN12	PIN12-S73D	Glass 40 mL	3	4 C, HCl	WA			N		VOA

Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time
[signature]	3-17-10	1815	[signature]	3/18/10	1730			
Received by (signature)	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**
[signature]	3-17-10	1815	[signature]	3/19/10	0900			

Login Sample Receipt Check List

Client: S.M. Stoller Corporation

Job Number: 280-1555-1
SDG Number: 10022892

Login Number: 1555
Creator: Green, Angel L
List Number: 1

List Source: TestAmerica Denver

Question	T / F/ NA	Comment
Radioactivity either was not measured or, if measured, is at or below background	True	
The cooler's custody seal, if present, is intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the sample IDs on the containers and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
VOA sample vials do not have headspace or bubble is <6mm (1/4") in diameter.	True	
If necessary, staff have been informed of any short hold time or quick TAT needs	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Is the Field Sampler's name present on COC?	True	
Sample Preservation Verified	True	

This portion can be removed for Recipient's records.

ate 3/18/10 FedEx Tracking Number 871281771841

ender's ame CUSTODY Phone

Company TESTAMERICA TAMPA

Address 6712 BENJAMIN RD STE 100

Dept/Floor/Suite/Room

ty TAMPA State FL ZIP 33634-4403

our Internal Billing Reference

ANALYTICAL REPORT

Job Number: 280-1619-1

SDG Number: 10022892

Job Description: PINELLAS MONITORING

For:
S.M. Stoller Corporation
2597 B 3/4 Road
Grand Junction, CO 81503

Attention: Mr. Steve Donivan

Approved for release.
Kae E Yoder
Project Manager II
4/16/2010 1:39 PM

Kae E Yoder
Project Manager II
kae.yoder@testamericainc.com
04/16/2010

The test results in this report relate only to the samples in this report and meet all requirements of NELAC, with any exceptions noted. Pursuant to NELAP, this report shall not be reproduced except in full, without the written approval of the laboratory. All questions regarding this report should be directed to the TestAmerica Denver Project Manager.

The Lab Certification ID# is E87667.

Reporting limits are adjusted for sample size used, dilutions and moisture content if applicable.

Table of Contents

Cover Title Page 1

Data Summaries 4

Report Narrative 4

Data Qualifiers 5

Executive Summary 6

Method Summary 9

Method / Analyst Summary 10

Sample Summary 11

Sample Datasheets 12

Surrogate Summary 93

QC Data Summary 96

QC Association Summary 118

Lab Chronicle 121

Manual Integration Summary 131

Organic Sample Data 141

GC/MS VOA 141

Method 8260B 141

Method 8260B QC Summary 142

Method 8260B Sample Data 173

Standards Data 448

Method 8260B ICAL Data 448

Method 8260B CCAL Data 662

Raw QC Data 741

Method 8260B Tune Data 741

Method 8260B Blank Data 777

Method 8260B LCS/LCSD Data 804

Table of Contents

Method 8260B MS/MSD Data 822

Method 8260B Run Logs 862

Method 8260B SIM 894

Method 8260B SIM QC Summary 895

Method 8260B SIM Sample Data 904

Standards Data 919

Method 8260B SIM ICAL Data 919

Method 8260B SIM CCAL Data 935

Raw QC Data 941

Method 8260B SIM Tune Data 941

Method 8260B SIM Blank Data 953

Method 8260B SIM LCS/LCSD Data 956

Method 8260B SIM MS/MSD Data 959

Method 8260B SIM Run Logs 965

Shipping and Receiving Documents 973

Client Chain of Custody 974

Sample Receipt Checklist 977

CASE NARRATIVE

Client: S.M. Stoller Corporation

Project: PINELLAS MONITORING - 10022892

Report Number: 280-1619-1

With exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. All laboratory quality control samples analyzed in conjunction with the samples in this project were within established control limits, with any exceptions noted. Calculations are performed before rounding to avoid round-off errors in calculated results.

This report includes reporting limits (RLs) less than TestAmerica Denver's practical quantitation limits. These reporting limits are being used specifically at the client's request to meet the needs of this project. Please note that data are not normally reported to these levels without qualification, since they are inherently less reliable and potentially less defensible than required by the current NELAC standards.

Results between the method detection limit (MDL) and reporting limit (RL) are flagged with a "J" qualifier to indicate an estimated value. These results are statistically less reliable than results greater than or equal to the RL and should be considered a qualitative value.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

RECEIPT

The samples were received on 3/23/2010; the samples arrived in good condition, properly preserved and on ice. The temperature of the cooler at receipt was 3.3 C.

GC/MS VOLATILES - SW846 8260B

Styrene was detected in the method blank associated with batch 280-9155 at a level that was above the method detection limit but below the reporting limit. The value should be considered an estimate, and has been flagged “J”. If the associated samples reported a result above the MDL and/or RL, the result has been “B” flagged.

Methylene chloride, a common laboratory contaminant, was detected in the method blanks associated with batches 280-9293 and 280-9544 at levels that were above the method detection limit but below the reporting limit. The values should be considered estimates, and have been flagged “J”. If the associated samples reported a result above the MDL and/or RL, the result has been “B” flagged.

1,1-Dichloroethene failed the recovery criteria high for the MS associated with batch 280-9155. The LCS was within control limits.

No other anomalies were encountered.

GC/MS VOLATILES - SW846 8260B-SIM

No anomalies were encountered.

DATA REPORTING QUALIFIERS

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Lab Section	Qualifier	Description
GC/MS VOA		
	B	Compound was found in the blank and sample.
	U	Indicates the analyte was analyzed for but not detected.
	F	MS or MSD exceeds the control limits
	J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
	F	RPD of the MS and MSD exceeds the control limits

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1619-1	**PIN06-0500**					
cis-1,2-Dichloroethene		0.26	J	1.0	ug/L	8260B
280-1619-2	**PIN21-0502**					
1,2,3-Trichlorobenzene		0.25	J	1.0	ug/L	8260B
280-1619-5	**PIN21-0505**					
Acetone		7.5	J	10	ug/L	8260B
280-1619-6	**PIN12-0509**					
cis-1,2-Dichloroethene		0.19	J	1.0	ug/L	8260B
280-1619-7	**PIN12-0521**					
Benzene		0.61	J	1.0	ug/L	8260B
1,1-Dichloroethane		0.79	J	1.0	ug/L	8260B
cis-1,2-Dichloroethene		6.5		1.0	ug/L	8260B
trans-1,2-Dichloroethene		0.64	J	1.0	ug/L	8260B
Trichloroethene		5.3		1.0	ug/L	8260B
Vinyl chloride		1.4		1.0	ug/L	8260B
280-1619-8	**PIN12-0527**					
Acetone		3.4	J	10	ug/L	8260B
280-1619-10	**PIN12-0530**					
cis-1,2-Dichloroethene		4.6		1.0	ug/L	8260B
Vinyl chloride		2.1		1.0	ug/L	8260B
280-1619-11	**PIN12-0531**					
cis-1,2-Dichloroethene		0.93	J	1.0	ug/L	8260B
280-1619-12	**PIN12-0532**					
cis-1,2-Dichloroethene		1.0	J	1.0	ug/L	8260B
Vinyl chloride		2.7		1.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1619-14	**PIN12-0534**					
Acetone		4.6	J	10	ug/L	8260B
cis-1,2-Dichloroethene		0.23	J	1.0	ug/L	8260B
280-1619-15	**PIN12-0535**					
Acetone		4.3	J	10	ug/L	8260B
280-1619-16	**PIN12-0536**					
Acetone		5.6	J	10	ug/L	8260B
280-1619-17	**PIN12-0540**					
Acetone		7.1	J	10	ug/L	8260B
1,1-Dichloroethane		0.67	J	1.0	ug/L	8260B
trans-1,2-Dichloroethene		0.83	J	1.0	ug/L	8260B
Vinyl chloride		14		1.0	ug/L	8260B
280-1619-18	**PIN12-0543**					
Acetone		6.5	J	10	ug/L	8260B
cis-1,2-Dichloroethene		1.1		1.0	ug/L	8260B
Vinyl chloride		0.81	J	1.0	ug/L	8260B
280-1619-19	**PIN12-0544**					
Acetone		8.5	J	10	ug/L	8260B
cis-1,2-Dichloroethene		0.49	J	1.0	ug/L	8260B
Vinyl chloride		0.82	J	1.0	ug/L	8260B
280-1619-20	**PIN12-0545**					
Acetone		8.0	J	10	ug/L	8260B
280-1619-21	**PIN12-0546**					
Acetone		4.8	J	10	ug/L	8260B
280-1619-23	**PIN12-0559-1**					
Acetone		7.4	J	10	ug/L	8260B
Methylene Chloride		0.35	J B	1.0	ug/L	8260B

EXECUTIVE SUMMARY - Detections

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Lab Sample ID / Analyte	Client Sample ID	Result / Qualifier		Reporting Limit	Units	Method
280-1619-24	**PIN12-0559-2**					
Acetone		7.2	J	10	ug/L	8260B
Methylene Chloride		0.34	J B	1.0	ug/L	8260B
280-1619-25	**PIN12-0559-3**					
Acetone		5.4	J	10	ug/L	8260B
Methylene Chloride		0.37	J B	1.0	ug/L	8260B
280-1619-27	**PIN12-0560-2**					
Toluene		0.33	J	1.0	ug/L	8260B
280-1619-28	**PIN12-0560-3**					
Acetone		7.4	J	10	ug/L	8260B
Toluene		0.36	J	1.0	ug/L	8260B
280-1619-32	**PIN99-2877**					
Methylene Chloride		1.3		1.0	ug/L	8260B
280-1619-33	**PIN99-2903**					
Toluene		0.19	J	1.0	ug/L	8260B
280-1619-36	**PIN23-SW01**					
Acetone		3.8	J	10	ug/L	8260B
Trichloroethene		0.39	J	1.0	ug/L	8260B
280-1619-37	**PIN99-2891**					
Methylene Chloride		1.1		1.0	ug/L	8260B
280-1619-38TB	**PIN99-2885**					
Methylene Chloride		1.2		1.0	ug/L	8260B

METHOD SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Description	Lab Location	Method	Preparation Method
Matrix: Water			
Volatile Organic Compounds (GC/MS)	TAL DEN	SW846 8260B	
Purge and Trap	TAL DEN		SW846 5030B
Volatile Organic Compounds (GC/MS-SIM)	TAL DEN	SW846 8260B SIM	
Purge and Trap	TAL DEN		SW846 5030B

Lab References:

TAL DEN = TestAmerica Denver

Method References:

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

METHOD / ANALYST SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Method	Analyst	Analyst ID
SW846 8260B	Meier, Greg P	GPM
SW846 8260B	Waterland, Hayley E	HEW
SW846 8260B SIM	Wolfe, Ashley E	AEW

SAMPLE SUMMARY

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
280-1619-1	PIN06-0500	Water	03/17/2010 1425	03/23/2010 0900
280-1619-2	PIN21-0502	Water	03/17/2010 1625	03/23/2010 0900
280-1619-2MS	PIN21-0502	Water	03/17/2010 1625	03/23/2010 0900
280-1619-2MSD	PIN21-0502	Water	03/17/2010 1625	03/23/2010 0900
280-1619-3	PIN21-0503	Water	03/17/2010 1600	03/23/2010 0900
280-1619-4	PIN21-0504	Water	03/18/2010 0955	03/23/2010 0900
280-1619-5	PIN21-0505	Water	03/18/2010 0910	03/23/2010 0900
280-1619-6	PIN12-0509	Water	03/17/2010 1340	03/23/2010 0900
280-1619-7	PIN12-0521	Water	03/17/2010 1515	03/23/2010 0900
280-1619-8	PIN12-0527	Water	03/18/2010 1135	03/23/2010 0900
280-1619-9	PIN12-0528	Water	03/18/2010 1035	03/23/2010 0900
280-1619-10	PIN12-0530	Water	03/17/2010 1333	03/23/2010 0900
280-1619-11	PIN12-0531	Water	03/17/2010 1503	03/23/2010 0900
280-1619-12	PIN12-0532	Water	03/17/2010 1535	03/23/2010 0900
280-1619-13	PIN12-0533	Water	03/18/2010 0947	03/23/2010 0900
280-1619-14	PIN12-0534	Water	03/18/2010 1030	03/23/2010 0900
280-1619-15	PIN12-0535	Water	03/18/2010 1432	03/23/2010 0900
280-1619-16	PIN12-0536	Water	03/18/2010 1519	03/23/2010 0900
280-1619-17	PIN12-0540	Water	03/17/2010 1425	03/23/2010 0900
280-1619-18	PIN12-0543	Water	03/17/2010 1405	03/23/2010 0900
280-1619-19	PIN12-0544	Water	03/17/2010 1613	03/23/2010 0900
280-1619-20	PIN12-0545	Water	03/18/2010 1116	03/23/2010 0900
280-1619-21	PIN12-0546	Water	03/18/2010 1535	03/23/2010 0900
280-1619-22	PIN12-0548	Water	03/17/2010 1515	03/23/2010 0900
280-1619-23	PIN12-0559-1	Water	03/18/2010 1055	03/23/2010 0900
280-1619-24	PIN12-0559-2	Water	03/18/2010 1135	03/23/2010 0900
280-1619-25	PIN12-0559-3	Water	03/18/2010 1205	03/23/2010 0900
280-1619-26	PIN12-0560-1	Water	03/18/2010 1345	03/23/2010 0900
280-1619-27	PIN12-0560-2	Water	03/18/2010 1410	03/23/2010 0900
280-1619-28	PIN12-0560-3	Water	03/18/2010 1440	03/23/2010 0900
280-1619-29	PIN12-0561-1	Water	03/18/2010 0900	03/23/2010 0900
280-1619-30	PIN12-0561-2	Water	03/18/2010 0945	03/23/2010 0900
280-1619-31	PIN12-0561-3	Water	03/18/2010 1015	03/23/2010 0900
280-1619-32	PIN99-2877	Water	03/17/2010 0800	03/23/2010 0900
280-1619-33	PIN99-2903	Water	03/18/2010 1300	03/23/2010 0900
280-1619-34	PIN37-S002	Water	03/18/2010 1345	03/23/2010 0900
280-1619-35	PIN12-S73B	Water	03/17/2010 1620	03/23/2010 0900
280-1619-36	PIN23-SW01	Water	03/18/2010 1330	03/23/2010 0900
280-1619-37	PIN99-2891	Water	03/17/2010 0800	03/23/2010 0900
280-1619-38TB	PIN99-2885	Water	03/17/2010 1200	03/23/2010 0900

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN06-0500**

Lab Sample ID: 280-1619-1
Client Matrix: Water

Date Sampled: 03/17/2010 1425
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8581.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1124		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1124			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.26	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN06-0500**

Lab Sample ID: 280-1619-1
Client Matrix: Water

Date Sampled: 03/17/2010 1425
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8581.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1124		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1124			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101		70 - 127
Toluene-d8 (Surr)	95		80 - 125
4-Bromofluorobenzene (Surr)	110		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN21-0502**

Lab Sample ID: 280-1619-2
Client Matrix: Water

Date Sampled: 03/17/2010 1625
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8582.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1144		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1144			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN21-0502**

Lab Sample ID: 280-1619-2
Client Matrix: Water

Date Sampled: 03/17/2010 1625
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8582.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1144		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1144			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.25	J	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104		70 - 127
Toluene-d8 (Surr)	94		80 - 125
4-Bromofluorobenzene (Surr)	109		78 - 118
Dibromofluoromethane (Surr)	105		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN21-0503**

Lab Sample ID: 280-1619-3
Client Matrix: Water

Date Sampled: 03/17/2010 1600
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8576.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 0943		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 0943			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN21-0503**

Lab Sample ID: 280-1619-3
Client Matrix: Water

Date Sampled: 03/17/2010 1600
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8576.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 0943		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 0943			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	99		70 - 127
Toluene-d8 (Surr)	97		80 - 125
4-Bromofluorobenzene (Surr)	109		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN21-0504**

Lab Sample ID: 280-1619-4
Client Matrix: Water

Date Sampled: 03/18/2010 0955
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8587.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1326		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1326			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN21-0504**

Lab Sample ID: 280-1619-4
Client Matrix: Water

Date Sampled: 03/18/2010 0955
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8587.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1326		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1326			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	102		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	108		78 - 118
Dibromofluoromethane (Surr)	105		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN21-0505**

Lab Sample ID: 280-1619-5
Client Matrix: Water

Date Sampled: 03/18/2010 0910
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8590.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1426		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1426			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	7.5	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN21-0505**

Lab Sample ID: 280-1619-5
Client Matrix: Water

Date Sampled: 03/18/2010 0910
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8590.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1426		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1426			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	102		70 - 127
Toluene-d8 (Surr)	97		80 - 125
4-Bromofluorobenzene (Surr)	109		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0509**

Lab Sample ID: 280-1619-6
Client Matrix: Water

Date Sampled: 03/17/2010 1340
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8577.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1004		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1004			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.19	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0509**

Lab Sample ID: 280-1619-6
Client Matrix: Water

Date Sampled: 03/17/2010 1340
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8577.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1004		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1004			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	100		70 - 127
Toluene-d8 (Surr)	97		80 - 125
4-Bromofluorobenzene (Surr)	109		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0521**

Lab Sample ID: 280-1619-7
Client Matrix: Water

Date Sampled: 03/17/2010 1515
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8578.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1024		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1024			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.61	J	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.79	J	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	6.5		0.15	1.0
trans-1,2-Dichloroethene	0.64	J	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0521**

Lab Sample ID: 280-1619-7
Client Matrix: Water

Date Sampled: 03/17/2010 1515
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8578.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1024		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1024			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	5.3		0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	1.4		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	108		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0527**

Lab Sample ID: 280-1619-8
Client Matrix: Water

Date Sampled: 03/18/2010 1135
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8588.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1346		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1346			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	3.4	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0527**

Lab Sample ID: 280-1619-8
Client Matrix: Water

Date Sampled: 03/18/2010 1135
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8588.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1346		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1346			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	102		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	109		78 - 118
Dibromofluoromethane (Surr)	105		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0528**

Lab Sample ID: 280-1619-9
Client Matrix: Water

Date Sampled: 03/18/2010 1035
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8589.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1406		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1406			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0528**

Lab Sample ID: 280-1619-9
Client Matrix: Water

Date Sampled: 03/18/2010 1035
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8589.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1406		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1406			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	110		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0530**

Lab Sample ID: 280-1619-10
Client Matrix: Water

Date Sampled: 03/17/2010 1333
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8579.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1044		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1044			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	4.6		0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0530**

Lab Sample ID: 280-1619-10
Client Matrix: Water

Date Sampled: 03/17/2010 1333
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8579.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1044		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1044			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	2.1		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	100		70 - 127
Toluene-d8 (Surr)	97		80 - 125
4-Bromofluorobenzene (Surr)	109		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0531**

Lab Sample ID: 280-1619-11
Client Matrix: Water

Date Sampled: 03/17/2010 1503
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8580.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1104		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1104			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.93	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0531**

Lab Sample ID: 280-1619-11
Client Matrix: Water

Date Sampled: 03/17/2010 1503
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8580.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1104		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1104			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	100		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	109		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0532**

Lab Sample ID: 280-1619-12
Client Matrix: Water

Date Sampled: 03/17/2010 1535
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8583.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1205		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1205			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	1.0	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0532**

Lab Sample ID: 280-1619-12
Client Matrix: Water

Date Sampled: 03/17/2010 1535
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8583.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1205		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1205			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	2.7		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	109		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0533**

Lab Sample ID: 280-1619-13
Client Matrix: Water

Date Sampled: 03/18/2010 0947
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8591.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1446		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1446			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0533**

Lab Sample ID: 280-1619-13
Client Matrix: Water

Date Sampled: 03/18/2010 0947
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8591.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1446		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1446			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	110		78 - 118
Dibromofluoromethane (Surr)	105		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0534**

Lab Sample ID: 280-1619-14
Client Matrix: Water

Date Sampled: 03/18/2010 1030
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8592.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1506		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1506			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	4.6	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.23	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0534**

Lab Sample ID: 280-1619-14
Client Matrix: Water

Date Sampled: 03/18/2010 1030
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8592.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1506		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1506			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	105		70 - 127
Toluene-d8 (Surr)	97		80 - 125
4-Bromofluorobenzene (Surr)	109		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0535**

Lab Sample ID: 280-1619-15
Client Matrix: Water

Date Sampled: 03/18/2010 1432
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8593.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1526		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1526			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	4.3	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0535**

Lab Sample ID: 280-1619-15
Client Matrix: Water

Date Sampled: 03/18/2010 1432
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8593.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1526		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1526			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	109		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0536**

Lab Sample ID: 280-1619-16
Client Matrix: Water

Date Sampled: 03/18/2010 1519
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8594.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1546		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1546			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	5.6	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0536**

Lab Sample ID: 280-1619-16
Client Matrix: Water

Date Sampled: 03/18/2010 1519
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8594.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1546		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1546			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	102		70 - 127
Toluene-d8 (Surr)	95		80 - 125
4-Bromofluorobenzene (Surr)	110		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0540**

Lab Sample ID: 280-1619-17
Client Matrix: Water

Date Sampled: 03/17/2010 1425
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8584.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1225		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1225			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	7.1	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.67	J	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.83	J	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0540**

Lab Sample ID: 280-1619-17
Client Matrix: Water

Date Sampled: 03/17/2010 1425
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8584.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1225		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1225			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	14		0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	100		70 - 127
Toluene-d8 (Surr)	95		80 - 125
4-Bromofluorobenzene (Surr)	107		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0543**

Lab Sample ID: 280-1619-18
Client Matrix: Water

Date Sampled: 03/17/2010 1405
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8585.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1245		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1245			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	6.5	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	1.1		0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0543**

Lab Sample ID: 280-1619-18
Client Matrix: Water

Date Sampled: 03/17/2010 1405
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8585.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1245		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1245			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.81	J	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103		70 - 127
Toluene-d8 (Surr)	95		80 - 125
4-Bromofluorobenzene (Surr)	109		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0544**

Lab Sample ID: 280-1619-19
Client Matrix: Water

Date Sampled: 03/17/2010 1613
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8586.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1305		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1305			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	8.5	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.49	J	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0544**

Lab Sample ID: 280-1619-19
Client Matrix: Water

Date Sampled: 03/17/2010 1613
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8586.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1305		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1305			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.82	J	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	107		70 - 127
Toluene-d8 (Surr)	94		80 - 125
4-Bromofluorobenzene (Surr)	109		78 - 118
Dibromofluoromethane (Surr)	106		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0545**

Lab Sample ID: 280-1619-20
Client Matrix: Water

Date Sampled: 03/18/2010 1116
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8595.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1606		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1606			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	8.0	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0545**

Lab Sample ID: 280-1619-20
Client Matrix: Water

Date Sampled: 03/18/2010 1116
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9293	Instrument ID:	MSV_Z
Preparation:	5030B		Lab File ID:	Z8595.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1606		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1606			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	110		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0546**

Lab Sample ID: 280-1619-21
Client Matrix: Water

Date Sampled: 03/18/2010 1535
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9544	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7786.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1926		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1926			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	4.8	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0546**

Lab Sample ID: 280-1619-21
Client Matrix: Water

Date Sampled: 03/18/2010 1535
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9544	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7786.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1926		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1926			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	108		70 - 127
Toluene-d8 (Surr)	95		80 - 125
4-Bromofluorobenzene (Surr)	96		78 - 118
Dibromofluoromethane (Surr)	93		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0548**

Lab Sample ID: 280-1619-22
Client Matrix: Water

Date Sampled: 03/17/2010 1515
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5068.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1624		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1624			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0548**

Lab Sample ID: 280-1619-22
Client Matrix: Water

Date Sampled: 03/17/2010 1515
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5068.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1624		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1624			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	109		70 - 127
Toluene-d8 (Surr)	102		80 - 125
4-Bromofluorobenzene (Surr)	99		78 - 118
Dibromofluoromethane (Surr)	102		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0559-1**

Lab Sample ID: 280-1619-23
Client Matrix: Water

Date Sampled: 03/18/2010 1055
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9544	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7787.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1947		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1947			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	7.4	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.35	J B	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0559-1**

Lab Sample ID: 280-1619-23
Client Matrix: Water

Date Sampled: 03/18/2010 1055
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9544	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7787.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1947		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1947			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	106		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	96		78 - 118
Dibromofluoromethane (Surr)	92		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0559-2**

Lab Sample ID: 280-1619-24
Client Matrix: Water

Date Sampled: 03/18/2010 1135
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9544	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7788.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 2007		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 2007			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	7.2	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.34	J B	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0559-2**

Lab Sample ID: 280-1619-24
Client Matrix: Water

Date Sampled: 03/18/2010 1135
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9544	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7788.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 2007		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 2007			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	108		70 - 127
Toluene-d8 (Surr)	98		80 - 125
4-Bromofluorobenzene (Surr)	97		78 - 118
Dibromofluoromethane (Surr)	94		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0559-3**

Lab Sample ID: 280-1619-25
Client Matrix: Water

Date Sampled: 03/18/2010 1205
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9544	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7789.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 2027		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 2027			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	5.4	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.37	J B	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0559-3**

Lab Sample ID: 280-1619-25
Client Matrix: Water

Date Sampled: 03/18/2010 1205
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9544	Instrument ID:	MSV_R1
Preparation:	5030B		Lab File ID:	R7789.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 2027		Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 2027			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	108		70 - 127
Toluene-d8 (Surr)	96		80 - 125
4-Bromofluorobenzene (Surr)	97		78 - 118
Dibromofluoromethane (Surr)	93		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0560-1**

Lab Sample ID: 280-1619-26
Client Matrix: Water

Date Sampled: 03/18/2010 1345
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2686.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 0907		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 0907			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0560-1**

Lab Sample ID: 280-1619-26
Client Matrix: Water

Date Sampled: 03/18/2010 1345
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2686.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 0907		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 0907			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	83		70 - 127
Toluene-d8 (Surr)	91		80 - 125
4-Bromofluorobenzene (Surr)	102		78 - 118
Dibromofluoromethane (Surr)	97		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0560-2**

Lab Sample ID: 280-1619-27
Client Matrix: Water

Date Sampled: 03/18/2010 1410
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2690.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1034		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1034			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0560-2**

Lab Sample ID: 280-1619-27
Client Matrix: Water

Date Sampled: 03/18/2010 1410
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2690.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1034		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1034			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.33	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	90		70 - 127
Toluene-d8 (Surr)	89		80 - 125
4-Bromofluorobenzene (Surr)	100		78 - 118
Dibromofluoromethane (Surr)	100		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0560-3**

Lab Sample ID: 280-1619-28
Client Matrix: Water

Date Sampled: 03/18/2010 1440
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2691.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1055		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1055			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	7.4	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0560-3**

Lab Sample ID: 280-1619-28
Client Matrix: Water

Date Sampled: 03/18/2010 1440
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2691.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1055		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1055			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.36	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103		70 - 127
Toluene-d8 (Surr)	87		80 - 125
4-Bromofluorobenzene (Surr)	99		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0561-1**

Lab Sample ID: 280-1619-29
Client Matrix: Water

Date Sampled: 03/18/2010 0900
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2692.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1116		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1116			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0561-1**

Lab Sample ID: 280-1619-29
Client Matrix: Water

Date Sampled: 03/18/2010 0900
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2692.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1116		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1116			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	79		70 - 127
Toluene-d8 (Surr)	93		80 - 125
4-Bromofluorobenzene (Surr)	102		78 - 118
Dibromofluoromethane (Surr)	94		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0561-2**

Lab Sample ID: 280-1619-30
Client Matrix: Water

Date Sampled: 03/18/2010 0945
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2693.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1138		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1138			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0561-2**

Lab Sample ID: 280-1619-30
Client Matrix: Water

Date Sampled: 03/18/2010 0945
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2693.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1138		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1138			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	80		70 - 127
Toluene-d8 (Surr)	88		80 - 125
4-Bromofluorobenzene (Surr)	101		78 - 118
Dibromofluoromethane (Surr)	94		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0561-3**

Lab Sample ID: 280-1619-31
Client Matrix: Water

Date Sampled: 03/18/2010 1015
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2694.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1200		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1200			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0561-3**

Lab Sample ID: 280-1619-31
Client Matrix: Water

Date Sampled: 03/18/2010 1015
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2694.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1200		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1200			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	83		70 - 127
Toluene-d8 (Surr)	89		80 - 125
4-Bromofluorobenzene (Surr)	101		78 - 118
Dibromofluoromethane (Surr)	95		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2877**

Lab Sample ID: 280-1619-32
Client Matrix: Water

Date Sampled: 03/17/2010 0800
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5069.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1645		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1645			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	1.3		0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2877**

Lab Sample ID: 280-1619-32
Client Matrix: Water

Date Sampled: 03/17/2010 0800
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5069.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1645		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1645			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	108		70 - 127
Toluene-d8 (Surr)	111		80 - 125
4-Bromofluorobenzene (Surr)	107		78 - 118
Dibromofluoromethane (Surr)	108		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2903**

Lab Sample ID: 280-1619-33
Client Matrix: Water

Date Sampled: 03/18/2010 1300
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2695.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1222		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1222			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2903**

Lab Sample ID: 280-1619-33
Client Matrix: Water

Date Sampled: 03/18/2010 1300
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2695.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1222		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1222			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.19	J	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	79		70 - 127
Toluene-d8 (Surr)	89		80 - 125
4-Bromofluorobenzene (Surr)	101		78 - 118
Dibromofluoromethane (Surr)	93		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN37-S002**

Lab Sample ID: 280-1619-34
Client Matrix: Water

Date Sampled: 03/18/2010 1345
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2696.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1243		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1243			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN37-S002**

Lab Sample ID: 280-1619-34
Client Matrix: Water

Date Sampled: 03/18/2010 1345
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2696.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1243		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1243			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	78		70 - 127
Toluene-d8 (Surr)	90		80 - 125
4-Bromofluorobenzene (Surr)	103		78 - 118
Dibromofluoromethane (Surr)	93		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S73B**

Lab Sample ID: 280-1619-35
Client Matrix: Water

Date Sampled: 03/17/2010 1620
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5070.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1706		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1706			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-S73B**

Lab Sample ID: 280-1619-35
Client Matrix: Water

Date Sampled: 03/17/2010 1620
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5070.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1706		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1706			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	111		70 - 127
Toluene-d8 (Surr)	98		80 - 125
4-Bromofluorobenzene (Surr)	97		78 - 118
Dibromofluoromethane (Surr)	103		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN23-SW01**

Lab Sample ID: 280-1619-36
Client Matrix: Water

Date Sampled: 03/18/2010 1330
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2697.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1305		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1305			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	3.8	J	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN23-SW01**

Lab Sample ID: 280-1619-36
Client Matrix: Water

Date Sampled: 03/18/2010 1330
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9458	Instrument ID:	MSV_H
Preparation:	5030B		Lab File ID:	H2697.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 1305		Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 1305			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.39	J	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	85		70 - 127
Toluene-d8 (Surr)	91		80 - 125
4-Bromofluorobenzene (Surr)	102		78 - 118
Dibromofluoromethane (Surr)	95		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2891**

Lab Sample ID: 280-1619-37
Client Matrix: Water

Date Sampled: 03/17/2010 0800
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5071.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1726		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1726			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	1.1		0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2891**

Lab Sample ID: 280-1619-37
Client Matrix: Water

Date Sampled: 03/17/2010 0800
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5071.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1726		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1726			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	106		70 - 127
Toluene-d8 (Surr)	110		80 - 125
4-Bromofluorobenzene (Surr)	102		78 - 118
Dibromofluoromethane (Surr)	104		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2885**

Lab Sample ID: 280-1619-38TB
Client Matrix: Water

Date Sampled: 03/17/2010 1200
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5072.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1747		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1747			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	1.2		0.32	1.0
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2885**

Lab Sample ID: 280-1619-38TB
Client Matrix: Water

Date Sampled: 03/17/2010 1200
Date Received: 03/23/2010 0900

8260B Volatile Organic Compounds (GC/MS)

Method:	8260B	Analysis Batch: 280-9155	Instrument ID:	MSV_R2
Preparation:	5030B		Lab File ID:	RR5072.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1747		Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1747			

Analyte	Result (ug/L)	Qualifier	MDL	RL
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	101		70 - 127
Toluene-d8 (Surr)	102		80 - 125
4-Bromofluorobenzene (Surr)	99		78 - 118
Dibromofluoromethane (Surr)	101		77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0530**

Lab Sample ID: 280-1619-10
Client Matrix: Water

Date Sampled: 03/17/2010 1333
Date Received: 03/23/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8860	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34023.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1231		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1231			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	91		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0540**

Lab Sample ID: 280-1619-17
Client Matrix: Water

Date Sampled: 03/17/2010 1425
Date Received: 03/23/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8860	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34026.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1347		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1347			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	72		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0543**

Lab Sample ID: 280-1619-18
Client Matrix: Water

Date Sampled: 03/17/2010 1405
Date Received: 03/23/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8860	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34027.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1412		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1412			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	77		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN12-0548**

Lab Sample ID: 280-1619-22
Client Matrix: Water

Date Sampled: 03/17/2010 1515
Date Received: 03/23/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8860	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34028.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1438		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1438			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	74		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Client Sample ID: **PIN99-2903**

Lab Sample ID: 280-1619-33
Client Matrix: Water

Date Sampled: 03/18/2010 1300
Date Received: 03/23/2010 0900

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Method:	8260B SIM	Analysis Batch: 280-8860	Instrument ID:	MSV_G2
Preparation:	5030B		Lab File ID:	g2_34029.D
Dilution:	1.0		Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1503		Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1503			

Analyte	Result (ug/L)	Qualifier	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	%Rec	Qualifier	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103		70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B Volatile Organic Compounds (GC/MS)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DBFM %Rec	DCA %Rec	TOL %Rec	BFB %Rec
280-1619-1	PIN06-0500	104	101	95	110
280-1619-2	PIN21-0502	105	104	94	109
280-1619-3	PIN21-0503	103	99	97	109
280-1619-4	PIN21-0504	105	102	96	108
280-1619-5	PIN21-0505	104	102	97	109
280-1619-6	PIN12-0509	103	100	97	109
280-1619-7	PIN12-0521	104	103	96	108
280-1619-8	PIN12-0527	105	102	96	109
280-1619-9	PIN12-0528	104	101	96	110
280-1619-10	PIN12-0530	103	100	97	109
280-1619-11	PIN12-0531	104	100	96	109
280-1619-12	PIN12-0532	104	101	96	109
280-1619-13	PIN12-0533	105	104	96	110
280-1619-14	PIN12-0534	104	105	97	109
280-1619-15	PIN12-0535	104	104	96	109
280-1619-16	PIN12-0536	104	102	95	110
280-1619-17	PIN12-0540	104	100	95	107
280-1619-18	PIN12-0543	104	103	95	109
280-1619-19	PIN12-0544	106	107	94	109
280-1619-20	PIN12-0545	104	103	96	110
280-1619-21	PIN12-0546	93	108	95	96
280-1619-22	PIN12-0548	102	109	102	99
280-1619-23	PIN12-0559-1	92	106	96	96
280-1619-24	PIN12-0559-2	94	108	98	97
280-1619-25	PIN12-0559-3	93	108	96	97
280-1619-26	PIN12-0560-1	97	83	91	102
280-1619-27	PIN12-0560-2	100	90	89	100
280-1619-28	PIN12-0560-3	103	103	87	99
280-1619-29	PIN12-0561-1	94	79	93	102

Surrogate	Acceptance Limits
DBFM = Dibromofluoromethane (Surr)	77-119
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127
TOL = Toluene-d8 (Surr)	80-125
BFB = 4-Bromofluorobenzene (Surr)	78-118

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B Volatile Organic Compounds (GC/MS)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DBFM %Rec	DCA %Rec	TOL %Rec	BFB %Rec
280-1619-30	PIN12-0561-2	94	80	88	101
280-1619-31	PIN12-0561-3	95	83	89	101
280-1619-32	PIN99-2877	108	108	111	107
280-1619-33	PIN99-2903	93	79	89	101
280-1619-34	PIN37-S002	93	78	90	103
280-1619-35	PIN12-S73B	103	111	98	97
280-1619-36	PIN23-SW01	95	85	91	102
280-1619-37	PIN99-2891	104	106	110	102
280-1619-38	PIN99-2885	101	101	102	99
MB 280-9155/5		101	99	104	102
MB 280-9293/4		101	94	96	110
MB 280-9458/4		95	84	89	101
MB 280-9544/3		96	103	102	99
LCS 280-9155/4		99	97	108	104
LCS 280-9293/3		97	96	104	105
LCS 280-9458/3		97	87	98	94
LCS 280-9544/7		97	98	105	102
280-1619-2 MS	PIN21-0502 MS	99	98	103	104
280-1619-26 MS	PIN12-0560-1 MS	97	91	96	95
280-1555-A-3 MS		96	97	100	99
280-1609-G-1 MS		97	104	101	96
280-1619-2 MSD	PIN21-0502 MSD	99	99	103	103
280-1619-26 MSD	PIN12-0560-1 MSD	96	90	97	99
280-1555-D-3 MSD		96	95	102	97
280-1609-G-1 MSD		98	110	99	97

Surrogate	Acceptance Limits
DBFM = Dibromofluoromethane (Surr)	77-119
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127
TOL = Toluene-d8 (Surr)	80-125
BFB = 4-Bromofluorobenzene (Surr)	78-118

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Surrogate Recovery Report

8260B SIM Volatile Organic Compounds (GC/MS-SIM)

Client Matrix: Water

Lab Sample ID	Client Sample ID	DCA %Rec
280-1619-10	PIN12-0530	91
280-1619-17	PIN12-0540	72
280-1619-18	PIN12-0543	77
280-1619-22	PIN12-0548	74
280-1619-33	PIN99-2903	103
MB 280-8860/4		91
LCS 280-8860/3		93
280-1619-10 MS	PIN12-0530 MS	97
280-1619-10 MSD	PIN12-0530 MSD	93

Surrogate	Acceptance Limits
DCA = 1,2-Dichloroethane-d4 (Surr)	70-127

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Method Blank - Batch: 280-9155

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-9155/5	Analysis Batch:	280-9155	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR5051.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1031			Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1031				

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Method Blank - Batch: 280-9155

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-9155/5
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 03/30/2010 1031
Date Prepared: 03/30/2010 1031

Analysis Batch: 280-9155
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_R2
Lab File ID: RR5051.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.428	J	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	99	70 - 127
Toluene-d8 (Surr)	104	80 - 125
4-Bromofluorobenzene (Surr)	102	78 - 118
Dibromofluoromethane (Surr)	101	77 - 119

Quality Control Results

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Lab Control Sample - Batch: 280-9155

Method: 8260B
Preparation: 5030B

Lab Sample ID:	LCS 280-9155/4	Analysis Batch:	280-9155	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR5048.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 0929			Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 0929				

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Benzene	5.00	5.52	110	77 - 120	
Bromodichloromethane	5.00	4.89	98	78 - 120	
Carbon tetrachloride	5.00	5.62	112	80 - 120	
Chlorobenzene	5.00	5.34	107	78 - 120	
Chloroform	5.00	5.07	101	78 - 120	
1,3-Dichlorobenzene	5.00	5.81	116	75 - 120	
1,1-Dichloroethane	5.00	5.39	108	77 - 120	
trans-1,2-Dichloroethene	5.00	5.57	111	80 - 120	
1,1-Dichloroethene	5.00	6.35	127	68 - 133	
1,2-Dichloropropane	5.00	4.92	98	76 - 120	
Ethylbenzene	5.00	5.75	115	78 - 120	
Methylene Chloride	5.00	5.22	104	71 - 120	
Tetrachloroethene	5.00	5.89	118	77 - 120	
Toluene	5.00	4.97	99	73 - 120	
1,1,1-Trichloroethane	5.00	5.17	103	78 - 120	
Trichloroethene	5.00	5.12	102	78 - 122	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	97	70 - 127
Toluene-d8 (Surr)	108	80 - 125
4-Bromofluorobenzene (Surr)	104	78 - 118
Dibromofluoromethane (Surr)	99	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-9155

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1555-A-3 MS	Analysis Batch:	280-9155	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR5054.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1133			Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1133				

MSD Lab Sample ID:	280-1555-D-3 MSD	Analysis Batch:	280-9155	Instrument ID:	MSV_R2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	RR5055.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/30/2010 1153			Final Weight/Volume:	20 mL
Date Prepared:	03/30/2010 1153				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	108	103	77 - 120	5	20		
Bromodichloromethane	94	95	78 - 120	1	20		
Carbon tetrachloride	111	97	80 - 120	14	21		
Chlorobenzene	103	102	78 - 120	1	20		
Chloroform	105	102	78 - 120	3	20		
1,3-Dichlorobenzene	107	105	75 - 120	2	20		
1,1-Dichloroethane	109	103	77 - 120	5	21		
trans-1,2-Dichloroethene	113	105	80 - 120	8	24		
1,1-Dichloroethene	134	117	68 - 133	13	20	F	
1,2-Dichloropropane	95	97	76 - 120	2	20		
Ethylbenzene	111	107	78 - 120	4	26		
Methylene Chloride	111	104	71 - 120	5	20		
Tetrachloroethene	112	102	77 - 120	9	20		
Toluene	97	99	73 - 120	2	20		
1,1,1-Trichloroethane	107	93	78 - 120	14	20		
Trichloroethene	100	96	78 - 122	4	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	97	95	70 - 127
Toluene-d8 (Surr)	100	102	80 - 125
4-Bromofluorobenzene (Surr)	99	97	78 - 118
Dibromofluoromethane (Surr)	96	96	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-9155

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1555-A-3 MS	Units: ug/L	MSD Lab Sample ID:	280-1555-D-3 MSD
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/30/2010 1133		Date Analyzed:	03/30/2010 1153
Date Prepared:	03/30/2010 1133		Date Prepared:	03/30/2010 1153

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual		MSD Result/Qual
Benzene	0.16	U	5.00	5.00	5.39		5.14
Bromodichloromethane	0.17	U	5.00	5.00	4.72		4.76
Carbon tetrachloride	0.19	U	5.00	5.00	5.56		4.84
Chlorobenzene	0.17	U	5.00	5.00	5.16		5.11
Chloroform	0.16	U	5.00	5.00	5.24		5.09
1,3-Dichlorobenzene	0.13	U	5.00	5.00	5.36		5.27
1,1-Dichloroethane	0.22	U	5.00	5.00	5.45		5.16
trans-1,2-Dichloroethene	0.15	U	5.00	5.00	5.67		5.25
1,1-Dichloroethene	0.23	U	5.00	5.00	6.68	F	5.87
1,2-Dichloropropane	0.18	U	5.00	5.00	4.77		4.86
Ethylbenzene	0.16	U	5.00	5.00	5.53		5.33
Methylene Chloride	0.66	J	5.00	5.00	6.21		5.88
Tetrachloroethene	0.20	U	5.00	5.00	5.62		5.12
Toluene	0.17	U	5.00	5.00	4.84		4.94
1,1,1-Trichloroethane	0.16	U	5.00	5.00	5.37		4.67
Trichloroethene	0.16	U	5.00	5.00	5.01		4.82

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Method Blank - Batch: 280-9293

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-9293/4	Analysis Batch:	280-9293	Instrument ID:	MSV_Z
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	Z8570.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 0743			Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 0743				

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.469	J	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Method Blank - Batch: 280-9293

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-9293/4	Analysis Batch:	280-9293	Instrument ID:	MSV_Z
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	Z8570.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 0743			Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 0743				

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	94	70 - 127
Toluene-d8 (Surr)	96	80 - 125
4-Bromofluorobenzene (Surr)	110	78 - 118
Dibromofluoromethane (Surr)	101	77 - 119

Quality Control Results

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Lab Control Sample - Batch: 280-9293

Method: 8260B
Preparation: 5030B

Lab Sample ID:	LCS 280-9293/3	Analysis Batch:	280-9293	Instrument ID:	MSV_Z
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	Z8568.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 0703			Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 0703				

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Benzene	5.00	5.42	108	77 - 120	
Bromodichloromethane	5.00	5.35	107	78 - 120	
Carbon tetrachloride	5.00	5.59	112	80 - 120	
Chlorobenzene	5.00	5.55	111	78 - 120	
Chloroform	5.00	5.28	106	78 - 120	
1,3-Dichlorobenzene	5.00	5.01	100	75 - 120	
1,1-Dichloroethane	5.00	5.29	106	77 - 120	
trans-1,2-Dichloroethene	5.00	5.36	107	80 - 120	
1,1-Dichloroethene	5.00	5.17	103	68 - 133	
1,2-Dichloropropane	5.00	5.33	107	76 - 120	
Ethylbenzene	5.00	5.85	117	78 - 120	
Methylene Chloride	5.00	5.20	104	71 - 120	
Tetrachloroethene	5.00	5.75	115	77 - 120	
Toluene	5.00	5.45	109	73 - 120	
1,1,1-Trichloroethane	5.00	5.47	109	78 - 120	
Trichloroethene	5.00	5.43	109	78 - 122	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	96	70 - 127
Toluene-d8 (Surr)	104	80 - 125
4-Bromofluorobenzene (Surr)	105	78 - 118
Dibromofluoromethane (Surr)	97	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-9293

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1619-2	Analysis Batch:	280-9293	Instrument ID:	MSV_Z
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	Z8573.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 0843			Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 0843				

MSD Lab Sample ID:	280-1619-2	Analysis Batch:	280-9293	Instrument ID:	MSV_Z
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	Z8574.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 0904			Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 0904				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	101	108	77 - 120	7	20		
Bromodichloromethane	102	109	78 - 120	6	20		
Carbon tetrachloride	108	117	80 - 120	8	21		
Chlorobenzene	105	113	78 - 120	8	20		
Chloroform	103	109	78 - 120	6	20		
1,3-Dichlorobenzene	100	101	75 - 120	1	20		
1,1-Dichloroethane	107	113	77 - 120	6	21		
trans-1,2-Dichloroethene	105	109	80 - 120	3	24		
1,1-Dichloroethene	101	105	68 - 133	4	20		
1,2-Dichloropropane	103	110	76 - 120	7	20		
Ethylbenzene	110	117	78 - 120	6	26		
Methylene Chloride	98	106	71 - 120	8	20		
Tetrachloroethene	109	115	77 - 120	6	20		
Toluene	105	111	73 - 120	6	20		
1,1,1-Trichloroethane	107	111	78 - 120	3	20		
Trichloroethene	104	110	78 - 122	6	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	98	99	70 - 127
Toluene-d8 (Surr)	103	103	80 - 125
4-Bromofluorobenzene (Surr)	104	103	78 - 118
Dibromofluoromethane (Surr)	99	99	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-9293

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1619-2	Units: ug/L	MSD Lab Sample ID:	280-1619-2
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/31/2010 0843		Date Analyzed:	03/31/2010 0904
Date Prepared:	03/31/2010 0843		Date Prepared:	03/31/2010 0904

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	0.16	U	5.00	5.00	5.06	5.41
Bromodichloromethane	0.17	U	5.00	5.00	5.11	5.44
Carbon tetrachloride	0.19	U	5.00	5.00	5.40	5.83
Chlorobenzene	0.17	U	5.00	5.00	5.23	5.66
Chloroform	0.16	U	5.00	5.00	5.14	5.43
1,3-Dichlorobenzene	0.13	U	5.00	5.00	5.01	5.04
1,1-Dichloroethane	0.22	U	5.00	5.00	5.33	5.67
trans-1,2-Dichloroethene	0.15	U	5.00	5.00	5.27	5.45
1,1-Dichloroethene	0.23	U	5.00	5.00	5.06	5.24
1,2-Dichloropropane	0.18	U	5.00	5.00	5.14	5.51
Ethylbenzene	0.16	U	5.00	5.00	5.49	5.84
Methylene Chloride	0.32	U	5.00	5.00	4.89	5.28
Tetrachloroethene	0.20	U	5.00	5.00	5.43	5.75
Toluene	0.17	U	5.00	5.00	5.24	5.57
1,1,1-Trichloroethane	0.16	U	5.00	5.00	5.36	5.54
Trichloroethene	0.16	U	5.00	5.00	5.20	5.49

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Method Blank - Batch: 280-9458

Method: 8260B
Preparation: 5030B

Lab Sample ID: MB 280-9458/4
Client Matrix: Water
Dilution: 1.0
Date Analyzed: 04/01/2010 0846
Date Prepared: 04/01/2010 0846

Analysis Batch: 280-9458
Prep Batch: N/A
Units: ug/L

Instrument ID: MSV_H
Lab File ID: H2685.D
Initial Weight/Volume: 20 mL
Final Weight/Volume: 20 mL

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.32	U	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Method Blank - Batch: 280-9458

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-9458/4	Analysis Batch:	280-9458	Instrument ID:	MSV_H
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	H2685.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 0846			Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 0846				

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	84	70 - 127
Toluene-d8 (Surr)	89	80 - 125
4-Bromofluorobenzene (Surr)	101	78 - 118
Dibromofluoromethane (Surr)	95	77 - 119

Quality Control Results

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Lab Control Sample - Batch: 280-9458

Method: 8260B
Preparation: 5030B

Lab Sample ID:	LCS 280-9458/3	Analysis Batch:	280-9458	Instrument ID:	MSV_H
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	H2683.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 0802			Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 0802				

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Benzene	5.00	5.29	106	77 - 120	
Bromodichloromethane	5.00	4.98	100	78 - 120	
Carbon tetrachloride	5.00	5.98	120	80 - 120	
Chlorobenzene	5.00	5.30	106	78 - 120	
Chloroform	5.00	5.48	110	78 - 120	
1,3-Dichlorobenzene	5.00	4.84	97	75 - 120	
1,1-Dichloroethane	5.00	5.38	108	77 - 120	
trans-1,2-Dichloroethene	5.00	5.52	110	80 - 120	
1,1-Dichloroethene	5.00	5.60	112	68 - 133	
1,2-Dichloropropane	5.00	4.94	99	76 - 120	
Ethylbenzene	5.00	5.51	110	78 - 120	
Methylene Chloride	5.00	5.04	101	71 - 120	
Tetrachloroethene	5.00	5.45	109	77 - 120	
Toluene	5.00	5.08	102	73 - 120	
1,1,1-Trichloroethane	5.00	5.96	119	78 - 120	
Trichloroethene	5.00	4.97	99	78 - 122	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	87	70 - 127
Toluene-d8 (Surr)	98	80 - 125
4-Bromofluorobenzene (Surr)	94	78 - 118
Dibromofluoromethane (Surr)	97	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-9458

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1619-26	Analysis Batch:	280-9458	Instrument ID:	MSV_H
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	H2687.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 0929			Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 0929				

MSD Lab Sample ID:	280-1619-26	Analysis Batch:	280-9458	Instrument ID:	MSV_H
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	H2688.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	04/01/2010 0950			Final Weight/Volume:	20 mL
Date Prepared:	04/01/2010 0950				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	102	110	77 - 120	8	20		
Bromodichloromethane	104	114	78 - 120	9	20		
Carbon tetrachloride	106	114	80 - 120	7	21		
Chlorobenzene	100	109	78 - 120	9	20		
Chloroform	106	114	78 - 120	7	20		
1,3-Dichlorobenzene	100	106	75 - 120	7	20		
1,1-Dichloroethane	109	116	77 - 120	7	21		
trans-1,2-Dichloroethene	102	105	80 - 120	2	24		
1,1-Dichloroethene	93	98	68 - 133	5	20		
1,2-Dichloropropane	101	113	76 - 120	11	20		
Ethylbenzene	100	111	78 - 120	10	26		
Methylene Chloride	88	93	71 - 120	6	20		
Tetrachloroethene	102	112	77 - 120	9	20		
Toluene	103	111	73 - 120	8	20		
1,1,1-Trichloroethane	107	113	78 - 120	5	20		
Trichloroethene	98	106	78 - 122	8	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	91	90	70 - 127
Toluene-d8 (Surr)	96	97	80 - 125
4-Bromofluorobenzene (Surr)	95	99	78 - 118
Dibromofluoromethane (Surr)	97	96	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-9458

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1619-26	Units: ug/L	MSD Lab Sample ID:	280-1619-26
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	04/01/2010 0929		Date Analyzed:	04/01/2010 0950
Date Prepared:	04/01/2010 0929		Date Prepared:	04/01/2010 0950

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	0.16	U	5.00	5.00	5.08	5.51
Bromodichloromethane	0.17	U	5.00	5.00	5.22	5.72
Carbon tetrachloride	0.19	U	5.00	5.00	5.32	5.68
Chlorobenzene	0.17	U	5.00	5.00	4.99	5.46
Chloroform	0.16	U	5.00	5.00	5.29	5.69
1,3-Dichlorobenzene	0.13	U	5.00	5.00	4.98	5.32
1,1-Dichloroethane	0.22	U	5.00	5.00	5.43	5.81
trans-1,2-Dichloroethene	0.15	U	5.00	5.00	5.12	5.24
1,1-Dichloroethene	0.23	U	5.00	5.00	4.66	4.90
1,2-Dichloropropane	0.18	U	5.00	5.00	5.07	5.63
Ethylbenzene	0.16	U	5.00	5.00	5.01	5.54
Methylene Chloride	0.32	U	5.00	5.00	4.41	4.66
Tetrachloroethene	0.20	U	5.00	5.00	5.12	5.60
Toluene	0.17	U	5.00	5.00	5.14	5.57
1,1,1-Trichloroethane	0.16	U	5.00	5.00	5.35	5.65
Trichloroethene	0.16	U	5.00	5.00	4.89	5.30

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Method Blank - Batch: 280-9544

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-9544/3	Analysis Batch:	280-9544	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7765.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1212			Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1212				

Analyte	Result	Qual	MDL	RL
Acetone	1.9	U	1.9	10
Benzene	0.16	U	0.16	1.0
Bromobenzene	0.17	U	0.17	1.0
Bromochloromethane	0.10	U	0.10	1.0
Bromodichloromethane	0.17	U	0.17	1.0
Bromoform	0.19	U	0.19	1.0
Bromomethane	0.21	U	0.21	1.0
2-Butanone (MEK)	2.0	U	2.0	5.0
n-Butylbenzene	0.14	U	0.14	1.0
sec-Butylbenzene	0.17	U	0.17	1.0
tert-Butylbenzene	0.16	U	0.16	1.0
Carbon disulfide	0.45	U	0.45	1.0
Carbon tetrachloride	0.19	U	0.19	1.0
Chlorobenzene	0.17	U	0.17	1.0
Dibromochloromethane	0.17	U	0.17	1.0
Chloroethane	0.41	U	0.41	1.0
Chloroform	0.16	U	0.16	1.0
Chloromethane	0.30	U	0.30	1.0
2-Chlorotoluene	0.17	U	0.17	1.0
4-Chlorotoluene	0.21	U	0.21	1.0
1,2-Dibromo-3-Chloropropane	0.47	U	0.47	1.0
Dibromomethane	0.17	U	0.17	1.0
1,2-Dichlorobenzene	0.15	U	0.15	1.0
1,3-Dichlorobenzene	0.13	U	0.13	1.0
1,4-Dichlorobenzene	0.16	U	0.16	1.0
Dichlorodifluoromethane	0.31	U	0.31	1.0
1,1-Dichloroethane	0.22	U	0.22	1.0
1,2-Dichloroethane	0.13	U	0.13	1.0
cis-1,2-Dichloroethene	0.15	U	0.15	1.0
trans-1,2-Dichloroethene	0.15	U	0.15	1.0
1,1-Dichloroethene	0.23	U	0.23	1.0
1,2-Dichloropropane	0.18	U	0.18	1.0
1,3-Dichloropropane	0.22	U	0.22	1.0
2,2-Dichloropropane	0.18	U	0.18	1.0
cis-1,3-Dichloropropene	0.16	U	0.16	1.0
trans-1,3-Dichloropropene	0.19	U	0.19	1.0
1,1-Dichloropropene	0.19	U	0.19	1.0
Ethylbenzene	0.16	U	0.16	1.0
Hexachlorobutadiene	0.12	U	0.12	1.0
2-Hexanone	1.7	U	1.7	5.0
Isopropylbenzene	0.19	U	0.19	1.0
4-Isopropyltoluene	0.20	U	0.20	1.0
Methylene Chloride	0.536	J	0.32	1.0

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Method Blank - Batch: 280-9544

Method: 8260B
Preparation: 5030B

Lab Sample ID:	MB 280-9544/3	Analysis Batch:	280-9544	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7765.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1212			Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1212				

Analyte	Result	Qual	MDL	RL
4-Methyl-2-pentanone	0.98	U	0.98	5.0
Naphthalene	0.22	U	0.22	1.0
n-Propylbenzene	0.16	U	0.16	1.0
Styrene	0.17	U	0.17	1.0
1,1,1,2-Tetrachloroethane	0.21	U	0.21	1.0
1,1,2,2-Tetrachloroethane	0.21	U	0.21	1.0
Tetrachloroethene	0.20	U	0.20	1.0
Toluene	0.17	U	0.17	1.0
1,2,3-Trichlorobenzene	0.21	U	0.21	1.0
1,2,4-Trichlorobenzene	0.21	U	0.21	1.0
1,1,1-Trichloroethane	0.16	U	0.16	1.0
1,1,2-Trichloroethane	0.27	U	0.27	1.0
Trichloroethene	0.16	U	0.16	1.0
Trichlorofluoromethane	0.29	U	0.29	1.0
1,2,3-Trichloropropane	0.33	U	0.33	1.0
1,2,4-Trimethylbenzene	0.15	U	0.15	1.0
1,3,5-Trimethylbenzene	0.16	U	0.16	1.0
Vinyl chloride	0.40	U	0.40	1.0
Xylenes, Total	0.19	U	0.19	1.0
1,2-Dibromoethane	0.18	U	0.18	1.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	103	70 - 127
Toluene-d8 (Surr)	102	80 - 125
4-Bromofluorobenzene (Surr)	99	78 - 118
Dibromofluoromethane (Surr)	96	77 - 119

Quality Control Results

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Lab Control Sample - Batch: 280-9544

Method: 8260B
Preparation: 5030B

Lab Sample ID:	LCS 280-9544/7	Analysis Batch:	280-9544	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7763.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1131			Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1131				

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
Benzene	5.00	4.21	84	77 - 120	
Bromodichloromethane	5.00	4.21	84	78 - 120	
Carbon tetrachloride	5.00	4.28	86	80 - 120	
Chlorobenzene	5.00	4.25	85	78 - 120	
Chloroform	5.00	4.34	87	78 - 120	
1,3-Dichlorobenzene	5.00	4.34	87	75 - 120	
1,1-Dichloroethane	5.00	4.34	87	77 - 120	
trans-1,2-Dichloroethene	5.00	4.19	84	80 - 120	
1,1-Dichloroethene	5.00	4.57	91	68 - 133	
1,2-Dichloropropane	5.00	4.37	87	76 - 120	
Ethylbenzene	5.00	4.42	88	78 - 120	
Methylene Chloride	5.00	4.83	97	71 - 120	
Tetrachloroethene	5.00	4.09	82	77 - 120	
Toluene	5.00	4.24	85	73 - 120	
1,1,1-Trichloroethane	5.00	4.26	85	78 - 120	
Trichloroethene	5.00	4.02	80	78 - 122	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	98	70 - 127
Toluene-d8 (Surr)	105	80 - 125
4-Bromofluorobenzene (Surr)	102	78 - 118
Dibromofluoromethane (Surr)	97	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-9544

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1609-G-1 MS	Analysis Batch:	280-9544	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7769.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1335			Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1335				

MSD Lab Sample ID:	280-1609-G-1 MSD	Analysis Batch:	280-9544	Instrument ID:	MSV_R1
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	R7770.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/31/2010 1356			Final Weight/Volume:	20 mL
Date Prepared:	03/31/2010 1356				

	% Rec.						
Analyte	MS	MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
Benzene	97	100	77 - 120	3	20		
Bromodichloromethane	96	99	78 - 120	3	20		
Carbon tetrachloride	112	118	80 - 120	5	21		
Chlorobenzene	94	96	78 - 120	1	20		
Chloroform	101	104	78 - 120	3	20		
1,3-Dichlorobenzene	91	94	75 - 120	3	20		
1,1-Dichloroethane	101	105	77 - 120	4	21		
trans-1,2-Dichloroethene	103	110	80 - 120	7	24		
1,1-Dichloroethene	120	129	68 - 133	6	20		
1,2-Dichloropropane	97	102	76 - 120	5	20		
Ethylbenzene	101	102	78 - 120	1	26		
Methylene Chloride	99	106	71 - 120	7	20		
Tetrachloroethene	101	102	77 - 120	1	20		
Toluene	99	100	73 - 120	1	20		
1,1,1-Trichloroethane	107	114	78 - 120	6	20		
Trichloroethene	94	89	78 - 122	1	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	104	110	70 - 127
Toluene-d8 (Surr)	101	99	80 - 125
4-Bromofluorobenzene (Surr)	96	97	78 - 118
Dibromofluoromethane (Surr)	97	98	77 - 119

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-9544

Method: 8260B
Preparation: 5030B

MS Lab Sample ID:	280-1609-G-1 MS	Units: ug/L	MSD Lab Sample ID:	280-1609-G-1 MSD
Client Matrix:	Water		Client Matrix:	Water
Dilution:	1.0		Dilution:	1.0
Date Analyzed:	03/31/2010 1335		Date Analyzed:	03/31/2010 1356
Date Prepared:	03/31/2010 1335		Date Prepared:	03/31/2010 1356

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
Benzene	0.16	U	5.00	5.00	4.86	5.02
Bromodichloromethane	0.17	U	5.00	5.00	4.78	4.93
Carbon tetrachloride	0.19	U	5.00	5.00	5.59	5.89
Chlorobenzene	0.17	U	5.00	5.00	4.72	4.78
Chloroform	0.16	U	5.00	5.00	5.03	5.19
1,3-Dichlorobenzene	0.13	U	5.00	5.00	4.56	4.72
1,1-Dichloroethane	0.22	U	5.00	5.00	5.06	5.25
trans-1,2-Dichloroethene	0.15	U	5.00	5.00	5.15	5.50
1,1-Dichloroethene	1.9		5.00	5.00	7.95	8.40
1,2-Dichloropropane	0.18	U	5.00	5.00	4.87	5.12
Ethylbenzene	0.16	U	5.00	5.00	5.07	5.12
Methylene Chloride	0.32	U	5.00	5.00	4.95	5.29
Tetrachloroethene	0.20	U	5.00	5.00	5.05	5.08
Toluene	0.17	U	5.00	5.00	4.95	5.02
1,1,1-Trichloroethane	0.16	U	5.00	5.00	5.37	5.68
Trichloroethene	13		5.00	5.00	17.5	17.3

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Method Blank - Batch: 280-8860

Method: 8260B SIM
Preparation: 5030B

Lab Sample ID:	MB 280-8860/4	Analysis Batch:	280-8860	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_34017.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0955			Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0955				

Analyte	Result	Qual	MDL	RL
1,4-Dioxane	0.75	U	0.75	2.0

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	91	70 - 127

Lab Control Sample - Batch: 280-8860

Method: 8260B SIM
Preparation: 5030B

Lab Sample ID:	LCS 280-8860/3	Analysis Batch:	280-8860	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_34015.D
Dilution:	1.0	Units:	ug/L	Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 0851			Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 0851				

Analyte	Spike Amount	Result	% Rec.	Limit	Qual
1,4-Dioxane	5.00	4.09	82	25 - 141	

Surrogate	% Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	93	70 - 127

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

Matrix Spike/
Matrix Spike Duplicate Recovery Report - Batch: 280-8860

Method: 8260B SIM
Preparation: 5030B

MS Lab Sample ID:	280-1619-10	Analysis Batch:	280-8860	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_34024.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1256			Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1256				

MSD Lab Sample ID:	280-1619-10	Analysis Batch:	280-8860	Instrument ID:	MSV_G2
Client Matrix:	Water	Prep Batch:	N/A	Lab File ID:	g2_34025.D
Dilution:	1.0			Initial Weight/Volume:	20 mL
Date Analyzed:	03/26/2010 1322			Final Weight/Volume:	20 mL
Date Prepared:	03/26/2010 1322				

Analyte	% Rec. MS	% Rec. MSD	Limit	RPD	RPD Limit	MS Qual	MSD Qual
1,4-Dioxane	105	106	25 - 141	1	20		

Surrogate	MS % Rec	MSD % Rec	Acceptance Limits
1,2-Dichloroethane-d4 (Surr)	97	93	70 - 127

Matrix Spike/
Matrix Spike Duplicate Data Report - Batch: 280-8860

Method: 8260B SIM
Preparation: 5030B

MS Lab Sample ID:	280-1619-10	Units:	ug/L	MSD Lab Sample ID:	280-1619-10
Client Matrix:	Water			Client Matrix:	Water
Dilution:	1.0			Dilution:	1.0
Date Analyzed:	03/26/2010 1256			Date Analyzed:	03/26/2010 1322
Date Prepared:	03/26/2010 1256			Date Prepared:	03/26/2010 1322

Analyte	Sample Result/Qual		MS Spike Amount	MSD Spike Amount	MS Result/Qual	MSD Result/Qual
1,4-Dioxane	0.75	U	5.00	5.00	5.25	5.29

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:280-8860					
LCS 280-8860/3	Lab Control Sample	T	Water	8260B SIM	
MB 280-8860/4	Method Blank	T	Water	8260B SIM	
280-1619-10	PIN12-0530	T	Water	8260B SIM	
280-1619-10MS	Matrix Spike	T	Water	8260B SIM	
280-1619-10MSD	Matrix Spike Duplicate	T	Water	8260B SIM	
280-1619-17	PIN12-0540	T	Water	8260B SIM	
280-1619-18	PIN12-0543	T	Water	8260B SIM	
280-1619-22	PIN12-0548	T	Water	8260B SIM	
280-1619-33	PIN99-2903	T	Water	8260B SIM	
Analysis Batch:280-9155					
LCS 280-9155/4	Lab Control Sample	T	Water	8260B	
MB 280-9155/5	Method Blank	T	Water	8260B	
280-1555-A-3 MS	Matrix Spike	T	Water	8260B	
280-1555-D-3 MSD	Matrix Spike Duplicate	T	Water	8260B	
280-1619-22	PIN12-0548	T	Water	8260B	
280-1619-32	PIN99-2877	T	Water	8260B	
280-1619-35	PIN12-S73B	T	Water	8260B	
280-1619-37	PIN99-2891	T	Water	8260B	
280-1619-38TB	PIN99-2885	T	Water	8260B	

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:280-9293					
LCS 280-9293/3	Lab Control Sample	T	Water	8260B	
MB 280-9293/4	Method Blank	T	Water	8260B	
280-1619-1	PIN06-0500	T	Water	8260B	
280-1619-2	PIN21-0502	T	Water	8260B	
280-1619-2MS	Matrix Spike	T	Water	8260B	
280-1619-2MSD	Matrix Spike Duplicate	T	Water	8260B	
280-1619-3	PIN21-0503	T	Water	8260B	
280-1619-4	PIN21-0504	T	Water	8260B	
280-1619-5	PIN21-0505	T	Water	8260B	
280-1619-6	PIN12-0509	T	Water	8260B	
280-1619-7	PIN12-0521	T	Water	8260B	
280-1619-8	PIN12-0527	T	Water	8260B	
280-1619-9	PIN12-0528	T	Water	8260B	
280-1619-10	PIN12-0530	T	Water	8260B	
280-1619-11	PIN12-0531	T	Water	8260B	
280-1619-12	PIN12-0532	T	Water	8260B	
280-1619-13	PIN12-0533	T	Water	8260B	
280-1619-14	PIN12-0534	T	Water	8260B	
280-1619-15	PIN12-0535	T	Water	8260B	
280-1619-16	PIN12-0536	T	Water	8260B	
280-1619-17	PIN12-0540	T	Water	8260B	
280-1619-18	PIN12-0543	T	Water	8260B	
280-1619-19	PIN12-0544	T	Water	8260B	
280-1619-20	PIN12-0545	T	Water	8260B	
Analysis Batch:280-9458					
LCS 280-9458/3	Lab Control Sample	T	Water	8260B	
MB 280-9458/4	Method Blank	T	Water	8260B	
280-1619-26	PIN12-0560-1	T	Water	8260B	
280-1619-26MS	Matrix Spike	T	Water	8260B	
280-1619-26MSD	Matrix Spike Duplicate	T	Water	8260B	
280-1619-27	PIN12-0560-2	T	Water	8260B	
280-1619-28	PIN12-0560-3	T	Water	8260B	
280-1619-29	PIN12-0561-1	T	Water	8260B	
280-1619-30	PIN12-0561-2	T	Water	8260B	
280-1619-31	PIN12-0561-3	T	Water	8260B	
280-1619-33	PIN99-2903	T	Water	8260B	
280-1619-34	PIN37-S002	T	Water	8260B	
280-1619-36	PIN23-SW01	T	Water	8260B	

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
Sdg Number: 10022892

QC Association Summary

Lab Sample ID	Client Sample ID	Report Basis	Client Matrix	Method	Prep Batch
GC/MS VOA					
Analysis Batch:280-9544					
LCS 280-9544/7	Lab Control Sample	T	Water	8260B	
MB 280-9544/3	Method Blank	T	Water	8260B	
280-1609-G-1 MS	Matrix Spike	T	Water	8260B	
280-1609-G-1 MSD	Matrix Spike Duplicate	T	Water	8260B	
280-1619-21	PIN12-0546	T	Water	8260B	
280-1619-23	PIN12-0559-1	T	Water	8260B	
280-1619-24	PIN12-0559-2	T	Water	8260B	
280-1619-25	PIN12-0559-3	T	Water	8260B	

Report Basis
T = Total

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1619-1 **Client ID:** PIN06-0500

Sample Date/Time: 03/17/2010 14:25 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-B-1		280-9293		03/31/2010 11:24	1	TAL DEN	GPM
A:8260B	280-1619-B-1		280-9293		03/31/2010 11:24	1	TAL DEN	GPM

Lab ID: 280-1619-2 **Client ID:** PIN21-0502

Sample Date/Time: 03/17/2010 16:25 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-B-2		280-9293		03/31/2010 11:44	1	TAL DEN	GPM
A:8260B	280-1619-B-2		280-9293		03/31/2010 11:44	1	TAL DEN	GPM

Lab ID: 280-1619-2 **Client ID:** PIN21-0502

Sample Date/Time: 03/17/2010 16:25 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-2 MS		280-9293		03/31/2010 08:43	1	TAL DEN	GPM
A:8260B	280-1619-A-2 MS		280-9293		03/31/2010 08:43	1	TAL DEN	GPM

Lab ID: 280-1619-2 **Client ID:** PIN21-0502

Sample Date/Time: 03/17/2010 16:25 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-2 MSD		280-9293		03/31/2010 09:04	1	TAL DEN	GPM
A:8260B	280-1619-A-2 MSD		280-9293		03/31/2010 09:04	1	TAL DEN	GPM

Lab ID: 280-1619-3 **Client ID:** PIN21-0503

Sample Date/Time: 03/17/2010 16:00 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-3		280-9293		03/31/2010 09:43	1	TAL DEN	GPM
A:8260B	280-1619-A-3		280-9293		03/31/2010 09:43	1	TAL DEN	GPM

Lab ID: 280-1619-4 **Client ID:** PIN21-0504

Sample Date/Time: 03/18/2010 09:55 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-4		280-9293		03/31/2010 13:26	1	TAL DEN	GPM
A:8260B	280-1619-C-4		280-9293		03/31/2010 13:26	1	TAL DEN	GPM

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1619-5 **Client ID:** PIN21-0505

Sample Date/Time: 03/18/2010 09:10 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-5		280-9293		03/31/2010 14:26	1	TAL DEN	GPM
A:8260B	280-1619-C-5		280-9293		03/31/2010 14:26	1	TAL DEN	GPM

Lab ID: 280-1619-6 **Client ID:** PIN12-0509

Sample Date/Time: 03/17/2010 13:40 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-6		280-9293		03/31/2010 10:04	1	TAL DEN	GPM
A:8260B	280-1619-C-6		280-9293		03/31/2010 10:04	1	TAL DEN	GPM

Lab ID: 280-1619-7 **Client ID:** PIN12-0521

Sample Date/Time: 03/17/2010 15:15 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-7		280-9293		03/31/2010 10:24	1	TAL DEN	GPM
A:8260B	280-1619-C-7		280-9293		03/31/2010 10:24	1	TAL DEN	GPM

Lab ID: 280-1619-8 **Client ID:** PIN12-0527

Sample Date/Time: 03/18/2010 11:35 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-8		280-9293		03/31/2010 13:46	1	TAL DEN	GPM
A:8260B	280-1619-C-8		280-9293		03/31/2010 13:46	1	TAL DEN	GPM

Lab ID: 280-1619-9 **Client ID:** PIN12-0528

Sample Date/Time: 03/18/2010 10:35 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-9		280-9293		03/31/2010 14:06	1	TAL DEN	GPM
A:8260B	280-1619-C-9		280-9293		03/31/2010 14:06	1	TAL DEN	GPM

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1619-10 **Client ID:** PIN12-0530

Sample Date/Time: 03/17/2010 13:33 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-10		280-9293		03/31/2010 10:44	1	TAL DEN	GPM
A:8260B	280-1619-A-10		280-9293		03/31/2010 10:44	1	TAL DEN	GPM
P:5030B	280-1619-C-10		280-8860		03/26/2010 12:31	1	TAL DEN	AEW
A:8260B SIM	280-1619-C-10		280-8860		03/26/2010 12:31	1	TAL DEN	AEW

Lab ID: 280-1619-10 MS **Client ID:** PIN12-0530

Sample Date/Time: 03/17/2010 13:33 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-D-10 MS		280-8860		03/26/2010 12:56	1	TAL DEN	AEW
A:8260B SIM	280-1619-D-10 MS		280-8860		03/26/2010 12:56	1	TAL DEN	AEW

Lab ID: 280-1619-10 MSD **Client ID:** PIN12-0530

Sample Date/Time: 03/17/2010 13:33 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-D-10 MSD		280-8860		03/26/2010 13:22	1	TAL DEN	AEW
A:8260B SIM	280-1619-D-10 MSD		280-8860		03/26/2010 13:22	1	TAL DEN	AEW

Lab ID: 280-1619-11 **Client ID:** PIN12-0531

Sample Date/Time: 03/17/2010 15:03 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-11		280-9293		03/31/2010 11:04	1	TAL DEN	GPM
A:8260B	280-1619-C-11		280-9293		03/31/2010 11:04	1	TAL DEN	GPM

Lab ID: 280-1619-12 **Client ID:** PIN12-0532

Sample Date/Time: 03/17/2010 15:35 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-12		280-9293		03/31/2010 12:05	1	TAL DEN	GPM
A:8260B	280-1619-A-12		280-9293		03/31/2010 12:05	1	TAL DEN	GPM

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1619-13 **Client ID:** PIN12-0533

Sample Date/Time: 03/18/2010 09:47 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-13		280-9293		03/31/2010 14:46	1	TAL DEN	GPM
A:8260B	280-1619-C-13		280-9293		03/31/2010 14:46	1	TAL DEN	GPM

Lab ID: 280-1619-14 **Client ID:** PIN12-0534

Sample Date/Time: 03/18/2010 10:30 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-14		280-9293		03/31/2010 15:06	1	TAL DEN	GPM
A:8260B	280-1619-A-14		280-9293		03/31/2010 15:06	1	TAL DEN	GPM

Lab ID: 280-1619-15 **Client ID:** PIN12-0535

Sample Date/Time: 03/18/2010 14:32 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-15		280-9293		03/31/2010 15:26	1	TAL DEN	GPM
A:8260B	280-1619-C-15		280-9293		03/31/2010 15:26	1	TAL DEN	GPM

Lab ID: 280-1619-16 **Client ID:** PIN12-0536

Sample Date/Time: 03/18/2010 15:19 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-16		280-9293		03/31/2010 15:46	1	TAL DEN	GPM
A:8260B	280-1619-C-16		280-9293		03/31/2010 15:46	1	TAL DEN	GPM

Lab ID: 280-1619-17 **Client ID:** PIN12-0540

Sample Date/Time: 03/17/2010 14:25 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-17		280-9293		03/31/2010 12:25	1	TAL DEN	GPM
A:8260B	280-1619-A-17		280-9293		03/31/2010 12:25	1	TAL DEN	GPM
P:5030B	280-1619-D-17		280-8860		03/26/2010 13:47	1	TAL DEN	AEW
A:8260B SIM	280-1619-D-17		280-8860		03/26/2010 13:47	1	TAL DEN	AEW

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1619-18 **Client ID:** PIN12-0543

Sample Date/Time: 03/17/2010 14:05 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-D-18		280-9293		03/31/2010 12:45	1	TAL DEN	GPM
A:8260B	280-1619-D-18		280-9293		03/31/2010 12:45	1	TAL DEN	GPM
P:5030B	280-1619-A-18		280-8860		03/26/2010 14:12	1	TAL DEN	AEW
A:8260B SIM	280-1619-A-18		280-8860		03/26/2010 14:12	1	TAL DEN	AEW

Lab ID: 280-1619-19 **Client ID:** PIN12-0544

Sample Date/Time: 03/17/2010 16:13 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-19		280-9293		03/31/2010 13:05	1	TAL DEN	GPM
A:8260B	280-1619-A-19		280-9293		03/31/2010 13:05	1	TAL DEN	GPM

Lab ID: 280-1619-20 **Client ID:** PIN12-0545

Sample Date/Time: 03/18/2010 11:16 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-20		280-9293		03/31/2010 16:06	1	TAL DEN	GPM
A:8260B	280-1619-C-20		280-9293		03/31/2010 16:06	1	TAL DEN	GPM

Lab ID: 280-1619-21 **Client ID:** PIN12-0546

Sample Date/Time: 03/18/2010 15:35 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-21		280-9544		03/31/2010 19:26	1	TAL DEN	HEW
A:8260B	280-1619-A-21		280-9544		03/31/2010 19:26	1	TAL DEN	HEW

Lab ID: 280-1619-22 **Client ID:** PIN12-0548

Sample Date/Time: 03/17/2010 15:15 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-22		280-9155		03/30/2010 16:24	1	TAL DEN	HEW
A:8260B	280-1619-A-22		280-9155		03/30/2010 16:24	1	TAL DEN	HEW
P:5030B	280-1619-C-22		280-8860		03/26/2010 14:38	1	TAL DEN	AEW
A:8260B SIM	280-1619-C-22		280-8860		03/26/2010 14:38	1	TAL DEN	AEW

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1619-23 **Client ID:** PIN12-0559-1

Sample Date/Time: 03/18/2010 10:55 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-23		280-9544		03/31/2010 19:47	1	TAL DEN	HEW
A:8260B	280-1619-C-23		280-9544		03/31/2010 19:47	1	TAL DEN	HEW

Lab ID: 280-1619-24 **Client ID:** PIN12-0559-2

Sample Date/Time: 03/18/2010 11:35 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-24		280-9544		03/31/2010 20:07	1	TAL DEN	HEW
A:8260B	280-1619-A-24		280-9544		03/31/2010 20:07	1	TAL DEN	HEW

Lab ID: 280-1619-25 **Client ID:** PIN12-0559-3

Sample Date/Time: 03/18/2010 12:05 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-B-25		280-9544		03/31/2010 20:27	1	TAL DEN	HEW
A:8260B	280-1619-B-25		280-9544		03/31/2010 20:27	1	TAL DEN	HEW

Lab ID: 280-1619-26 **Client ID:** PIN12-0560-1

Sample Date/Time: 03/18/2010 13:45 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-26		280-9458		04/01/2010 09:07	1	TAL DEN	GPM
A:8260B	280-1619-A-26		280-9458		04/01/2010 09:07	1	TAL DEN	GPM

Lab ID: 280-1619-26 MS **Client ID:** PIN12-0560-1

Sample Date/Time: 03/18/2010 13:45 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-B-26 MS		280-9458		04/01/2010 09:29	1	TAL DEN	GPM
A:8260B	280-1619-B-26 MS		280-9458		04/01/2010 09:29	1	TAL DEN	GPM

Lab ID: 280-1619-26 MSD **Client ID:** PIN12-0560-1

Sample Date/Time: 03/18/2010 13:45 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-B-26 MSD		280-9458		04/01/2010 09:50	1	TAL DEN	GPM
A:8260B	280-1619-B-26 MSD		280-9458		04/01/2010 09:50	1	TAL DEN	GPM

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1619-27 **Client ID:** PIN12-0560-2

Sample Date/Time: 03/18/2010 14:10 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-27		280-9458		04/01/2010 10:34	1	TAL DEN	GPM
A:8260B	280-1619-A-27		280-9458		04/01/2010 10:34	1	TAL DEN	GPM

Lab ID: 280-1619-28 **Client ID:** PIN12-0560-3

Sample Date/Time: 03/18/2010 14:40 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-28		280-9458		04/01/2010 10:55	1	TAL DEN	GPM
A:8260B	280-1619-A-28		280-9458		04/01/2010 10:55	1	TAL DEN	GPM

Lab ID: 280-1619-29 **Client ID:** PIN12-0561-1

Sample Date/Time: 03/18/2010 09:00 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-29		280-9458		04/01/2010 11:16	1	TAL DEN	GPM
A:8260B	280-1619-C-29		280-9458		04/01/2010 11:16	1	TAL DEN	GPM

Lab ID: 280-1619-30 **Client ID:** PIN12-0561-2

Sample Date/Time: 03/18/2010 09:45 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-30		280-9458		04/01/2010 11:38	1	TAL DEN	GPM
A:8260B	280-1619-C-30		280-9458		04/01/2010 11:38	1	TAL DEN	GPM

Lab ID: 280-1619-31 **Client ID:** PIN12-0561-3

Sample Date/Time: 03/18/2010 10:15 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-31		280-9458		04/01/2010 12:00	1	TAL DEN	GPM
A:8260B	280-1619-A-31		280-9458		04/01/2010 12:00	1	TAL DEN	GPM

Lab ID: 280-1619-32 **Client ID:** PIN99-2877

Sample Date/Time: 03/17/2010 08:00 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-32		280-9155		03/30/2010 16:45	1	TAL DEN	HEW
A:8260B	280-1619-A-32		280-9155		03/30/2010 16:45	1	TAL DEN	HEW

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
SDG: 10022892

Laboratory Chronicle

Lab ID: 280-1619-33 **Client ID:** PIN99-2903

Sample Date/Time: 03/18/2010 13:00 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-33		280-9458		04/01/2010 12:22	1	TAL DEN	GPM
A:8260B	280-1619-C-33		280-9458		04/01/2010 12:22	1	TAL DEN	GPM
P:5030B	280-1619-D-33		280-8860		03/26/2010 15:03	1	TAL DEN	AEW
A:8260B SIM	280-1619-D-33		280-8860		03/26/2010 15:03	1	TAL DEN	AEW

Lab ID: 280-1619-34 **Client ID:** PIN37-S002

Sample Date/Time: 03/18/2010 13:45 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-34		280-9458		04/01/2010 12:43	1	TAL DEN	GPM
A:8260B	280-1619-C-34		280-9458		04/01/2010 12:43	1	TAL DEN	GPM

Lab ID: 280-1619-35 **Client ID:** PIN12-S73B

Sample Date/Time: 03/17/2010 16:20 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-B-35		280-9155		03/30/2010 17:06	1	TAL DEN	HEW
A:8260B	280-1619-B-35		280-9155		03/30/2010 17:06	1	TAL DEN	HEW

Lab ID: 280-1619-36 **Client ID:** PIN23-SW01

Sample Date/Time: 03/18/2010 13:30 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-36		280-9458		04/01/2010 13:05	1	TAL DEN	GPM
A:8260B	280-1619-C-36		280-9458		04/01/2010 13:05	1	TAL DEN	GPM

Lab ID: 280-1619-37 **Client ID:** PIN99-2891

Sample Date/Time: 03/17/2010 08:00 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-C-37		280-9155		03/30/2010 17:26	1	TAL DEN	HEW
A:8260B	280-1619-C-37		280-9155		03/30/2010 17:26	1	TAL DEN	HEW

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
SDG: 10022892

Laboratory Chronicle

Lab ID: **280-1619-38** **Client ID:** **PIN99-2885**

Sample Date/Time: 03/17/2010 12:00 Received Date/Time: 03/23/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1619-A-38		280-9155		03/30/2010 17:47	1	TAL DEN	HEW
A:8260B	280-1619-A-38		280-9155		03/30/2010 17:47	1	TAL DEN	HEW

Lab ID: **MB** **Client ID:** **N/A**

Sample Date/Time: N/A Received Date/Time: N/A

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	MB 280-9155/5		280-9155		03/30/2010 10:31	1	TAL DEN	HEW
A:8260B	MB 280-9155/5		280-9155		03/30/2010 10:31	1	TAL DEN	HEW
P:5030B	MB 280-9293/4		280-9293		03/31/2010 07:43	1	TAL DEN	GPM
A:8260B	MB 280-9293/4		280-9293		03/31/2010 07:43	1	TAL DEN	GPM
P:5030B	MB 280-9544/3		280-9544		03/31/2010 12:12	1	TAL DEN	HEW
A:8260B	MB 280-9544/3		280-9544		03/31/2010 12:12	1	TAL DEN	HEW
P:5030B	MB 280-9458/4		280-9458		04/01/2010 08:46	1	TAL DEN	GPM
A:8260B	MB 280-9458/4		280-9458		04/01/2010 08:46	1	TAL DEN	GPM
P:5030B	MB 280-8860/4		280-8860		03/26/2010 09:55	1	TAL DEN	AEW
A:8260B SIM	MB 280-8860/4		280-8860		03/26/2010 09:55	1	TAL DEN	AEW

Lab ID: **LCS** **Client ID:** **N/A**

Sample Date/Time: N/A Received Date/Time: N/A

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	LCS 280-9155/4		280-9155		03/30/2010 09:29	1	TAL DEN	HEW
A:8260B	LCS 280-9155/4		280-9155		03/30/2010 09:29	1	TAL DEN	HEW
P:5030B	LCS 280-9293/3		280-9293		03/31/2010 07:03	1	TAL DEN	GPM
A:8260B	LCS 280-9293/3		280-9293		03/31/2010 07:03	1	TAL DEN	GPM
P:5030B	LCS 280-9544/7		280-9544		03/31/2010 11:31	1	TAL DEN	HEW
A:8260B	LCS 280-9544/7		280-9544		03/31/2010 11:31	1	TAL DEN	HEW
P:5030B	LCS 280-9458/3		280-9458		04/01/2010 08:02	1	TAL DEN	GPM
A:8260B	LCS 280-9458/3		280-9458		04/01/2010 08:02	1	TAL DEN	GPM
P:5030B	LCS 280-8860/3		280-8860		03/26/2010 08:51	1	TAL DEN	AEW
A:8260B SIM	LCS 280-8860/3		280-8860		03/26/2010 08:51	1	TAL DEN	AEW

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
SDG: 10022892

Laboratory Chronicle

Lab ID: **MS** **Client ID:** **N/A**

Sample Date/Time: 03/17/2010 10:57 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-A-3 MS		280-9155		03/30/2010 11:33	1	TAL DEN	HEW
A:8260B	280-1555-A-3 MS		280-9155		03/30/2010 11:33	1	TAL DEN	HEW
P:5030B	280-1609-G-1 MS		280-9544		03/31/2010 13:35	1	TAL DEN	HEW
A:8260B	280-1609-G-1 MS		280-9544		03/31/2010 13:35	1	TAL DEN	HEW

Lab ID: **MSD** **Client ID:** **N/A**

Sample Date/Time: 03/17/2010 10:57 Received Date/Time: 03/19/2010 09:00

Method	Bottle ID	Run	Analysis Batch	Prep Batch	Date Prepared / Analyzed	Dil	Lab	Analyst
P:5030B	280-1555-D-3 MSD		280-9155		03/30/2010 11:53	1	TAL DEN	HEW
A:8260B	280-1555-D-3 MSD		280-9155		03/30/2010 11:53	1	TAL DEN	HEW
P:5030B	280-1609-G-1 MSD		280-9544		03/31/2010 13:56	1	TAL DEN	HEW
A:8260B	280-1609-G-1 MSD		280-9544		03/31/2010 13:56	1	TAL DEN	HEW

Lab References:
TAL DEN = TestAmerica Denver

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1619-1

SDG No.: 10022892

Instrument ID: MSV_H Analysis Batch Number: 8342

Lab Sample ID: IC 280-8342/2 Client Sample ID:

Date Analyzed: 03/23/10 10:08 Lab File ID: H2475.D GC Column: DB-624 ID: 0.53(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Chloromethane	2.12	Analyte not Identified by the Data System	meierg	03/23/10 12:39
Chloroethane	2.57	Analyte not Identified by the Data System	meierg	03/23/10 12:39
Dibromomethane	7.18	Analyte not Identified by the Data System	meierg	03/23/10 12:39
4-Chlorotoluene	12.95	Baseline Event	meierg	03/23/10 12:59

Lab Sample ID: IC 280-8342/3 Client Sample ID:

Date Analyzed: 03/23/10 10:30 Lab File ID: H2476.D GC Column: DB-624 ID: 0.53(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Vinyl acetate	4.35	Analyte not Identified by the Data System	meierg	03/23/10 13:31

Lab Sample ID: IC 280-8342/4 Client Sample ID:

Date Analyzed: 03/23/10 10:51 Lab File ID: H2477.D GC Column: DB-624 ID: 0.53(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Chloroethyl vinyl ether	7.81	Analyte not Identified by the Data System	meierg	03/23/10 12:56

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1619-1

SDG No.: 10022892

Instrument ID: MSV_H Analysis Batch Number: 9458

Lab Sample ID: CCV 280-9458/2 Client Sample ID:

Date Analyzed: 04/01/10 07:19 Lab File ID: H2682.D GC Column: DB-624 ID: 0.53(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Acetone	3.26	Split Peak	meierg	04/01/10 07:46

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1619-1

SDG No.: 10022892

Instrument ID: MSV_R1 Analysis Batch Number: 8905

Lab Sample ID: IC 280-8905/2 Client Sample ID:

Date Analyzed: 03/26/10 07:45 Lab File ID: R7591.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
trans-1,3-Dichloropropene	9.34	Analyte not Identified by the Data System	tinkhams	03/26/10 11:59

Lab Sample ID: IC 280-8905/4 Client Sample ID:

Date Analyzed: 03/26/10 08:27 Lab File ID: R7593.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Ethanol	5.33	Analyte not Identified by the Data System	tinkhams	03/26/10 11:58

Lab Sample ID: IC 280-8905/10 Client Sample ID:

Date Analyzed: 03/26/10 10:39 Lab File ID: R7599.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Chloroethyl vinyl ether	8.79	Analyte not Identified by the Data System	waterlah	03/30/10 08:37

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1619-1

SDG No.: 10022892

Instrument ID: MSV_R2 Analysis Batch Number: 8473

Lab Sample ID: IC 280-8473/5 Client Sample ID:

Date Analyzed: 03/24/10 09:17 Lab File ID: RR4926.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Acetonitrile	4.57	Split Peak	waterlah	03/24/10 10:00

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1619-1

SDG No.: 10022892

Instrument ID: MSV_Z Analysis Batch Number: 8731

Lab Sample ID: IC 280-8731/3 Client Sample ID:

Date Analyzed: 03/25/10 08:09 Lab File ID: Z8449.D GC Column: DB-624 ID: 0.53(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Hexanone	8.32	Split Peak	meierg	03/25/10 10:14

Lab Sample ID: IC 280-8731/8 Client Sample ID:

Date Analyzed: 03/25/10 09:51 Lab File ID: Z8454.D GC Column: DB-624 ID: 0.53(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Chloromethane	1.53	Analyte not Identified by the Data System	meierg	03/25/10 10:15

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1619-1

SDG No.: 10022892

Instrument ID: MSV_H Analysis Batch Number: 8342

Lab Sample ID: IC 280-8342/2 Client Sample ID:

Date Analyzed: 03/23/10 10:08 Lab File ID: H2475.D GC Column: DB-624 ID: 0.53(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Chloromethane	2.12	Analyte not Identified by the Data System	meierg	03/23/10 12:39
Chloroethane	2.57	Analyte not Identified by the Data System	meierg	03/23/10 12:39
Dibromomethane	7.18	Analyte not Identified by the Data System	meierg	03/23/10 12:39
4-Chlorotoluene	12.95	Baseline Event	meierg	03/23/10 12:59

Lab Sample ID: IC 280-8342/3 Client Sample ID:

Date Analyzed: 03/23/10 10:30 Lab File ID: H2476.D GC Column: DB-624 ID: 0.53(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Vinyl acetate	4.35	Analyte not Identified by the Data System	meierg	03/23/10 13:31

Lab Sample ID: IC 280-8342/4 Client Sample ID:

Date Analyzed: 03/23/10 10:51 Lab File ID: H2477.D GC Column: DB-624 ID: 0.53(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Chloroethyl vinyl ether	7.81	Analyte not Identified by the Data System	meierg	03/23/10 12:56

WA/G

8260B

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1619-1

SDG No.: 10022892

Instrument ID: MSV_H Analysis Batch Number: 9458

Lab Sample ID: CCV 280-9458/2 Client Sample ID:

Date Analyzed: 04/01/10 07:19 Lab File ID: H2682.D GC Column: DB-624 ID: 0.53(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Acetone	3.26	Split Peak	meierg	04/01/10 07:46

4/9

8260B

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1619-1

SDG No.: 10022892

Instrument ID: MSV_R1 Analysis Batch Number: 8905

Lab Sample ID: IC 280-8905/2 Client Sample ID:

Date Analyzed: 03/26/10 07:45 Lab File ID: R7591.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
trans-1,3-Dichloropropene	9.34	Analyte not Identified by the Data System	tinkhams	03/26/10 11:59

Lab Sample ID: IC 280-8905/4 Client Sample ID:

Date Analyzed: 03/26/10 08:27 Lab File ID: R7593.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Ethanol	5.33	Analyte not Identified by the Data System	tinkhams	03/26/10 11:58

Lab Sample ID: IC 280-8905/10 Client Sample ID:

Date Analyzed: 03/26/10 10:39 Lab File ID: R7599.D GC Column: DB-624 ID: 0.25(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Chloroethyl vinyl ether	8.79	Analyte not Identified by the Data System	waterlah	03/30/10 08:37

W 4/9

8260B

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1619-1

SDG No.: 10022892

Instrument ID: MSV_R2 Analysis Batch Number: 8473

Lab Sample ID: IC 280-8473/5 Client Sample ID:

Date Analyzed: 03/24/10 09:17 Lab File ID: RR4926.D GC Column: DB-624 ID: 0.32(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Acetonitrile	4.57	Split Peak	waterlah	03/24/10 10:00

4/9

8260B

GC/MS VOA MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica Denver Job No.: 280-1619-1

SDG No.: 10022892

Instrument ID: MSV_Z Analysis Batch Number: 8731

Lab Sample ID: IC 280-8731/3 Client Sample ID:

Date Analyzed: 03/25/10 08:09 Lab File ID: Z8449.D GC Column: DB-624 ID: 0.53(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
2-Hexanone	8.32	Split Peak	meierg	03/25/10 10:14

Lab Sample ID: IC 280-8731/8 Client Sample ID:

Date Analyzed: 03/25/10 09:51 Lab File ID: Z8454.D GC Column: DB-624 ID: 0.53(mm)

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Chloromethane	1.53	Analyte not Identified by the Data System	meierg	03/25/10 10:15

4/4/4

8260B

Shipping and Receiving Documents

3.3
AG
IMI
3/23/10

Stoller
Legacy Management Team

Chain of Custody / Sample Submittal Form

RIN: 10022892
COC: 10022892.5.1
Sampler(s): baer, lombardi, atkinson, caballero, moe, walters

Project: Pinellas Monitoring
Purchase Order: 3864
Cost Number: 1-502-1-06-509-4-02
Turnaround (Days): 28
Matrix: WA - Water

Laboratory: TestAmerica Denver
Address: 4955 Yarrow Street
Arvada, Colorado 80002
Phone: 303 736 0100

Ship #	Ticket	Sample Date	Time	Site	Location	Container	# Cont.	Preservation	Matrix	Comp.	Grab	Filtered	QC	Analysis
5	IDW 096	03/17/2010	14:25	PIN06	PIN06-0500	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 229	03/17/2010	16:25	PIN21	PIN21-0502	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 230	03/17/2010	16:00	PIN21	PIN21-0503	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 231	03/18/2010	9:55	PIN21	PIN21-0504	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 232	03/18/2010	9:10	PIN21	PIN21-0505	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 097	03/17/2010	13:40	PIN12	PIN12-0509	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 103	03/17/2010	15:15	PIN12	PIN12-0521	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 107	03/18/2010	11:35	PIN12	PIN12-0527	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 108	03/18/2010	10:35	PIN12	PIN12-0528	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 110	03/17/2010	13:33	PIN12	PIN12-0530	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
5	IDW 111	03/17/2010	15:03	PIN12	PIN12-0531	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 112	03/17/2010	15:35	PIN12	PIN12-0532	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 113	03/18/2010	9:47	PIN12	PIN12-0533	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 114	03/18/2010	10:30	PIN12	PIN12-0534	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 115	03/18/2010	14:32	PIN12	PIN12-0535	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 116	03/18/2010	15:19	PIN12	PIN12-0536	Glass 40 mL	3	4 C, HCl	WA			N		VOA

Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time
	3-19-10	1210		3/22/10	7W			
Received by (signature)	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**
	3-19-10	1210		3/23/10	0900			

Stoller
Legacy Management Team

Chain of Custody / Sample Submittal Form

RIN: 10022892
COC: 10022892.5.2
Sampler(s): baer, lombardi, atkinson, caballero, moe, walters

Project: Pinellas Monitoring
Purchase Order: 3864 **Cost Number:** 1-502-1-06-509-4-02
Turnaround (Days): 28 **Matrix:** WA - Water

Laboratory: TestAmerica Denver
Address: 4955 Yarrow Street
Arvada, Colorado 80002
Phone: 303 736 0100

Ship #	Ticket	Sample Date	Time	Site	Location	Container	# Cont.	Preservation	Matrix	Comp.	Grab	Filtered	QC	Analysis
5	IDW 120	03/17/2010	14:25	PIN12	PIN12-0540	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
5	IDW 123	03/17/2010	14:05	PIN12	PIN12-0543	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
5	IDW 124	03/17/2010	16:13	PIN12	PIN12-0544	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 125	03/18/2010	11:16	PIN12	PIN12-0545	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 126	03/18/2010	15:35	PIN12	PIN12-0546	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 128	03/17/2010	15:15	PIN12	PIN12-0548	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
5	IDW 148	03/18/2010	10:55	PIN12	PIN12-0559-1	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 149	03/18/2010	11:35	PIN12	PIN12-0559-2	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 150	03/18/2010	12:05	PIN12	PIN12-0559-3	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 151	03/18/2010	13:45	PIN12	PIN12-0560-1	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 152	03/18/2010	14:10	PIN12	PIN12-0560-2	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 153	03/18/2010	14:40	PIN12	PIN12-0560-3	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 154	03/18/2010	9:00	PIN12	PIN12-0561-1	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 155	03/18/2010	9:45	PIN12	PIN12-0561-2	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 156	03/18/2010	10:15	PIN12	PIN12-0561-3	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 245	03/17/2010	8:00	PIN99	PIN99-2877	Glass 40 mL	3	4 C, HCl	WA			N		VOA

Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time	Relinquished by (signature)	Date	Time
[signature]	3-19-10	1210	[signature]	3/22/10	1200			
Received by (signature)	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**
[signature]	3-19-10	1210	[signature]	3/23/10	0900			

Stoller
Legacy Management Team

Chain of Custody / Sample Submittal Form

RIN: 10022892
COC: 10022892.5.3
Sampler(s): baer, lombardi, atkinson, caballero, moe, walters

Project: Pinellas Monitoring
Purchase Order: 3864
Cost Number: 1-502-1-06-509-4-02
Turnaround (Days): 28
Matrix: WA - Water

Laboratory: TestAmerica Denver
Address: 4955 Yarrow Street
Arvada, Colorado 80002
Phone: 303 736 0100

Ship #	Ticket	Sample Date	Time	Site	Location	Container	# Cont.	Preservation	Matrix	Comp.	Grab	Filtered	QC	Analysis
5	IEX 517	03/18/2010	13:00	PIN99	PIN99-2903	Glass 40 mL	4	4 C, HCl	WA			N		VOA,Dioxane
5	IDW 234	03/18/2010	13:45	PIN37	PIN37-S002	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 203	03/17/2010	16:20	PIN12	PIN12-S73B	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 233	03/18/2010	13:30	PIN23	PIN23-SW01	Glass 40 mL	3	4 C, HCl	WA			N		VOA
5	IDW 259	3/17/10	0800	PIN99	PIN99-2891	"	3	"	WA			N		VOA
5	IDW 253	↓	1200	PIN99	PIN99-2885	"	3	"	WA			N		VOA

Reliquished by (signature)	Date	Time	Reliquished by (signature)	Date	Time	Reliquished by (signature)	Date	Time
[signature]	3-19-10	1210	[signature]	3/22/10	[illegible]			
Received by (signature)	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**	**Received by (signature)**	**Date**	**Time**
[signature]	3-19-10	1210	[signature]	3/23/10	0900			

Login Sample Receipt Check List

Client: S.M. Stoller Corporation

Job Number: 280-1619-1
SDG Number: 10022892

Login Number: 1619
Creator: Green, Angel L
List Number: 1

List Source: TestAmerica Denver

Question	T / F/ NA	Comment
Radioactivity either was not measured or, if measured, is at or below background	True	
The cooler's custody seal, if present, is intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the sample IDs on the containers and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
VOA sample vials do not have headspace or bubble is <6mm (1/4") in diameter.	True	
If necessary, staff have been informed of any short hold time or quick TAT needs	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Is the Field Sampler's name present on COC?	True	
Sample Preservation Verified	True	

Loc: 280

fedex.com 1.800.GoFedEx 1.800.463.3339

FedEx Express **US Airbill**

Issue Number [illegible]

FedEx TRK# 0215 8712 8177 1966

TUE - 23 MAR A1
PRIORITY OVERNIGHT

80002
CO-US
DEN

XH WHHA

RECIPIENT: PEEL HERE

1 From This portion can be removed for Recipient's records.

Date 3/22/10 FedEx Tracking Number 871281771966

Sender's Name BUSTODI Phone 813 585 7122

Company TESTAMERICA TAMPA

Address 6712 BENJAMIN RD STE 100 Dept./Floor/Suite/Room

City TAMPA State FL ZIP 33634-4403

2 Your Internal Billing Reference

3 To

Recipient's Name SAMPLE RECEIVER Phone 303 736 0100

Company TEST AMERICA

HOLD Weekday — Print FedEx location address below. NOT available for FedEx First Overnight.

HOLD Saturday — Print FedEx location address below. Available ONLY for FedEx Priority Overnight and FedEx 2Day to select locations.

Address — We cannot deliver to P.O. boxes or P.O. ZIP codes. Dept./Floor/Suite/Room

Address 4955 YARROW ST

Print FedEx location address here if HOLD option is selected.

City ARVADA State CO ZIP 80002

0414646911

8712 8177 1966

Emp# 631384 22MAR10 TPFA

Large Pak, and FedEx Sturdy Pak.

6 Special Handling and Delivery Signature Options

☐ SATURDAY Delivery — NOT available for FedEx Standard Overnight, FedEx First Overnight, FedEx Express Saver, or FedEx 3Day Freight.

☐ No Signature Required — Package may be left without obtaining a signature for delivery.

☐ Direct Signature — Someone at recipient's address may sign for delivery. *Fee applies.*

☐ Indirect Signature — If no one is available at recipient's address, someone at a neighboring address may sign for delivery. For residential deliveries only. *Fee applies.*

Does this shipment contain dangerous goods?

One box must be checked.

☑ No ☐ Yes — As per attached Shipper's Declaration. ☐ Yes — Shipper's Declaration not required. ☐ Dry Ice — Dry Ice, 9, UN 1845 ____ x ____ kg

Dangerous goods (including dry ice) cannot be shipped in FedEx packaging or placed in a FedEx Express Drop Box.

☐ Cargo Aircraft Only

7 Payment *Bill to:*

Enter FedEx Acct. No. or Credit Card No. below.

Obtain Recip. Acct. No. ☐

☐ Sender — Acct. No. in Section 1 will be billed. ☑ Recipient ☐ Third Party ☐ Credit Card ☐ Cash/Check

Total Packages 1 Total Weight 54 LBS Credit Card Auth.

Our liability is limited to $100 unless you declare a higher value. See the current FedEx Service Guide for details.

553

Rev. Date 2/08 Part #158279 ©1994-2008 FedEx PRINTED IN U.S.A. SRS