

ER PROGRAM DATA ASSESSMENT
SUMMARY REPORT FORM

Batch No. 8908S036 Site 881 Hillside
 Laboratory Accu-Labs Research No. of Samples/Matrix 6/Water
 Method Standard Methods Reviewer Org. TechLaw, Inc.
 Sample Numbers SW070005, SW070005D, SW070005FB, G52870889002, G05870889002, G08870889002

Data Assessment Summary

	<u>Oil/ Grease</u>	<u>Alkalinity</u>	<u>Chloride</u>	<u>Nitrate/ Nitrite</u>	<u>Sulfide/ Sulfate</u>	<u>Gravimetric</u>	<u>Comments</u>
1. <u>Holding Times</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>A</u>	<u>Action Item 1</u>
2. <u>Calibrations</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>V</u>	
3. <u>Blanks</u>	<u>V</u>	<u>V</u>	<u>A</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>Action Item 2</u>
4. <u>Lab Control Sample Results</u>	<u>N/A</u>	<u>V</u>	<u>N/A</u>	<u>N/A</u>	<u>N/A</u>	<u>V</u>	
5. <u>Duplicate Sample Results</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>V</u>	
6. <u>Matrix Spike Sample Results</u>	<u>V</u>	<u>N/A</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>N/A</u>	
7. <u>Sample Verification</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>V</u>	
8. <u>Other QC</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>V</u>	<u>V</u>	
9. <u>Overall Assessment</u>	<u>V</u>	<u>V</u>	<u>A</u>	<u>V</u>	<u>V</u>	<u>A</u>	<u>Data is valid, or acceptable with qualifications</u>

V = Data had no problems.

A = Data acceptable but qualified due to problems.

R = Data rejected.

X = Problems, but do not affect data.

N/A = Not applicable.

Data Quality: Data contained in this batch were reviewed and found to be valid, or acceptable with qualifications. Acceptable qualified data may be used provided that individual values impacted by the "Action Items" listed below are appropriately flagged.
 (Refer to attached Results Summary Tables).

ADMIN REVIEWED FOR CLASSIFICATION

By R. B. Hoffman (u)
 Date 2-11-90

REVIEWED FOR CLASSIFICATION/UCM
 By George H. Salloch
 Date 6/27/90


s036/rk23j
 A-DU01-000088

Action Items: 1) The TSS and TDS values are estimated (J) because holding times were exceeded.

2) The Chloride values for G05870889002 and G08870889002 are estimated (J) because there was no evidence of a blank analysis after the run.

Comments: None

Note: Data Summary Tables are attached.



Reviewer Signature

03/12/90

Date

